



www.chemasianj.org



LABS: Laboratory Automation and Batch Scheduling – A Modular Open Source Python Program for the Control of Automated Electrochemical Synthesis with a Web Interface

Maximilian M. Hielscher^{+,[a]} Maurice Dörr^{+,[a]} Johannes Schneider,^[a] and Siegfried R. Waldvogel^{*(a, b)}

Dedicated to Prof. Dr. Dennis Curran on the occasion of his 70th birthday.

Abstract: With LABS, an open source Python-based lab software is established that enables users to orchestrate automated synthesis setups. The software consists of a user-friendly interface for data input and system monitoring. A flexible backend architecture enables the integration of multiple lab devices. The software allows users to easily modify experimental parameters or routines and switch between different lab devices. Compared to previously

published projects, we aim to provide a more widely applicable and easily customizable automation software for any experimental setup. The usefulness of this tool was demonstrated in the oxidative coupling of 2,4-dimethylphenol to the corresponding 2,2'-biphenol. In this context, the suitable electrolysis parameters for flow electrolysis were optimized by way of design of experiments.

Introduction

In recent years, electro-organic chemistry has developed into a cost-friendly synthesis alternative compatible with green chemistry principles^[1] for the preparation of small molecules, the liberation of value-added products from waste streams, and the sustainable production of platform oxidizers.^[2] The use of electrons as a reagent often facilitates the generation of radical intermediates, which can offer shortcuts from multi-step reaction sequences to a single direct step.^[3,4] The direct (or mediated) electron transfer to or from the electrode often allows a metal-free catalytic reaction to be carried out.^[4,5] Typical setups for electrochemical reactions are either batch-type mode,^[6] or electrochemical flow cells through which the electrolyte is pumped, either in a single pass, or pumped multiple times through the flow cell using a reservoir.^[7] Electrochemical reactions in flow setups often exhibit better scalability

and higher space-time yields, and the (typically) shorter interelectrode distances allow more sustainable syntheses using significantly smaller amounts of supporting electrolyte. [8,9] However, electrochemical reactions in flow require more peripheral equipment such as pumps, valves, and fraction collectors, and are therefore often operated with only a single electrolyzer, whereas for screening in batch a variety of systems exist that allow parallel operation of several reactions. A review article for high-throughput screening in electro-organic synthesis was compiled by J. C. Vantourout and co-workers. [10] With regards to electrochemical flow cells, methods like 3D printing prototyping and miniaturization to speed up development.[11] Additional reaction parameters for electrolysis in flow, such as flow rate, can complicate the optimization. This challenge can be overcome systematically by methodological approaches such as design of experiments or Bayesian $optimization.^{\tiny [12]}$

Automation of the setup for electrolysis in flow can help to increase the efficiency of screening runs, while at the same time observables such as the cell voltage can be continuously monitored. Analysis of this data can provide information on reproducibility and status of the electrodes, e.g., fouling or corrosion events.^[17] Due to the growing importance of electronic laboratory notebooks, flexible control loops capable of automated processing of measurement data are required.

Automated systems are increasingly being established not only for commercial applications, but also as free and open source software. ^[18] In particular, if the functional scope of the control software is to be changed, open source projects offer great advantages over closed source. One such advantage is that changes can be implemented, tested, and made generally available to the community by the user. Since 2016 several groups have reported projects dealing with laboratory automation (Figure 1). These are commonly built using LabView or

- [a] M. M. Hielscher,⁺ M. Dörr,⁺ J. Schneider, Prof. Dr. S. R. Waldvogel Department of Chemistry Johannes Gutenberg University Duesbergweg 10–14, 55128 Mainz (Germany) E-mail: waldvogel@uni-mainz.de
- [b] Prof. Dr. S. R. Waldvogel Institute of Biological and Chemical Systems – Functional Molecular Systems (IBCS-FMS) Hermann-von-Helmholtz-Platz 1, 76344 Eggenstein-Leopoldshafen (Ger-
- [+] These authors contributed equally to this work.
- Supporting information for this article is available on the WWW under https://doi.org/10.1002/asia.202300380
- This manuscript is part of an Indo-German Workshop (IGW) special collec-
- © 2023 The Authors. Chemistry An Asian Journal published by Wiley-VCH GmbH. This is an open access article under the terms of the Creative Commons Attribution License, which permits use, distribution and reproduction in any medium, provided the original work is properly cited.








Figure 1. Recent projects towards automated synthesis platforms.[13,14–16]

Python and the majority were published in an open source fashion. While these projects are all very inspiring and impressive contributions, they unfortunately did not lead to the sought after communalization and widespread availability of automated reaction setups. Our initial contributions to this field were published in 2021; a semi-automated setup for electrolysis using a reservoir with variable electrolysis parameters in flow cells with an automated fraction collector. [19] However, no purging step was possible with this setup and the control was implemented in LabView as a closed source file. In contrast to single setup automation, Aspuru Guzik et al. emphasized the orchestration of multiple setups and the use of AI informed experimental planning. However, they did not report the actual code that controlled the experimental setup itself.[14,16] Cronin et al. elegantly showed the applicability of a formal and general description of chemical experiments in a machine-readable format.[20] As with the previously mentioned work, the focus was not on the communication to the physical devices.[15] More recently, Cambié et al. published their project Flowchem on GitHub; it is the only open source code among those discussed that is actively maintained in 2023. [21,22] Its aim was to expose an interface that makes the devices' functionality available to other software (a RESTful API) for every given device in a setup, wherein the flow of commands in the experiment has to be implemented by the user in a programming language of their choice.[22] This is undeniably valuable to users with programming skills, however, for the uninitiated it can be a roadblock to implementing these systems. While the Chemputer abstracts the setup on an even higher level and Flowchem on a similar level than LABS, many of the above-mentioned projects lack generalizability or abstraction layers in regard to the devices used in the setups. Their accordingly high degree of specialization is a disadvantage for broader application, and adaptation and subsequent application of the experimental setup is often difficult, especially if software components need to be refactored on a major scale. Furthermore, the use of LabView as basis for the software adds upfront cost and the need to familiarize oneself with LabView in order to run and adapt the programs. In our project, we aim to close this gap by providing a free and open source software base for sequence-based control of arbitrary experimental setups. At the same time, we want to provide the user with a modern, centralized user interface where experimental conditions can be planned and running experiments can be monitored. Furthermore, an administrative view allows for the adjustment of general parameters (e.g., a flush step between two different runs or within one run).

Although automated chemistry is of interest to the wider chemical community, it is outside of many traditional chemists' skillset to operate and adapt custom solutions to these software packages. Therefore, we aim to provide a package that enables the facile building of automated systems from the supported devices with negligible prior knowledge needed. Since only the configuration for the experimental setup has to be defined, little to no programming knowledge is required.

Results and Discussion

Our approach to implement the system is based on two components (Figure 2): the web-based user interface provides an interface for the user while the backend deals with the logic for controlling the setups. The backend communicates with the connected devices based on the asynchronous network framework Twisted, and communicates with the user interface in a many-to-one relationship.[23] The use of Twisted for device communication has been pioneered by Ley et al. with their project octopus.[24] This framework is especially well suited, since it is designed to allow a simple implementation of networking protocols independent of the physical connection like RS232 and Ethernet. A setup class organizes the execution of the scheduled experiments and can stop the sequence if necessary. The experiment sequence is provided by the experiment class, which reads the commands to be executed from a configuration file in YAML format. This can be changed by the user, and individual sequences can be combined into larger sequences, so that any intermediate steps, such as purging routines, can be actualized in addition to the actual experiment to be carried out. A device driver class provides generic adapters for communication with the devices, such as pumps or power supply units, which provide device-specific functions.






Figure 2. Representation of the link between the user frontend and the backends at the experimental stations.

The adapter then accesses a driver file that contains the model-specific control commands and overwrites the generic functions of the adapter so that any device can be used, provided they can map the required basic functions. The user interface is a *Flask*-based^[25] Python application that provides a web management platform for users and experimental parameters, as well as monitoring of the various backends. Flask is a popular choice for web development because of its simplicity and flexibility, allowing developers to quickly create scalable and efficient web applications. The rich ecosystem of Flask already provides additional functionalities such as database integration, authentication, API development. It provides a web server so that the user interface can be accessed with common web browsers.

The user interface is designed in a modular fashion, the functionality modules are organized as so-called blueprints, so far there is the module "Users", which provides the functionality to create different users in a database, to authenticate and authorize them to use the user interface. The module "Monitoring" offers an overview page of all integrated backends with their current state, as well as a detailed view, which displays the queued experiments of a selected station in table form. Also displayed there are the observables (which are defined in the configuration file of the station) as line plots (voltage, current, temperature), progress bars (remaining dosing time of the pump, applied charge), or as data tables (positions of the switching valves over time). The module "Experiments" allows the hierarchical management of the experimental data. First, an experimental design is created, which is assigned to a user and

a backend. In this design, individual stages can be created, each of which is linked to a method (the experimental sequence to be carried out) provided by the corresponding backend. For example, additional cleaning steps can be added between experiments. Individual runs can then be created within a stage. The stage overview page offers a down- and upload function (Excel) to add, update, or delete the runs of a stage. Any number of backends can be added on an administrative level. The currently loaded configuration can be downloaded from the stations to the user interface's database. For each method known at the particular device, parameters can be set as static and do not have to be defined for each run (e.g., the volume of purging solvent). This should enhance reproducibility in multiuser scenarios.

The configuration syntax (Figure 3) was designed to implement the following features: multichannel devices can be used as one device with the channel as a parameter in each command or each channel as a separate device. Experiments can contain other experiments to allow high reusage of common processes, like purging a compartment or tube. A full description of the configuration file can be found in the supporting information.

To test the sequence of the methods in a dry run, we created virtual "dummies" of the devices used. These behave like the devices in terms of device communication. The readouts can be counting down times or randomized values, the cell voltage of a current source, etc. As of now, the core functions for the devices we use are programmed into the





Methods Config File Example

```
The "experiments" node contains a list of
  listen port: 11123
                                          subexperiment:
                                                                                                                     experiment names. The elements are the
  log_level: info
                                           conditions:
                                                                                                                     experiments, conditions, abort conditions, parameters, observables (measured values
Application API web port and station
                                                                                                                        the devices that are
log level
                                                                                                                                                 output to the
                                            stopconditions:
                                                                                                                     frontend) and commands.
   devices:
                                           parameters:
                                                                                                                     If a parameter is entered under Parameters
                                              valveport: [int,
                                                                                                                     it can be passed from the frontend. Then
                                                                                         Parameter from
       driver: reglo ico
                                           observables:
                                                                                         user input
                                                                                                                     this value can be referred to in curly
       address: COM1
                                              - [ pump1, remaining_time, int, s ]
                                                                                                                     brackets in an experiment. Alternatively, the
       channel: 1
                                                [ valve1, position, int,
     valve1:
                                                                                                                     be entered directly
                                            commands:
       driver: azura vu
                                                [ valve1, set_position, [ "{valveport}" ], { } ]
       address: 192.168.0.3:10123
                                                                                                                     Observables (measured values from the instruments) entered here will be passed to
                                                [ pump1, dispense, [ ], { rate: '0.1', volume: '0.05' } ]
     psu:
       driver: tdk lambda zplus
                                         example run:
       address: COM2
                                           conditions:
                                                                                                              Stop condition (Here: if voltage readout on device
 List of the names of devices to
                                            1_min_wait: [ TimeCondition, [ 1_min_wait, 60 ], { } ]
                                                                                                              psu is over {voltage_limit} (user defined in frontend,
 connect during startup. The device
                                            stopconditions:
                                                                                                              otherwise defined here as value) stop the run
 name can be used as a reference to
                                              - [ ObservableGreaterOrEqualValueCondition, [ max_voltage, psu, voltage, "{voltage_limit}" ], { } ]
 this device in the different routines
                                            parameters:
 (on the right).
 Each device needs an address and a
                                              amount of charge: [float, C]
                                                                                                            Parameter from
driver file name. If a device supports multiple channels, the device can be
                                            observables:
                                                                                                            user input
                                                [ psu, current, float, A
                                                                                                                                        Static parameter
 added in the following way
                                                  psu, voltage, float, V
                                              - [ psu, amount of charge, float, C ]
    multiChannelDevice1:
                                            commands:
                                                  subexperiment, { valveport: 2 } 1
                                                  psu, wait, [ 1_min_wait ], { } ]
                                              - [ psu, output constant_current, [ ], { amount_of_charge: "{amount_of_charge}}", current: 0.2 } ]
                                                                Wait condition (Here: Flooding
      channel: 2
                                                                the cell prior to electrolysis)
```

Figure 3. Configuration file example. Left: General settings and device configuration; demonstrated features: channel of a multichannel device is treated as a single device, COM- and TCP/IP ports can be used. Top right: First experiment named *subexperiment* (demonstrated features: experiment has a variable parameter named *valveport*). Bottom right: Second experiment named *example_run* (demonstrated features: condition 1_min_wait used in wait command, stop-condition utilized to cancel experiment if voltage is above configurable threshold, parameter valve port handed to *subexperiment*).

dummies. The function for generating random numbers can be coupled with a user-defined function that simulates, for example, a time-dependent increase in voltage. In this way, more complex stop conditions for the automated run observation can be tested efficiently prior to the actual experiment. For the devices used in this project we provide virtual devices in a simple framework. We employed this to test changes in the configuration before running it in the laboratory. This "virtual experimental setup" can also be integrated into the user interface to control it. This facilitates the further development of the user interface, as no real devices are needed to test new features.

Due to our experience with studying dehydrogenative C–C coupling of and with phenols, we chose the anodic oxidative homocoupling reaction of 2,4-dimethylphenol as an example for the first test runs. [19,26,27,28] This electrochemical transformation to the corresponding 2,2'-biphenol is particularly challenging as it is very prone to side reactions, such as the formation of polycyclic by-products. [28,29] Therefore, our first experimental setup (Figure 4) consisted of a dead-end selector switching valve (*Knauer*) to which 2,4-dimethylphenol, pyridine, hexafluoro-isopropanol (HFIP), and isopropanol reservoirs were connected. The common outlet connects to one channel of a four-channel Reglo ICC peristaltic pump with independent channel control (*Ismatec*) into a mixing chamber, in our case a 50 mL *Schott* glass screw-cap bottle with a lid with a plug-in insert for four 1/8" OD tubes. An additional hole was drilled in

the center of the lid to allow flushing with compressed air. The compressed air was set to 1.5 bar using a pressure regulator and fed in via a switchable normally closed solenoid valve. An Arduino Uno microcontroller (Arduino) opens and closes the valve via a relay. The reservoir was clamped at a slight angle to enable full drainage via the second channel of the pump. This channel leads to a flow electrolysis cell with a 2×6 cm² anode made of boron-doped diamond on silica and a stainless-steel cathode block with temperature control. A spacer made of PTFE adjusted the electrode gap to 0.25 mm.^[9] The outlet of the flow cell was connected to the common port of another switching valve, from where the flow can be directed back to the reservoir, to a waste canister, or to a fraction collector. The fraction collector used was a low-cost Lambda Omnicoll fraction collector (Lambda Instruments), which can be operated with different sized collection vessels.

In previous studies on electrochemical conversions in flow with active electrode areas of 2×6 cm² peristaltic pumps were typically used. These offer the possibility to work in divided cells with separate channels for anolyte and catholyte solutions, or to operate several cells in parallel with one pump. In the past, we used PharMed® BPT tubing as peristaltic pump tubing, which has a high mechanical stability but only moderate chemical stability. We therefore also tested peristaltic tubing made of fluororubber (FPM, Viton® *Optubus*). However, these were not mechanically stable in the original cassettes that came with the pumps and ruptured within the first few minutes.


1861471x, 2023, 14, Downloaded from https://onlinelbrary.wiley.com/doi/10.1002/asia.202300380 by Karlsruher Institut F. Wiley Online Library on [13/02/2024]. See the Terms and Conditions (https://onlinelibrary.wiley.com/rems-and-conditions) on Wiley Online Library for rules of use; OA articles are governed by the applicable Creative Commons License

Figure 4. Left: Scheme of the automated flow setup enabling continuous and cyclic electrolysis. Right: Picture of the setup in use, the components are labelled. Electrolyte flow goes from right to left. Picture on the right by J. Schneider.

Consequently, we equipped the pump with PharMed® BPT tubing for the initial experiments (Scheme 1).

We started with a test series of the same electrolysis conditions to investigate the overall behavior of the system in terms of yield (Conditions: j=10 mA/cm², Q=0.9 F, v=1.99 mL/min (15×Single pass rate), C(1)=1.25 mol/L, vol.%(Pyridine)=5%, T=25 °C). This included a purge phase in which the system was purged with acetone then primed with HFIP. The electrolyte was then prepared in the reservoir, electrolyzed in cyclic flow, and collected by the fraction collector.

By starting with a new, pristine BDD anode, we observed a sharp increase in yield in the first three experiments (red area, Figure 5). We attribute this to the known change in surface termination from H-type to OH-termination. [30] Runs 7, 11, and 15 were each the first run after the cell was idle overnight during the initial test phase where the interface was open to the air. From previous studies, deposit formation (fouling) is known to occur while electrolysis, and its modification by oxidation may contribute to the observed effect.[31] The system subsequently behaved in a largely stable manner over 15 electrolysis reactions with respect to the yield of coupling product (green area). We equipped the pump with new peristaltic tubing and began submission of all runs necessary for the factorial design Table 1) in sections mixed with the control experiments (orange area). The system initially continued to behave stable in terms of yield, but a steady downward trend was observed after control experiment 30. The anode did not show any deposit formation or fouling during a



Scheme 1. Anodic dehydrogenative C–C homocoupling reaction of 2,4-dimethylphenol (1) to 3,3',5,5'-Tetramethyl-2,2'-biphenol (2).




Figure 5. Yield of the control experiment with increasing number of experiments of the automation system. The annotations at the markers indicate the actual run number of the data point. In the green highlighted area, the yields of the consecutive control experiments are plotted. In the brown highlighted area, the yields of the control experiments that were performed between the experiments for the DoE design are plotted. The number next to the markers indicates the total consecutive experiment number associated with the data point.

visual inspection, and it was concluded that the drift could be explained by the deterioration of the peristaltic tubing.

Since our default peristaltic tubing choice was not suitable for continuous operation, we equipped the peristaltic pump with a new cassette that provided control over the contact pressure of the tubing. This allowed us to use the fluorocarbon rubber (FPM) for a longer operating time. We repeated the experimental design but halved the concentration range from 0.75–1.25 mol/L to 0.38–0.63 mol/L. This forced us to reduce the flow multiplier in the high setting from 15 to 10 times, as the



Table 1. Fractional factorial 2⁶⁻² design for the investigation of the homocoupling of 2,4-dimethylphenol. Each experiment was performed in duplicate, with a total of two central points.

Parameter	Unit	Low level	High level
Current density	mA/cm ²	20	40
Applied charge	F	0.70	0.90
Flowrate multiplier ^[a]	-	5	15
Concentration of 2,4-dimethylphenol	mol/L	0.75	1.25
Concentration of pyridine	vol.%	5	10
Temperature	°C	10	25

[a] We calculated the flowrate needed for a single pass electrolysis and multiplied it by this parameter value.

flow rates were otherwise greater than the pumps could handle. The 32 experiments of the design were divided into 4 blocks of eight experiments each. Three central point experiments were performed before each block. For this, a new tube was inserted into each pump channel. The dosing tube was calibrated to 1 mL/min, and the tube at the flow cell was calibrated to 6 mL/min. After each block of 11 experiments, the flow rate was determined, the tubing was changed and recalibrated. Since flow rate deviations to lower but also higher flow rates were determined, we determined a correction factor for all experiments. This was determined from a linear fit between the actual flow rate before and after each block of experiments (See SI for more information). The concentrations of 1 and the yields of 2 were corrected using this factor. In addition, the deviation of the dosing tube was included as a covariate in the regression and analysis of variance.

No parameter could be excluded by ANOVA from the obtained regression model ($R^2 = 74\%$) with $\alpha = 10\%$, however, current density, charge quantity, and the flow rate multiplier are below the significance threshold. However, these are involved in two-parameter interactions. The model obtained, subject to the corrections discussed, shows values in agreement with our previous studies with respect to the main effects and the curvature analysis (Figure 6).




Figure 6. Main effects plot for the yield of 2. The yield was determined using GC-FID with n-octylbenzene as internal reference material.

The optimum of the applied charge is equal to the central point of the model at 0.8 *F*. Higher concentrations of 1 favor the dehydrodimerization reaction, in agreement with the previously performed reactions at 1.25 mol/l of phenol 1. The positive effect of the higher temperature could be an artifact since there was not sufficient condensation of the solvent in the reservoir. Turbulence in the tube with the cathodically generated hydrogen facilitates the evaporation of the solvent.

We have refrained from expanding the model to a CCD due to the poor suitability of the pump. Nonetheless, automation with LABS has worked successfully. In our next project, we want to implement a driver for piston pumps to replace the peristaltic pumps.

Conclusion

We have developed and successfully applied LABS, a Pythonbased software for automating laboratory setups. It can simultaneously control any number of devices, to run any combination of supported devices, in this work an electrochemical flow setup capable of single-pass and cyclic electrolysis. The focus of this project goes beyond the previously published frameworks. With only the configuration of the setup and experiments left to the user, we enable scientists with less programming experience to get in touch with automated synthesis. A centralized user front-end allows the submission of experiments and the monitoring of any number of experimental stations. The ability to implement stop conditions allows for safe, unattended operation. The open programming allows the user to customize LABS to use additional sensors or measurement devices for their specific use case to enable more complex feedback loops. The program code for front and backend has been published as open source on GitHub.[31-33] The continuous operation has enabled us to increase our sample throughput for electrosynthetic screening in flow cells enormously.

This project lays the foundation for future implementation of advanced autonomous systems, including methods like Bayesian optimization. As such we will base upcoming projects on LABS and gradually support more devices not limited to but showcased in electrochemical setups. The use of an API interface for communication between the two modules also allows the control of third stations, provided that they offer corresponding API interfaces. Our main goal is to make LABS easier accessible and widely applicable in many areas of chemistry. Enabling future releases of LABS to be compatible with $Cronins\ \chi DL^{[20,34]}$ for example, would vastly improve the accessibility of LABS and the reusability of procedures run by LABS driven setups.

Data and materials availability

Full details of the experimental methods and equipment used are given in the electronic supporting information. The program code for the controller was published in three separate





repositories: LABS-Backend, [32] LABS-Frontend, [31] and LABS-DeviceDummies. [33]

Acknowledgements

Funded by the Deutsche Forschungsgemeinschaft (DFG, German Research Foundation) under FOR 2982/1–UNODE (WA1276/24-1). Support by Forschungsinitiative Rheinland-Pfalz within the frame of SusInnoScience was very helpful. We would like to express our sincere thanks to R. M. Hamill for the effort she put into improving the language of this manuscript. We would like to thank F. Weyer for fruitful discussions regarding programming. Open Access funding enabled and organized by Projekt DEAL.

Conflict of Interests

The authors declare no conflict of interest.

Data Availability Statement

The data that support the findings of this study are openly available in LABS-Backend: Beta-Publication_submission at https://doi.org/10.5281/zenodo.7879589, reference number 7879589, Labs-Frontend: Beta-Publication submission at https://doi.org/10.5281/zenodo.7879586, reference number: 7879586, LABS-DeviceDummys at https://doi.org/10.5281/zenodo.7879590, reference number 7879590.

Keywords: electrochemistry • automation • anodic oxidation • cheminformatics • phenol coupling reaction

- a) P. Anastas, N. Eghbali, Chem. Soc. Rev. 2010, 39, 301–312; b) B. A. Frontana-Uribe, R. D. Little, J. G. Ibanez, A. Palma, R. Vasquez-Medrano, Green Chem. 2010, 12, 2099.
- [2] a) A.-K. Seitz, P. J. Kohlpaintner, T. van Lingen, M. Dyga, F. Sprang, M. Zirbes, S. R. Waldvogel, L. J. Gooßen, Angew. Chem. Int. Ed. 2022, 61, e202117563; Angew. Chem. 2022, 134, e202117563; b) N. Kurig, J. Meyers, F. J. Holzhäuser, S. Palkovits, R. Palkovits, ACS Sustainable Chem. Eng. 2021, 9, 1229–1234; c) S. Arndt, D. Weis, K. Donsbach, S. R. Waldvogel, Angew. Chem. Int. Ed. 2020, 59, 8036–8041; Angew. Chem. 2020, 132, 8112–8118; d) D. Pollok, S. R. Waldvogel, Chem. Sci. 2020, 11, 12386–12400; e) S. Möhle, M. Zirbes, E. Rodrigo, T. Gieshoff, A. Wiebe, S. R. Waldvogel, Angew. Chem. Int. Ed. 2018, 57, 6018–6041; Angew. Chem. 2018, 130, 6124–6149; f) S. R. Waldvogel, B. Janza, Angew. Chem. Int. Ed. 2014, 53, 7122–7123; Angew. Chem. 2014, 126, 7248–7249.
- [3] a) J. L. Röckl, D. Pollok, R. Franke, S. R. Waldvogel, Acc. Chem. Res. 2020, 53, 45–61; b) J. L. Röckl, D. Schollmeyer, R. Franke, S. R. Waldvogel, Angew. Chem. Int. Ed. 2020, 59, 315–319; Angew. Chem. 2020, 132, 323–327; c) S. R. Waldvogel, S. Lips, M. Selt, B. Riehl, C. J. Kampf, Chem. Rev. 2018, 118, 6706–6765; d) A. Wiebe, S. Lips, D. Schollmeyer, R. Franke, S. R. Waldvogel, Angew. Chem. Int. Ed. 2017, 56, 14727–14731; Angew. Chem. 2017, 129, 14920–14925; e) E. J. Horn, B. R. Rosen, P. S. Baran, ACS Cent. Sci. 2016, 2, 302–308.
- [4] M. Yan, Y. Kawamata, P. S. Baran, Chem. Rev. 2017, 117, 13230–13319.
- [5] A. Wiebe, T. Gieshoff, S. Möhle, E. Rodrigo, M. Zirbes, S. R. Waldvogel, Angew. Chem. Int. Ed. 2018, 57, 5594–5619; Angew. Chem. 2018, 130, 5694–5721.
- [6] D. Pollok, B. Gleede, A. Stenglein, S. R. Waldvogel, Aldrichimica Acta 2021, 54, 3–15.

- [7] a) N. Amri, T. Wirth, J. Org. Chem. 2021, 86, 15961–15972; b) N. Amri, T. Wirth, Synlett 2020, 31, 1894–1898.
- [8] a) E. Sato, M. Fujii, H. Tanaka, K. Mitsudo, M. Kondo, S. Takizawa, H. Sasai, T. Washio, K. Ishikawa, S. Suga, J. Org. Chem. 2021, 86, 16035–16044; b) M. Elsherbini, T. Wirth, Acc. Chem. Res. 2019, 52, 3287–3296.
- [9] C. Gütz, A. Stenglein, S. R. Waldvogel, Org. Process Res. Dev. 2017, 21, 771–778.
- [10] A. G. Wills, S. Charvet, C. Battilocchio, C. C. Scarborough, K. M. P. Wheelhouse, D. L. Poole, N. Carson, J. C. Vantourout, Org. Process Res. Dev. 2021, 25, 2587–2600.
- [11] N. Kurig, J. Meyers, E. Richter, S. Palkovits, R. Palkovits, Chem. Ing. Tech. 2022, 94, 786–790.
- [12] a) E. Sato, G. Tachiwaki, M. Fujii, K. Mitsudo, T. Washio, S. Takizawa, S. Suga, Org. Process Res. Dev. 2023; b) M. Dörr, M. M. Hielscher, J. Proppe, S. R. Waldvogel, ChemElectroChem 2021, 8, 2621–2629; c) M. M. Hielscher, B. Gleede, S. R. Waldvogel, Electrochim. Acta 2021, 368, 137420; d) M. Dörr, J. L. Röckl, J. Rein, D. Schollmeyer, S. R. Waldvogel, Chem. Eur. J. 2020, 26, 10195–10198; e) M. Santi, J. Seitz, R. Cicala, T. Hardwick, N. Ahmed, T. Wirth, Chem. Eur. J. 2019, 25, 16230–16235; f) R. Möckel, J. Hille, E. Winterling, S. Weidemüller, T. M. Faber, G. Hilt, Angew. Chem. Int. Ed. 2018, 57, 442–445; Angew. Chem. 2018, 130, 450–454.
- [13] a) Z. Crandall, K. Basemann, L. Qi, T. L. Windus, React. Chem. Eng. 2022, 7, 416–428; b) C. J. van der Westhuizen, J. Du Toit, N. Neyt, D. Riley, J.-L. Panayides, Digital Discovery 2022, 1, 596–604; c) S. Chatterjee, M. Guidi, P. H. Seeberger, K. Gilmore, Nature 2020, 579, 379–384; d) N. Cherkasov, Y. Bai, A. J. Expósito, E. V. Rebrov, React. Chem. Eng. 2018, 3, 769–780; e) D. E. Fitzpatrick, C. Battilocchio, S. V. Ley, Org. Process Res. Dev. 2016, 20, 386–394.
- [14] L. M. Roch, F. Häse, C. Kreisbeck, T. Tamayo-Mendoza, L. P. E. Yunker, J. E. Hein, A. Aspuru-Guzik, Sci. Robot. 2018, 3.
- [15] S. Steiner, J. Wolf, S. Glatzel, A. Andreou, J. M. Granda, G. Keenan, T. Hinkley, G. Aragon-Camarasa, P. J. Kitson, D. Angelone et al., *Science* 2019, 363.
- [16] L. M. Roch, F. Häse, C. Kreisbeck, T. Tamayo-Mendoza, L. P. E. Yunker, J. E. Hein, A. Aspuru-Guzik, PLoS One 2020, 15, e0229862.
- [17] T. Wirtanen, T. Prenzel, J.-P. Tessonnier, S. R. Waldvogel, Chem. Rev. 2021, 121, 10241–10270.
- [18] M. Seifrid, R. Pollice, A. Aguilar-Granda, Z. Morgan Chan, K. Hotta, C. T. Ser, J. Vestfrid, T. C. Wu, A. Aspuru-Guzik, Acc. Chem. Res. 2022, 55, 2454–2466.
- [19] M. M. Hielscher, B. Gleede, S. R. Waldvogel, *Electrochim. Acta* 2021, 368, 137420.
- [20] A. J. S. Hammer, A. I. Leonov, N. L. Bell, L. Cronin, JACS Au 2021, 1, 1572– 1587
- [21] a) GitHub, "ChemputerSoftware/client at master-croningp/Chemputer-Software", can be found under https://github.com/croningp/ChemputerSoftware, 2023; b) GitHub, "GitHub aspuru-guzik-group/ChemOS", can be found under https://github.com/aspuru-guzik-group/ChemOS, 2023; c) GitHub, "GitHub RxnRover/RxnRover", can be found under https://github.com/RxnRover/RxnRover, 2023; d) SourceForge, "Open-FlowChem", can be found under https://sourceforge.net/projects/open-flowchem/, 2023
- [22] GitHub, "GitHub cambiegroup/flowchem: Flowchem is an application to simplify the control of instruments and devices commonly found in chemistry labs", can be found under https://github.com/cambiegroup/ flowchem, 2023.
- [23] "Twisted", can be found under https://twisted.org/, 2023.
- [24] a) GitHub, "richardingham/octopus: Real-time laboratory automation and monitoring in Python", can be found under https://github.com/ richardingham/octopus, 2023; b) R. J. Ingham, C. Battilocchio, J. M. Hawkins, S. V. Ley, Beilstein J. Org. Chem. 2014, 10, 641–652.
- [25] "Welcome to Flask Flask Documentation (2.2.x)", can be found under https://flask.palletsprojects.com/en/2.2.x/, 2023.
- [26] a) M. Selt, S. Mentizi, D. Schollmeyer, R. Franke, S. R. Waldvogel, Synlett 2019, 30, 2062–2067; b) B. Riehl, K. Dyballa, R. Franke, S. Waldvogel, Synthesis 2016, 49, 252–259.
- [27] M. Selt, R. Franke, S. R. Waldvogel, Org. Process Res. Dev. 2020, 24, 2347– 2355.
- [28] I. M. Malkowsky, U. Griesbach, H. Pütter, S. R. Waldvogel, Eur. J. Org. Chem. 2006, 2006, 4569–4572.
- [29] a) J. Barjau, G. Schnakenburg, S. R. Waldvogel, Angew. Chem. Int. Ed. 2011, 50, 1415–1419; Angew. Chem. 2011, 123, 1451–1455; b) I. M. Malkowsky, C. E. Rommel, K. Wedeking, R. Fröhlich, K. Bergander, M. Nieger, C. Quaiser, U. Griesbach, H. Pütter, S. R. Waldvogel, Eur. J. Org. Chem. 2006, 2006, 241–245.





186141x, 2023, 14, Downloaded from https://onlinelbtrary.wiley.com/doi/10.1002/asia.202300380 by Karlscher Institut F., Wiley Online Library on [13/02/2024]. See the Terms and Conditions (https://onlinelibrary.wiley.com/terms-and-conditions) on Wiley Online Library for rules of use; OA articles are governed by the applicable Creative Commons License

- [30] a) S. Palmas, A. M. Polcaro, A. Vacca, M. Mascia, F. Ferrara, J. Appl. Electrochem. 2006, 37, 63–70; b) I. Duo, C. Levy-Clement, A. Fujishima, C. Comninellis, J. Appl. Electrochem. 2004, 34, 935–943.
- [31] J. Fangmeyer, A. Behrens, B. Gleede, S. R. Waldvogel, U. Karst, Angew. Chem. Int. Ed. 2020, 59, 20428–20433; Angew. Chem. 2020, 46, 20608– 20613.
- [32] M. M. Hielscher, M. Dörr, LABS-Frontend 2023, DOI 10.5281/ZENO-DO.7879586.
- [33] M. Dörr, M. M. Hielscher, LABS-Backend 2023, DOI 10.5281/ZENO-DO.7879588.
- [34] M. Dörr, M.M. Hielscher, LABS-DeviceDummys 2023, DOI 10.5281/ ZENODO.7879590.
- [35] "Cronin Group/Chemputer/XDL GitLab", can be found under https://gitlab.com/croningroup/chemputer/xdl, 2023.

Manuscript received: April 30, 2023 Revised manuscript received: May 27, 2023 Accepted manuscript online: June 3, 2023 Version of record online: June 12, 2023