


## The WMO GAW-VOC Net-Work in 2010



**Rainer Steinbrecher**  
And  
**Stephan Thiel**

Helmig, D., Bottenheim J., Galbally I.E., Lewis A., Milton M., Penkett S., Plass-Duelmer C., Read K. Reimann S., Steinbrecher R., Tans P., Thiel S. (2009): The WMO-GAW Volatile Organic Compound Program *Eos Trans. AGU*, 90(52), 513–514.

<http://imk-ifu.fzk.de/wcc-voc/>


Source: WRI

## VOC – volatile organic compounds

- Impact tropospheric chemistry
- Contribute in the presence of  $\text{NO}_x$  to tropospheric ozone-/particle formation
- Scavenge the OH radical, the most important atmospheric cleaning agent



By increasing tropospheric ozone / particle levels and the life-time of other climate relevant trace gases, e.g methane, VOC force climate change.



Target compound in the GAW network (WMO-Report 172, 2007)

# GAW Network for VOC

## Current status:

- Global coverage only achieved for NMHC based on the NOAA-GMD Glass Flask Sampling Program

## Principle Set-Up for NMHC-Analysis in Air Samples

- Cryogenically:
- L = 300 mm, i.d. = 1.5 mm
  - Silcosteel-tube; no filling
  - at -30°C

**Dryers:**

Sample Vol.: 400 ml  
Sample Flow: 100 ml/min


**Detectors**

Alternatively (mainly for checks):  
Mg(ClO<sub>4</sub>)<sub>2</sub>

Temp. Progr. (\*):

- 2.5 min isothermal
- 3.5 K/min to -13°C
- 20 K/min to 8°C
- 5 K/min to 70°C
- 10 K/min to 240°C
- 12.2 min isothermal

50 min



**Separation**

Varian 3600CX


Column B:  
CP-SIL 5 CB  
L = 50 m, i.d. 0.25mm, df 1 µm

Column A:  
- PLOT column Gaspro GSC  
- L = 17 m, i.d. 0.32 mm  
• Restriction capillary:  
• L = 1 m; i.d. 0.1 mm, deactivated


Up to 8 canisters:

**Sample pre-concentration trap (SPT; Varian)**

- 6 samples,
- 1 NCAR-canister,
- 1 calibration canister  
- (n-butane/benzene)
- L = 8 cm, i.d. = 2.1 cm
- Carboxipack® BHT
- at -120°C (liquid N<sub>2</sub>)
- Desorption at 200°C
- sample is back-flushed
- sample transferred split-less to columns.



# GAW Network for VOC




(Helmig et al., 2009)

## Current status:

- Global coverage only achieved for NMHC based on the NOAA-GMD Glass Flask Sampling Program

### Components: Integrated Atmospheric Observations System





## The WCC-VOC within GAW




**Global Atmosphere Watch (GAW)**

### GAW Central Facilities:



## Traceability of Calibrations and Audits





## Round Robin Exercises

- Evaluate Results on the basis of data quality objectives
- Report findings to the participants
- Enquire reasons for deviations in bilateral meetings
- Suggest joint measures to improve quality
- Check progress by repeating QA/QC experiments



## Audits

- Report discovered discrepancies to station staff
- Take possibilities to solve detected problems on-site
- Define an action list in the final audit meeting with station staff to timely solve encountered problems.
- Check progress by repeating audit

# GAW-VOC Targets

Ethane	Acetone
Propane	DMS
Acetylene	Benzene
Isoprene	Toluene
Formaldehyde	Iso-Butane
Monoterpenes	n-Butane
Acetonitrile	Iso-Pentane
Methanol	n-Pentane
Ethanol	



a large number of individual species should be measured:


- nonmethane hydrocarbons (NMHC),

- monoterpenes (MTs)  
e.g.  $\alpha$ -pinene, limonene

- oxyVOCs

- dimethylsulfid (DMS)

- acetonitril (ACT)



**GAW-Scale of standards**

WMO Report 171; 2007



## *Task*

**Establishment of a Central Calibration Laboratory (CCL) for VOCs for the WMO Global Atmosphere Watch (GAW) network.**

## *Problem*

**Due to the large number of compounds involved, the task exceeds the capacities of a single laboratory (institution).**

## *Solution*

**The responsibilities for the individual compound are shared among several laboratories (institutions) and four National Metrology Institutes (NMIs) are working together to form the CCL.**



**In cooperation with BIPM and the CCQM Gas Analysis Working Group (GAWG) a concept for the future CCL for VOCs has been set up and is being implemented.**

# VOC Central Calibration Laboratory (CCL)

Ethane	Acetone
Propane	DMS
Acetylene	Benzene
Isoprene	Toluene
Formaldehyde	Iso-Butane
Monoterpenes	n-Butane
Acetonitrile	Iso-Pentane
Methanol	n-Pentane
Ethanol	

## Responsibilities (Status 2010)

NMHC



NPL (National Physical Laboratory, GB)

MTs



NIST (National Institute of Standard and Technology, USA)

DMS, ACT



KRISs (Korea Research Institute of Standards and Science, South Korea)



oxyVOC

VSL (Dutch Metrological Institute, NL)

➤ **Develop quality control procedures.**

➤ **Ensure traceability of standards used in the WCC-VOC.**

➤ **Conduct performance and system audits at stations.**

➤ **Perform round-robin experiments (inter-comparisons).**


➤ **Support a network-wide quality review.**

➤ **Provide training and long-term technical consulting to station scientists and technicians (e.g. through the GAW Training and Education Centre GAWTEC).**

# Tracability of the WCC-VOC



Participating in EUROMET-886 VOC inter-comparison of the CIPM-CCQM Gas Analysis Working Group



# NMHC Standards in WCC QA/QC Measures



## WCC-NMHC **Secondary** and **Laboratory-/Working** standards (Status 2010)

Compound	GAW/ppb	uncertainty 2σ/ppb	Apel/Riemer /ppb	uncertainty 2σ/ppb	Ambient air/ppb	uncertainty 2σ/ppb
Ethane	2.7	0.05	13.51	0.58	1.25	0.05
Ethine	2.66	0.05	7.55	0.33	1.02	0.05
Propane	2.67	0.05	12.13	0.53	0.53	0.04
i-Butane	2.68	0.05	5.97	0.47	0.49	0.09
n-Butane	2.6	0.05	11.11	0.98	1.17	0.12
i-Pentane	2.59	0.05	7.79	0.32	1.72	0.08
n-Pentane	2.63	0.05	9.35	0.39	0.47	0.05
Isoprene	2.6	0.05	5.34	0.23	n.r.	n.r.
Benzene	2.62	0.05	2.26	0.16	0.36	0.03
Toluene	2.59	0.05	3.52	0.41	0.74	0.08

## GAW Stations and VOC Central Facilities (Status 2003)

- Representing GAW, EMEP, CAPMoN and LBA environment monitoring programs
- 7 countries (Brazil, Canada (2 labs), Czech Republic, Finland, Germany (2 labs; 3 instruments), Ireland, and Slovakia)
- 9 different stations/laboratories
- 10 different instruments (off-line and on-line)

# Round-Robin Exercises and Audits

## Results (Status 2003)

VOC	Participants									
	A	B	C	D	E	F	G	H	I	J
ethane	3.8	-64.2	-3.4	-1.2			-2.1	-4.4	-1.1	
ethylene	5.2	-73.5	5.2	7.3			-16.5	-7.5	-2.3	
acetylene	-13.8	-54.7		4.0				-22.4	-25.1	
propane	9.6	1.0	-0.1	2.1	-27.6	-3.0	-2.7	-9.1	-1.0	
propylene	8.0	-1.9	5.4	11.5	-66.0	2.3	-15.7	-2.3	1.1	
i-butane	9.4	81.5	8.4	11.0	-33.7	-10.1	1.8	-4.3	2.5	8.1
n-butane	6.7	78.8	-0.2	5.7	-30.6	-9.0	-2.6	-3.5	-0.7	
1-butene	4.2	137.3	6.4	11.2					3.2	
t-2-butene	3.4	47.0	7.5	9.2		-18.1			-2.8	12.7
i-butene	6.9		2.7	10.3				2.3	-3.8	
c-2-butene	-2.4		1.8	4.3		-7.6		-12.7	-7.4	3.2
i-pentane	-14.3		-2.3	5.2	-42.9		-5.8	-12.1		-2.5
n-pentane	-26.3		-1.5	3.7		-0.2	-0.6			
isoprene	5.5	-98.0	-6.2	-17.0	-78.8	-16.7	-2.5	5.2	1.4	-1.2
t-2-pentene	-52.2	-22.9	92.3	6.6		-6.8	-4.0	-29.3	-11.6	23.8
c-2-pentene	-20.8	19.8	3.6	-0.6		-6.4	-1.7	-12.2	-5.4	1.0
2-me pentane			5.9			-17.2	1.2			
3-me-pentane			0.7	13.3		-7.9	-4.6			2.3
n-hexane	-27.3	236.6	-3.0	4.3	-40.8	-11.5	0.2	-30.7	-12.8	-0.6
benzene	6.3	208.2	-5.5	2.9		0.5	-0.4	-14.7	-5.2	
cyclohexane	51.4									-22.7
n-heptane	5.1		-6.5	4.3	-45.0	6.7	3.5		-1.4	-0.5
toluene	27.2	-79.8	-5.1	10.1	-22.2	16.6		-6.5		10.1
et-benzene			1.1	-3.0	75.5	1.1		-21.1	3.0	5.3
m,p-xylene			4.1	-2.4	34.0	-14.3		-16.3	7.4	0.4
o-xylene		1529.6	5.1		228.2	-22.9		-28.5		
1,3,5 trime-benzene			-29.0							-16.6
1,2,4 trime-benzene			30.0							0.8

➤ **Approx. only 50% of the labs perform reasonably well.**

*Deviation in % from the WCC-VOC reference values (Standard CC154935)*

**Results that did not meet the DQOs are shown in red**

Rappengluck B., Apel E., Bauerfeind M, Bottenheim J., Brickell P., Cavolka P., CechJ., Gatti L., Hakola H., Honzak J., Junek R., Martin D., Noone C., Plass-Dulmer Ch., Travers D., Wang D. (2006): The first VOC intercomparison exercise within the Global Atmosphere Watch (GAW), Atmospheric Environment, 40, 7508-7527,

## Results (Status 2003)

**Sometimes large differences (up to a factor of two) from the target mole fractions became obvious.**



**Strong efforts are needed to harmonise VOC measurements in environmental monitoring networks.**

***Key topic identified:***

**Harmonise the calibration standards.**



## GAW Stations and VOC Central Facilities: (Status 2010)

- Global (*in situ*): Jungfrauoch, Hohenpeißenberg, Cap Verde
- Regional: (*in situ*) Rigi, Egbert
- Central Facilities (*flasks*): Analysis, sampling and transport/storage of air samples
  - Institute of Alpine and Arctic Research (INSTAAR)  
(Global Monitoring Division (GMD) network, National Oceanic and Atmospheric Administration (NOAA), Boulder CO, USA)
  - Environmental Science and Technology Centre,  
Environment Canada, Ottawa, Canada (global station Alert)
  - Max Plank Institute for Chemistry Mainz  
(CARIBIC Aircraft Atmospheric Monitoring Program)

# Round-Robin Exercises and Audits

## Results (Status 2010)

Compound	A	B	C	D	E	F	G	H
Ethane	0.37	-0.78	-	-2.21	0.37	-1.36	-	6.57
Ethine	-1.13	-1.47	-	-14.98	-	-	-	6.07
Propane	0.00	-0.20	-12.72	-7.38	-0.37	-0.48	-	5.71
i-Butane	0.00	-0.61	-8.61	-2.32	0.00	-0.86	-12.64	5.64
n-Butane	0.00	-0.68	-8.77	-4.28	3.47	-	-6.82	5.37
i-Pentane	-0.77	-0.38	-9.98	-11.62	3.09	-0.54	-3.45	4.58
n-Pentane	-1.14	-0.54	-9.08	-2.70	0.57	-0.64	-11.54	4.52
Isoprene	-0.77	-0.51	27.45	-3.29	-6.73	0.10	-	-1.93
Benzene	0.38	-0.72	20.38	-0.85	-0.94	0.32	-8.98	1.67
Toluene	-8.11	-0.81	31.61	-2.84	-1.74	-1.28	-3.57	-0.32

- Not reported

***Deviation in % from the WCC-VOC reference values (Standard D296263)***

- Within Data Quality Objective
- Outside Data Quality Objective
- near Data Quality Objective

- Finalise the setup of the CCL for VOC
- QA/QC missions to the GAW global stations Pallas Finland (PAL) and Cape Grim, Australia (CGO).
- Intra-laboratory QA/QC procedures for NMHC analysis of the WCC-VOC
- Setup of an analysis system for oxyVOC
- **Announcement:**  
Joint GAW-VOC/CCQM-GAWG Workshop in Helsinki, Finland, June 29-30, 2010.

- **Recent inter-comparisons and audits in the GAW-VOC network on NMHC show good results but there is space for improvements.**
- **Establishing of the CCL for VOC in co-operation with international NMIs, BIPM and GAW-VOC is on an excellent way and should further be promoted.**
- **Further information about WCC-VOC activities are available on the web.**
- **The next step of QA/QC measures in the GAW-VOC network is in focus (other VOC).**

# Thank you for your attention and the



# for funding and



# thanks to all GAW stations people for their excellent co-operation

