

**KERNFORSCHUNGSZENTRUM  
KARLSRUHE**


Juli 1975

KFK 1905/II  
EUR 5204 e

Projekt Spaltstoffflußkontrolle

**The Interlaboratory Experiment IDA-72  
on Mass Spectrometric Isotope Dilution Analysis  
Volume II**

Edited by  
W. Beyrich, E. Drosselmeyer



**GESELLSCHAFT  
FÜR  
KERNFORSCHUNG M.B.H.**

**KARLSRUHE**

Als Manuskript vervielfältigt

Für diesen Bericht behalten wir uns alle Rechte vor

GESELLSCHAFT FÜR KERNFORSCHUNG M. B. H.  
KARLSRUHE

KERNFORSCHUNGSZENTRUM KARLSRUHE

KFK 1905/II

EUR 5204e

Projekt Spaltstoffflußkontrolle

The Interlaboratory Experiment IDA-72  
on Mass Spectrometric Isotope Dilution Analysis

Volume II


Edited by

W. Beyrich \* and E. Drosselmeyer \*\*

---

\* EURATOM, delegated to the Institute for Applied Systems Analysis and Reactor Physics (IASR), Gesellschaft für Kernforschung (GfK), Karlsruhe, Germany

\*\* IASR, GfK, Karlsruhe, Germany



The Interlaboratory Experiment IDA-72  
on Mass Spectrometric Isotope Dilution Analysis

Volume II of the report on the IDA-72 experiment contains papers written by different authors on a number of special topics connected with the preparation, performance and evaluation of the interlaboratory test. In detail the sampling procedures for active samples of the reprocessing plant and the preparation of inactive reference and spike solution from standard material are described as well as new methods of sample conditioning by evaporation.

An extra chapter is devoted to the chemical sample treatment as a preparation for mass spectrometric analysis of the U and Pu content of the solutions. Special topics are also methods for mass discrimination corrections,  $\alpha$ -spectrometer measurements as a supplement for the determination of Pu-238 and the comparison of concentration determinations by mass spectrometric isotope dilution analysis with those performed by X-ray fluorescence spectrometry.

The last part of this volume contains papers connected with the computerized statistical evaluation of the high number of data.

Der Interlaboratoriumstest IDA-72 zur  
massenspektrometrischen Isotopenverdünnungsanalyse

Band II des Berichtes über das IDA-72-Experiment enthält Beiträge von verschiedenen Autoren zu einer Anzahl besonderer Themen im Zusammenhang mit der Vorbereitung, Durchführung und Auswertung des Interlaboratoriumstests. Im einzelnen sind sowohl das Probenahmeverfahren für die aktiven Proben aus der Wiederaufarbeitungsanlage und die Vorbereitung inaktiver Referenzlösung und Spikelösung aus Standardmaterial beschrieben als auch neue Methoden der Probenbehandlung durch Eindampfen.

Ein eigenes Kapitel beschäftigt sich mit der chemischen Probenbehandlung als Vorbereitung für die massenspektrometrische Analyse des U- und Pu-Gehaltes der Lösungen. Weitere Themen sind die Methoden zur Korrektur der Massendiskriminierung,  $\alpha$ -spektrometrische Messungen als ergänzende Methode zur Bestimmung von Pu-238 und der Vergleich der Konzentrationsbestimmungen durch massenspektrometrische Isotopenverdünnungsanalyse mit solchen durch Röntgenfluoreszenzspektrometrie.

Der letzte Teil des Bandes enthält Arbeiten zur statistischen Auswertung der Vielzahl von Daten mit einer Großrechenanlage.

<u>Contents of Volume II</u>	<u>Page</u>
Preface and Acknowledgement	10
1. Compilation of Constants Used (W. Beyrich and E. Drosselmeyer)	11
1.1 Nuclide Masses of the Uranium and Plutonium Isotopes	12
1.2 Avogadro Constant	12
1.3 Half Life Times of the Plutonium Isotopes	13
2. Sampling for the IDA-72 Experiment at EUROCHEMIC (R. Berg and R. Reynders)	14
2.1 Introduction	17
2.2 Sampling	18
2.3 Measurements and Treatments Performed on IDA-72 Samples by EUROCHEMIC	20
3. Sample Preparation and Assay at CBNM for the IDA-72 Experiment (Y. Le Duigou, P. de Bièvre, J. Brulmans and W. Leidert)	29
3.1 Introduction	31
3.2 Samples for the Standard Experiment	33
3.3 Additional Experiments	41

	<u>Page</u>
4. Sample Preparation for the Dry Spike Technique (H. Frittum)	49
5. Preparation of Samples for the Aluminium-Capsule- Experiment in the Framework of the IDA-72 Experiment (K. Kammerichs and H. Lohner)	52
5.1 Objectives	53
5.2 Sample Preparation	53
5.3 Dissolution of the Samples	54
6. Chemical Treatment of Samples and the Attempt of Interpreting the Results (E. Mainka)	57
7. Corrections for Mass Discrimination (W. Beyrich)	71
7.1 Introduction	72
7.2 Inquiry on the Methods Used	73
8. Determination of Pu-238 by $\alpha$ -Spectrometry for the IDA-72 Experiment (A. Cricchio)	75
8.1 Introduction	76
8.2 Procedures	76
8.3 Results	77
8.4 Discussion	79
9. IDA-72: Concentration Determination by X-Ray Fluorescence Spectrometry (K. Matern, A.v. Baeckmann, W. Beyrich, E. Drosselmeyer)	80

	<u>Page</u>
10. The Application of Outlier Criteria in the Evaluation of Measuring Results for IDA-72 (G. Bork and H. Frick)	83
10.1 Introduction	85
10.2 Theoretical Considerations	86
10.3 Practical Application of the Dixon and Chauvenet Criterion	88
10.4 Numerical Results of the Comparison of the Dixon and Chauvenet Criterion	95
10.5 Resulting Considerations for the Application of Outlier Criteria to the Data of the IDA-72 Experiment	100
11. Comments to the Computer Program for the Evaluation of the Standard Experiment and Some Additional Experiments in the Framework of IDA-72 (W. Beyrich and E. Drosselmeyer)	103
11.1 Introduction	105
11.2 Organisation of the Computer Program for the Evaluation of the Standard Experiment	105
11.3 Programs for the Evaluation of Some Additional Experiments	124
11.4 Computer Program	127
12. Example of Evaluation-Results for One Laboratory and the Results for the Group of Laboratories	185
12.1 Results of One Laboratory for the Standard Experiment	185
12.2 Results of One Laboratory for the Dry Spike and Aluminium Capsule Experiments	205
12.3 Results of the Group of Laboratories	210



13. IDA 72-Measurements in One Additional Laboratory  
(E. Drosselmeyer)

229

References

261

<u>Contents of Volume I</u>	<u>Page</u>
Preface (D. Gupta)	10
Acknowledgement	12
List of Participants	13
1. Introduction	19
2. Performance of the Experiment	23
2.1 Survey of the Experimental Design	23
2.2 Sampling and Sample Preparation	28
2.3 Packing and Transportation	34
2.4 Instruments and Measurement Techniques Used by the Laboratories	39
2.5 Reporting of Data and Their General Treatment	41
3. Evaluation of the Standard Experiment	47
3.1 Layout and Participation	47
3.2 Isotopic Ratio Determinations by Mass Spectrometry	50
3.2.1 Uranium Measurements	50
3.2.2 Plutonium Measurements	58
3.2.3 Calculation of Estimates for the Relative Standard Deviations of Error Components	67
3.3 Determination of Pu-238	73
3.3.1 Participation	73
3.3.2 $\alpha$ -Spectrometric Determination	73
3.3.3 Mass Spectrometric Determination	82

	<u>Page</u>
3.4 Calculated Isotopic Compositions	86
3.4.1 Uranium	86
3.4.2 Plutonium	90
3.5 Calculated Concentrations	94
3.5.1 Method of Calculation	94
3.5.2 Uranium	96
3.5.3 Plutonium	100
3.5.4 Calculation of Estimates for Relative Standard Deviations of Error Components	104
3.5.5 Concentration Ratio Pu-239/U-238	105
3.5.6 Considerations on the Absolute Concentration Values Obtained	107
4. Evaluation of the Self Spike Experiment	111
4.1 Layout and Participation	111
4.2 Concentration Determinations	113
4.3 Calculation of Estimates for the Relative Standard Deviations of Error Components	123
4.4 Compilation of Basic Data for Calculating Estimates of Total Errors in Isotope Dilution Analysis	126
5. Investigations on New Techniques of Sample Conditioning	129
5.1 The Dry Spike Experiment	129
5.1.1 Layout and Participation	129
5.1.2 Concentration Determinations	129
5.2 The Aluminium-Capsule Experiment, Part I	138
5.2.1 Layout and Participation	138
5.2.2 Concentration Determinations	138

	<u>Page</u>
5.3 The Aluminium-Capsule Experiment, Part II	146
5.3.1 Layout and Participation	146
5.3.2 Concentration Determinations	148
5.4 Compilation of Results	158
6. The Aging Experiments	160
6.1 General Survey	160
6.2 Aging Experiment I	161
6.2.1 Objective	161
6.2.2 Layout and Sample Preparation	161
6.2.3 Results	163
6.3 Aging Experiment II (M. Bonnevie-Svendsen)	170
6.3.1 Objective	170
6.3.2 Measurements	170
6.3.3 Results	170
6.3.4 Additional Experiment	171
6.3.5 Discussion	173
6.4 Aging Experiment III	181
6.4.1 Objective	181
6.4.2 Layout and Sample Preparation	181
6.4.3 Results	181
7. The Analytical Efforts of the Laboratories Participating in the IDA-72 Experiment	187

	<u>Page</u>
8. Report on the IDA-72 Meeting	190
8.1 Agenda of the Meeting	190
8.2 List of Participants	193
8.3 Working Groups	195
8.3.1 Conclusions of the Working Group A - Chemical Sample Treatment	195
8.3.2 Conclusions of the Working Group B - Dry Spike and Aluminium Capsule Techniques	198
8.3.3 Conclusions of the Working Group C - Mass Spectrometry Measurements	200
8.3.4 Conclusions of the Working Group D - Statistical Evaluation of Data	203
8.3.5 Conclusions of the Working Group E - $\alpha$ -Spectrometric Pu-238 Determination	206
8.3.6 Conclusions of the Working Group F - Handling of Sampling Material for Shipment	208
8.3.7 Conclusions of the Working Group G - Spiking Procedures and Standards	210
8.3.8 Conclusions of the Working Group H - Aging	212
8.4 An Accurate Procedure to Safeguard the Fissile Material Content of Input and Output Solutions of Reprocessing Plants (P. de Bièvre and J. van Audenhove)	215
8.5 Summary of the Final Plenum Discussion	219
9. Summary on the Results and Experiences of the Experiment IDA-72	223
9.1 Errors Involved in Mass Spectrometric Isotope Dilution Analysis	223
9.2 Capability of New Techniques for Sample Conditioning	226

	<u>Page</u>
9.3 Stability of Liquid Samples	226
9.4 General Experiences	226
10. Conclusions and Recommendations	227
10.1 Analytical Aspects	227
10.2 Safeguards Aspects	228
10.3 General	229
References	230

## Preface and Acknowledgement

In this second volume on the IDA-72 experiment reports on different special topics connected with the preparation, organisation, performance and evaluation of the interlaboratory test are collected.

In this context it becomes evident that the execution of the experiment would not have been possible without the contributions of all those who have been actively involved. This experiment is based on the common knowledge, experience and actual work of a very active group of those interested in the method under study. The editors would like to thank all authors for their kind collaboration.

The papers collected in this volume are based on the collaboration of the participating laboratories with the evaluation group. The participating laboratories and the names of all those, who have been actively involved in the experiment, are given in Volume I.

The Editors

1. Compilation of Constants Used

by

W. Beyrich\* and E. Drosselmeyer\*\*

\* EURATOM, delegated to the Institute for Applied Systems Analysis and Reactor Physics (IASR), Gesellschaft für Kernforschung (GfK), Karlsruhe, Germany

\*\* IASR, GfK, Karlsruhe, Germany



### 1.1 Nuclide Masses of the Uranium and Plutonium Isotopes

In the following, the constants used by the evaluation staff of the IDA-72 experiment are given.

U-233: 233.0395

U-234: 234.0409

U-235: 235.0439

U-236: 236.0457

U-238: 238.0508

Pu-238: 238.0495

Pu-239: 239.0522

Pu-240: 240.0540

Pu-241: 241.0567

Pu-242: 242.0587

They are taken from Ref. /1/.

### 1.2 Avogadro Constant

$$N_A = 6.022045(31) \cdot 10^{23} \text{ mol}^{-1}$$

This value, recommended by CODATA, is given in Ref. /2/.

### 1.3 Half Life Times of the Plutonium Isotopes

#### 1.3.1 $\alpha$ -Decay:

$$\tau(238) = 87.7 \text{ years}$$

$$\tau(239) = 2.44 \cdot 10^4 \text{ years}$$

$$\tau(240) = 6.58 \cdot 10^3 \text{ years}$$

The value for Pu-238 is given in Ref. /3/, the values for Pu-239 and Pu-240 are given in Ref. /4/.

#### 1.3.2 $\beta$ -Decay:

$$\tau(241) = 15.10 \pm 0.14 \text{ years.}$$

This value was determined by measuring the ratios Pu-241/Pu-240 and Pu-241/Pu-242 during 6.65 years, it is given in Ref. /5/.

2. Sampling for the IDA-72 Experiment at EUROCHEMIC

by

R. Berg<sup>\*</sup> and R. Reynders  
EUROCHEMIC, Mol, Belgium

\* Now at the Gesellschaft zur Wiederaufarbeitung von  
Kernbrennstoffen, Karlsruhe, Germany

Table of Contents

	<b>Abstract</b>
<b>2.1</b>	<b>Introduction</b>
<b>2.2</b>	<b>Sampling</b>
<b>2.2.1</b>	<b>General</b>
<b>2.2.2</b>	<b>The IDA-72 Samples</b>
<b>2.3</b>	<b>Measurements and Treatments Performed on IDA-72 Samples by EUROCHEMIC</b>
<b>2.3.1</b>	<b>Density Determinations</b>
<b>2.3.2</b>	<b>Weighing of the Samples</b>
<b>2.3.3</b>	<b>Dilution of the Samples</b>
<b>2.3.4</b>	<b>Error Sources and Precision Estimates</b>

Abstract

The sampling procedure used for the IDA-72 experiment is described in detail.

Input tank homogenization procedures, descriptions of the sampling system, the density measurement system, the weighing of aliquots, and the dilution system are given.

Error estimates for each step in the sampling procedure are given.

## 2.1 Introduction

The samples that served for the IDA-72 experiment came from the reactor TRINO, Vercellese, a PWR with net output of 247 MWe.

Two elements were dissolved, with a total quantity of  $\sim 600$  kg of uranium.

Burn-up was estimated from the reactor to be 21.400 MWd/t and the initial enrichment was 3.897 wt % U-235.

About 0.5 % of the total quantity of uranium and about 1 % of the total plutonium was material recycled with the acid used for dissolution.

The recycled material originated from another type of reactor (CANDU-type).

Chemical decladding is used at EUROCHEMIC and prior to the dissolution the stainless steel canning was removed with  $H_2SO_4$ .

Less than 1 % of uranium and  $\sim 0.5$  % of plutonium was lost to the decladding solution.

After dissolution of the uranium dioxide with nitric acid, the dissolver solution was cooled down and then transferred by steam jet to the input tank used in the experiment (221.4), a cylindrical slab tank with a nominal volume of 3000 l. This transfer was performed  $\sim 36$  hours prior to sampling.

During standing, gentle air-sparging took place.

At sampling, the solution contained approximately 223 g/l U, 1.8 g/l Pu, 1.9 M nitric acid and an activity of about 250 Ci/l.

## 2.2 Sampling

### 2.2.1 General

The sampling procedure started 05.00h on 21.6.72 with one of the authors present to supervise this important part of the operation.

Vigorous air sparging of the solution in the tank was started at 05.00h and the first sample was drawn about 40 minutes later. Sparging for half an hour is more than sufficient to produce a homogeneous solution in the tank. During sampling gentle air sparging took place .

The sampling system can briefly be described as follows /Fig. 2-1/.

A sample bottle, penicillin type of 5ml total volume, closed with a rubber septum and an aluminium cap, is pierced by a double needle sitting on a "needle block". By aid of a vacuum system and an air lift, the solution to be sampled enters the sample bottle through one needle and returns to the tank via the other needle. In this way, the lines leading to and from the tank and the rest of the sampling system are efficiently rinsed with the solution to be sampled. The first sample bottle in a series of samples is so flushed for 10 minutes on the needle block.

The EUROCHEMIC experience with such a system is very good and for years no cross contamination or inhomogeneities have been detected when homogenization and sampling have been properly executed. Needless to say, the sampling is performed remotely in a heavily shielded blister. Manipulations are done with tongs.





### 2.2.2 The IDA-72 Samples

Prior to the sampling for the IDA-72 experiment, the samples required by EUROCHEMIC for their normal input analyses were taken.

The IDA-72 experiment required 8 samples as shown in Fig. 2-2. Also shown in this figure are the times when the samples were taken and other information pertinent to the experiment.

All samples were transferred to the analytical laboratory by pneumatic post where they were immediately stored in upright position awaiting further treatment as requested and described below.

## 2.3 Measurements and Treatments Performed on IDA-72 Samples by EUROCHEMIC

### 2.3.1 Density Determinations

A set of homogeneous samples, representative for the whole batch, was a prime requirement, both for assessing the batch at EUROCHEMIC and for the IDA-72 experiment. Thus, it was decided to use the normal EUROCHEMIC procedure for input sampling and draw the IDA-72 samples at the same time.

The homogeneity of an input batch is established by density measurements on different samples.

As can be seen from Fig. 2-2, density was measured on samples 2,3,A,B,C and F. Bottle 1 is never used in the input procedure as the risk for contamination from previous batches is highest in this sample.

Date	Sample	EUROCHEMIC Input Samples			IDA - 72 Samples								
		1	2	3	A	B	C	D	E <sub>1</sub>	E <sub>2</sub>	E <sub>3</sub>	E <sub>4</sub>	F
21.6.1972	Time of sampling [h.min]	05.35	05.45	05.55	06.35	06.30	06.25	-	06.40	06.45	06.50	06.55	07.00
	Density at 30°C [g/ml]	-	1.3728	-	1.3728	1.3727	-	-	-	-	-	-	1.3719
	Aliquot weight [g]	-	0.9861	-	2.2149	2.0970	-	-	-	-	-	-	0.5546
	Time of dilution [h.min]	-	08.00	-	10.30	14.30	-	-	-	-	-	-	16.00
	Approx.dil. factor	-	-	-	161	171	-	-	-	-	-	-	20
	Volume after dilution [ml]	-	-	-	262.46	262.46	-	-	-	-	-	-	-
	Weight after dilution [g]	-	-	-	-	-	-	-	-	-	-	-	8.8857
	Density of dil. 30°C [g/ml]	-	-	-	1.1622	1.1622	-	-	-	-	-	-	1.1705
22.6.1973	Density at 30°C [g/ml]	-	-	-	-	-	1.3729	-	-	-	-	-	-
	Aliquot weight [g]	-	-	-	-	-	3.7326	-	-	-	-	-	-
	Time of dilution [h.min]	-	-	-	-	-	10.00	-	-	-	-	-	-
	Density of dil. 30°C [g/ml]	-	-	-	-	-	1.1636	-	-	-	-	-	-
	Time of filling D [h.min]	-	-	-	-	-	-	10.30	-	-	-	-	-

Dilution liquid : 5.0 M HNO<sub>3</sub> with density : 1.1609 at 30°C.  
Time Sequence and Measurements Performed at EUROCHEMIC for IDA-72.

Figure 2-2

All density values were consistent and the homogeneity of the batch and the samples thus established.

The density was measured by the vibrating capillary method, whose principle is shown in Fig. 2-3 and can be described as follows:

A capillary in the form of a V is filled with the liquid whose density is to be measured. The tip of the capillary can vibrate freely whereas the entry- and exit- points of the capillary are fixed in a large mass of stainless steel.

The resonance frequency of the vibrating capillary depends on the mass of the capillary and consequently will change with the density of the filling solution since the volume remains constant. Thus a proportionality exists between the resonance frequency and the density of the solution. The resonance frequency (or rather the time necessary for a fixed number of vibrations at the resonance frequency) is measured electronically.

Two magnets are fixed to the capillary, one is used to set the system into vibrations, the other is used to record the number of vibrations.


The excitation and detection are fully electronic.

The capillary is isolated and the measurements take place at  $30 \pm 0.05$  °C.

Calibration and control of the instrument is done prior to and after each density measurement with ambient air (corrections for humidity, barometric pressure and box underpressure are applied) and distilled water.

Less frequently, standards with other densities are measured to verify the well established, excellent linearity.

Aliquots can be filled directly from the densitometer into preweighed aliquot containers, specially designed with the purpose of directly delivering the aliquot to the dilution vessels.



LEGEND:

1. Magnet on oscillating glass tube.
2. Thermal insulation containing heating wires and thermistors.
3. Fixing point of oscillating tube.
4. 5-way Hamilton valve.
5. 3-way Hamilton valve.
6. Glass extraction tube or aliquot vessel.
7. Syringe.
8. Needle block
9. Needle with bypass for air and filter.
10. Sample bottle.
11. Storage bottle for bidist. water.
12. Storage bottle for acetone p.a.
13. Motor driven air pump.

Fig. 2-3.

SCHEMATIC FLOWSHEET OF THE DENSITY DETERMINATION  
IN A SHIELDED BOX

### 2.3.2 Weighing of the Samples

Prior to the dilution step, the aliquot is weighed on a remotely installed analytical balance, type Metter H-16. The balance is controlled daily with a standard weight, weighing approximately the same as a filled aliquot container.

Less frequently, standard weights are measured to control the balance.

With precise density and weight determinations, the volume is also known precisely.

The aliquot size can be varied according to wish by for example counting the number of drops to the aliquot vessel as the drop size is well known.

### 2.3.3 Dilution of the Samples

Following aliquotation, dilution is immediately performed in a shielded box adjacent to the balance box. The dilution system in use at EUROCHEMIC is shown schematically in Fig. 2-4.

With the aliquot transfer vessel in position above the dilution vessel and about 50 ml of dilution liquid (5  $\underline{M}$   $\text{HNO}_3$  in the case of IDA-72) present in the dilution vessel, filling takes place slowly via the aliquot vessel.

Following several rinse steps, the teflon cone is firmly placed into the aliquot container and approximately 200 ml of the dilution liquid passes through the aliquot vessel. The aliquot vessel is then removed and the solution stirred magnetically. The dilution vessel bulb is flushed with a special spray head and then slowly filled via a special line until contact is made between Pt-electrodes and the solution.

After thorough mixing, about 50 ml solution passes the outlet system to ensure an effective rinse of this part of the system. The rest of the solution serves for aliquots as needed.




Fig. 2-4.  
EUROCHEMIC HOT SAMPLES  
DILUTION SYSTEM

#### 2.3.4 Error sources and Precision Estimates

The IDA-72 samples A and B originated through the following steps:

- A) Sampling
- B) Density measurements and aliquotation
- C) Weighing of aliquots
- D) Dilution of aliquots
- E) Weighing of diluted samples (CBNM)
- f) Spiking of diluted samples (CBNM)

The points E) and F) are treated in the next chapter

The points A) - D) are shortly discussed below.

- A) The errors introduced by drawing the samples (tank inhomogeneities and cross contamination from sampling system) are believed to be small compared to other error sources.

This is concluded from long term EUROCHEMIC input experience.

- B) The density measurements performed on a series of input samples rarely exceed a range of 0.03 % and the contribution from the density measurement to the overall error in the EUROCHEMIC input procedure is small.

Careful rinsing and control of the densitometer also makes cross contamination from previous samples unlikely.

- C) The remote weighing step is performed with an overall precision of 0.01 % (standard deviation on a single aliquot weight determination).

- D) The most critical step in the EUROCHEMIC input procedure is in our opinion the remote dilution step where the following phenomena may contribute to undetected errors:

- Incomplete transfer of aliquot to dilution vessel.
- Undetected leaks from bottom valve of dilution vessel during the dilution step.
- Incorrect volume reading by Pt-electrodes or leaking solenoid valve system.
- Inhomogeneities in diluted aliquot solution.
- Cross contamination from previous solutions caused by insufficient rinsing of outlet system.

It goes without saying that all of the above steps are carefully controlled and extreme care is executed during a dilution.

The calibration of the dilution vessel is performed outside the shielded box by weighing the vessel filled with 1 M  $\text{HNO}_3$  (the dilution liquid normally used at EUROCHEMIC) to be nearest 0.01 g.

The volume is controlled by the Pt-electrode system with the solenoid valve and feeding bottles at the same positions relative to the dilution vessel as used in the shielded box.

Calibration results give a relative standard deviation of 0.03 % for a single measurement.

Keeping in mind that calibration is executed under optimum conditions, our best estimate of the remote dilution error is 0.1-0.2 % (relative standard deviation).

This figure is in good agreement with evaluated EUROCHEMIC experience on input analyses where the dilution step is part of the procedure.

The table below summarizes this chapter.



PRECISION ESTIMATES FOR IDA-72.

STEP	OPERATION	ERROR SOURCES	PRECISION *
A	Sampling	Inhomogeneities Sampling error	0.1 %
B	Density and aliquotation	Instrumental errors Cross contamination Evaporation	0.02 %
C	Weighing of aliquots	Remote weighing error Evaporation	0.1 %
D	Dilution of aliquots	Transfer errors Inhomogeneities Instrumental errors	0.1-0.2 %

\* Standard deviation of a single measurement.

3. Sample Preparation and Assay at CBNM for the

IDA - 72 Experiment

by

Y. Le Duigou, P. De Bièvre, J. Brulmans, W. Leidert  
Central Bureau for Nuclear Measurements, EURATOM,  
Geel, Belgium

Table of Contents

- 3.1 Introduction
- 3.2 Samples for the Standard Experiment
  - 3.2.1 Reference Solution R
  - 3.2.2 Mixed Spike Solution S
  - 3.2.3 Spiked Reference Solution  $R_s$
  - 3.2.4 Spiked Active Feed Solution  $A_s$
  - 3.2.5 Spiked Active Feed Solution  $B_s$
- 3.3 Additional Experiments
  - 3.3.1 Dry Spike Experiment
  - 3.3.2 Aluminium Capsules Techniques Experiment
  - 3.3.3 Distribution of Reference Solution for the Self Spike Experiment
  - 3.3.4 Distribution of Mixed Spike Solution for the Aging Experiment
  - 3.3.5 Preparation of Inactive Samples for the Aging Experiment III
  - 3.3.6 Preparation of Active Samples for the Aging Experiment III
  - 3.3.7 Preparation of Samples for the Aging Experiment I
  - 3.3.8 Preparation of Samples for the Aging Experiment II

Acknowledgements

### 3.1 Introduction

It was decided at the Vienna analytical working group meeting on 6 and 7 December 1971 that the samples (including the spiked samples) and reference solutions for the IDA-72 experiment would be prepared by CBNM (Geel) with the assistance of EUROCHEMIC and SCK/CEN(Mol).

This chapter reports on the preparation and assay of these samples. The total experiment is illustrated in a scheme given in Fig. 3-1.

IDA - 72 EXPERIMENT




Fig. 3-1

### 3.2 Samples for the Standard Experiment:

The standard experiment involves:

1. preparation, definition and aliquoting of a reference solution R.
2. preparation, definition and aliquoting of a mixed spike solution S.
3. preparation and aliquoting of a spiked reference solution  $R_s$ .
4. aliquoting and spiking of an active feed solution A.
5. aliquoting and spiking of an active feed solution B.


#### 3.2.1 Reference Solution R

The required characteristics of this solution were the following:  
1 mg U and 10  $\mu$ g Pu per ml of a 5 molar nitric acid solution.

The R solution was prepared by mixing weighed fractions of a plutonium reference solution and of a uranium reference solution. The chemically well-characterized NBS materials largely used by analytical chemists working in the nuclear field were considered to be the most appropriate materials for the preparation of an U and Pu reference solution.

The plutonium solution (1282) was a 5 molar nitric acid solution containing  $0.56445 \pm 0.00005$  g of NBS 949 c plutonium in  $573.424 \pm 0.005$  g of solution. The plutonium metal was dissolved in HCl. After addition of 10 ml of concentrated nitric acid the solution was evaporated to moistness. This operation was repeated three times. The nitrate was finally dissolved in 50 ml of 5 M nitric acid and the solution was transferred one day later in a weighed flask and diluted up to 500 ml. This solution was weighed just before use. The uranium solution (1283) was a 5 molar nitric acid solution of  $292.694 \pm 0.003$  g containing the uranium equivalent to  $1.1720 \pm 0.0001$  g of NBS uranium oxide 950 a, ignited one hour at  $900^\circ\text{C}$  before weighing. The uranium oxide was dissolved in  $\text{HNO}_3$  1:1.

Fig.3-2IDA - 72 MIXED SPIKE AND REFERENCE SOLUTIONS



The solution was evaporated to moistness. The nitrate was finally dissolved in 50 ml of 5M nitric acid and the solution was transferred one day later in a weighed flask and diluted to 250 ml. The solution was weighed just before use.

The reference solution R was obtained by mixing  $2.5110 \pm 0.0005$  g of the Pu solution (1282) with  $93.912 \pm 0.001$  g of the U solution (1283) and adding 5 molar nitric acid to a total weight of  $287.402 \pm 0.003$  g. The different steps of the preparative work are summarized in the scheme given in Fig. 3-2.

The Pu concentration was calculated (see foot-note (1)) to be 8.599  $\mu$ g Pu per g of solution. NBS certification for the material 949 c is  $99.99 \pm 0.06$  % Pu.

The U concentration (see foot-note (2)) was calculated to be 1.1088 mg U per g of solution. NBS certification for material 950 a is  $99.94 \pm 0.02$  %  $U_3O_8$ .

Taking into account the uncertainties on the weighings and on the NBS materials, the uncertainty on both U and Pu concentrations was less than  $\pm 0.1$  %.

The isotopic composition of the NBS 950 a material measured at CBNM is in atom %:

U-234	0.0055
U-235	$0.7203 \pm 0.0007$
U-238	99.2742

The isotopic composition of the NBS 949 c material measured at

---

(1) Pu concentration calculation=  $\frac{0.56445 \cdot 0.9999 \cdot 2.5110}{573.424 \cdot 287.402} = 8.599$   
 $10^{-6}$  g Pu/g solution

(2) U concentration calculation=  $\frac{1.1720 \cdot 0.8480 \cdot 0.9994 \cdot 93.912}{292.694 \cdot 287.402} = 1.1088$   
 $10^{-3}$  g U/g solution



CBNM and valid for January 20, 1972 is in atom %:

Pu-238	0.0039
Pu-239	97.3553 $\pm$ 0.0090
Pu-240	2.5652
Pu-241	0.0724
Pu-242	0.0032

According to the approximate figures given in Fig. 3-1 aliquots of the solution R have been used for the standard experiment, the self-spike experiment, and the aging experiment III.

### 3.2.2 Mixed Spike Solution S

The required characteristics of this solution were the following: 1 mg of U, highly enriched in U-233, and 10  $\mu$ g of Pu, highly enriched in Pu-242, per ml of a 5 molar nitric acid solution. The solution S was prepared by mixing weighed fractions of chemically and isotopically well-defined plutonium and uranium solutions.

The different steps of the preparative work are summarized in the diagram given in Fig. 3-2.

The Pu solution (1290) was a 5 molar nitric acid solution. About 80 ml of plutonium oxide were dissolved in concentrated nitric acid containing a drop of hydrofluoric acid. The solution was evaporated to moistness and the nitrate dissolved in 20 ml of 5M nitric acid. One day later the solution was diluted to 80 ml.

The major impurities in the plutonium oxide measured by emission spectrography were 350 ppm of iron and 20 ppm of copper.

The solution (1290) of enriched Pu-242 was assayed on a diluted fraction (1291) by triplicate isotope dilution mass spectrometry as follows:

sampling unknown 242 Pu solution (1291) (g)	adding fractions of NBS 949 c solution (1285) (g)	$^{239}\text{Pu}/^{242}\text{Pu}$ atomic ratio observed in blend	concentration of the unknown solu- tion ( $\mu\text{g/g}$ solution)
1. 1.16488	1.17276	0.8534	26.975
2. 1.16598	1.12694	0.8193	$\frac{26.975}{26.975}$
3* 1.13342	1.11562	0.8350	26.956
<u>Accuracy:</u>			
$\pm$ 0.00005	$\pm$ 0.00005	$\pm$ 0.0013	$\pm$ 0.054 ( $\pm 0.2\%$ )

\* The third determination was performed one week later to check additionally whether nothing happened to the solution during the sampling.

The isotopic composition (atom %) of the material was determined to be:

Pu-238	0.003
Pu-239	0.020
Pu-240	0.084
Pu-241	0.075
Pu-242	$99.818 \pm 0.010$

and is valid for July 27, 1973.

The concentration of the original Pu-242 solution (1290) used to blend the S solution was obtained from the dilution factor:

$$26.975 \mu\text{g/g} \cdot \frac{280.652 \text{ g}}{7.4293 \text{ g}} = 1.0190 \pm 0.0020 \text{ mg Pu/g sol.}$$

The NBS 949 c solution (1285) used in this assay was prepared by dissolving  $564.45 \pm 0.05$  mg Pu NBS 949 c metal in  $\text{HNO}_3$  to a solution of 573.424 g (1282) and diluting to:

$$\frac{564.45}{573.424} \cdot \frac{3.7125}{157.845} = 23.152 \mu\text{g Pu/g sol.}$$

The uranium solution (1292) was a 5 molar nitric acid solution prepared under the same conditions as the U reference solution. This material was rather impure and contained about 1% of iron as a major impurity, magnesium and plutonium were present respectively at the 500 and 250 ppm levels.

The material was delivered only one day before use and it was not possible to analyze and purify it before the preparation.

The enriched U-233 solution (1292) was assayed by triplicate isotope dilution mass spectrometry as follows:

U-233 solution (1292) (g)	adding NBS 950 a oxide (g)	U-233/U-238 atomic ratio observed in blend	concentration of the unknown U-233 solution (mg U/g solution)
1. 1.16794	0.040006	0.053180	1.5371
2. 1.17200	0.040366	0.052907	<u>1.5376</u>
			Mean: 1.5374
3* 1.14695	0.039582	0.052861	1.5393
<u>Accuracy:</u>			
+ 0.00005	+ 0.000004	+ 0.000080	+ 0.0031 (+ 0.2%)

\* The third determination was performed one week later to check additionally whether nothing happened to the solution during the samplings.

The isotopic composition (in atom %) of the material was determined to be :

U-233	97.779 ± 0.010
U-234	0.018
U-235	0.040
U-236	0.003
U-238	2.163

An isotope dilution mass spectrometry concentration measurement of 0.397 µg of plutonium per g of solution has also been performed on the solution 1292. This plutonium concentration has to be taken into account when calculating the plutonium concentration in the mixed spike solution S.

The mixed spike solution S was obtained by mixing  $3.5480 \pm 0.0005$  g of the plutonium solution (1290) with  $197.482 \pm 0.002$  g of the uranium solution (1292) and adding 5 molar nitric acid to a total weight of  $366.238 \pm 0.003$  g.

The Pu concentration was calculated (see foot-note (1)) to be  $10.085 \mu\text{g Pu}$  per g of solution. The U concentration was calculated (see foot-note (2)) to be  $0.8290 \text{ mg U}$  per g of solution. Taking into account the uncertainties on the weighings and on the assay of the uranium and plutonium solutions, the uncertainty on both U and Pu concentrations was less than  $\pm 0.25 \%$ . The isotopic composition (atom %) of the plutonium in the mixed spike solution S was determined to be:

Pu-238	0.0044
Pu-239	1.3084
Pu-240	0.5923
Pu-241	0.1007
Pu-242	97.9942

The aliquoting for the standard experiment, the dry spike experiment, the aluminium capsule technique experiment, the aging effect experiments I to III was done according to the approximate figures given in Fig. 3-1. A fraction of the Pu-242 solution (1290) was diluted to about  $20 \mu\text{g Pu/ml}$  and 23 aliquots (vials 1 to 23) of this solution were taken for distribution as pure Pu-242 spike solution (239/242 and 240/242 atomic ratio measurements). A fraction of the U-233 solution (1292) was diluted to about  $20 \mu\text{g U/ml}$  and 23 aliquots (vials 24 to 46) were taken for distribution as pure U-233 spike solution (238/233 atomic ratio measurements).

---

(1) Pu concentration calculation: 
$$\frac{1019.0 \cdot 3.5480}{366.238} + \frac{0.397 \cdot 197.482}{366.238}$$
  

$$= 10.085 \mu\text{g Pu/g solution}$$

(2) U concentration calculation: 
$$\frac{1.5374 \cdot 197.482}{366.238} = 0.8290 \text{ mg U/g solution}$$

### 3.2.3 Spiked Reference Solution R<sub>s</sub>

The spiked reference solution R<sub>s</sub> was prepared at CBNM by mixing 45.943 ± 0.001 g of the R solution (see Par. 3.2.1) with 46.628 ± 0.001 g of the S solution (see Par. 3.2.2).

The R<sub>s</sub> solution, after addition of hydroxylamine (about 20 times the theoretical quantity needed to reduce Pu) was heated to 80°C for about 45 minutes and finally split up into 23 aliquots placed in the vials 70 to 92.

At the same time 23 aliquots of the unspiked reference solution R were transferred in vials 47 to 69.

### 3.2.4 Spiked Active Feed Solution A<sub>s</sub>

A fraction of a diluted active feed solution A prepared by Eurochemic was spiked immediately after preparation, mixing 52.154 ± 0.002 g of this solution with 47.114 ± 0.002 g of the spike solution S. For this purpose an aliquot of the spike solution (about 50 ml) was carefully transported from CBNM to the EUROCHEMIC plant. A sensitive balance and its anti-vibration unit was transported from CBNM to EUROCHEMIC for the weighing operations. After addition of hydroxylamine and heating the spiked solution was aliquoted and bottled in vials 203 to 225 (23 samples). At the same time 23 aliquots of the unspiked A solution were transferred to vials 180 to 202.

### 3.2.5 Spiked Active Feed Solution B<sub>s</sub>

A similar preparation like in Par. 3.2.4 was done in the EUROCHEMIC plant using a diluted active feed solution B prepared by EUROCHEMIC. An aliquot of 47.157 ± 0.002 g of the spike solution S was added to 55.377 g ± 0.002 g of the solution B.

Aliquoting in vials 249 to 271 (23 samples) was performed after addition of hydroxylamine and heating.

At the same time 23 aliquots of the unspiked solution B were transferred to vials 226 to 248.

### 3.3 Additional Experiments

#### 3.3.1 Dry Spike Experiment

The preparation of the samples for this experiment was done by chemists of the IAEA, Vienna, working at the CBNM laboratories. The equipment (fume hoods, balances, vials...), spike solution S and assistance were provided by CBNM. An aliquot of the diluted active feed solution A used for this preparation was transported from EUROCHEMIC to CBNM.

#### 3.3.2 Aluminium Capsules Techniques Experiment

The preparation of the samples for this experiment was done by chemists of the Transuranium Institute, Karlsruhe, working at the CBNM laboratories. As in Par. 3.3.1, equipment, solutions and assistance were provided by CBNM. An aliquot of the diluted active feed solution A used for this preparation was transported from EUROCHEMIC to CBNM.

#### 3.3.3 Distribution of Reference Solution for the Self Spike Experiment

Thirteen vials (no. 93-104 and 150) were filled each with about 8 ml of the reference solution R (see Par. 3.2.1), weighed and packed for distribution to the different laboratories participating in the self spike experiment. The weights of the vials filled with solution R, with plastic stopper, but without screw cap, and the masses of solution are given in Table 3-1.

Table 3-1

No. of vial	Weight of vial [g]	Mass of solution [g]
93	22.5001	9.5040
94	22.5413	9.3981
95	22.3310	9.1245
96	22.6440	9.2376
97	22.6709	9.3399
98	22.3672	9.1754
99	22.0576	8.9606
100	21.9315	8.9100
101	22.3996	9.2702
102	22.4460	9.2026
103	22.3297	9.1721
104	22.6141	9.2310
150	20.4164	7.2997

The uncertainty on the masses of solution was  $\pm 0.0001$  g

## 3.3.4 Distribution of Mixed Spike Solution for the Aging Experiment

Eight vials (no. 105 to 111) were filled, each with about 8 ml of the spike solution S (see Par. 3.2.3) and packed for distribution to three different laboratories participating in the aging experiments I to III. The weight of the vials filled with solution S, with plastic stopper, but without screw-cap, and the masses of solution are given in Table 3-2.

Table 3-2

No. of vial	Weight of vial [g]	Mass of solution [g]
105	22.2211	9.2505
106	22.6158	9.2154
107	22.1564	9.2459
108	22.3651	9.2145
109	22.1940	9.1893
110	21.9502	8.9101
111	22.2852	9.0783

The uncertainty on the masses of solution was  $\pm 0.0001$  g.



## 3.3.5 Preparation of Inactive Samples for the Aging Experiment III

This experiment was planned to obtain information on the aging of material over a short time period (2-10 days) on a U - Pu solution without fission products. The reference solution R, (see Par. 3.2.1), was used for this experiment.

About 1.2 g of the spike solution S (see Par. 3.2.3) was weighed in advance (date 14.06.72) in each of the three vials no. 137-138-139. Two, six and ten days after the date of preparation (13.06.72) of the solution R about 1 g of this solution R was added in each vial. The date quoted for the preparation of the reference solution R was the date of mixing of the Pu and U reference solutions. These vials were closed and packed after addition of hydroxylamine and heating to 80°C like it was done for the solution R<sub>s</sub> in the standard experiment. The masses of R and S solutions and dates of preparation are given in Table 3-3.

Table 3-3

No. of vial	Mass of solution [g]	Date	Mass of solution R [g]
137	1.2124 $\pm$ 0.0001	15.06.72	1.1214 $\pm$ 0.0001
138	1.2179 $\pm$ 0.0001	19.06.72	1.1490 $\pm$ 0.0001
139	1.2073 $\pm$ 0.0001	23.06.72	1.1302 $\pm$ 0.0001

The fraction of the solution R remaining in the vial 149 after the preparation was weighed. The weight of the vial filled with solution was 18.5674 g. The mass of solution in the vial was 5.5692  $\pm$  0.0001 g.

## 3.3.6 Preparation of Active Samples for the Aging Experiment III

This experiment had the same aim as the aging experiment on the R solution (see Par. 3.3.5), but instead of R, the diluted active feed solution A prepared by EUROCHEMIC (date 21.06.72) was used. The spiking was done under the same conditions as for the solution R.

The masses of A and S solutions and dates of preparation are given in Table 3-4.

Table 3-4

No. of vial	Mass of solution S [g]	Date	Mass of solution A [g]
140	1,2133 $\pm$ 0,0001	23,06,72	1,6660 $\pm$ 0,0001
141	1,2170 $\pm$ 0,0001	26,06,72	1,6459 $\pm$ 0,0001
142	1,2005 $\pm$ 0,0001	30,06,72	1,5797 $\pm$ 0,0001

The fraction of A solution remaining in the vial 229 was weighed. The weight of the vial filled with solution was 18,8989 g.

## 3.3.7 Preparation of Samples for the Aging Experiment I

It was intended to study the dependency of aging effects on the molarity of the acid solution over a time interval of some months.

A fraction of the spike solution S (see Par. 3.2.3) was weighed (14.06.72) in the three vials no. 143 - 144 - 145.

At a same date (22.06.72) a fraction of each of three different active feed solutions D (molarity 2.5, 5 and 8 M) prepared by EUROCHEMIC was added to the corresponding vial. After addition of hydroxylamine and heating to 80°C the vials were closed and packed.

The masses of S and D solutions are given in Table 3-5.

Table 3-5

No. of vial	Mass of solution S [g]	Mass of solution D [g]
143	2.8614 + 0.0001	1.7170 + 0.0001
144	2.8424 + 0.0001	2.1704 + 0.0001
145	2.8392 + 0.0001	2.1422 + 0.0001

Vials containing aliquots of the different solutions D were weighed and packed for distribution to the laboratory performing the experiment I. The weights of the vials filled with solution are given in Table 3-6. The laboratory was asked to spike aliquots of these D solutions from time to time using the delivered spike solution S (see Par. 3-4).

Table 3-6

No. of vial	Reference	Molarity	Weight of vial with solution
272	D <sub>u</sub> -1	2.5	20.9345
273	D <sub>u</sub> -2	2.5	18.2492
275	D <sub>u</sub> -3	5	21.1122
276	D <sub>u</sub> -4	5	19.0012
278	D <sub>u</sub> -5	8	22.5244
279	D <sub>u</sub> -6	8	18.9098

## 3.3.8 Preparation of Samples for the Aging Experiment II

It was intended to study the dependency of aging effects on the Pu concentration over a time interval of some months. Different aliquots of the spike solution S were weighed (14.06.72) in the three vials no.146 - 147 and 148.

At the same date (22.06.72) a fraction of each of three different active feed solutions C (concentration 2, 10 and 20  $\mu\text{g Pu/ml}$ ) prepared by EUROCHEMIC was added to the corresponding vial. After addition of hydroxylamine and heating to  $80^{\circ}\text{C}$  the vials were closed and packed. The masses of S and C solutions are given in Table 3-7.

Table 3-7

No. of vial	Mass of solution S [g]	Mass of solution C [g]
146	1.3379 $\pm$ 0.0001	6.2205 $\pm$ 0.0001
147	2.8640 $\pm$ 0.0001	2.4032 $\pm$ 0.0001
148	4.0156 $\pm$ 0.0001	1.9800 $\pm$ 0.0001

Vials containing aliquots of the different solutions C were weighed and packed for distribution to the laboratory performing the experiment II (further spiking from time to time as in Par. 3-7).

The weights of the vials filled with solution are given in Table 3-8.

Table 3-8

No. of vial	Reference	Concentration [ $\mu\text{g Pu/ml}$ ]	Weight of vial with solution
281	C <sub>u</sub> -1	2	21.4450
282	C <sub>u</sub> -2	2	20.3377
284	C <sub>u</sub> -3	10	21.9872
285	C <sub>u</sub> -4	10	19.7012
287	C <sub>u</sub> -5	20	22.0266
288	C <sub>u</sub> -6	20	19.6082

Acknowledgements

The sample preparation for the IDA-72 experiment was coordinated by Y. Le Duigou.

The determination of the isotopic compositions as well the assay of the spike solution by isotope dilution mass spectrometry were performed by P. De Bièvre.

The mass determinations were done by J. Brulmans.

The entire chemical preparation, aliquoting and packing of the samples was done by W. Leidert.

The participation of many CBNM collaborators, for some of them more than one month, was required to carry out this task. To all personnel involved many thanks.

The authors wish to thank specially M. Gallet for his assistance in the mass spectrometry measurements, M. Hendrickx for the weighing during the preparation of the samples, M. De Buck for his assistance in the health physics and transport and M. van Hengel for his help in many manipulations.

The authors also want to thank M. Moret, Head of the Division Sample Preparation and Assay, for his interest in this work, M. Lauer for many valuable contributions when planning this work and ordering the spike materials, M. Debus for useful discussions and M. Berthelot for the spectrographic analysis of the spike materials.

Y. Le Duigou wants to thank specially Miss Drosselmeyer, MM. Beyrich, Berg, Frittum, Koch and all the persons he came into contact with, during the execution of this work, for the fruitful discussions and the valuable collaboration.

4. Sample Preparation

for the

"Dry Spike Technique"

by

H. Frittm

International Atomic Energy Agency  
Department of Safeguards and Inspection  
Division of Development  
Vienna, Austria

The dry spike technique is part of the experiment and consists of two main parts:

- a) preparation of the "dry spike"
  - b) spiking of the safeguards sample
- 
- a) the calibrated spike solution (preparation is described in Chapter 3) was used
    - 1) an empty Sovirel tube was tare weighed on a semi-micro balance
    - 2) 1 ml of the spike solution containing 10 micrograms Pu-242 and 1 mg U-233 was measured by a multidosimate E 415 burette and filled into the tube
    - 3) the tube was gross weighed on a semi-micro balance
    - 4) 20 spikes were prepared in this manner
    - 5) the spikes were placed in an aluminium block and carefully evaporated to dryness which required about 20 hours
    - 6) the vials with the dried spikes were tare weighed on a semi-micro balance

In the case of a safeguards inspection operation the inspector brings his set of calibrated "dry spikes" to the facility (i.e. reprocessing plant)

- b) 7) a 1 ml portion of the diluted sample was added to each spike by an Eppendorf pipette
- 8) the tubes were gross weighed on a semi-micro balance
- 9) samples and spikes were evaporated according to 5)
- 10) the vials containing dry spike and dry sample mixture were closed and prepared for shipment

The subsequent treatment for the "dry spiked" sample will be the addition of 2 ml 5-8M nitric acid and heating until gentle boiling to dissolve spike and sample. The valency adjustment is made according to the individual procedures of the laboratory followed by their own separation procedure.



5. Preparation of Samples for the Aluminium-Capsule-Experiment

in the Framework of the IDA-72 Experiment

by

K. Kammerichs and H. Lohner

EURATOM-European Institute for Transuranium Elements,  
Karlsruhe, Germany

## 5.1 Objectives

The storage-stability of samples taken from dissolved irradiated fuels is influenced by radiolysis which may lead firstly to a loss of solvent and a change of concentration in the sample, and secondly a plating of plutonium on the walls of the container may occur because of a depletion of acid. To avoid these problems we tried, as part of the IDA-72 experiment, the described technique of sample conditioning. An aliquot of the sample or spike solution was weighed in an aluminium capsule. The solution was then carefully dried, precaution being taken that no spattering during evaporation had occurred. Later, the aluminium capsule is closed. For the subsequent analysis the container together with the evaporated sample is dissolved. This gives the advantage that all the material of the sample is made available for the analysis. Storage of samples over longer periods becomes possible.

## 5.2 Sample Preparation

For the experiment three kinds of samples had to be prepared:

- a) 16 capsules containing an aliquot of about 1 ml sample solution "A" and 1 ml spike solution.
- b) 16 capsules containing an aliquot of about 1 ml sample solution "A"
- c) 8 capsules containing an aliquot of 1 ml spike solution.

Procedures used for a) :

1. 8 preweighed aluminium-capsules were inserted into an aluminium-block and filled carefully with 1 ml of sample solution "A". (Eppendorf pipettes 1 ml)
2. Each of the capsules was weighed again.
3. The aluminium-block was heated up to 80°C. To speed up the evaporation a stream of filtered air was passed over the surface of the solutions /Fig. 5-1/. About two hours were needed to completely evaporate the samples.
4. After a cooling time of half an hour the capsules were weighed again.
5. 1 ml of the spike solution was pipetted into each capsule which was weighed once more.
6. The drying-step(3.)was repeated.
7. The capsules were carefully closed with pincers, put into plastic cans and welded into plastic bags.

Procedures used for b) and c) :

For the capsules referring to point b) and c) the procedure described above was applied except for the steps 4. to 6.

### 5.3 Dissolution of the Samples

The Al-capsule is placed in a cleaned, preweighed conical glass flask of 300 - 500 ml. A few drops of saturated  $\text{Hg}(\text{NO}_3)_2$  solution are placed on the can and 10 ml of 7 M  $\text{HNO}_3$  are added. To initiate the reaction the flask is warmed up gently. The reaction starts suddenly. After the reaction becomes less vigorous, the flask is filled up to about 20 ml with 7 M  $\text{HNO}_3$  and is heated until complete dissolution of the capsule. (This can take up to 5 hours.) Finally, 10 ml  $\text{H}_2\text{O}$  are added to prevent crystallisation of aluminium-nitrate during cooling. The flask is weighed and an aliquot is taken for analysis.




Fig. 5-1 Setup for Evaporation of Sample Solutions

Further conditioning of the sample can follow any appropriate standard procedure, which is not affected by excessive aluminium nitrate. We are applying a method described in /6/.

Reagents

Hg (NO<sub>3</sub>)<sub>2</sub> p.a.

7 M HNO<sub>3</sub> quartz distilled

H<sub>2</sub>O quartz distilled

6. Chemical Treatment of Samples and the Attempt  
of Interpreting the Results

by

E. Mainka

Institute for Radiochemistry,  
Gesellschaft für Kernforschung,  
Karlsruhe, Germany

It is known that in mass spectroscopic isotope dilution analysis the exchange of isotopes between the sample and spike solutions constitutes a fundamental condition as soon as chemical separation steps are required in the analysis.

While such an exchange is relatively easily attained in the case of uranium, difficulties are more frequently encountered in the case of plutonium.

The polymerization of Pu-IV ions which may be present both in the sample and in the spike solutions quite often inhibits complete oxydation and reduction, respectively. However, the process is based on the possibility of an isotope exchange between the two initial solutions brought about by specific redox steps. Good working conditions can be achieved in a number of ways.

The significant steps of analysis for the chemical preparation techniques used in the IDA 72 experiment by the individual laboratories are summarized in Table 6-1.

It appears from the table that both ferrous solutions and hydroxylamine solutions are used for reduction. Most of the laboratories use  $\text{Na}_2\text{NO}_2$  for oxydation, two of them use  $\text{HClO}_4$ .

To separate uranium and plutonium the ion exchange reaction is preferrably employed. Four laboratories apply solvent extraction.

A relationship cannot be established between the methods and the quality of results, the more so, since in most cases own analytical rules are applied and detailed data would be required to allow a judgement to be made. Only six laboratories rely on prescriptions taken from the literature.

It should be mentioned in this context that unlike the previous rules the new issue of ASTM E 267-70 published in 1972 includes a detailed discussion in the chapter devoted to isotope exchange. No laboratory observed this prescription.

Table 6-1

Code Nr.	Reduction	Oxidation	Separation	Lit. Data
2	without chem. prep.	-	-	without literature data
3	NH <sub>2</sub> OH · HCl in HCl after addition of tracer	HNO <sub>3</sub> -	Ion Exchange 8 M HNO <sub>3</sub> - System	without literature data
4	ferrous-ammonium- sulfate allow to stand for 5 min	NaNO <sub>2</sub> -	Ion Exchange 8 M HNO <sub>3</sub> - System	ASTM E 267-67 T
5	2NH <sub>2</sub> · OH · SO <sub>4</sub> allow to stand by 60°C 1 hour	HNO <sub>3</sub> -	Ion Exchange 8 M HNO <sub>3</sub> - System	without literature data
6	NH <sub>2</sub> OH · HCl allow to stand for 10 min	NaNO <sub>4</sub> -	Ion Exchange 8 M HNO <sub>3</sub> - System	without literature data
7	NH <sub>2</sub> OH · HCl in HCl evaporate to dryness	HNO <sub>3</sub> -	Ion Exchange 8 M HNO <sub>3</sub> - System	Webster et al, Anal. Chim.Acta 24 (1961) and Hemlin et al, Anal.Chim.Acta 33 1547 (1961)
8	ferrous-ammonium- sulfate allow to stand for 5 min	NaNO <sub>2</sub>	Ion Exchange 8 M HNO <sub>3</sub> - System	GEAP 5354
10	Red-Ox-System not described	-	Solvent Extraction	without literature data
12 <sup>+</sup>	without Red-Ox-System Oxidized with HClO <sub>4</sub>	HClO <sub>4</sub>	Ion Exchange 12 M HCl- System	IAEA/SM 149 (71)

<sup>+</sup> Spiked sample solutions, which contained NH<sub>2</sub>OH, were evaporated twice with 16 M HNO<sub>3</sub> before the chemical treatment to destroy organic residues.



Table 6-1 cont.

Code Nr.	Reduction	Oxidation	Separation	Lit. Data
13	NH <sub>2</sub> OH · HCl no comment	NaNO <sub>2</sub>	Ion Exchange 8 M HNO <sub>3</sub> <sup>-</sup> System	without literature data
14	HClO <sub>4</sub> evaporation	HClO <sub>4</sub> -	Ion Exchange 12 M HCl- System	without literature data
15	NH <sub>2</sub> OH · HCl allow to stand for 20 min by 60°C	NaNO <sub>2</sub>	Solvent Extraction HNO <sub>3</sub>	without literature data
17	ferrous solution allow to stand for 5 min	NaNO <sub>2</sub> -	Ion Exchange 7.5 M HNO <sub>3</sub> <sup>-</sup> System	ASTM E 321-67 T
18	Fe <sub>2</sub> -Ion and NH <sub>2</sub> OH no comment	NaNO <sub>2</sub>	Ion Exchange 7-8 M HNO <sub>3</sub> <sup>-</sup> System	without literature data
19	NH <sub>2</sub> OH · HCl ferrous sulfamat stabilized	NaNO <sub>2</sub>	Solvent Extraction	
20	ferrous solution NaNO <sub>2</sub> allow to stand for 5 min	NaNO <sub>2</sub>	Ion Exchange 7.5 M HNO <sub>3</sub> <sup>-</sup> System	ASTM E 321-67 T
21	NH <sub>2</sub> OH · HCl no comment	HNO <sub>3</sub>	Ion Exchange 7 M HNO <sub>3</sub> <sup>-</sup> System	without literature data
23	-	-	no separation	without literature data

It need not be stressed that with the small analytical quantities involved there is a particularly high risk of cross contamination.

In this report an attempt is made to analyse the results with a view to these two sources of error.

Since errors in the isotope exchange are anticipated in spiked solutions only and, as already said, the exchange of plutonium isotopes raises some problems, only results for plutonium in spiked solutions will be discussed here.

Tables 6-2 to 6-4 show the atom ratios Pu-242/239, 240/239, and 241/239 listed together with the respective relative standard deviations RSD of the laboratory mean values - calculated from the values of the three "runs" - and the differences RD between the laboratory mean values and the mean of all laboratory mean values relative to this mean of all mean values. The means of all laboratory mean values are calculated without the values in brackets. The reason for this omission was that in these cases the ratio Pu-242/239 was an outlier for the laboratory in question and it seemed meaningful to do the calculation for the same group of laboratories for all the isotopic ratios. For a better understanding of the results, the correlation of the atom ratio and the relative standard deviations of the three "runs" for the different laboratories are shown in Figures 6-1 to 6-3.

It seems that three different sources of error can be concluded from these values.

In the case of laboratory 8 it is seen that the RSD of the laboratory mean varies much more than for other laboratories. The striking fact is that in this case all isotope ratios show deviations of the same order of magnitude.

If the possibility also exists to observe the deviation of the individual scans, it becomes obvious that they introduce a great error into the analysis. A great deviation of the scans should indicate poor performance of the mass spectrometer.

Evidently, the results of laboratory 21, sample  $R_s$ , reveal a different error. In this case the relative standard deviation within the laboratory for Pu-242/Pu-239 takes a very good value (0.10%). All other values determined statistically, such as the RSD of the run and even of the scans, are very good. When, however, the result is compared to the mean value calculated from the values submitted by the other laboratories, a deviation of 40% is found. Since the other isotope values for the  $R_s$  sample are also considerably shifted - Pu-240/239=100% and 241/239=337% - this should be a typical example of cross contamination which, however, can hardly be discovered by the laboratory itself.

Table 6-2 Sample A<sub>s</sub>

Lab Code	242/239	RSD (%) Lab Mean	RD (%) <sup>**</sup>	240/239	RSD (%) Lab Mean	RD (%) <sup>**</sup>	241/239	RSD (%) Lab Mean	RD (%) <sup>**</sup>
2	1.302	0.23	+ 0.00	0.2351	0.13	0.26	(0.1299)	0.93	2.36
3	1.307	0.07	0.38	0.2355	0.19	0.43	0.1263	0.35	- 0.47
4	(0.818)	8.82	-37.17	(0.2286)	0.51	- 2.52	(0.1230)	3.06	- 3.07
5	1.298	0.80	- 0.31	0.2340	0.20	- 0.21	0.1257	0.49	- 0.95
6	1.301	0.32	- 0.08	0.2353	0.10	0.34	0.1270	0.17	0.08
7	1.304	0.21	0.15	0.2351	0.16	0.26	0.1274	0.24	0.39
8	(0.935)	3.50	-28.42	(0.2328)	1.03	- 0.72	(0.1289)	1.31	1.58
10	1.321	0.43	1.46	0.2349	0.14	0.17	0.1281	0.50	0.95
12	1.304	0.18	0.15	0.2345	0.22	+ 0.00	0.1268	0.13	- 0.08
13	1.302	0.01	+ 0.00	0.2337	0.06	- 0.34	0.1261	0.21	- 0.63
14	1.309	0.07	0.54	0.2341	0.13	- 0.17	0.1279	0.17	0.79
15	1.293	0.23	- 0.69	0.2337	0.09	- 0.34	0.1273	0.02	0.32
17	1.297	0.13	- 0.38	0.2348	0.44	0.13	0.1263	0.40	- 0.47
18	1.304	0.04	0.15	0.2341	0.15	- 0.17	0.1267	0.18	- 0.16
19	1.306	0.10	0.31	0.2351	0.05	0.26	0.1268	0.19	- 0.08
20	1.310	0.06	0.61	0.2336	0.25	- 0.38	0.1251	0.53	- 1.42
21	1.277	0.22	- 1.92	0.2344	0.15	- 0.04	0.1255	0.18	- 1.10
	1.302 <sup>*</sup> )			0.2345 <sup>*</sup> )			0.1269 <sup>*</sup> )		

\* ) Mean of laboratory means calculated without values in brackets. For the ratio 241/239 the value from laboratory 2 was also omitted because this laboratory had not performed a separation of Am.

\*\* ) Difference between laboratory mean and mean of all laboratory means relative to the mean of all laboratory means.

Table 6-3 Sample B<sub>s</sub>

Lab Code	242/239	RSD (%) Lab Mean	RD (%) <sup>**</sup>	240/239	RSD (%) Lab Mean	RD (%)	241/239	RSD (%) Lab Mean	RD (%)
2	1.302	0.33	0.85	0.2350	0.05	0.34	(0.1344)	0.43	6.25
3	1.295	0.05	0.31	0.2364	0.20	0.94	0.1269	0.13	0.32
4	(0.898)	9.12	- 30.44	(0.2348)	0.34	0.26	(0.1264)	0.31	- 0.08
5	1.281	0.55	- 0.77	0.2336	0.33	- 0.26	0.1251	0.48	- 1.11
6	(1.166)	0.34	- 9.68	(0.2333)	0.71	- 0.38	(0.1262)	0.97	- 0.24
7	1.286	0.21	- 0.39	0.2327	0.24	- 0.64	0.1256	0.42	- 0.71
8	(1.122)	3.30	- 13.09	(0.2375)	0.61	1.41	(0.1276)	1.36	0.87
10	1.300	0.21	0.70	0.2340	0.08	- 0.09	0.1270	0.23	0.40
12	1.288	0.15	- 0.23	0.2342	0.04	+ 0.00	0.1268	0.09	0.24
13	1.291	0.03	+ 0.00	0.2341	0.03	- 0.04	0.1259	0.26	- 0.47
14	1.291	0.16	+ 0.00	0.2339	0.17	0.13	0.1275	0.20	0.79
15	1.275	0.15	- 1.24	0.2325	0.63	- 0.73	0.1270	0.39	0.40
17	1.296	0.25	0.39	0.2345	0.10	0.13	0.1267	0.15	0.16
18	1.287	0.06	0.31	0.2335	0.14	- 0.30	0.1262	0.12	- 0.24
19	1.292	0.07	0.08	0.2351	0.07	0.38	0.1267	0.16	0.16
20	1.298	1.06	0.54	0.2355	1.53	0.56	0.1262	2.31	- 0.24
21	(1.206)	0.32	- 6.58	(0.2342)	0.12	+ 0.00	(0.1247)	0.25	- 1.42

1.291<sup>\*</sup>

0.2342

0.1265

<sup>\*</sup>) Mean of laboratory means calculated without values in brackets. For the ratio 241/239 the value from laboratory 2 was also omitted because this laboratory had not performed a separation of Am.

<sup>\*\*</sup>) Difference between laboratory mean and mean of all laboratory means relative to the mean of all laboratory means.

Table 6-4 Sample R<sub>s</sub>

Lab Code	242/239	RSD (%) Lab Mean	RD (%) <sup>**)</sup>	240/239	RSD (%) Lab Mean	RD (%)	241/239	RSD (%) Lab Mean	RD (%)
2	1.181	0.39	1.20	0.03324	0.16	0.73	(0.0025)	4.88	21.36
3	1.173	0.07	0.51	0.03324	0.16	0.73	0.0020	4.75	- 2.91
4	(1.115)	0.88	- 4.46	(0.03292)	1.19	- 0.24	-	-	-
5	1.154	0.21	- 1.11	0.03338	0.75	1.15	0.0021	1.60	1.94
6	(1.033)	1.13	-11.48	(0.03350)	0.12	1.52	(0.0020)	0.97	- 2.91
7	1.166	0.16	- 0.09	0.03275	0.91	- 0.76	0.0020	1.36	- 2.91
8	(1.067)	3.39	- 8.57	(0.03587)	4.37	8.70	(0.0023)	3.58	11.65
10	1.174	0.28	0.60	0.03289	0.09	- 0.33	0.0020	0.35	- 2.91
12	1.168	0.11	0.09	0.03290	0.19	- 0.30	0.0022	0.94	6.80
13	1.168	0.04	0.09	0.03277	0.33	- 0.70	0.0019	2.49	- 7.77
14	1.171	0.17	0.34	0.03312	0.23	0.36	0.0020	0.13	- 2.91
15	1.160	0.37	- 0.60	0.03288	0.40	- 0.36	0.0020	3.39	- 2.91
17	1.172	0.05	0.43	0.03294	0.26	- 0.18	0.0020	0.46	- 2.91
18	1.164	0.02	- 0.26	0.03291	0.08	- 0.27	0.0020	0.21	- 2.91
19	1.163	0.09	- 0.34	0.03303	0.13	0.09	0.0020	0.16	- 2.91
20	1.160	0.32	- 0.60	(0.03227)	0.41	- 2.21	(0.0017)	2.68	- 17.48
21	(0.701)	0.10	-39.93	(0.06662)	0.48	101.9	(0.0091)	0.14	337.2

1.167<sup>\*)</sup>

0.03300

0.00206

<sup>\*)</sup> Mean of laboratory means calculated without values in brackets. For the ratio 241/239 the value from laboratory 2 was also omitted because this laboratory had not performed a separation of Am.

<sup>\*\*)</sup> Difference between laboratory mean and mean of all laboratory means relative to the mean of all laboratory means.

Fig.6-1 Relative Standard Deviations of the Laboratory Mean Values





Fig. 6-2 Relative Standard Deviations of the Laboratory Mean Values






Fig. 6-3 Relative Standard Deviations of the Laboratory Mean Values

Sample  $R_s$   
 Means of lab. means for the isotopic ratios  
 ————  $242/239 = 1.167$   
 - - - - -  $240/239 = 0.03300$   
 .....  $241/239 = 0.00206$



The results from laboratory 4, especially of the samples A<sub>s</sub> and B<sub>s</sub>, differ in another way with respect to their deviations. The relative standard deviation for the Pu-242/239 isotope ratio amounts to 8.8 and 9.1 %, respectively, within the laboratory, while the other isotope ratios show RSD values of 0.51 % and 0.34 % for Pu-240/239 and of 3.06 % and 0.31 % for Pu-241/239. In this case the analysis of variances for the single measurements of the Pu-242/239 ratio gives 0.5 % and 0.8 %, respectively, for the scan components and 15.3 % and 15.8 % for the run components of the RSD. Here it is noticeable that the deviations of the runs amount to 0.8 % and 0.4 %, respectively, for the Pu-240/239 isotope ratios, i.e., they attain the usual orders of magnitude, so that these results are an indication of unsatisfactory isotope exchange.

Similar results are reported by another laboratory. Therefore, the comments will be quoted here:

"While the analysis of all the samples in the standard experiment went on smoothly there was some problem in the Pu-242/Pu-239 ratio of plutonium samples A (spiked) and B (spiked). Following the ASTM-E-321-67<sup>T</sup> procedure, the three separate aliquots from each sample were subjected to the procedure which consisted of

- (1) addition of ferrous ammonium sulphate
- (2) addition of sodium nitrite and
- (3) anion exchange separation in 7.5 M nitric acid.

The mass spectrometric scans of plutonium showed that two chemical preparations from sample A (spiked) and two from sample B (spiked) gave atom ratios Pu-242/Pu-239 which agreed well between themselves. The third preparations from both samples gave very different Pu-242/Pu-239 ratios as can be seen from the separate data sheets. In the fourth chemical preparation both these samples were prepared following the same procedure but more carefully and the Pu-242/Pu-239 ratio was again different. This tempted us to suspect that either the ASTM procedure was inadequate or Pu-242 in the dissolver solutions did not completely exchange with the other isotopes, since the values of only Pu-242/Pu-239 had shown difference (and other ratios were alright). Therefore, we decided to subject the small portions of the samples still remaining in the vials to a thorough procedure which would bring about depolymerisation of any polymer that may exist and also complete exchange between

the spike Pu-242 and other isotopes in case it was not complete earlier. To achieve this objective, ferrous ammonium sulphate and sodium nitrite were added to the vials and the solutions evaporated, concentrated nitric acid was added and the ferrous-nitrite treatment was repeated before anion exchange separation was carried out. The values of Pu-242/Pu-239 obtained from these solutions agreed very well with the values in the first two analyses. This shows that there was no polymer sticking on to the walls of the vial. We do not completely understand the reason for this type of results but the values of Pu-242/Pu-239 obtained in first, second and fifth chemical separations seem to be reliable and correct since the last sample was obtained by directly treating the solution in the vial itself. We would have to wait till we know the experience and results of others."

The laboratory did well to distrust the two values. If in the case of the ratio Pu-242/Pu-239 all five runs instead of three runs are taken into consideration the run component of the RSD rises from 0.13 % to 12 % for sample A<sub>s</sub> and from 0.41 % to 11 % for sample B<sub>s</sub>. In contrast to this the RSD components for the determination of the ratio Pu-240/Pu-239 remain in the same order of magnitude also if one includes all five runs, i.e. the values are 0.74 % (3 runs) and 0.89 % (5 runs) for the sample A<sub>s</sub> and 0.13 % (3 runs) and 0.16 % (5 runs) for the sample B<sub>s</sub>. Since the spike solution contained practically no Pu-240 and this effect remains restricted to the ratio Pu-242/Pu-239 it is evident that the difficulty lies in the mixture of sample and spike plutonium.

It is not easy to interpret the results provided by laboratory 6. It is surprising that for the high Pu-242/239 isotope ratio the relative standard deviation of the scans is of the order of magnitude of 2.3 to 3.3 % for all spiked samples \*). This laboratory gave the information that the amount of available Pu was much smaller than normally used, which explains the moderate reproducibility. This depends on the quality of the mass-spectrometer. Nevertheless, the results of the sample A<sub>s</sub> is good. With respect to the sample B<sub>s</sub> the RSD is equally good (0.33 %) for the Pu-242/239 ratio within the laboratory, so that in this case there was

---

\*) It should be noted that for the laboratory the normal numerical analysis of scan values is not fully adequate since the special scanning procedure of this laboratory would actually require a grouping of values before the calculations.

neither a reason for the laboratory to reject the values; also the analysis of variances gave negligible error components for the runs. Contrary to that, the respective values for the  $R_s$ -sample were of the order of 1 % for the high abundant Pu-242 isotope which means that the results of this very laboratory show the limits set to the assignment of errors.

As appears from the short report, it is difficult to establish a relationship between the chemical working technique and the results. It seems, however, that somewhat better information can be obtained by statistical evaluation of the individual results.

Major deviations of the individual values revealed by the relative standard deviation of the scans might be an indication of faulty performance of the mass spectrometer.

Large variations between the individual runs, which are observed only between the spike isotope and the main isotope, should in most cases be attributable to poor exchange of isotopes. Cross contamination should be the most difficult thing to discover inside one laboratory.

7. Corrections for Mass Discrimination

by

W. Beyrich

EURATOM, delegated to the Institute for Applied Systems  
Analysis and Reactor Physics, Gesellschaft für Kernforschung,  
Karlsruhe, Germany

## 7.1 Introduction

At the beginning of the IDA-experiment, the participating laboratories were requested to report the isotopic ratios corrected for mass discrimination according to the method normally used,

As far as concentration determinations are concerned, errors due to mass discrimination are compensated under the condition that the spike solution is calibrated by the same individual laboratory which performs the analysis. With the exception of the "Self spike experiment", this was not the case in IDA-72. Nevertheless, in the evaluation of the other parts of the experiment, compensation of this correction factor could be obtained by considering the results of the A- and B-solutions relative to those of the R-sample /Vol. I, Par. 3.5.1/.

However, corrections for mass discrimination are always necessary

- if the spike solution used for the isotope dilution analysis is not calibrated by the laboratory itself and if no analysis of a suitable known reference sample exists, and
- for all isotopic ratio or relative isotopic abundance determinations.

The exact determination of the correction factors for plutonium is specifically complicated by the lack of suitable isotopic standards for this element.

In this context it should be mentioned that the error caused by the mass discrimination effect (in the wide sense as it is usually understood) can be corrected only partly by the calibration with standards, as errors due to changes in the instrumental operating conditions from the filament loading with the standard to the one with the sample remain unaffected /Vol. I, Chapter 8/.

## 7.2 Inquiry on the Methods Used.

In order to obtain a survey on the methods used by the individual laboratories questionnaires were sent to the participants. They were answered by 17 laboratories. The questions together with the information obtained is given in the following. It corresponds to the final status on which this report is based. It was not possible to indicate the individual laboratories which used the various methods as this would lead to a partial revealing of the codes.

A. Were the isotopic ratios reported for IDA-72 corrected for mass discrimination?

	Yes	No	No information
Number of laboratories:	17 <sup>*</sup> )	-	2 (labs 16 + 23)

B. Are the correction factors applied in your laboratory considered as independent of the magnitude of the isotopic ratio?

	Yes	No	No information
Number of laboratories:	10	2	7

C. Uranium measurements:

Correction methods used:	Number of laboratories
a) Calibration with NBS-standards (005, 010, 020, 050, 350, 500, 930 and 950 A were reported)	12
b) Correction with $\sqrt{\frac{m_1}{m_2}}$	1
c) No information on the method applied	4

---

<sup>\*</sup>) Laboratories 4 and 20 for U-measurements only.

The values reported for the correction of the U-233/U-238 ratio are:

- 1.0 % in average,

- 2.0 % at maximum.

D. Plutonium measurements.

Correction methods used:	Number of laboratories
a) Calibration with U-standards:	7
b) Calibration with NBS Pu-948	2
c) Correction with $\sqrt{\frac{m_1}{m_2}}$	2
d) No information on the method applied	4

The values reported for the correction of the Pu-242/Pu-239 ratio are:

+ 0.6 % in average

+ 1.2 % at maximum.

8. Determination of Pu-238 by  $\alpha$ -Spectrometry  
for the IDA - 72 Experiment

by

A. Cricchio

European Institute for Transuranium Elements, EURATOM,  
Karlsruhe, Germany



## 8.1 Introduction

In the framework of the IDA-72 experiment, it was decided to perform  $\alpha$ -spectrometric determination of Pu-238 in each of the following spiked and unspiked samples:

- active feed solution A
- active feed solution B
- reference solution R

The laboratories participating in the experiment were asked to perform three repetitive measurements for the unspiked samples  $A_u$ ,  $B_u$  and  $R_u$ . In the case of the spiked samples  $A_s$ ,  $B_s$  and  $R_s$ , the laboratories had to divide the samples into three parts and to perform one measurement for each of the three parts. Consequently a total of six results were given by each of the 11 laboratories which performed the measurements.

This chapter reports and discusses the procedures and the results of the Pu-238  $\alpha$ -spectrometric determination.

## 8.2 Procedures

A questionnaire was sent to the laboratories involved in the  $\alpha$ -spectrometric measurements for information on procedures and instruments used. The questions dealt with:

- method of alpha source preparation
- material used for the source support
- type and manufacturer of the instrument used
- detector resolution (FWHM)
- reproducibility of the measurements.

From the answers it results that the most frequent method used for the preparation of the alpha source is the direct evaporation of a solution drop (7 laboratories). Three laboratories used the electro-deposition method; one laboratory (code 18) used the method of Kirby.

As source support, six laboratories used stainless steel discs, three tantalum discs and two platinum ones.

Pre-amplifiers, linear amplifiers and multi-channel pulse height analyzers were used by the laboratories for the measurements. The equipment is commercially available with the exception of a few laboratories which used home made amplifiers.

Surface barrier detectors were used by all the laboratories. The detector resolution is, on average, about 20 KeV at 5.50 MeV. Detectors with lower resolution ( $\sim$  40 KeV) were used by three laboratories, one laboratory has used a very high resolution detector (8-9 KeV).

Only a few laboratories answered the question concerning the reproducibility of the measurements. Some laboratories have given a relative average deviation of  $\pm 1$  %.

### 8.3 Results

As it is known, the alpha decay energies of Pu-239 and Pu-240 overlap so that, after separation of the Am-241, only the alpha activity ratios Pu-238/(Pu-239 + Pu-240) can be determined. However, by use of the ratio Pu-240/Pu-239 determined by mass spectrometry, the isotopic ratio Pu-238/Pu-239 can be calculated.

For each laboratory, the mean of the alpha activity ratio Pu-238/(Pu-239 + Pu-240) and the Pu-238/Pu-239 isotopic ratio are listed in the report on the evaluation of data /Volume I, Par. 3.3.2/. Here, for brevity, for the alpha activity ratio and the calculated isotopic ratio, only the means of laboratory means together with the precision and the interlaboratory deviation are reported (Table 8-1).

A discussion of the results is given in the next paragraph.

Tab.8-1 Results of the Alpha Spectrometric Measurements.

Sample	Mean of lab. means for $\alpha$ -activity ratio $\frac{\text{Pu-238}}{\text{Pu-239+Pu-240}}$	Calculated isotopic ratio  Pu-238/Pu-239	Precision  RSD [%]	Interlab. deviation  RSD [%]
A and B unspiked	2.191	0.01460	1.13	0.66
A and B spiked	2.145	0.01440	1.41	1.14
R unspiked	0.0117	0.00005	4.08	33.1
R spiked	0.0243	0.00010	4.49	15.7

Values calculated after rejection of outlier values as described  
in Vol. I, Par. 3.3.2.

#### 8.4 Discussion

From the data reported in Table 8-1, it is evident that, in general, higher precision and lower interlaboratory deviations are obtained for the unspiked samples A, B and R. Moreover, a discrepancy between the alpha activity ratio for spiked and unspiked samples can be noted. The discrepancy, according to the layout of the standard experiment, cannot be attributed totally to the Pu-239 and Pu-240 content in spiked samples.

Generally speaking, the methods and procedures used by the different laboratories for the Pu-238 determination are equivalent and, therefore, should not be the source of very important differences. Nevertheless, some deviations may be caused by the different methods of treatment of alpha spectra data. Unfortunately, this point was not included in the questionnaire so that no information is available for further discussion.

The larger deviation observed for the spiked samples may be explained, perhaps, by the fact that these samples were divided into three parts. For each part a separate treatment was performed. In this case the risk of a cross contamination of the samples is increased; moreover, the Am-241 separation may be slightly different from sample to sample and, consequently, larger deviations can be obtained for the Pu-238 activity.

9. IDA-72: Concentration Determination by  
X-Ray Fluorescence Spectrometry

performed by

K. Matern <sup>1)</sup>

reported by

A.v. Baeckmann <sup>1)</sup>, W. Beyrich <sup>2)</sup>, E. Drosselmeyer <sup>3)</sup>

---

<sup>1)</sup> Institute of Radiochemistry (IRCh), Gesellschaft für Kernforschung (GfK), Karlsruhe, Germany

<sup>2)</sup> EURATOM, delegated to the Institute of Applied Systems Analysis and Reactor Physics (IASR), GfK, Karlsruhe, Germany

<sup>3)</sup> IASR of GfK, Karlsruhe, Germany

In addition to the A, B and C samples used for the standard and aging experiments four further samples E<sub>1</sub> to E<sub>4</sub> were taken at EUROCHEMIC in the same sampling procedure on June 21st, 1972 /Chapter 2, Fig. 2-2/. On July 5th, the Institute of Radiochemistry of the GfK determined the uranium and plutonium concentrations on this undiluted sample material by X-ray fluorescence spectrometry. These analyses were performed without any chemical separation within one day /7/.

The laboratory reported 6 determinations for each of both elements, indicating one of the uranium determinations as an outlier. The means of these X-ray determinations and their relative standard deviations are compiled in the table together with the mean of the values calculated from the "best" concentration values obtained in the standard experiment by isotope dilution analyses. For the calculation of the concentrations of the undiluted sample material, the data on the dilution procedure, given in Chapter 2, Fig. 2-2 were used. Reference is also made to Vol. I, Par. 3.5.6.

All values are valid for 22°C (density 1.3790 g/ml).

Table: Concentration of Undiluted Sample Material

	Uranium		Plutonium	
	mean concentration [mg U/g sol.]	RSD of mean [%]	mean concentration [μg Pu/g sol.]	RSD of mean [%]
Standard exp. (mean of samples A and B)	164.11		1310.6	
x-ray fluorescence spectrometry	166.79	0.14	1309.1	0.25
Rel. deviation %	+ 1.6		- 0.1	

The positive deviation of the X-ray value for uranium might indicate that there was evaporation due to radiolysis of the high active undiluted sample material during the storage and transportation time of two weeks. However, this explanation does not fit with the good agreement of the plutonium results which are also confirmed by the process analysis of EUROCHEMIC /Vol. I, Par. 3.5.6/.

10. The Application of Outlier Criteria in the  
Evaluation of Measuring Results for IDA-72

by

G. Bork and H. Frick

Institute for Applied Systems Analysis and Reactor Physics,  
Gesellschaft für Kernforschung  
Karlsruhe, Germany



Table of Contents

10.1	Introduction
10.2	Theoretical Considerations
10.2.1	Dixon Criterion
10.2.2	Chauvenet Criterion
10.2.3	Comparison
10.3	Practical Application of
10.3.1	Dixon Criterion
10.3.2	Chauvenet Criterion
10.4	Numerical Results of the Comparison of the Dixon and Chauvenet Criterion
10.5	Resulting Considerations for the Application of Outlier Criteria to the Data of the IDA-72 Experiment

## 10.1 Introduction

At the begin of planning the IDA-72 experiment it was necessary to decide how many repetitions of measurements would be necessary and possible in order to allow a meaningful statistical evaluation under the constraint of reasonable efforts.

It was quite clear that the only point in the course of evaluation where a larger number of repetitions was possible was the number of mass spectrometer scans made for one filament loading (run). Already the number of such filament loadings would have to be limited to a maximum of three.

For the number of scans, i.e. of single determinations of all the isotopic ratios of one sample of a U or Pu solution loaded on a filament, it was decided to take 10 as a reasonable compromise.

This number would also allow to check the given series for wrong numbers which could originate from mistakes in reading or/and transmitting the data or other sources of error.

Since we did not know how often such mistakes would occur we decided to evaluate finally 8 scan values in order to have an orthogonal system which is important for the aim of analysis of variances. By this decision we had the possibility to omit outlier values or to take a random choice in order to reduce the number of 10 given values to 8 accepted values.

For discrimination of outlier the two criteria of Dixon /8-10/ and Chauvenet /11-14/ were taken into account. Other criteria are reported in Ref. /15/ and /16/. If no outlier could be detected just the first 8 of the 10 given values were taken - which means a random choice since the given measuring results were not ordered.

For studying the influence of the application of the outlier criteria the measurement results of 5 laboratories for the standard experiment were taken. This means that about 25 % of all given values were the basis for the decision which criterion should be applied for the whole data material.

The aim of the application of outlier criteria was defined as to detect strongly deviating measurement results with a low probability of error. Since it is the general objective of the experiment to analyse the accuracy of the method and the sources of errors it was important not to simulate a false precision by omission of too many "outliers".

## 10.2 Theoretical Considerations

### 10.2.1 Dixon Criterion

It is assumed that the measurements  $x_i$  are normally distributed. These values  $x_i$  are put into ascending order:  $x_1 < x_2 \dots < x_n$ . For test purposes quotients  $r_{ij}$  of differences between these ordered values involving  $x_n$  or  $x_1$  are used (see below). The distribution functions of  $r_{ij}$  depend on  $n$  and cannot be given by an explicit expression. The criterium for elimination of a value is  $r_{ij} > r_{1-\alpha}$ .  $r_{1-\alpha}$  is the significance threshold and is determined by  $F_{r_{ij}}(r_{1-\alpha}) = \alpha$  for a given probability of error  $\alpha$ . In this context "error" means that a measurement is excluded by mistake in spite of its belonging to the series.  $F_{r_{ij}}$  are the distribution functions of  $r_{ij}$ .

### 10.2.2 Chauvenet Criterion

Also in this case it is assumed that the measurements  $x_i$  are normally distributed. Their expectation value shall be  $\mu$  and the variance  $\sigma^2$ . In this case the expression  $\frac{x_i - \mu}{\sigma}$  has a standardized normal distribution, i.e.

$$E\left(\frac{x_i - \mu}{\sigma}\right) = 0 \quad \text{and}$$

$$\text{var}\left(\frac{x_i - \mu}{\sigma}\right) = 1$$

The distribution function of  $\frac{x_i - \mu}{\sigma}$  is  $\Phi = \frac{1}{\sqrt{2\pi}} \int_{-\infty}^x e^{-\frac{t^2}{2}} dt$ . The probability that  $\frac{x_i - \mu}{\sigma}$  deviates from 0 by more than  $x_c$  is given by

$$\begin{aligned} P\left(\left|\frac{x_i - \mu}{\sigma}\right| > x_c\right) &= 1 - P\left(-x_c \leq \frac{x_i - \mu}{\sigma} \leq x_c\right) \\ &= 1 - (\Phi(x_c) - \Phi(-x_c)) \\ &= 1 - (2\Phi(x_c) - 1) \\ &= 2(1 - \Phi(x_c)) \end{aligned}$$

A measurement is rejected if  $\left| \frac{x_i - \mu}{\sigma} \right| > x_c$ . The probability that the condition  $\left| \frac{x_i - \mu}{\sigma} \right| > x_c$  is fulfilled, which means that also a correct measurement will be eliminated by mistake, is named  $P$ . For this latter reason it is also called probability of error. By the definition it is clear that  $P$  corresponds to the probability of error  $\alpha$  defined in the case of the Dixon criterion. It is a convention to put  $P = \frac{1}{2n}$  where  $n$  is the number of measurements  $x_i$ . So we have:

$$\frac{1}{2n} = 2 (1 - \Phi(x_c))$$

and

$$x_c = U \left( 1 - \frac{1}{4n} \right)$$

where  $U$  is the inverse function of  $\Phi$ .

### 10.2.3 Comparison

The premise for the application of the Chauvenet criterion is the knowledge of  $\mu$  and  $\sigma$ . This condition is normally not fulfilled and, therefore,  $\mu$  is replaced by  $\bar{x} = \frac{1}{n} \sum_{i=1}^n x_i$  and  $\sigma^2$  by  $s^2 = \frac{1}{n-1} \sum_{i=1}^n (x_i - \bar{x})^2$  as approximation. This procedure seems plausible because the expectation value of  $\bar{x}$  is

$$E(\bar{x}) = \mu,$$

and that of  $s^2$  is

$$E(s^2) = \sigma^2.$$

It seems doubtful whether or not this procedure is justified in cases like ours, where only 10 values  $x_i$  exist and so one strongly deviating value  $x_i$  has a big influence on  $\bar{x}$  and  $s^2$ . However, when the Dixon criterion is to be applied the knowledge of mean value and variance of the distribution of the measurements is not necessary since they do not affect the distribution function of the  $r_{ij}$ .

### 10.3 Practical Application of the Dixon and Chauvenet Criterion

#### 10.3.1 Dixon Criterion

The rules for the application of this criterion are given in Ref. /8/.

In the following the application procedure is described stepwise:

1. step: The measurement values are ordered so that

$$x_1 \leq x_2 \leq \dots \leq x_n$$

2. step: A probability of error  $\alpha$  that a value is excluded by mistake is chosen.

3. step: Corresponding to the number of measurements  $n$  the following quotients of differences  $r_{ij}$  are calculated:

For	$r_{ij}$
$3 \leq n \leq 7$	$r_{10}$
$8 \leq n \leq 10$	$r_{11}$
$11 \leq n \leq 13$	$r_{21}$
$14 \leq n \leq 25$	$r_{22}$

with the following expressions for the  $r_{ij}$

$r_{ij}$	$x_1$ suspicious	$x_n$ suspicious
$r_{10}$	$\frac{x_2 - x_1}{x_n - x_1}$	$\frac{x_n - x_{n-1}}{x_n - x_1}$
$r_{11}$	$\frac{x_2 - x_1}{x_{n-1} - x_1}$	$\frac{x_n - x_{n-1}}{x_n - x_2}$
$r_{21}$	$\frac{x_3 - x_1}{x_{n-1} - x_1}$	$\frac{x_n - x_{n-2}}{x_n - x_2}$
$r_{22}$	$\frac{x_3 - x_1}{x_{n-2} - x_1}$	$\frac{x_n - x_{n-2}}{x_n - x_3}$

4. step: From the following table the numerical value  $r_{1-\alpha}$  for the specific number of measurements is taken:

Table for  $r_{1-\alpha}$

$r_{ij}$	n	$1-\alpha=$ .70	$1-\alpha=$ .80	$1-\alpha=$ .90	$1-\alpha=$ .95	$1-\alpha=$ .98	$1-\alpha=$ .99	$1-\alpha=$ .995
$r_{10}$	3	.684	.781	.886	.941	.976	.988	.994
	4	.471	.560	.679	.765	.846	.889	.926
	5	.373	.451	.557	.642	.729	.780	.821
	6	.318	.386	.482	.560	.644	.698	.740
	7	.261	.344	.434	.507	.586	.637	.680
$r_{11}$	8	.318	.385	.479	.554	.631	.683	.725
	9	.288	.352	.441	.512	.587	.635	.677
	10	.265	.325	.409	.477	.551	.597	.639
$r_{21}$	11	.391	.442	.517	.576	.638	.679	.713
	12	.370	.419	.490	.546	.605	.642	.675
	13	.351	.399	.467	.521	.578	.615	.649
$r_{22}$	14	.370	.421	.492	.546	.602	.641	.674
	15	.353	.402	.472	.525	.579	.616	.647
	16	.338	.386	.454	.507	.559	.595	.624
	17	.325	.373	.438	.490	.542	.577	.605
	18	.314	.361	.424	.475	.527	.561	.589
	19	.304	.350	.412	.462	.514	.547	.575
	20	.295	.340	.401	.450	.502	.535	.562
	21	.287	.331	.391	.440	.491	.524	.551
	22	.280	.323	.382	.430	.481	.514	.541
	23	.274	.316	.374	.421	.472	.505	.532
	24	.268	.310	.367	.413	.464	.497	.524
	25	.262	.304	.360	.406	.457	.489	.516

5. step: The suspicious measurement result  $x_1$  resp.  $x_n$  is regarded as an outlier, if  $r_{ij} > r_{1-\alpha}$ .

For example: The series of measurement results is, already ordered:

$$x_1 = 0.3850$$

$$x_2 = 0.3898$$

$$x_3 = 0.3928$$

$$x_4 = 0.3931$$

$$x_5 = 0.3935$$

$$x_6 = 0.3949$$

$$x_7 = 0.3958$$

$$x_8 = 0.3961$$

$$x_9 = 0.3962$$

$$x_{10} = 0.3998$$

$$n = 10$$

$$\bar{x}_{(10)} = 0.3937$$

First of all the value  $x_1$  seems suspicious.

Since  $n = 10$ , the expression  $r_{11} = \frac{x_2 - x_1}{x_{n-1} - x_1}$

has to be calculated. It becomes

$$r_{11} = \frac{0.3898 - 0.3850}{0.3962 - 0.3850} = 0.429 .$$

If the probability of error is given by  $\alpha = 0.1$  one has to compare this value of  $r_{11}$  with the value  $r_{1-\alpha} = 0.409$  from the table given in step 4. In this example is  $r_{11} > r_{1-\alpha}$  so that the value  $x_1$  is regarded as an outlier.

The next suspicious value is  $x_{10}$ .

The calculation yields

$$r_{11} = \frac{x_{10} - x_9}{x_{10} - x_2} *) = \frac{0.3998 - 0.3962}{0.3989 - 0.3928} *) = 0.514 .$$

---

\*) This value was named  $x_3$  in the original series of results. After rejection of  $x_1$  the series which is checked now begins with the value originally named  $x_2$  .

$$\bar{X}_{(10)} = 0.3937$$

$$\bar{X}_{(9)} = 0.3946$$

$$\bar{X}_{(8)} = 0.3940$$




Fig.10-1: Example for the Application of the Dixon Criterion.



In the table  $r_{1-\alpha} = 0.441$  is found for a series consisting of the residual 9 values. Since  $r_{11} > r_{1-\alpha}$  also for this case, also the value  $x_{10}$  is regarded as an outlier.

From Fig 10-1 one could guess that also the value  $x_2$  could be an outlier. The calculation leads to

$$r_{11} = \frac{x_3 - x_2}{x_8 - x_2} = \frac{0.3928 - 0.3898}{0.3961 - 0.3898} = 0.476$$

The corresponding  $r_{1-\alpha}$  for 8 values is 0.479 so that in this case  $r_{11} < r_{1-\alpha}$  ; i.e.  $x_2$  is not regarded as an outlier.

### 10.3.2 Chauvenet - Criterion

In the following also the application of this criterion is described stepwise.

1. step : The algebraic mean of the n given measurement results is calculated:

$$\bar{x} = \frac{1}{n} \sum_{i=1}^n x_i$$

2. step : The corresponding standard deviation is calculated:

$$s = \sqrt{\sum_{i=1}^n (\bar{x} - x_i)^2 / (n-1)}$$

3. step : The measurement for which the expression  $|\bar{x} - x_i|$  gives the highest value is named  $x^*$  and is looked upon as suspicious.

4. step : A value  $x_c$  has to be found which is determined by  $\Phi(x_c) = 1 - \frac{1}{4n}$  .

The function

$$\Phi(x) = \frac{1}{\sqrt{2\pi}} \cdot \int_{-\infty}^x e^{-\frac{t^2}{2}} dt$$

is found in several collections of mathematical tables, as e.g. in Ref [17],

5. step : If  $\left| \frac{x^* - \bar{x}}{s} \right| > x_c$  ,

the value  $x^*$  is considered as an outlier.

There exists another form of the rules given above. In this modification the suspected value  $x^*$  is not included in the calculation of algebraic mean and standard deviation according to steps 1 and 2. By doing so the difference between the suspicious value and the calculated "mean value" is growing greater and the standard deviation smaller so that the criterion becomes more sensitive (Chauvenet II).

Example:

As an example for the application of the Chauvenet criterion the same series of measuring results is taken, which was used as an example for the application of the Dixon criterion in Par. 10.3.2.

The values are - in original order

0.3949  
 0.3961  
 0.3935  
 0.3962  
 0.3998  
 0.3850  
 0.3958  
 0.3898  
 0.3931  
 0.3928

The calculation is done for both versions of the Chauvenet criterion given above.

	Chauvenet I	Chauvenet II
1. step	$\bar{x}_{(10)} = 0.3937$	$\bar{x}_{(9)} = 0.3946$
2. step	$s_{(10)} = 0.0040$	$s_{(9)} = 0.0030$
3. step	$x^* = 0.3850$ $\bar{x}_{(10)} - x^* = 0.0087$	$x^* = 0.3850$ $\bar{x}_{(9)} - x^* = 0.0096$
4. step	$\Phi(x_c) = 1 - \frac{1}{40} = 0.9750$ $x_c = 1.960$	$\Phi(x_c) = 1 - \frac{1}{36} = 0.9722$ $x_c = 1.914$
5. step	$\frac{0.0087}{0.0040} = 2.177 > 1.96$	$\frac{0.0096}{0.0030} = 3.20 > 1.914$

The suspicious value proves to be an outlier in both cases.

The same calculation is then carried out for the largest measurement value (in this calculation the outlier found before is excluded) :

Chauvenet I	Chauvenet II
$x^* = 0.3998$	$x^* = 0.3998$
$\bar{x}(9) = 0.3946$	$\bar{x}(8) = 0.3940$
$s(9) = 0.0028$	$s(8) = 0.0021$
$x_c = 1.914$	$x_c = 1.862$
$\frac{x^* - \bar{x}}{s} = 1.8571 < x_c$	$\frac{x^* - \bar{x}}{s} = 2.7619 > x_c$
The suspicious value is <u>no</u> outlier.	The suspicious value <u>is</u> an outlier.

From this example it becomes evident that the version "Chauvenet II" is the stricter criterion.

#### 10.4 Numerical Results of the Comparison of the Dixon and Chauvenet Criteria.

To begin with, a computer program was written which allowed the application of the three possible criteria (Dixon, Chauvenet I and II). First the difference between the two versions of the Chauvenet criterion was tested. As a result of the application of these criteria to 120 series of measurements it came out that for the version I (suspicious value included in the calculation of mean value and standard deviation) one has the following result:

15.5 %	of the series show	1 outlier
8.1 %	"	2 outlier
2.4 %	"	3 or more outliers.

In contrast to this for the version II (suspicious value excluded from the calculation of mean value and standard deviation) one gets 3 or more outliers for more than 50 % of these same 120 series.

The empirical comparison of the effect of application of the Dixon criterion (with  $\alpha = 0.1$ ) and the Chauvenet criterion (Version I) to the results of 5 laboratories for the standard experiment concerned 603 series of measurements and showed the following picture:

No. of series of measurements	Type of sample	No. of outliers Dixon (a)	No. of outliers Chauvenet I (b)	b/a
253	Unspiked sample solution	79	95	1.20
350	Spiked sample solution	68	125	1.84
Total; 603		147	220	1.50

At a first view this result is astonishing since one could have expected from a comparison of the probabilities of errors (Dixon :  $\alpha = 0.1$  ; Chauvenet :  $P = 0.050$  for  $n=10$ ,  $P = 0.056$  for  $n=9$ , and  $P = 0.063$  for  $n=8$ ) that the application of the Dixon criterion would yield more outliers than the application of the Chauvenet criterion.

The explanation could be that in the case of the Chauvenet criterion 10, 9 or 8 measurements have an influence, whereas in the case of the Dixon criterion only 3 measurements are taken into account for the single decision.

For all measurements of one laboratory the influence of the elimination of outliers according to both criteria on the mean values of each series was studied. The resulting shift in the mean values varies between 0.01 and 0.5 % in relation to the mean values of the uncorrected series of measurements. This large spread is easily explained by the fact that the influence on the mean value is small in cases where maximum and minimum are outliers. In all cases where one or two measurements at one end of the series are excluded, the influence of this elimination is relatively large. Apart from this it comes out that in approximately one third of the studied cases the mean values of the corrected series lie outside the region of the mean values of the two other parallel runs for this isotopic ratio and sample. This shows that without regarding the outliers one gets a wrong impression about the distribution of the run mean values.

#### Special Cases:

a) If the Chauvenet criterion is applied consecutively without the necessary precautions, it can result that too many values of one measurement series prove as outliers. This is shown in the following example:

The ordered sequence of measurements is:

2.274  
2.282  
2.289  
2.296  
2.300  
2.300  
2.301  
2.302  
2.305  
2.307

1.) At the beginning the minimum value is suspicious. The application of the Chauvenet criterion (see above) yields:

$$\begin{aligned}x_1^* &= 2.274 \\ \bar{x}(10) &= 2.2956 \\ s(10) &= 9.576 \cdot 10^{-3} \\ \frac{x_1^* - \bar{x}(10)}{s(10)} &= 2.25 & x_c = 1.96\end{aligned}$$

Since  $2.25 > 1.99$ , the value  $x_1^*$  is an outlier.

2.) A repetition of the application of the criterion gives:

$$\begin{aligned}x_2^* &= 2.282 \\ \bar{x}(9) &= 2.2980 \\ s(9) &= 6.697 \cdot 10^{-3} \\ \frac{x_2^* - \bar{x}(9)}{s(9)} &= 2.39 & x_c = 1.914\end{aligned}$$

Again  $2.39 > 1.914$  so that  $x_2^*$  is also an outlier.

3.) Now  $x_3^*$  is looked upon as suspicious.

$$\begin{aligned}x_3^* &= 2.289 \\ \bar{x}(8) &= 2.300 \\ s(8) &= 4.984 \cdot 10^{-3} \\ \frac{x_3^* - \bar{x}(8)}{s(8)} &= 2.20 & x_c = 1.86\end{aligned}$$

Also this value is an outlier.

4.) Another application leads to:

$$x_4^* = 2.3070$$

$$\bar{x}(7) = 2.3015$$

$$s(7) = 4.660 \cdot 10^{-3}$$

$$\frac{x_4^* - \bar{x}(7)}{s(7)} = 1.17$$

$$x_c = 1.79$$

Finally this measurement proves to be no outlier.

In this case  $\sim 30\%$  of the measurements are eliminated. It seems justified to check if the premises for the application of the criterion were given, i.e. if the values were normally distributed. For this test statistical procedures are known/Ref. 11/. In our example we can follow from the measuring procedure in the mass spectrometer that a normal distribution cannot be assumed since there is a high probability of systematic errors (e.g. the ion beam intensity of the instrument depends on the temperature of the filament, which is not constant in course of time).

b) In parallel also the application of the Dixon criterion may lead to problematic results, as the following example shows.

For the very low isotopic ratios one gets results in the order of  $10^{-6}$ . A real series of measurements then looks like

0.000001  
 0.000001  
 0.000002  
 0.000002  
 0.000002  
 0.000001  
 0.000001  
 0.000001  
 0.000001  
 0.000003  
 0.000002



because of the limited sensitivity of the measuring system, i.e. there is only one significant digit. When the Dixon criterion is applied to such a series the value 0.000003 is excluded as an outlier even if the probability of error is put  $\alpha = 0.05$ .

#### 10.5 Resulting Considerations for the Application of Outlier Criteria to the Data of the IDA - 72 Experiment.

The application of an objective criterion which can be used in a computer program with the aim to eliminate strongly deviating (wrong) measurements and to find errors of typing and data transmission seemed inevitable. The alternatives taken into consideration were

- Dixon criterion
- Chauvenet criterion, version I
- " " , version II .

The version II of the Chauvenet was excluded because it yields too many outliers - see above. Also the normal version I of the Chauvenet criterion gives 1.5 times more outliers than the Dixon criterion and so did not fulfill the condition of rejecting strongly deviating values only. In addition to this the premise that the measurements should be normally distributed is more important for the Chauvenet criterion than for the Dixon criterion, because in the calculation of the distribution function  $F_{r_{ij}}$  of the quotients of differences  $r_{ij}$ , the expectation value  $\mu$  and the variance  $\sigma^2$  of the measurements  $x_i$  are not used in contrast to the calculations necessary for the application of the Chauvenet criterion. Since this condition of a normal distribution cannot be fulfilled strictly by the real measurements we came to the conclusion that the Dixon criterion is the more appropriate one.

In the case of the Dixon criterion it is easy to change the sensitivity of the criterion by adjusting the probability of error.

In our study for the standard experiment results from four laboratories the numbers of outliers found under the assumption of a certain  $\alpha$  were plotted against these  $\alpha$ -values. Fig. 10-2 shows that in three of four cases the curves indicate a region where the number of outliers is independent of  $\alpha$ . This region (A) lies between  $\alpha=0.0$  and  $\alpha=0.02$ ; above  $\alpha=0.02$  (region B) the relation between  $\alpha$  and the number of outliers seems to be linear. For the application of the criterion with an  $\alpha$ -value from region A it is plausible that false measurements are excluded by this choice as it was defined as our aim in the application of an outlier criterion. So the value  $\alpha=0.01$  was chosen in order to be sure not to be in region B where the variance of the measurements could be affected by omission of too many outliers (see above). Our decision should be looked upon as practically oriented for the actual experiment - it cannot be a general judgement on the criteria or a basis of procedures for other experiments.




Fig.10-2: Number of outliers as function of the error probability  $\alpha$ .

11. Comments to the Computer Program for the Evaluation  
of the Standard Experiment and Some Additional  
Experiments in the Framework of IDA-72

by

W. Beyrich \* and E. Drosselmeyer \*\*

\* EURATOM, delegated to the Institute for Applied Systems Analysis  
and Reactor Physics (IASR), Gesellschaft für Kernforschung (GfK),  
Karlsruhe, Germany

\*\* IASR, GfK, Karlsruhe, Germany

Table of Contents

11.1	Introduction
11.2	Organisation of the Computer Program for the Evaluation of the Standard Experiment.
11.2.1	General Information and Comments on the Single Parts of the Program
11.2.2	MAIN program
11.2.3	Block Data HANNA
11.2.4	Subroutine DIXON
11.2.5	" MINL
11.2.6	" COR
11.2.7	" LAB1
11.2.8	" LAB2
11.2.9	" ALPHA
11.2.10	" LAB3
11.2.11	" LAB4
11.2.12	" LAB5
11.2.13	" LAB6
11.3	Programs for the Evaluation of Some Additional Experiments
11.3.1	General
11.3.2	Subroutine DRY
11.4	Computer Program

## 11.1 Introduction

This report is written in order to allow a better understanding of the computer program and the results printed out for each participating laboratory. An example for the output list is given in Chapter 12.

## 11.2 Organisation of the Computer Program for the Evaluation of the Standard Experiment

The program is written in FORTRAN IV for an IBM 370/185 computer. It is organized in the form of one main program and a number of subroutines for special calculations. A typical run with the data from about 20 laboratories ( 40 000 numbers) takes about 1 minute, and about 480 000 locations of storages are needed.

In the following a list of the parts of the program system is given with very short indications on what is done in these parts.

MAIN	reading of data cards and general program organisation
HANNA	block data subroutine, necessary for data organisation
DIXON	application of outlier criteria
MINL	ordering of a series of input data (in double precision), necessary for DIXON
COR	application of special corrections for the data from one laboratory
LAB1	calculation of the mean value of eight scan values with variance, standard deviation and relative standard deviation
LAB2	calculation of the mean value of three runs with "double stage" variance analysis
ALPHA	reading of $\alpha$ -measurement results, calculation of their mean values
LAB3	calculation of the isotopic composition of the different samples, inclusion of $\alpha$ -measurements

- LAB4        calculation of U- and Pu-concentrations
- LAB5        "triple stage" variance analysis: scan, run and  
interlaboratory error components
- LAB6        "double stage" variance analysis: run and  
interlaboratory error components

### 11.2.1 General Information and Comments on the Single Parts of the Program

In some of the tables comprising results of the computer calculations, abbreviations for sample types and isotopic ratios have been used. Since these refer to more than one subroutine of the program, they are listed and explained here.

With respect to the isotopic ratios the abbreviations stand for the following ratios:

83	U-238/ U-233
<hr/>	
92	Pu-239/ Pu-242
2	Pu-240/ Pu-242
<hr/>	
38	U-233/ U-238
48	U-234/ U-238
58	U-235/ U-238
68	U-236/ U-238
<hr/>	
89	Pu-238/ Pu-239
9	Pu-240/ Pu-239
19	Pu-241/ Pu-239
29	Pu-242/ Pu-239

The type of the samples is indicated by letter combinations which mean:

US	uranium	spike solution
PS	plutonium	spike solution
AU	unspiked A	sample solution
BU	unspiked B	sample solution
RU	unspiked R	sample solution
AS	spiked	A sample solution
BS	spiked	B sample solution
RS	spiked	R sample solution

In the following the single parts of the program, which is listed in Par.11.4, are described in an order which follows the order of the calculations.

At the same time the produced output is explained so that this paper can serve as a guide for reading the results of

- a) the single laboratories
- b) the group of all laboratories.

Typical output lists are given in Chapter 12.

In some of the program lists comments are given, marked by the letter "C" in the first column of the respective line, in order to explain the procedure. In addition some statements which were necessary for testing purposes have been made "dummy" by putting a "C" in column 1 of the punch cards.

### 11.2.2 MAIN Program

The calculations for the IDA-72 standard experiment begin with the main program MAIN. In this part of the program, which is central for all calculations, mainly all data are read in, and the other subroutines of the program, which are written for special calculations, are called.



In particular:

- a) at the beginning of MAIN the data for the correction for the AM-241 decay of Pu-241 are read in (statement 0041 etc. - the internal statement numbers are given at the beginning of the lines in the program listing).

The time intervals between sample preparation and Am separation which have to be known for this correction can be found in this program. They are given in days. The code numbers of the respective laboratories are given as M = ... (e.g. M = 2 in statement 0042).

The three indices of the variable DEC are this code number, a number indicating the type of sample (3 corresponds to AU, 4 to BU and 5 to RU, 6 to AS, 7 to BS and 8 to RS), and the last index gives the number of the run (cf. e.g. statement 0043 from which one can see that the decay time of the sample material AS for the first run in laboratory 2 was 106 days). The two indices of DC are the laboratory code number and the number indicating the type of sample - in these simpler cases the chemical treatment for all runs of the same sample was performed at the same day.

- b) A typical example of a data sheet as they were sent to the evaluation group by the laboratory is given in Fig. 11-1. Fig. 11-2 shows how the same data for the ratios Pu-238/Pu-239 and Pu-240/Pu-239 have been put on punched cards for the computer input (the columns 73-80 of these cards are meaningless in this context). The information that no mass spectrometric measurement of the ratio Pu-238/Pu-239 was performed is transmitted by punching the 9.9 in columns 12 to 13 of the first card.

At the beginning of the data of each lab the sentence

NEW SERIES OF MEASURING RESULTS

is printed on top of a new page.

- c) For some laboratories (code no. 15 and 18) this part of the program differs from the normal procedure. For laboratory 18 the mass discrimination and other correction factors were given separately and are applied by means of the computer program. For doing so, an extra subroutine COR was written, see 11.2.6.

For laboratory 15 the data cards were punched in a different format.

- d) In the next part of the program each series of - normally 10 - measurements for one specific isotopic ratio of one specific sample is checked for the actual number of measurements.

Interesting are e.g. the following possibilities:

- α) no measurements at all
- β) only one measurement
- γ)  $\leq$  8 measurements
- δ) 10 measurements.

Standard ExperimentSample: A (spiked) - PlutoniumCode: 15 \*(For  $\alpha$ -spectrometric data, see II-39)Chemical preparation No.: 2Date: 9.8.1972Date of MS-measurement: 14.8.1972

Scan No.:	Atomic Ratio			
	238/239	240/239	241/239	242/239
1	0.-----	0. <u>2332</u>	0. <u>1261</u>	1. <u>288</u>
2	-	0.2339	0.1259	1.293
3	-	0.2339	0.1270	1.301
4	-	0.2357	0.1267	1.301
5	-	0.2334	0.1258	1.283
6	-	0.2335	0.1262	1.297
7	-	0.2327	0.1267	1.300
8	-	0.2340	0.1274	1.300
9	-	0.2347	0.1255	1.297
10	-	0.2354	0.1257	1.295

Don't use this space!

Remark:

Fig. 11-1 Example for a Data Form Sheet

\* The data are published with the kind permission of the laboratory



After this a line is printed giving the code number of the lab, the type of the sample, the number of the run, and the measured ratio, e.g. :

3 BS2 48

- e) If there are more than 8 input values, their number is reduced to 8 by applying the Dixon outlier criterion or by omitting the last input values if no outliers can be found.

The subroutine DIXON, which is called in the statement with the internal No. 0291, and its output is shortly described under 11.2.5, the underlying mathematical procedure is dealt with in Chapter 10.

- f) After this selection of input values the "Am-241 corrections" are applied to the input values of the measured Pu-241/Pu-239 ratios. (internal statements 0295 - 0352).
- g) After that the run mean values of the series of measurements are calculated in subroutine LAB1, see 11.2.7. The results of this subroutine are passed back to the main program and then a table is printed as described under 11.2.7.
- h) In the following (internal statements 0358 etc.) the subroutines LAB2, LAB3 and LAB4 which are concerned with the results of the single laboratories are called. The calculations for the group of laboratories are done in the subroutines LAB5 and LAB6.
- i) In addition a combined table of laboratory mean values for isotopic ratios (results of LAB2) is printed in the main program.
- j) Parts of the additional experiments in the framework of IDA-72 are also connected with the standard experiment via the main program.

### 11.2.3 BLOCK DATA

In the program listing the MAIN is followed by a BLOCK DATA subroutine. This part of the program is necessary for data organisation in the computer.

It produces no output.

### 11.2.4 Subroutine DIXON

In this subroutine each series of measuring results consisting of more than 8 values is checked for outliers. The criterion of DIXON /Chapter 10/ is adjusted in such a way that only rather large deviations of single measurements are "discovered" - so the application of the test has more or less the function of discovering errors in data transmission.

The output of this subroutine is sent to the single laboratories. For each series of measurements the values are put into ascending order by the subroutine MINL /11.2.5/. The resulting series is printed. If an outlier at the lower or upper end of this ordered series of data is found, a comment is printed, saying e.g. ;

THE MAXIMUM VALUE  $S(10) = 0.000300$  IS CANCELLED

The check for an outlier value at the end of this ordered series is repeated. If an outlier is found in a series of 8 remaining values an extra comment is printed:

\*\*\*\*\* THE 8TH INPUT VALUE IS AN OUTLIER \*\*\*\*\*,

but this outlier value is kept as a member of the series of measuring results.

If the series of measuring results consists of more than 8 values after the check for outliers the first 8 of the original input list modified by omission of the eventual outliers are taken as output values for this subroutine.

For cases with outliers the 8 resulting values after rejecting the outliers are listed in original order under omission of the outliers, the number of outliers is mentioned in an extra line. The results are passed to the main program.

#### 11.2.5 Subroutine MINL

This subroutine was written to put double precision numbers into ascending order.

It is called from subroutine DIXON, passes the results to DIXON, and produces no output.

#### 11.2.6 Subroutine COR

This subroutine was written to apply mass discrimination and some other corrections to the data from the laboratory with code no. 18.

The subroutine is called after the application of the Dixon outlier criterion. After multiplication with the necessary correction factors (internal statement no. 0064 etc.) the same calculations as in the program MAIN are done for the data of this laboratory up to calling the subroutine LAB1.

The results of COR and LAB1 are returned to MAIN.

#### 11.2.7 Subroutine LAB1

The results of this subroutine are printed in a table with 9 columns. In the first column the abbreviations for the isotopic ratios are given. In the second column the type of the sample is indicated. The run numbers in the third column are self-explaining.

The fourth column gives the run mean value of the ratio and sample indicated in the respective line. The formula for this mean is

$$\bar{X}_{jk} = \frac{1}{8} \sum_{i=1}^8 X_{ijk}$$

$X_{ijk}$  are the single scan results (the data sent to us on the distributed forms, an example is given above as Fig. 11-1) after application of the Dixon-criterion.

The indices stand for

- i - number of the scan
- j - number of the run
- k - number of the laboratory.

The expression 0.38950000 D-02 e.g. stands for  $0.3895 \cdot 10^{-2}$ . The letter D indicates that the calculations are done in double precision, that means with about 16 decimal digits.

After the calculation of the mean value, the scan variance  $\sigma_{jk}^2$ , the corresponding standard deviation  $SD_{jk}$  and relative standard deviation  $RSD_{jk}$  (in %) given in columns 5, 6, 7 are calculated.

The formulae are:

$$\sigma_{jk}^2 = \frac{1}{7} \sum_{i=1}^8 (\bar{X}_{jk} - X_{ijk})^2$$

$$SD_{jk} = \sqrt{\sigma_{jk}^2}$$

$$RSD_{jk} = (SD / \bar{X}_{jk}) \cdot 100$$



Column 8 gives the number of scans which was supplied by the laboratory for the run; in cases where this number is  $\leq 8$ , no outlier criteria are applied and the calculations described above are adjusted for the smaller number. In special, if this number is 0, one will find 0.0 in columns 4-7 in order to indicate that there were no measurements available. Sometimes a variance value in the order of  $\sigma_{jk}^2 = 0.1 \cdot 10^{-40}$  is found in column 5. This is the result of the computer calculation for cases where all values in a series of measurements are equal. This result should be read as 0.0, the calculation of the standard deviation and relative standard deviation is meaningless in these cases.

Finally, column 9 gives the number of outliers found for each run. The total number of outliers is given in the title of the table.

#### 11.2.8 Subroutine LAB2

The subroutine LAB2 calculates for each laboratory the mean value (MEANA), variance (SG2), standard deviation (SDG), and relative standard deviation (RSDG) for all measured isotopic ratios of the distributed samples. In addition it gives run and scan components of variances, standard deviations and relative standard deviations.

The results of this subroutine are printed in a table for each laboratory.

This table gives the abbreviation for the measured isotopic ratio in the first column, the name of the sample under consideration (see 11.2.1) in the second column.

The third column gives the laboratory mean value which is calculated from the results of subroutine LAB1 as:

$$\bar{\bar{X}}_k = \frac{1}{3} \sum_{j=1}^3 \bar{X}_{jk}$$

(Again this formula is given for the "normal" case with 8 scans and 3 runs, for other cases the formula was adjusted.)

The variance for the laboratory mean value, given in column 4, is calculated as

$$\sigma_k^2 = \frac{1}{6} \sum_{j=1}^3 (\bar{\bar{X}}_k - \bar{X}_{jk})^2$$

the standard deviation (column 5) and relative standard deviation (column 6, in %) are

$$SD_k = \sqrt{\sigma_k^2}$$

$$RSD_k = (SD_k / \bar{\bar{X}}_k) \cdot 100 .$$

The formulae for the estimates of the run and scan components (column 7 - 12) are given in the report on the Mol III-experiment / 18/.

In addition to these tables there also exists a table where the results of the subroutine LAB2 for all labs are combined so that one can easily make a comparison.

#### 11.2.9 Subroutine ALPHA

In this subroutine first of all a test number indicating whether the laboratory under consideration has performed  $\alpha$ -measurements (test number = + 7.) or not (test number = + 8.) is read in. If  $\alpha$ -results are available, the normal input data of the laboratory are followed by 6 extra punch cards containing the information about three repeated measurements of the activity ratios Pu-238/(Pu-239 + Pu-240) of the six samples AU, BU, RU, AS, BS, and RS.

The results of this subroutine are printed in a little table. This gives for each laboratory, which has performed  $\alpha$ -measurements, and for each sample the mean value of the measured  $\alpha$ -activity (column 2), the variance, standard deviation and relative standard deviation of the single measurements (column 3-5), the relative standard deviation of the mean value (column 6) and the calculated ratio Pu-238/Pu-239.

The RSD of the mean value (in %) is calculated as

$$\text{RSD (mean)} = \frac{100}{\bar{X}} \sqrt{\frac{\sum_{i=1}^n (X_i - \bar{X})^2}{n(n-1)}}$$

$$\text{where } \bar{X} = \frac{1}{n} \sum_{i=1}^n X_i$$

whereas the RSD of the single measurements (in %) is defined as

$$\text{RSD (single measurement)} = \frac{100}{\bar{X}} \sqrt{\frac{\sum_{i=1}^n (X_i - \bar{X})^2}{n-1}}$$

#### 11.2.10 Subroutine LAB3

This subroutine calculates the isotopic composition of the samples in atom percentages and weight percentages for each sample and each laboratory. The results of alpha spectrometry are included if the ratio Pu-238/Pu-239 has been determined by this method.

In order to give an indication of how the values are calculated, an example is given here. The content of U-235 (in atom percents) for sample AU is e.g. calculated as

$$\text{ATU 235} = \frac{X_k(\text{AU58}) \cdot 100}{(X_k(\text{AU38}) + X_k(\text{AU48}) + X_k(\text{AU58}) + X_k(\text{AU68}) + 1)}$$

In this formula  $X_k(\text{AU58})$  stands for the mean value of the ratio (U-235/Pu-238) for the unspiked A-sample (AU) which the laboratory k obtained. For the other isotopic ratios of U and Pu the same nomenclature is used and the formula is applied correspondingly.

For the case of the inclusion of  $\alpha$ -measurements the ratio Pu-238/Pu-239 comes from the subroutine ALPHA /see 11.2.9/, all other ratios are the results of mass spectrometry which have been used before.

For the calculation of weight percentages each isotopic ratio in the "prototype" formula above has to be replaced by its atomic percentage multiplied by the atomic weight of the corresponding isotope.

The atomic weights given in Chapter I have been used (one can find them in the program in the statements with internal no. 0068-0072 for U and in the statements 0084-0088 for Pu). Also here an extra calculation is made, if results of  $\alpha$ -spectrometry are available.

The tables printed in this subroutine should be self-explaining. The same abbreviations as above are used, cf. 11.2.1.

#### 11.2.11 Subroutine LAB4

Also this subroutine calculates results for single laboratories. It has been written in order to get four different types of results:

- a) U-concentrations
- b) Pu-concentrations calculated on the basis of mass spectrometry data only
- c) Pu-concentrations calculated using  $\alpha$ -spectrometry data for Pu-238
- d) U/Pu ratios

In the output tables one finds for uranium and plutonium first the concentrations of the reference isotope (U-238 or Pu-239, respectively) in atoms/g solution and then the element concentrations (in mg U/g solution or  $\mu\text{g}$  Pu/g solution, respectively).

The Pu-results are sometimes given in two versions, see b) and c) above.

In all these blocks the results are given in the following order:

results of the 1st run : line 1

" " " 2nd " : " 2

" " " 3rd " : " 3

mean values of the

results above : " 4

These results are given for the samples A, B and R.

In addition, in columns 4 and 5 so-called "calibrated" values for samples A and B can be found. As has been pointed out in the evaluation paper /Vol.I, Par. 3.5.1/ seems meaningful to calibrate the results on the A and B samples by multiplying them by the ratio RC

$$RC = \frac{\text{concentration of R-sample stated by the CBNM}}{\text{concentration of R-sample measured by the laboratory}}$$

The calculations for the concentration values are done as follows:

First the concentrations of the reference isotopes (U-238, Pu-239) on the basis of the single runs are calculated.

For this the following formula is applied:

$$C = \frac{1 - RR_s}{R - R_u} \cdot \frac{G_s}{G_u} \cdot S$$

with (the uranium case is given for example):

- R = Ratio U-233/U-238 of the spiked sample, corresponding run mean value from subroutine LAB1.
- $R_u$  = Ratio U-233/U-238 of the unspiked sample, laboratory mean value from subroutine LAB2.
- $R_s$  = Ratio U-238/U-233 of the mixed spike solution, laboratory mean value from subroutine LAB2 \*)
- $G_s$  and  $G_u$  = Masses of the aliquots of spike and unspiked solution, as follows /cf. also Par. 3.2.3 to 3.2.5/

Sample	$G_s$ [g]	$G_u$ [g]
A	47.114	52.154
B	47.157	55.377
R	46.628	45.943

- S = Number of atoms U-233 per gram of mixed spike solution
- =  $2,0931 \cdot 10^{18}$ , calculated from the data given in Par. 3.2.2.

From these three "run concentrations" the laboratory mean value given in the next line of the output and its variance,  $\sigma^2 = \frac{1}{6} \sum_{j=1}^3 (c_{jk} - \bar{c}_k)^2$  standard deviation and relative standard deviation are calculated.

\*) In the case of Pu a fixed value 0.0134 as measured by the CBNM was used for all calculations for all the laboratories /Par. 3.2.2 and Vol. I, Par. 3.5.1/.

For getting the element concentrations these numbers are multiplied by a "weight factor" which is (for uranium as example, again):

$$\begin{aligned}
 W &= 233.0395 \cdot \text{lab mean value of the ratio U-233/U-238 in the unspiked sample} \\
 &+ 234.0409 \cdot \begin{array}{c} \text{"} \\ \dots\dots\dots \end{array} \text{U-234/U-238} \dots\dots\dots \text{"} \\
 &+ 235.0439 \cdot \begin{array}{c} \text{"} \\ \dots\dots\dots \end{array} \text{U-235/U-238} \dots\dots\dots \text{"} \\
 &+ 236.0457 \cdot \begin{array}{c} \text{"} \\ \dots\dots\dots \end{array} \text{U-236/U-238} \dots\dots\dots \text{"} \\
 &+ 238.0508
 \end{aligned}$$

For plutonium the corresponding values have to be inserted.

For the nuclide masses used cf. Par. 1.1.

In the 5th line of the block concerning the concentrations of the reference isotopes, the relative standard deviations of the mean value for samples A, B and R are given. For the element concentrations in addition two lines, giving the variance and standard deviation, are inserted. This information is given for the samples A, B and R.

If any laboratory has not measured the Pu-238/Pu-239 ratio by mass-spectrometry, a note is printed before the Pu-concentrations are given.

These concentrations are then calculated as if the Pu-238 concentration were = 0. for this sample.

If  $\alpha$ -measurements exist, an extra block of results, giving the Pu-concentrations calculated after inclusion of these measurements, is printed. It has the same appearance as the block for the Pu concentrations based on mass-spectrometer measurements only.

After the end of the results of this subroutine the Pu/U-element-ratios for the three samples A, B and R are printed. They are calculated from the "calibrated" values for the A and B samples. The first line gives the results of mass-spectrometer measurements, in the second line  $\alpha$ -measurements are included, if they have been performed.

#### 11.2.12 Subroutine LAB5

This subroutine uses the results of all laboratories in order to determine the estimates of scan, run, and interlaboratory error components for the measurements of the isotopic ratios for the different samples.

Since the results of this subroutine belong to the group b) defined in Chapter 11.2.1 the output table was not distributed to the laboratories. It is now included in Chapter 12.3. For each sample and ratio this output table gives the number of laboratories, runs and scans taken into account for the calculations. In some cases results of particular laboratories have been omitted for reasons given in Vol.I,Par.3.2.3 (too large deviations e.g.). In the following columns of the table the calculated overall mean values, followed by variance, standard deviation and relative standard deviation for scan-, run- and interlaboratory-components are printed.

This table gives one aspect of the results of the whole experiment in a very compressed form on one page.

The formulae used for the calculation can be found in the report of the Mol III-experiment /19/.

#### 11.2.13 Subroutine LAB6

This subroutine was written in parallel to subroutine LAB5. It gives an analysis of variances in two stages, that means beginning with the run mean values. It seemed useful to check the rather complicated calculations of LAB5 by this simpler analysis, made under the assumption that in general the scan components of the overall error can be neglected. Also this subroutine belongs to group b) - results of the group of laboratories. An example of output is included in Chapter 12.3. The meaning of the numbers in the single columns is (from left to right):



Code for the isotopic ratio (see Par. 11.2.1), code for the sample (see Par. 11.2.1), mean of laboratory mean values for the isotopic ratio, variance of this mean, corresponding standard deviation, corresponding relative standard deviation; interlaboratory component of the variance, corresponding standard deviation, corresponding relative standard deviation; run component of the variance, corresponding standard deviation and relative standard deviation.

This subroutine corresponds to the subroutine LAB2 in which the calculations for a twofold analysis of variances are performed for the measurements inside one laboratory.

### 11.3 Programs for the Evaluation of Some Additional Experiments.

#### 11.3.1 General

Since the group of the additional experiments: self spike experiment, dry spike experiment, aluminium-capsule experiments, aging experiments and X-ray fluorescence comprises rather different parts it was not so evident that also these experiments should be evaluated by means of a computer program.

We came to the following conclusion:

For experiments with a larger group of participants (self spike, dry spike and aluminium-capsule) all the series of measurement results were checked for outliers by application of the DIXON subroutine.

For the self spike experiment the following calculations were done without a computer program /cf. Volume I, Chapter 4), because a lot of information was laboratory-dependent.

The remaining calculations for the dry spike and aluminium-capsule experiments were done in a subroutine DRY (see below). The rest of the other additional experiments were evaluated in collaboration of the laboratories performing these experiments and the evaluation group /Volume I, Chapter 8; this Volume, Chapter 9/.

### 11.3.2 Subroutine DRY

For the dry spike and aluminium-capsule experiments described in Vol. I, Chapter 5 the data and results coming from the standard experiment and these additional experiments had to be combined. This was done in the subroutine DRY.

This subroutine is called from the MAIN program used for the evaluation of the standard experiment (internal statement no. 432 etc.). The data needed as input of the subroutine DRY are put in on data cards following the block of all data cards for all laboratories participating in the standard experiment. They are read in by the subroutine DRY. Some information necessary for corrections is given in the MAIN program.

A list of the subroutine DRY is given in Chapter 11.4.

This subroutine produces output which follows the block of output of the standard experiment.

An example is included in Chapter 12.2.

In the first line of this output the code number of the laboratory is given. In the next block the input values are listed. Each line corresponds to one run consisting of 10 scans for one sample and isotopic measurement. In each line the three last numbers give the aliquot weights of spike and sample and the number of the respective spiked sample. The order is: Three series (runs) of isotopic ratios U-235/U-238 for the first sample /cf. Volume I, Chapter 1.2 and 5/, three series of isotopic ratios U-235/U-238 for the second sample, three series of isotopic ratio Pu-242/Pu-239 for the first sample, and three series of isotopic ratios Pu-242/Pu-239 for the second sample.

Although also in this subroutine the subroutine DIXON (see Par. 11.2.4) is called the output normally produced by DIXON has been suppressed in this case. The next block gives the run mean values together with variances, standard deviations and relative standard deviations in the order indicated in the output list. The uranium and plutonium concentrations (given as number of atoms/g solution) which the laboratory got for the three runs of the standard experiment are printed in the next block. This allows a comparison with the results of the additional experiment which are given in the following block.

After this the mean numbers of atoms for uranium and plutonium in sample A-1 and A-2 are given, calculated firstly on the basis of the measured isotopic ratios and secondly on the resulting numbers of atoms. Due to different steps in the calculation these number can differ slightly. Unlike to the standard experiment, these values are different for each laboratory and sample.

The formulae used for the calculations in this subroutine are the same which are used for the calculations of those concentrations which are determined in the standard experiment. These formulae are given in Par. 11.2.11 above:

$$C = \frac{1 - RR_s}{R - R_u} \cdot \frac{G_s}{G_u} \cdot S$$

The weights of the aliquots of the spike solution ( $G_s$ ) and of the unspiked sample solution ( $G_u$ ) are included in the input data for this subroutine, see above.  $S$  is the number of atoms of the spike isotope (U-233 and Pu-242, respectively) per gram of the spike solution; for  $R_s$ , the ratio main isotope / spike isotope in the mixed spike solution (U-233 /U-238 and Pu-242/Pu-239 respectively), and for  $R_u$ , the ratio spike isotope/main isotope in the unspiked sample solution (U-233/U-238 and Pu-242/Pu-239, respectively) the laboratory mean values from the standard experiment are used.

$R$  is the isotopic ratio spike isotope/main isotope (U-233/U-238 and Pu-242/Pu-239, respectively) measured as run mean value in the dry spike experiment.

This same subroutine DRY has been used for the evaluation of the first part of the aluminium-capsule experiment. Also the output has the same appearance, in this case one should read "aluminium -capsule experiment" instead of "dry spike experiment".

#### 11.4 Computer Program

### Corrigenda

Due to a change in computer operation some lines of comments given in this program were truncated. The complete lines are given here for these cases; in addition some errors in the comments are indicated.

Page	Statement No. before comment	Complete text of comment
144	0052	Evaluation of extrema and decision, where an outlier is more probable
	0057	This parameter is a hint to the left part of the flow sheet; an outlier ...
156	0001	The input values for this subroutine are ratios of alpha-activities
158	0001	Input values are the mean values over scans and runs = lab mean values .... MEAN (M3, K3, LL3), where .... J3 ... should be skipped
159	0042	In this vector the order of isotopes is: 7-239, 8-238, 9-240, 10-241, 11-242
160	0055	Calculation of atom percentages with inclusion of alpha-measurements
162	0001	This subroutine calculates the U and Pu concentrations and U/Pu ratios for ....
163	0027	K4 indicates the sample type; order of sample types: US, PS, AU, BU, RU, AS, BS, RS .... Isotopic ratios are given in the order 83, 92, 02, 38, 48, 58, 68, 89, 09, 19, 29
164	0045	The values K0 define the concentration of the reference isotope, e.g. U-238,
	0060	In the formulae for C the results MEANV from LAB1 and MAIN are used
	0064	In the vectors C and CON the indices K4 = 3,4,5 stand for the original ...
165	0085	Lab. 22 has not measured the Pu spike, the BCMN value is taken.
166	0102	Special cases of missing measurements of ratio 242/239 and missing runs
171	0026	MM(K,LL) is the number of labs, for which standard experiment results are ... .... SS(K,LL) is the number of runs in all labs, in which the isotopic ratio ...
	0035	For the US sample only one isotopic ratio and for the PS sample only two ....

FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 74346

17/52/56

```

C
C
C
C
C
0001 REAL TYP(8)/*US ',*PS ',*AU ',*BU ',*RU ',*AS ',*BS ',*RS
1'/
0002 REAL*8 SIGMA,SIGMAV( 8,3,11)/ 264*-3./,SQI,SQIV( 8,3,11)/ 264*-3
1./,MEAN,SDV(8,3,11)/264*-3./,RSDV(8,3,11)/264*-3.
2 / ,X(10),S(10),Y(10),Z(10),ZZ(10),A(8),F(3),
3 WALPHA(21,8)/168*0./ ,SUMV(21,8,3,11),SD,RSD,SUMM
0003 REAL *8 MEANV (21,8,3,11)/5544*-3./,MEAN2(21,8,11)/1848*0./
0004 INTEGER IR(11) /83,92,02,38,48,58,68,89,09,19,29/,IRE(21)/21*0/
0005 INTEGER * 2
0006 MVAL( 8,3,11)/ 264*8/,SUMCUT(21) ,NAUS( 8,3,11),LVAL(8,3,11)
0007 DIMENSION IEXP(21),ILAB(21),SAMP(21),IRUN(21), IREL(21),ITYPE(21)
0008 INTEGER * 2 IFILIV(21,8,11)/1848*0/,MVALV(21,8,3,11)
COMMON /HANNA/ LINKS,MX,NVAL
1/ITEMS/SG2V(21,8,11),SDGV(21,8,11),RSDGV(21,8,11),RSDCOV(21,8,11),
1SIGA2V(21,8,11),SDAV(21,8,11),RSDAV(21,8,11),SIGE2V(21,8,11),SDEV
2(21,8,11),RSDEV(21,8,11),DIFV(21,8,11)
3/DRYA/ KO(21,3,5,2),KOAL(21,3,5,2)
0009 REAL*8 KO,KOAL,MDC,FACT
0010 REAL*8 SG2V,SDGV,RSDGV,RSDCOV,SIGA2V,SCAV,RSDAV,SIGE2V,SDEV,RSDEV,
1DIFV,DEC(22,8,3)/528*1./,DC(22,8),DICO(2)
C
0011 1000 FORMAT (I2)
0012 1001 FORMAT (I1,I2,1X,A2,I1,1X,I2,5F10.6)
0013 1002 FORMAT (/I3,2X,A2,I1,1X,I2)
0014 1003 FORMAT (F11.8)
0015 1004 FORMAT (/ 'NEW SERIES OF MEASURING RESULTS'/)
0016 1005 FORMAT (' TYP(I=1...8)=' ,8(A2,2X),', SAMP(M) AND ITYPE(M)=' ,A2,I3)
0017 1006 FORMAT('1',//9X,' STANDARD EXPERIMENT - RESULTS OF PROGRAM LAB1
1',/ , 9X,' MEAN VALUE,VARIANCE,SC AND RSD FOR THE ISOTOPIC RATIOS
2 OF EACH RUN',/, 9X,' (8 SCANS/RUN WERE CONSIDERED)',/, 'OLABRATORY
3 CODE NO:',I5, 5X,' TOTAL NUMBER OF OUTLIERS FOR THIS LAB=',I5,/
4 ,30X,'THE PU-241 VALUES OF THE UNSPIKED SAMPLES HAVE BEEN CORRECT
5ED FOR AM-241 DECAY'/'OISCT.RATIO SAMPLE RUN RUN-MEAN
6SCAN-VARIANCE SCAN-SD',9X,'SCAN-RSD(%) NO.OF MEASUREMENTS NO
7.OF OUTLIERS')
0018 1007 FORMAT (' THE VALUE X(II=' ,I2,')=' ,F10.6,' WAS CANCELLED, BECAUSE
1IT INDICATED END OF DATA')
0019 1008 FORMAT (//,' NVAL WAS SET =' ,I4,///,' *****')
0020 1009 FORMAT (' NVAL IS =' ,I5)
0021 1010 FORMAT (I10,A9,I5,D17.8,D16.8,D16.8,D16.8,I12,9X,I3,D16.4)
0022 1011 FORMAT (' AFTER EVALUATION THE VALUES Y(I) ARE: ',8F10.6)
0023 1012 FORMAT (' AFTER EVALUATION THE VALUES ZZ(I) ARE: ',8F10.6)

```

FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 74346

17/52/56

```

0024      1013 FORMAT (' AFTER EVALUATION THE VALUES Z(I) ARE: ',8F10.6)
0025      1014 FORMAT (' A SERIES OF X(I) COULD NOT BE MEASURED')
0026      1015 FORMAT ('OLABNUM,J,K,L,LL=',5I4)
0027      1016 FORMAT (' FOR LABNUM=',I3,' THE CODE NUMBER FOR THE ISOTOPIC RATIO
          1 WAS SET=',I3)
0028      1017 FORMAT (' FOR J,K,L,LL=',4I4,' THE VALUES FOR MEAN, SIGMA, SDV AND
          1 RSDV ARE:',F10.7,F14.10,2F10.6)
0029      1018 FORMAT (' A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE')
0030      1019 FORMAT ('1')
0031      1020 FORMAT (3F7.3,59X)
0032      1021 FORMAT (3F7.3,/)
0033      1022 FORMAT (I1,I2,1X,A2,I1,1X,I2,5(F9.6,1X))
0034      1023 FORMAT ( I2)
0035      1024 FORMAT (///)
0036      1025 FORMAT (' RATIO',I3,' SAMPLE ',A2)
0037      1026 FORMAT (' LAB',I3,2X,D10.4,4X,D10.4,2X,D9.3,2X,D9.3,3X,D10.4,2X,
          1D9.3,2X,D9.3,3X,D10.4,2X,D9.3,2X,D9.3)
0038      1027 FORMAT ('RESULTS OF LAB2 FOR ALL THE LABS'/' LAB      MEAN',10X,'V
          1ARIANCE      STAND.DEV. RSD(%) '
          1      5X,'VARIANCE      SD',9X,'RSD(%)',6X,'VARI
          2ANCE      SD',9X,'RSD(%)'/23X,'LAB MEAN      LAB MEAN      LAB MEAN', 4
          3X,'RUN',9X,'RUN',8X,'RUN',9X,'SCAN',8X,'SCAN',7X,'SCAN')
0039      1028 FORMAT ('IFOR LAB 16 THE SUBROUTINE LAB4 IS MEANINGLESS.'/' NO CON
          1CENTRATIONS CAN BE CALCULATED,BECAUSE NO SPIKED SAMPLES HAVE BEEN
          2MEASURED.')
0040      1029 FORMAT (I2,F10.5)
C
C
C
C      CORRECTIONS FOR AM-241 DECAY
C      THESE CORRECTIONS ARE MADE AT THE EARLIEST POSSIBILITY
C      SINCE THEY DEPEND ON LAB NO. ETC.,THIS HAS TO BE DONE IN THE MAIN PROGRAM.
C      DEC OR DC ARE THE TIME INTERVALS BETWEEN SAMPLE PREPARATION
C      IN GEEL/MOL AND CHEMICAL SAMPLE TREATMENT IN THE LABS.
C
0041      MM=0
0042      M=2
0043      DEC(M,3,1)=106.
0044      DEC(M,3,2)=132.
0045      DEC(M,3,3)=132.
C
0046      DEC(M,4,1)=106.
0047      DEC(M,4,2)=106.
0048      DEC(M,4,3)=106.
C
0049      DEC(M,5,1)=279.
0050      DEC(M,5,2)=288.
0051      DEC(M,5,3)=294.

```

FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 74346

17/52/56

```

      C
0052      DEC(M,6,1)=106.
0053      DEC(M,6,2)=125.
0054      DEC(M,6,3)=132.

      C
0055      DEC(M,7,1)=140.
0056      DEC(M,7,2)=153.
0057      DEC(M,7,3)=153.

      C
0058      DEC(M,8,1)=115.
0059      DEC(M,8,2)=133.
0060      DEC(M,8,3)=117.
0061      202 M=3
0062      DC(M,3)=80.
0063      DC(M,4)=80.
0064      DC(M,5)=234.
0065      DC(M,6)=83.
0066      DC(M,7)=83.
0067      DC(M,8)=91.
0068      GO TO 1
0069      203 M=4
0070      DC(M,3)=106.
0071      DC(M,4)=106.
0072      DC(M,5)=260.
0073      DC(M,6)=107.
0074      DC(M,7)=90.
0075      DC(M,8)=113.
0076      GO TO 1
0077      204 M=5
0078      DC(M,3)=107.
0079      DC(M,4)=100.
0080      DC(M,5)=261.
0081      DC(M,6)=111.
0082      DC(M,7)=111.
0083      DC(M,8)=122.
0084      GO TO 1
0085      205 M=6
0086      DC(M,3)=98.
0087      DC(M,4)=98.
0088      DC(M,5)=252.
0089      DC(M,6)=111.
0090      DC(M,7)=112.
0091      DC(M,8)=125.
0092      GO TO 1
0093      206 M=7
0094      DC(M,3)=91.
0095      DC(M,4)=98.
0096      DC(M,5)=262.
```



FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 74346

17/52/56

```
0097          DC(M,6)=89.
0098          DC(M,7)=99.
0099          DC(M,8)=113.
0100          GO TO 1
0101          207 M=8
0102          DC(M,3)=86.
0103          DC(M,4)=86.
0104          DC(M,5)=240.
0105          DC(M,6)=85.
0106          DC(M,7)=85.
0107          DC(M,8)=93.
0108          GO TO 1
0109          209 M=10
0110          DC(M,3)=73.
0111          DC(M,4)=73.
0112          DC(M,5)=253.
0113          DC(M,6)=76.
0114          DC(M,7)=77.
0115          DC(M,8)=86.
0116          GO TO 1
0117          211 M=12
0118          DC(M,3)=120.
0119          DC(M,4)=120.
0120          DC(M,5)=260.
0121          DC(M,6)=124.
0122          DC(M,7)=126.
0123          DC(M,8)=118.
0124          GO TO 1
0125          212 M=13
0126          DC(M,3)=104.
0127          DC(M,4)=100.
0128          DC(M,5)=265.
0129          DC(M,6)=91.
0130          DC(M,7)=98.
0131          DC(M,8)=119.
0132          GO TO 1
0133          213 M=14
0134          DC(M,3)=185.
0135          DC(M,4)=191.
0136          DC(M,5)=345.
0137          DC(M,6)=198.
0138          DC(M,7)=202.
0139          DC(M,8)=213.
0140          GO TO 1
0141          214 M=15
0142          DEC(M,3,1)=41.
0143          DEC(M,3,2)=41.
0144          DEC(M,3,3)=41.
```

FORTRAN IV G1 RELEASE 2.0

MAIN

CATE = 74346

17/52/56

```

      C
0145      DEC(M,4,1)=41.
0146      DEC(M,4,2)=41.
0147      DEC(M,4,3)=41.
      C
0148      DEC(M,5,1)=195.
0149      DEC(M,5,2)=195.
0150      DEC(M,5,3)=245.
      C
0151      DEC(M,6,1)=49.
0152      DEC(M,6,2)=49.
0153      DEC(M,6,3)=49.
      C
0154      DEC(M,7,1)=53.
0155      DEC(M,7,2)=63.
0156      DEC(M,7,3)=53.
      C
0157      DEC(M,8,1)=57.
0158      DEC(M,8,2)=57.
0159      DEC(M,8,3)=57.
0160      215 M=16
0161      DC(M,3)=133.
0162      DC(M,4)=192.
0163      DC(M,5)=410.
0164      GO TO 1
0165      216 M=17
0166      DC(M,3)=117.
0167      DC(M,4)=110.
0168      DC(M,5)=264.
0169      DC(M,6)=117.
0170      DC(M,7)=123.
0171      DC(M,8)=133.
0172      GO TO 1
0173      217 M=18
0174      DC(M,3)=245.
0175      DC(M,4)=245.
0176      DC(M,5)=428.
0177      DC(M,6)=156.
0178      DC(M,7)=154.
0179      DC(M,8)=274.
0180      GO TO 1
0181      218 M=19
0182      DC(M,3)=154.
0183      DC(M,4)=155.
0184      DC(M,5)=331.
0185      DC(M,6)=154.
0186      DC(M,7)=154.
0187      DC(M,8)=176.

```

FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 74346

17/52/56

```

0188          GO TO 1
0189          219 M=20
0190             DEC(M,3,1)=84.
0191             DEC(M,3,2)=108.
0192             DEC(M,3,3)=108.
              C
0193             DEC(M,4,1)=101.
0194             DEC(M,4,2)=108.
0195             DEC(M,4,3)=108.
              C
0196             DEC(M,5,1)=238.
0197             DEC(M,5,2)=262.
0198             DEC(M,5,3)=262.
              C
0199             DEC(M,6,1)=83.
0200             DEC(M,6,2)=107.
0201             DEC(M,6,3)=107.
              C
0202             DEC(M,7,1)=83.
0203             DEC(M,7,2)=107.
0204             DEC(M,7,3)=107.
              C
0205             DEC(M,8,1)=91.
0206             DEC(M,8,2)=115.
0207             DEC(M,8,3)=115.
0208          220 M=21
0209             DC(M,3)=108.
0210             DC(M,4)=108.
0211             DC(M,5)=262.
0212             DC(M,6)=107.
0213             DC(M,7)=107.
0214             DC(M,8)=115.
0215             GO TO 1
              C
0216          221 M=22
0217             DC(M,3)=345.
0218             DC(M,4)=345.
0219             DC(M,5)=495.
0220             DC(M,6)=348.
0221             DC(M,7)=348.
0222             DC(M,8)=356.
              C
              C
0223             1 DO 726 IB=1,3
0224             DO 726 IT=3,8
0225             726 DEC(M,IT,IB)=DC(M,IT)
              C
              C          SPECIAL CASES,INSERTED HERE FOR SIMPLER PROGRAMMING

```

```

FORTRAN IV G1  RELEASE 2.0          MAIN          DATE = 75015          13/37/58

0226          DEC(4,7,3)=107.
0227          DEC(17,6,3)=131.
0228          DEC(17,7,3)=131.
0229          DEC(18,7,3)=266.
0230          DEC(19,8,3)=182.
              C
0231          MM=MM+1
0232          GO TO (203,204,205,206,207,209,211,212,213,214,216,217,218,219,221
              1,222) ,MM
0233          222 CCNTINUE
              C
              C
              C          READING OF THE NUMBER OF LABS, TAKEN INTO ACCOUNT
0234          READ(5,1000) NLAB
              C          FOR SHORTER TEST OUTPUT
              C          NLAB=2
              C          PRINT 1002,NLAB
              C
              C          READING OF DATA CARDS,ORDER PRESCRIBED BY DATA SHEETS
              C          DO-LOOP OVER THE DIFFERENT LABS
0235          DO 100 M=1,NLAB
0236          PRINT 1004
0237          READ (5,1023) ILAB(M)
0238          SUMOUT(M)=0
              C          DO-LOOP OVER THE DIFFERENT SAMPLES,MORE THAN 8 BECAUSE OF U AND PU
0239          DO 99 IA=1,14
              C          FOR LAB 23 ONLY URANIUM HAS BEEN MEASURED. SO THE PU SPIKE SOLUTION
              C          AND ALL PU RATIOS OF THE OTHER SAMPLES HAVE TO BE SKIPPED
0240          IF(ILAB(M).NE.23) GO TO 20
0241          IRMELA =IA
0242          GO TO (21,99,21,21,21,99,99,99,21,21,21,99,99,99),IRMELA
              C          FOR LAB 16 ONLY SPIKE SOLUTIONS AND UNSPIKED SAMPLES HAVE BEEN MEASURED
0243          20 IF(ILAB(M).NE.16) GC TO 21
0244          IF(IA.GT.8) GO TO 996
              C          DO-LOOP OVER THE DIFFERENT RUNS
0245          21 DO 98 IB=1,3
              C          DO LOOP OVER THE DIFFERENT ISOTOPIIC RATIOS
0246          IRAT=4
0247          DO 97 IC=1,4
0248          LINKS=0
0249          MX=10
0250          NVAL=8
0251          IF(ILAB(M).EQ.15) GO TO 31
0252          IF(ILAB(M).EQ.18) GO TO 33
0253          30 READ (5,1001) IEXP(M),ILAB(M),SAMP(M),IRUN(M),IREL(M),(X(I),I=1,5)
0254          GO TO 32
0255          31 READ (5,1022) IEXP(M),ILAB(M),SAMP(M),IRUN(M),IREL(M),(X(I),I=1,5)
              C

```

```

FORTRAN IV G1  RELEASE 2.0          MAIN          DATE = 75015          13/37/58

      C   CHECKING OF THE ARTIFICIAL END OF FILE
0256      32 IF(IEXP(M).EQ.9) GO TO 101
0257      GO TO 34
0258      33 CALL COR(ILAB,SAMP,IRUN,IREL,X,IB,M,ITYPE,IEXP,      Y,A,NAUT,SUMD
      IUT,MEAN,SIGMA,SD,RSD,SQI,SUMM,DEC)
      NOUT=NAUT
0259      GO TO 79
0260      C   'TRANSLATION' OF THE SAMPLE TYPE
0261      34 DO 2 I=1,8
0262      II=I
0263      IF(SAMP(M).EQ.TYP(I)) GO TO 3
0264      2 CONTINUE
0265      3 ITYPE(M)=II
      PRINT 1005,(TYP(I),I=1,8),SAMP(M),ITYPE(M)
      C
      C   SOMETIMES SOME RATIOS WERE NOT MEASURED
      C   IF THEY COULD NOT BE MEASURED,X(1) IS PUT IN AS -1.
      C   IF IT WAS DECIDED NOT TO MEASURE THEM (BECAUSE E.G. FOR PU 238 ALPHA
      C   SPECTROMETRY WAS PREFERRED), X(1) IS PUT IN AS 9.9
0266      IF(X(1).GT.9.) GO TO 76
0267      IF(X(1).LT.0.) GO TO 77
      C
      C   TEST IF THERE ARE LESS THAN 5 MEASURING RESULTS
0268      DO 44 I=1,5
0269      II=I
0270      IF(X(I).GT.9.) GO TO 45
0271      44 CONTINUE
0272      GO TO 46
0273      45 MX=II-1
      C   PRINT 1007,II,X(II)
0274      GO TO 47
0275      46 IF(ILAB(M).EQ.15) GO TO 466
0276      READ (5,1001) IEXP(M),ILAB(M),SAMP(M),IRUN(M),IREL(M),(X(I),I=6,10
      1)
0277      GO TO 467
0278      466 READ (5,1022) IEXP(M),ILAB(M),SAMP(M),IRUN(M),IREL(M),(X(I),I=6,10
      1)
      C
      C   TEST IF THERE ARE LESS THAN 10 MEASURING RESULTS
0279      467 DO 4 I=6,MX
0280      II=I
0281      IF(X(I).GT.9.) GO TO 5
0282      4 CONTINUE
0283      GO TO 47
0284      5 MX=II-1
      C   PRINT 1007,II,X(II)
0285      47 PRINT 1002,ILAB(M),SAMP(M),IRUN(M),IREL(M)
      CC   PRINT 1003,(X(I),I=1,II)

```

FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 75015

13/37/58

```

C
C   X(I) ARE THE ISOTOPIC RATIOS IN ORIGINAL CRDER
C   Y(I) ARE THE ISOTOPIC RATIOS FOR FURTHER CALCULATIONS, COMING FROM
C   DIXON
C   ZZ(I) ARE THE ISOTOPIC RATIOS FOR FURTHER CALCULATIONS, COMING FROM
C   CHAUV1
C   Z(I) ARE THE ISOTOPIC RATIOS FOR FURTHER CALCULATIONS, COMING FROM
C   CHAUV2
C   IF THERE ARE LESS THAN 8 INPUT VALUES, NO OUTLIER CRITERIA ARE APPLIED
C
0286   49 IF(MX.GE.NVAL) GO TO 6
0287     NOUT=0
0288     NVAL=MX
C     PRINT 1008,NVAL
0289     DO 66 I=1,NVAL
0290   66 A(I)=X(I)
0291     GO TO 75
0292   6 CALL DIXON(X,Y,NOUT)
0293     SUMOUT(M)=SUMOUT(M)+NOUT
C     THE CHOSEN RESULTS ARE PASSED TO THE NEXT PROGRAM UNIT
C     PRINT 1011,(Y(I),I=1,NVAL)
C     CALL CHAUV2(X,Z)
C     PRINT 1013,(Z(I),I=1,NVAL)
0294     DO 7 I=1,NVAL
0295     7 A(I)=Y(I)
C
C
C
C     THE AM-241 CORRECTIONS APPLY ONLY TO THE RATIO 241/239
C     OF THE SAMPLES AU,BU,RU WITH ITYPE(M) = 3,4,5
0296   75 IF(IREL(M).NE.19) GO TO 729
0297     IF(ITYPE(M).LT.3) GO TO 729
0298     IF(ITYPE(M).GT.8) GO TO 729
0299     IT=ITYPE(M)
0300     MIT=ILAB(M)
0301     FACT= (DEC(MIT,IT,IB) * 0.693)/(15.1 * 365.)
0302   727 DO 728 I=1,NVAL
C     THE INDECES HAVE THE FOLLOWING MEANING:
C     M-LAB, IT-SAMPLE, IB-RUN
0303     A(I)=A(I) * DEXP(FACT)
0304   728 CONTINUE
0305   729 CONTINUE
C
C
C
C     FOR THE LABS 5 AND 13 MASS DISCRIMINATION CORRECTIONS
C     HAVE TO BE APPLIED 16.4.74
0306   IF(ILAB(M).EQ.5) GO TO 730

```

```

FORTRAN IV G1  RELEASE 2.0                MAIN                DATE = 75015        13/37/58

0307          IF(ILAB(M).EQ.13) GO TO 740
0308          GO TO 754

C
C
C   LAB 5
0309          730 MDC=1.
0310          IF(IREL(M).NE.83) GO TO 731
0311          MDC=1.00247
0312          731 IF(IREL(M).NE.38) GO TO 732
0313          MDC=0.99754
0314          732 IF(IREL(M).NE.48) GO TO 733
0315          MDC=0.99803
0316          733 IF(IREL(M).NE.58) GO TO 734
0317          MDC=0.99853
0318          734 IF(IREL(M).NE.68) GO TO 735
0319          MDC=0.99902
0320          735 DO 736 I=1,NVAL
0321          736 A(I)=A(I) * MDC
0322          GO TO 754

C
C   LAB 13 -- IN THIS CASE THE CORRECTIONS PARTLY
C   DEPEND ON THE ISOTOPIIC ABUNDANCIES
0323          740 MDC=1.
C   URANIUM
0324          IF(IREL(M).NE.83) GO TO 741
0325          MDC=1.0036
0326          741 IF(IREL(M).NE.38) GO TO 742
0327          MDC=0.9871
0328          742 IF(IREL(M).NE.48) GO TO 743
0329          MDC=0.99989
0330          743 IF(IREL(M).NE.58) GO TO 744
0331          IF(ITYPE(M).LT.8) MDC=0.99785
0332          IF(ITYPE(M).EQ.5) MDC=0.99979
0333          IF(ITYPE(M).EQ.8) MDC=0.99979
0334          744 IF(IREL(M).NE.68) GO TO 745
0335          MDC=0.99994

C
C   PLUTONIUM
0336          745 IF(IREL(M).NE.92) GO TO 747
0337          MDC=0.99991
0338          747 IF(IREL(M).NE.02) GO TO 748
0339          MDC=0.99992
0340          748 IF(IREL(M).NE.09) GO TO 749
0341          IF(ITYPE(M).LT.8) MDC=1.00213
0342          IF(ITYPE(M).EQ.5) MDC=1.0009
0343          IF(ITYPE(M).EQ.8) MDC=1.0009
0344          749 IF(IREL(M).NE.19) GO TO 750
0345          IF(ITYPE(M).LT.8) MDC=1.00335

```

FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 75015

13/37/58

```

0346          IF(ITYPE(M).EQ.5) MDC=1.00008
0347          IF(ITYPE(M).EQ.8) MDC=1.00012
0348      750 IF(IREL(M).NE.29) GO TO 751
0349          IF(ITYPE(M).LT.5) MDC=1.00216
0350          IF(ITYPE(M).GT.5) MDC=1.00847
0351      751 DO 752 I=1,NVAL
0352          752 A(I)=A(I) * MDC
C
0353      754 CONTINUE
C
C
0354          IF (NVAL.GT.1) GO TO 755
0355          MEAN = A(1)
0356          GO TO 788
C
C      LAB1 IS CALLED ONLY IF THERE ARE MORE THAN ONE SCANNING RESULTS PER
C      RATIO
0357      755 CALL LAB1 (A,MEAN ,SIGMA ,SD ,RSD,SQI,SUMM)
0358          GO TO 79
C
C      IF NO MEASURING RESULTS EXIST,NVAL MUST BE SET =0 .
C      MEAN,SIGMA,SD AND RSD ARE SET =0.  IN THIS CASE.
0359      76  NVAL=0
0360          MX=0
0361          NOUT=0
0362          PRINT 1002,ILAB(M),SAMP(M),IRUN(M),IREL(M)
0363          PRINT 1018
0364          GO TO 78
0365      77  NVAL=0
0366          NOUT=0
0367          MX=0
0368          PRINT 1002,ILAB(M),SAMP(M),IRUN(M),IREL(M)
0369          PRINT 1014
0370      78  MEAN=0.
0371      788 SIGMA=0.
0372          SD=0.
0373          RSD=0.
C
0374      79  J=IEXP (M)
0375          K=ITYPE (M)
0376          L=IRUN (M)
0377          DO 8 I=1,11
0378          II=I
0379          IF (IREL(M).EQ.IR(I))GO TO 9
0380      8  CONTINUE
0381      9  IRE (M)=II
C      PRINT 1016,M,IRE (M)
C      LL=IRE (M)
0382      C      PRINT 1015,M,J,K,L,LL

```



FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 75015

13/37/58

```

C      M IS THE RUNNING NUMBER FOR THE LAB
C      J IS THE NUMBER OF THE PART OF THE EXPERIMENT
C      K IS THE CODE FOR THE SAMPLE TYPE, E.G. FOR US OR BU, SEE REAL TYP
C      L IS THE RUN-INDEX
C      LL IS THE CODE FOR THE ISOTOPIC RATIO, SEE REAL*8 IR(11)
C      THE NAMES GET A 'V' AT THE END TO INDICATE THAT THEY ARE VECTORS
0383      MEANV(M,K,L,LL)=MEAN
0384      SIGMAV( K,L,LL)=SIGMA
0385      SDV( K,L,LL)=SD
0386      RSDV( K,L,LL)=RSD
0387      SQIV( K,L,LL)=SQI
0388      NAUS( K,L,LL)=NOUT
0389      SUMV(M,K,L,LL) = SUMM
C      PRINT 1017,J,K,L,LL,MEANV(M,K,L,LL),SIGMAV( K,L,LL),SDV( K,L,LL)
C      1,RSDV( K,L,LL)
0390      IF (NVAL.LT.8) GO TO 95
0391      GO TO 96
0392      95 CONTINUE
C      95 PRINT 1009,NVAL
0393      96 MVAL( K,L,LL)=NVAL
0394      LVAL(K,L,LL)=MX
0395      966 MVALV(M,K,L,LL)=NVAL
C
C      CONSIDERATIONS FOR SPECIAL CASES OF INPUT DATA
C      FOR US-SAMPLES ONLY ONE ISOTOPIC MEASUREMENT (38/33)
0396      IF (ITYPE(M).EQ.1) IRAT=1
C      FOR PS-SAMPLES ONLY TWO ISOTOPIC MEASUREMENTS(39/42 AND 40/42)
0397      IF (ITYPE(M).EQ.2) IRAT=2
0398      IF (IRAT.EQ.1C) GO TO 98
0399      97 CONTINUE
0400      98 CONTINUE
0401      99 CONTINUE
0402      996 PRINT 1006,ILAB(M) ,SUMOUT(M)
C
C      DO LOOPS FOR THE OUTPUT LIST I
C      THE ORDER WAS PRESCRIBED FOR THIS LIST
C      DO LOOP OVER THE DIFFERENT ISOTOPIC RATIOS
0403      DO 999 LL=1,11
C      DO LOOP OVER THE DIFFERENT SAMPLE TYPES
0404      DO 998 K=1,8
C      DO LOOP OVER THE DIFFERENT RUNS
0405      DO 997 L=1,3
C      THE VECTOR MEANV IS READ IN AS REAL*8 WITH ALL COMPONENTS=-3.
C      TRICK FOR STEERING THE OUTPUT
0406      IF (MEANV(M,K,L,LL).LT.-2.) GO TO 997
0407      PRINT 1010 , IR(LL),TYP(K),L,MEANV(M,K,L,LL),SIGMAV( K,L,LL),
      1SDV(K,L,LL),RSDV(K,L,LL),LVAL(K,L,LL),NAUS(K,L,LL)
0408      997 CONTINUE

```

FORTRAN IV G1 RELEASE 2.0

MAIN

DATE = 75015

13/37/58

```

0409          998 CONTINUE
0410          999 CONTINUE
0411          CALL LAB2 (MEANV, SQIV, M, IR, TYP, ILAB, MVAL, MEAN2, IFILIV)
0412          CALL LAB3 (MEAN2, M, IR, TYP, ILAB, NALPHA, WALPHA)
0413          IF (ILAB(M).EQ.16) GO TO 80
0414          CALL LAB4 (MEAN2, M, MEANV, NALPHA, WALPHA, ILAB, TYP)
0415          GO TO 100
0416          80 PRINT 1028
0417          100 CONTINUE
0418          CALL LAB5(NLAB, MVALV, IFILIV, MEANV, SUMV, IR, TYP, ILAB)
0419          CALL LAB6(MEANV, IR, TYP, ILAB, MEAN2, IFILIV, NLAB)

C
C PRINTING OF A COMBINED TABLE OF RESULTS OF LAB2 FOR ALL LABS
C IF (NLAB.GT.0) GO TO 101
C NONSENSE CONDITION IN ORDER TO SKIP THE FOLLOWING PRINT STATEMENTS
0420          PRINT 1027
0421          L2=1
C DO-LOOP OVER THE ISOTOPIC RATIOS
0422          DO 199 L=1,13
0423             LL2=L
0424             IF(L.EQ.2) GO TO 199
0425             IF(L.EQ.3) GO TO 199
0426             IF(L.EQ.8) LL2=2
0427             IF(L.EQ.9) LL2=3
0428             IF(L.GT.9) LL2=L-2
C EXTRA BLOCK FOR PLUTONIUM RESULTS
0429             IF(L.EQ.8) PRINT 1024
C DO-LOOP OVER THE SAMPLE TYPES
0430             DO 198 K2=1,8
C TRICK FOR ORGANISATION OF THE OUTPUT LIST
0431             IF (MEANV( 2,K2,LL2).LT.-2.) GO TO 198
0432             PRINT 1025 ,IR(LL2),TYP(K2)
0433             DO 197 M2=1,NLAB
0434             PRINT 1026,ILAB(M2),MEAN2(M2,K2,LL2),SG2V(M2,K2,LL2),SDGV(M2,K2,LL
12),RSDGV(M2,K2,LL2),          SIGA2V(M2,K2,LL2),SDAV(M2,K2
2,LL2),RSCAV(M2,K2,LL2) ,SIGE2V(M2,K2,LL2),SDEV(M2,K2,LL2), RSDEV
3(M2,K2,LL2)
0435             197 CONTINUE
0436             198 CONTINUE
0437             199 CONTINUE

C
C AT THIS POINT THE ADDITIONAL EXPERIMENTS WILL BE CONNECTED WITH
C THE STANDARD EXPERIMENT.
0438          200 IDER=0
0439          PRINT 1002,IDER
C CORRECTION FACTORS FOR LAB 18 - DRY SPIKE EXPERIMENT
0440          DICO(1)=0.9946
0441          DICO(2)=1.0033

```

```
FORTRAN IV G1  RELEASE 2.0          MAIN          DATE = 75015          13/37/58

0442          DO 103  M=1,NLAB
0443          READ (5,1029)  ILAB(M),TEST
0444          IF(TEST.GT.7.9) GO TO 103
0445          CALL DRY(M,MEAN2,ILAB,DICO)
0446          103 CONTINUE
          C
          C    CORRECTION FACTORS FOR LAB 18 - ALU-I-EXPERIMENT
0447          DICO(1)=1.
0448          DICO(2)=1.006
0449          DO 104 M=1,NLAB
0450          READ(5,1029)  ILAB(M),TEST
0451          IF(TEST.GT.7.9) GO TO 104
0452          CALL CRY(M,MEAN2,ILAB,DICO)
0453          104 CONTINUE
0454          PRINT 1002, IDER
0455          IDER=IDER + 1
0456          PRINT 1002, IDER
          C    FOR INCLUSION OF ALU - -EXPERIMENT
          C    IF(IDER.EQ.1) GO TO 200
          C
0457          101 CONTINUE
          C
0458          STOP
0459          END
```

FORTRAN IV G1 RELEASE 2.0

BLK DATA

DATE = 75015

13/37/58

0001

**BLOCK DATA**C  
C  
C  
C  
C

0002

COMMON /HANNA / LINKS,MX,NVAL  
 1/ITEMS/SG2V(21,8,11),SDGV(21,8,11),RSDGV(21,8,11),RSDCOV(21,8,11),  
 1SIGA2V(21,8,11),SDAV(21,8,11),RSDAV(21,8,11),SIGE2V(21,8,11),SDEV  
 2(21,8,11),RSDEV(21,8,11),DIFV(21,8,11)  
 3/DRYA/ KO(21,3,5,2),KOAL(21,3,5,2)

0003

REAL \* 8 SG2V/1848\*C./,SDGV/1848\*0./,RSDGV/1848\*0./,RSDCOV/1848\*0.  
 1/,SIGA2V/1848\*0./,SDAV/1848\*0./,RSDAV/1848\*0./,SIGE2V/1848\*C./,  
 2SDEV/1848\*0./,RSDEV/1848\*0./,DIFV/1848\*0./,KO,KOAL

0004

END

FORTRAN IV G1 RELEASE 2.0

DIXON

DATE = 75015

13/37/58

```

0001      SUBROUTINE DIXON(X,Y,NOUT)
          C
          C
          C
          C
          C
          C
          C      APPLICATION OF THE DIXON CRITERION FOR OUTLIERS
          C
0002      COMMON /HANNA / LINKS,MX,NVAL
0003      REAL *8  X(10),Y(10),S(10),F(3),T(10),D,DIX,R,D1,DN,DN1,DN2
          C
0004      1001 FORMAT (' DN1 IS =0.')
```

```

0005      1002 FORMAT (' DN2 IS =0.')
```

```

0006      1003 FORMAT (/, ' APPLICATION OF THE DIXON CRITERION')
```

```

0007      1004 FORMAT (' R=',F10.6, ' LINKS=',I2, ' NX=',I3)
```

```

0008      1005 FORMAT (' THE MAXIMUM VALUE S(',I3,')=',F9.6, ' IS CANCELLED')
```

```

0009      1006 FORMAT (' THE MINIMUM VALUE S(1)=',F9.6, ' IS CANCELLED')
```

```

0010      1007 FORMAT (' THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):
          1',/,40X,8F10.6)
```

```

0011      1008 FORMAT (' THE ORDERED VALUES S(I) ARE: ',10F9.6)
```

```

0012      1009 FORMAT (' D1,DN,DN1,DN2,D= ',5F10.6)
```

```

0013      1010 FORMAT(' ***** THE 8TH INPUT VALUE IS AN OUTLIER*****')
```

```

0014      1011 FORMAT (' THE RESULTING VALUES X(I) ARE',10X,8F10.6)
```

```

0015      1012 FORMAT (/' BEGIN OF A NEW TEST SERIES WITH DIFFERENT ALPHA')
```

```

0016      1013 FORMAT (' R,DIX AND K ARE:',2F14.6,I3)
```

```

0017      1014 FORMAT (' THE NUMBER OF OUTLIERS IS',I5)
```

```

0018      1015 FORMAT (' THE DIXON INPUT VALUES ARE',10F9.6)
          C
0019      NX=MX
          C      PRINT 1003
0020      DO 5 II=1,NX
0021      5 T(II)=X(II)
          C      PRINT 1015,(T(II),II=1,NX)
0022      6 CONTINUE
0023      DO 7 I=1,3
0024      7 F(I)=99.
0025      K=0
          C      ORDERING OF INPUT VALUES
          C      T(I) ARE THE ISOTOPIC RATIOS ORDERED, BEGINNING WITH THE SMALLEST
          C      THE VECTOR T(I) HAS TO BE DEFINED BEFORE THE ORDERING PROCEDURE
          C      NXX=NX
0026      8 CALL MINL(T,NXX)
0027      PRINT 1008,(T(I),I=1,NX)
0028      NX=MX
0029      LINKS=0
0030      NOUT=0
0031      DO 88 II=1,NX
0032
```

```

FORTRAN IV G1  RELEASE 2.0          DIXON          DATE = 75015          13/37/58

0033          88 S(II)=T(II)
0034          9  D1=S(2)-S(1)
0035          DN=S(NX)-S(NX-1)
0036          DN1=S(NX-1)-S(1)
0037          DN2=S(NX)-S(2)
0038          D=D1-DN
0039          C  PRINT 1009,D1,DN,DN1,DN2,D
0040          IF(DN1.LT.0.0000009) GO TO 112
0041          116 IF(DN2.LT.0.0000009) GO TO 113
0042          GO TO 111
0043          112 CONTINUE
0044          C 112 PRINT 1001
0045          IF(D1.GT.0.000000) GO TO 114
0046          GO TO 116
0047          114 R=0.5
0048          IF(D) 13,13,116
0049          113 CONTINUE
0050          C 113 PRINT 1002
0051          IF (DN.GT.0.000000) GO TO 115
0052          GO TO 23
0053          115 R=0.5
0054          LINKS=1
0055          GO TO 13
0056          C
0057          C  EVALUATION OF EXTREMA AND DECISION, WHERE AN OUTLIER IS MORE PROBABLE
0058          C  IF D IS POSITIVE, THE OUTLIER LIES AT THE LEFT END
0059          111 IF (D) 10,12,12
0060          12 R=D1/DN1
0061          GO TO 13
0062          10 R=DN/DN2
0063          LINKS=1
0064          C  THIS PARAMETER IS A HINT TO THE LEFT PART OF THE FLOWSHEET; AN OUTLIER
0065          C  WITH A HIGH VALUE IS ASSUMED, IF LINKS=1
0066          C 13 PRINT 1004,R,LINKS,NX
0067          13 CCNTINUE
0068          C
0069          C  APPLICATION OF THE CRITERION FOR 8,9 AND 10 INPUT VALUES
0070          IF(NX-9) 14,15,16
0071          C  THESE DIX-VALUES CORRESPOND TO THE PROBABILITY 1-ALPHA =0.95
0072          14 DIX=0.554
0073          K=3
0074          C  K IS THE NUMBER OF THE OUTLIER FOUND FOR THIS SERIES OF VALUES
0075          GO TO 17
0076          15 DIX=0.512
0077          K=2
0078          GO TO 17
0079          16 DIX=0.477
0080          K=1

```

FORTRAN IV G1 RELEASE 2.0

DIXON

DATE = 75015

13/31/58

```

0068      C 17 PRINT 1013,R,DIX,K
0069      17 CONTINUE
0070      IF (R-DIX) 26,18,18
0071      18 IF(K.EQ.3) PRINT 1010
0072      NOUT=NOUT+1
0073      IF (LINKS.NE.1) GO TO 20
0074      IF(K.EQ.3) GO TO 23
0075      19 PRINT 1005,NX,S(NX)
0076      C 19 CONTINUE
0077      F(K)=S(NX)
0078      NX=NX-1
0079      LINKS=0
0080      GO TO 9
0081      20 IF(K.EQ.3) GO TO 23
0082      PRINT 1006,S(1)
0083      F(K) =S(1)
0084      NN=NX-1
0085      DO 22 I=1,NN
0086      S(I)=S(I+1)
0087      22 CONTINUE
0088      NX=NX-1
0089      GO TO 9
0090      C
0091      C IF THERE ARE NO OUTLIERS, THE FIRST 8 VALUES ARE USED FROM HERE ON
0092      26 IF (K.NE.1) GO TO 23
0093      PRINT 1007,(X(I),I=1,NVAL)
0094      DO 27 I=1,NVAL
0095      27 Y(I)=X(I)
0096      GO TO 99
0097      23 II=0
0098      DO 25 I=1,MX
0099      DO 24 J=1,3
0100      IF(X(I).NE.F(J)) GO TO 24
0101      F(J) = 99.
0102      GO TO 25
0103      24 CONTINUE
0104      II=II+1
0105      Y(II)=X(I)
0106      25 CONTINUE
0107      PRINT 1011,(Y(I),I=1,NVAL)
0108      99 CONTINUE
0109      IF(NOUT.LE.0) GO TO 21
0110      INSEL=1
0111      PRINT 1014,NOUT
0112      21 CONTINUE
0113      RETURN
0114      END

```

FORTRAN IV G1 RELEASE 2.0

MINL

DATE = 75015

13/37/58

```

0001      SUBROUTINE MINL (T,NXX)
          C
          C
          C
          C
          C
          C      THIS SUBROUTINE PUTS DOUBLE PRECISION FLOATING POINT NUMBERS
          C      (VECTOR T(NXX) ) IN ASCENDING ORDER
          C
0002      REAL*8 R(10),MINI,KK,T(10)
          C
0003      1000 FORMAT (' NXX AND THE MINL INPUT VALUES ARE',I3,10F9.6)
          C
          C
          C      NOTE THE INPUT-VECTOR
0004      DO 10 I=1,NXX
0005      R(I) = T(I)
0006      10 CONTINUE
          C      PRINT 1000,NXX,(R(I),I=1,NXX)
          C      MARKER FOR SCANNING THE INPUT NUMBERS
0007      LOOP=0
0008      300 LOOP=LOOP+1
0009      L=0
0010      I=1
0011      N=1
          C      THE 1. INPUT VALUE IS EXPERIMENTALLY CALLED THE MINIMUM
0012      MINI = R(I)
0013      100 I=I+1
0014      KK=MINI-R(I)
          C      TEST IF THE NEXT NUMBER IS SMALLER THAN THE PRESENT MINIMUM
0015      IF(KK)200,200,20
0016      20 MINI=R(I)
0017      N=I
0018      200 L=I-NXX
0019      IF(L)100,30,100
          C      NOTE THE MINIMUM, CCNSTRUCT THE OUTPUT VECTOR
0020      30 T(LOOP)=MINI
0021      R(N)=10.
          C      NUMBERS WHICH HAVE BEEN TESTED ARE PUT OUT OF USE, 10 IS BIGGER
          C      THAN ALL ACTUAL INPUT VALUES
0022      LOOP=LOOP-NXX
0023      IF(LOOP)300,400,300
0024      400 CONTINUE
0025      RETURN
0026      END

```



FORTRAN IV G1 RELEASE 2.0

COR

DATE = 75015

13/37/58

0001

```
SUBROUTINE COR(ILAB, SAMP, IRUN, IREL, X, IB, M, ITYPE, IEXP,
Y, A, NAUT, SUMOUT, MEAN, SIGMA, SD, RSD, SQI, SUMM, DEC)
```

```
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
C
```

```
CORRECTIONS FOR LAB 18 SPECIALLY
FOR THIS LAB THE MEASURED VALUES ARE READ IN AS B(I) AND CORRECTED
BY MULTIPLICATION WITH THE FACTOR CO. IF CO=0, THE GIVEN INPUT VALUES ARE
CORRECT.
THE FACTOR CO IS GIVEN IN COLUMNS 61-70 OF THE SECOND INPUT CARD
FOR EACH MEASURING SERIES
IN ADDITION THEY ARE MULTIPLIED BY THE MASS DISCRIMINATION CORRECTION
FACTOR DICO WHICH DEPENDS ON THE ISOTOPIC RATIO.
THE ISOTOPIC RATIOS ARE READ IN AS IREL(M) AND ARE COMPARED WITH THE LIST
OF RATIOS GIVEN IN AS THE VECTOR IR(1-11). IREL(M) = 83 CORRESPONDS E.G.
TO THE RATIO U238/233 AND IS THE VECTOR COMPONENT IR(1).
THE CORRESPONDING CORRECTION FACTOR DICO IS TRANSMITTED AS MD(1)

CORRESPONDING COMMENTS CAN BE FOUND IN THE PARALLEL PART OF THE MAIN
PROGRAM
```

0002

0003

0004

0005

0006

0007

```
REAL*8 SIGMA, SQI, MEAN, X(10), Y(10), A(8), B(10)
DIMENSION ILAB(21), SAMP(21), IRUN(21), IREL(21), ITYPE(21), IEXP
1(21)
REAL TYP(8) / 'US ', 'PS ', 'AU ', 'BU ', 'RU ', 'AS ', 'BS ', 'RS
1' /
INTEGER IR(11) / 83, 92, 02, 38, 48, 58, 68, 89, 09, 19, 29 /
INTEGER * 2 SUMOUT (21)
REAL * 8 MD(11), CO, DICO, SD, RSD, SUMM, DEC(22, 8, 3), FACT
```

C

0008

0009

```
COMMON / HANNA / LINKS, MX, NVAL
1001 FORMAT (I1, I2, 1X, A2, I1, 1X, I2, 5F10.6)
```

C

0010

0011

0012

0013

0014

0015

0016

```
1000 FORMAT (I2)
1002 FORMAT (/ I3, 2X, A2, I1, 1X, I2)
1003 FORMAT (I1, I2, 1X, A2, I1, 1X, I2, 6F10.6)
1004 FORMAT (' DICO=', F10.4)
1005 FORMAT (' TYP(I=1....8)=' , 8(A2, 2X), ', SAMP(M) AND ITYPE(M)=' , A2, I3)
1006 FORMAT (F11.8)
1007 FORMAT (' THE VALUE X(II=' , I2, ')=' , F10.6, ' WAS CANCELLED, BECAUSE
1IT INDICATED END OF DATA')
1008 FORMAT (//, ' NVAL WAS SET =', I4, '//, ' *****')
1009 FORMAT (' CO=' , F10.6)
```

0017

0018

FORTRAN IV G1 RELEASE 2.0                      COR                      DATE = 75015                      13/37/58

```

0019      1014 FORMAT (' A SERIES OF X(I) COULD NOT BE MEASURED')
0020      1018 FORMAT (' A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE')
0021      1022 FORMAT (I1,I2,1X,A2,I1,1X,I2,5(F9.6,1X))
C
0022      READ (5,1001) IEXP(M), ILAB(M), SAMP(M), IRUN(M), IREL(M), (B(I),
           1I=1,5)
0023      34 DO 2 I=1,8
0024          II=I
0025          IF(SAMP(M).EQ.TYP(I)) GO TO 3
0026      2 CONTINUE
0027      3 ITYPE(M)=II
0028          IF(E(I).GT.9.) GO TO 76
0029          IF(B(I).LT.0.) GO TO 77
0030          DO 44 I=1,5
0031              II=I
0032              IF(B(I).GT.9.) GO TO 45
0033      44 CONTINUE
0034          GO TO 46
0035      45 MX=II-1
0036          GO TO 47
0037      46 CONTINUE
0038      READ (5,1003) IEXP(M),ILAB(M),SAMP(M),IRUN(M),IREL(M),(B(I),I=6,10
           1), CO
0039      467 DO 4 I=6,MX
0040          II=I
0041          IF(B(I).GT.9.) GO TO 5
0042      4 CONTINUE
0043          GO TO 47
0044      5 MX=II-1
0045      47 PRINT 1002,ILAB(M),SAMP(M),IRUN(M),IREL(M)
C
0046      47 CCNTINUE
C
0047      MD(1)=1.0056
0048      MD(2)=0.9940
0049      MD(3)=0.9960
0050      MD(4)=0.9946
0051      MD(5)=0.9956
0052      MD(6)=0.9967
0053      MD(7)=0.9878
0054      MD(8)=0.9989
0055      MD(9)=1.002
0056      MD(10)=1.004
0057      MD(11)=1.006
0058      DICO =1.
0059      DO 8 I=1,11
0060          IK=I
0061          IF(IREL(M).EQ.IR(I)) GO TO 9
           8 CONTINUE

```

```

FORTRAN IV G1  RELEASE 2.0                COR                DATE = 75015                13/37/58

0062          9 DICO=MD(IK)
              C PRINT 1004,DICO
0063          IF (CO.LE.0.) CO = 1.
              C PRINT 1009,CO
0064          DO 50 I=1,II
0065          50 X(I) = B(I) * CO * DICO
              C PRINT 1006,(B(I),I=1,II)
              C PRINT 1006,(X(I),I=1,II)
0066          IF(MX.GE.NVAL) GO TO 6
0067          NVAL=MX
0068          DO 66 I=1,NVAL
0069          66 A(I)=X(I)
0070          GO TO 75
0071          6 CALL DIXON(X,Y,NOU)
0072          NAUT=NOU
0073          SUMOUT(M)=SUMOUT(M)+NOU
0074          DO 7 I=1,NVAL
0075          7 A(I)=Y(I)

              C
              C
              C AM-CORRECTIONS,SEE MAIN PROGRAM
0076          75 IF(IREL(M).NE.19) GO TO 729
0077          IF(ITYPE(M).LT.3) GO TO 729
0078          IF(ITYPE(M).GT.8) GO TO 729
0079          IT=ITYPE(M)
0080          MIT=ILAB(M)
0081          FACT=(DEC(MIT,IT,IB)*0.693)/(15.1 * 365.)
0082          727 DO 728 I=1,NVAL
0083          A(I)=A(I) * DEXP(FACT)
0084          728 CONTINUE
0085          729 CONTINUE

              C
              C
0086          IF (NVAL.GT.1) GO TO 755
0087          MEAN = A(1)
0088          GO TO 788
0089          755 CALL LAB1 (A,MEAN ,SIGMA ,SD ,RSD,SQI,SUMM)
0090          GO TO 79
0091          76 NVAL=0
0092          PRINT 1002,ILAB(M),SAMP(M),IRUN(M),IREL(M)
0093          PRINT 1018
0094          GO TO 78
0095          77 NVAL=0
0096          PRINT 1002,ILAB(M),SAMP(M),IRUN(M),IREL(M)
0097          PRINT 1014
0098          78 MEAN=0.
0099          788 SIGMA=0.
0100          SD=0.

```

FORTRAN IV G1 RELEASE 2.0

COR

DATE = 75015

13/37/58

0101  
0102  
0103  
0104

RSD=0.  
79 CONTINUE  
RETURN  
END

FORTRAN IV G1 RELEASE 2.0

LAB1

DATE = 75015

13/37/58

```

0001      SUBROUTINE LAB1 (A,MEAN , SIGMA ,SD ,RSD,SQI,SUMM)
C
C
C
C
C
C
C
C      THIS SUBROUTINE CALCULATES THE MEAN VALUE MEAN, THE VARIANCE SIGMA,THE
C      STANCARD DEVIATION SD, AND THE RELATIVE STANDARD DEVIATION RSD FOR EACH
C      ISOTOPIC RATIO AND EACH RUN
C      THIS SERIES OF RESULTS IS GIVEN FOR EACH LAB.
C      THE RESULTS OF THIS SUBROUTINE ARE PRINTED IN THE MAIN PROGRAM
C
0002      REAL*8 D,SIG,SIGMA,A(8),MEAN,SD,RSD,SQI,SUM,D,SUMM
0003      COMMON /HANNA / LINKS,MX,NVAL
C
0004      1000 FERMAT (' MEAN=',D12.6,' A(I)= ',8F10.6)
C
C
0005      22 SUM=0.
0006          SUMM=0.
0007          DO 3 I=1,NVAL
0008              SUMM=SUMM+A(I)**2
0009          3 SUM=SUM+A(I)
0010          4 MEAN=SUM/NVAL
0011              SQI=0.
0012              DO 6 I=1,NVAL
0013                  D=A(I) - MEAN
0014          6 SQI=SQI+D**2
0015              SIGMA=SQI/(NVAL-1)
0016              SD=DSQRT (SIGMA)
0017              IF (SD.LT.0.000000) GO TO 10
0018              IF (MEAN.LE.0) GO TO 9
0019              RSD=(SD/MEAN)* 100.
0020              GO TO 11
0021          9 PRINT 1000,MEAN,(A(I),I=1,NVAL)
0022          10 RSD=0.
0023          11 CONTINUE
0024      RETURN
0025      END

```

FORTRAN IV G1 RELEASE 2.0

LAB2

DATE = 75015

13/37/58

0001

SUBROUTINE LAB2 (MEANV,SQIV,M,IR,TYP,ILAB,MVAL,MEAN2,IFILIV)

C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C

THIS SUBROUTINE CALCULATES FOR EACH LAB THE MEAN, SIGMA, SD AND RSD  
FOR ALL THE ISOTOPIIC RATIOS OF EACH MEASURED SAMPLE.  
IN ADDITION IT GIVES THE RUN-AND SCAN-COMPONENTS OF SIGMA,SD AND RSD

M - LAB  
K2 - SAMPLE TYPE  
L2 - RUN  
LL2 - ISOTOPIIC RATIO

0002

```
REAL * 8 SQIV( 8,3,11),MEANV(21,8,3,11),MEANA,MEAN2(21, 8,11),SU
IM,SUMA,SQL,SG2,SDG,RSDG,SQA,SQR,SIGA2,SDA,SIGE2,SDE,RSDE,RSDCON,DI
2FF,RSCA
```

0003

```
DIMENSION ILAB(21),IR(11),TYP( 8)
```

0004

```
INTEGER * 2 IFILIV(21,8,11),MVAL(8,3,11)
```

0005

```
COMMON /HANNA / LINKS,MX,NVAL
1/ITEMS/SG2V(21,8,11),SDGV(21,8,11),RSDGV(21,8,11),RSDCCV(21,8,11),
ISIGA2V(21,8,11),SDAV(21,8,11),RSDAV(21,8,11),SIGE2V(21,8,11),SDEV
2(21,8,11),RSDEV(21,8,11),DIFV(21,8,11)
REAL*8 SG2V,SDGV,RSDGV,RSDCOV,SIGA2V,SDAV,RSDAV,SIGE2V,SDEV,RSDEV,
1DIFV
```

0006

C  
C  
C

0007

```
1000 FORMAT(//'RESULTS OF PROGRAM LAB2 FOR THE STANDARD EXPERIMENT'//
1' FOR EACH LAB AND FOR EACH SAMPLE AND ISOTOPIIC RATIO THE FOLLOWIN
3G VALUES ARE GIVEN:'// LAB MEAN(MEANA) WITH VARIANCE(SG2),STANDARD
4 DEVIATION(SDG) AND RELATIVE STANDARD DEVIATION(RSDG).'/ ' IN ADDIT
5ION VARIANCE COMPONENTS FOR RUN AND SCAN ARE GIVEN SEPARATELY .'/
6 ' (IF THE RUN OR SCAN COMPONENT OF ANY VARIANCE IS NEGATIVE ,
7 ' AND THEREFORE HAS TO BE CONSIDERED AS NEGLIGIBLE,
8 ' IT IS PUT = -5.0 (RUN) OR = -4.0 (SCAN), RESPECTIVELY.)'// )
```

0008

```
1001 FORMAT ('// LABORATORY CODE NO.:'
1,I3//' RAT SAMP LAB MEAN',5X,'VARIANCE STAND.DEV. RSD(%) ',
25X,'VARIANCE SD',9X,'RSD(%)',6X,' VARIANCE',4X,'SD',9X,'RSD(%)'
3/23X,'LAB MEAN LAB MEAN LAB MEAN',4X,'RUN',9X,'RUN',8X,
4 'RUN',9X,'SCAN',8X,'SCAN',7X,'SCAN')
```

0009

```
1002 FORMAT (I3,2X,A2,2X, D10.4,4X,D10.4,2X,D9.3,2X,D9.3,3X,D10.4,
12X, D9.3,2X,D9.3,3X,D10.4,2X,D9.3,2X,D9.3)
```

0010

```
1003 FORMAT (///)
```

0011

```
1004 FCRMAT (' SQA=',D11.5, ' SQR=',D11.5, ' NVAL=',I3, ' IFIL= ',I3)
```

```

FORTRAN IV G1  RELEASE 2.0          LAB2          DATE = 75015          13/37/58

0012          1005 FORMAT (' ALARM      ALARM      ALARM      ')
0013          1006 FORMAT (' THE RUN COMPONENT OF THE VARIANCE IS NEGATIVE,IT IS PUT
              1= -5.')
0014          1007 FORMAT (' THE SCAN COMPONENT OF THE VARIANCE IS NEGATIVE,IT IS PUT
              1=-4.')
C
C          PRINTING OF THE TITEL OF THE PAGE
0015          PRINT 1000
C          PRINTING OF THE LAB AND TITEL CF THE LIST
0016          PRINT 1001,ILAB(M)
C
C          THE DO LOOPS ARE ARRANGED IN THE ORDER WANTED FOR THE OUTPUT LIST
C          THE NCRMAL ORDER OF ISOTOPIC RATIOS IS: 83,92,02,38,48,58,68,89,09,19,29
C          THE ORDER WISHED HERE IS:                83,38,48,58,68,92,02,89,09,19,29
C
0017          L2=1
0018          M2=M
C          DO-LOOP OVER THE ISOTOPIC RATIOS
0019          DO 999 L=1,13
0020          LL2=L
0021          IF(L.EQ.2) GO TO 999
0022          IF(L.EQ.3) GO TO 999
0023          IF(L.EQ.8) LL2=2
0024          IF(L.EQ.9) LL2=3
0025          IF(L.GT.9) LL2=L-2
C          DO-LOOP OVER THE SAMPLE TYPES
0026          DO 598 K2=1,8
C          TRICK FGR ORGANISATION OF THE OUTPUT LIST
0027          IF(MEANV(M ,K2,L2,LL2).LT.-2.) GO TC 998
0028          IFIL=3
0029          IFILI=3
0030          SUM=0.
C          CALCULATION OF THE LAB MEAN
0031          DO 5 LA=1,IFIL
C          TEST,IF A RUN IS MISSING
0032          IF(MVAL( K2,LA,LL2).EQ.0) IFILI=IFILI-1
0033          5 SUM=SUM+MEANV(M ,K2,LA,LL2)
0034          IFILIV(M,K2,LL2) = IFILI
0035          84 IF(IFILI.EQ.0) GO TO 86
0036          MEANA=SUM/IFILI
0037          GO TO 87
0038          86 MEANA=0.
0039          87 MEAN2 (M ,K2,LL2)=MEANA
0040          SCL=0.
0041          SG2=0.
0042          SDG=0.
0043          RSDG=0.
C

```

FORTRAN IV G1 RELEASE 2.0

LAB2

DATE = 75015

13/37/58

```

0044          SQA=0.
0045          SQR=0.
0046          SIGA2=0.
0047          SDA=0.
0048          RSDA=0.

C
0049          SIGE2=0.
0050          SDE=0.
0051          RSDE=0.
0052          RSDCON=0.
0053          DIFF=0.
0054          88 NVAL=MVAL( K2,LL2)
0055          IF(NVAL.LT.2) GO TO 14
C          IF THERE ARE LESS THAN 2 MEASUREMENTS IN THE FIRST RUN,
C          NO ANALYSIS OF VARIANCES CAN BE MADE
C
C          CALCULATION OF THE VARIANCE
0056          SUMA=0.
0057          DO 7 LA=1,IFIL
0058          NVAL=MVAL(K2,LA,LL2)
0059          IF(NVAL.LT.2) GO TO 7
0060          SUMA=SUMA+(MEANV(M ,K2,LA,LL2)-MEANA)**2
0061          7 CONTINUE
0062          SQL=SUMA
0063          IF(IFILI.LE.1) GO TO 89
0064          SG2=SQL/(IFILI*(IFILI-1))
0065          SDG=DSQRT(SG2)
0066          IF(MEANA.LE.0.) GO TO 89
0067          RSDG=(SDG/MEANA)*100.
0068          89 CONTINUE

C
C          THE FOLLOWING CALCULATION IS O.K. FOR THE NORMAL CASE WITH 8 SCANS PER RUN
0069          SQA=8.*SQL
0070          DO 9 LA=1,IFIL
0071          9 SQR=SQR+SQIV( K2,LA,LL2)
C          PRINT 1004,SQA,SQR,NVAL,IFIL
0072          NVAL=8
0073          IF(IFILI.LE.1) GO TO 15
0074          SIGA2=(1./NVAL)*((SQA/(IFILI-1))-(SQR/(IFILI*(NVAL-1))))
0075          GO TO 18
0076          15 SIGA2=-5.
0077          18 IF (SIGA2.LT.0.00000000) GO TO 10
0078          SDA=DSQRT(SIGA2)
0079          IF(MEANA. LT.0.) GO TO 11
0080          RSDA=(SDA/MEANA)*100.
0081          GO TO 12
C 10 PRINT 1006

```



```

FORTRAN IV G1  RELEASE 2.0                LAB2                DATE = 75015                13/37/58

0082          10 CONTINUE
0083          SCA =-5.
0084          11 RSDA = -5.
0085          12 NUM=IFILI*NVAL
0086          IF((NUM-IFILI).LE.0) GO TO 16
0087          SIGE2=SQR/(NUM-IFILI)
0088          GO TO 19
0089          16 SIGE2=-4.
0090          19 IF(SIGE2.LT.0.00000000) GO TO 20
0091          SDE=DSQRT(SIGE2)
0092          IF(MEANA. LE.0.) GO TO 13
0093          RSDE=(SDE/MEANA)*100.
0094          IF(IFILI.EQ.0) GO TO 17
0095          IF(NUM.EQ.0) GO TO 17
0096          IF((SIGA2/IFILI +SIGE2/NUM).LE.0.) GO TO 17
0097          RSDCON=(100./MEANA)*DSQRT(SIGA2/IFILI +SIGE2/NUM)
C            THIS NUMBER IS CALCULATED FOR CONTROL
0098          DIFF=RSDG-RSDCON
0099          GO TO 14
0100          20 PRINT 1007
0101          SDE=-4.
0102          13 RSDE=-4.
0103          17 RSDCON=-4.
0104          DIFF=-4.
C            PRINTING OF THE RESULTS
0105          14 IF(L.EQ.8) PRINT 1003
C            EXTRA BLOCK FOR PLUTONIUM RESULTS
0106          PRINT 1002,IR(LL2),TYP(K2),MEANA,SG2,SDG,RSDG,SIGA2,SDA,RSD
1A,SIGE2,SDE,RSDE
0107          SG2V (M2,K2,LL2)=SG2
0108          SDGV (M2,K2,LL2)=SDG
0109          RSDGV (M2,K2,LL2)=RSDG
0110          RSDCOV(M2,K2,LL2)=RSDCON
0111          SIGA2V(M2,K2,LL2)=SIGA2
0112          SDAV (M2,K2,LL2)=SDA
0113          RSDAV (M2,K2,LL2)=RSDA
0114          SIGE2V(M2,K2,LL2)=SIGE2
0115          SDEV (M2,K2,LL2)=SDE
0116          RSDEV (M2,K2,LL2)=RSDE
0117          DIFV (M2,K2,LL2)=DIFF
0118          998 CCNTINUE
0119          999 CONTINUE
0120          RETURN
0121          END

```

FORTRAN IV G1 RELEASE 2.0

ALPHA

DATE = 75015

13/37/58

```

0001      SUBROUTINE ALPHA (M3,WALPHA,NALPHA,MEAN3,ILAB,TYP)
          C
          C
          C
          C
          C
          C      THIS SUBROUTINE READS THE RESULTS OF ALPHA SPECTROMETRY
          C      NORMAL RESULTS BEGIN WITH A CARD WITH +7. IN THE FIRST 3 COLUMNS.
          C      IF NO RESULTS ARE AVAILABLE,A CARD WITH +8. IN COLUMNS 1-3 IS READ
          C      THE VECTOR VALPHA HAS THREE INDICES.
          C      THE FIRST(M3)STANDS FOR THE LAB
          C      THE SECCND (K3) STANDS FOR THE SAMPLE TYPE
          C      THE THIRD (L3) STANDS FOR THE REPETITION OF MEASUREMENT
          C      THE INPUT VALUES FOR THIS SUBROUTINE ARE RATIOS OF ALPHA-ACTIVITI
          C      R(238/239+240)
          C
0002      REAL *8 WALPHA (21,8), MEAN3(21, 8,11), VALPHA(21,8,3)
0003      REAL*8 XSUM, XMEAN, XVAR, XSIG, RXSD,RSD,XSD
0004      DIMENSION ILAB(21),TYP(8)
          C
0005      1000 FORMAT (F10.5)
0006      1001 FGRMAT (10X,3 F10.5, 40X)
0007      1002 FORMAT (////' RESULTS OF SUBROUTINE ALPHA',/, 'OLABORATORY CODE NO.
          1: ' ,I2//' SAMPLE ALPHA-ACTIVITY VARIANCE SD R
          2SD(%) RSD(%) RATIO 89'/' MEAN VALUE
          3 OF SINGLE MEASUREMENT CF MEAN')
0008      1003 FORMAT (A6,D15.5,D17.5,2D12.4,2X,2D12.4)
0009      1004 FORMAT (' FOR SAMPLE ',A2,' NO ALPHA MEASUREMENTS; GIVEN VALUES AR
          1E DUMMIES')
0010      1005 FORMAT (' FOR SAMPLE ',A2,' ONLY ONE SINGLE ALPHA MEASUREMENT INST
          LEAD OF THREE RUNS; VARIANCE ETC. ARE SET DUMMY = 9.9')
          C
          C
0011      READ (5,1000) TEST
0012      IF (TEST.LT. 7.99) GO TO 10
0013      NALPHA = 0
0014      RETURN
0015      10 READ (5,1001) ((VALPHA(M3,K3,N3) ,N3=1,3),K3=3,8)
          C      PRINT 1001, ((VALPHA (M3,K3,N3),N3=1,3),K3=3,8)
0016      NALPHA = 1
          C      PRINTING OF THE TITLE OF THE TABLE
0017      PRINT 1002,ILAB(M3)
0018      DO 50 KI = 3,8
0019      XSUM = 0.
0020      NA=3
0021      DO 12 NI = 1,3
0022      IF (VALPHA(M3,KI,NI).GT.9.) GO TO 11

```

FORTRAN IV G1 RELEASE 2.0

ALPHA

DATE = 75015

13/37/58

```

0023      XSUM = XSUM + VALPHA(M3,KI,NI)
0024      GO TO 12
0025      11 NA=NI-1
0026      GO TO 13
0027      12 CONTINUE
0028      13 IF (NA.GE.1) GO TO 14
0029      PRINT 1004,TYP(KI)
0030      XMEAN =9.9
0031      GO TO 15
0032      14 XMEAN = XSUM/NA
0033      15 XVAR = 0.
0034      IF(NA.LT.1) GO TO 19
0035      IF (NA.EQ.1) GO TO 17
0036      DO 16 NI = 1,NA
0037      16 XVAR = XVAR + (XMEAN - VALPHA(M3,KI,NI))**2
0038      XSIG = XVAR/(NA-1)
0039      XSD = DSQRT (XSIG)
0040      RXSD = (XSD/XMEAN)*100.
C      RSD OF THE MEAN VALUE
0041      RSD=RXSD/SQRT(FLOAT(NA))
0042      GO TO 18
0043      17 PRINT 1005,TYP(KI)
0044      19 XSIG=9.9
0045      XSD=9.9
0046      RXSD=9.9
0047      RSD=9.9
0048      18 WALPHA (M3,KI ) = 87.7*XMEAN*((1./2.44D+04) + (MEAN3(M3 ,KI,9)/6.
158D+03))
0049      PRINT 1003,TYP(KI), XMEAN, XSIG, XSD, RXSD,RSD, WALPHA(M3,KI)
0050      50 CONTINUE
0051      RETURN
0052      END

```



```

FORTRAN IV G1  RELEASE 2.0          LAB3          DATE = 75015          13/37/58

0016          1011 FORMAT (///)
0017          1012 FORMAT(' ISOTOPIC COMPOSITION IN WEIGHT PERCENTAGES,CALCULATED IN
                1PROGRAM LAB3'/' INCLUDING ALPHA RESULTS' // ' PLUTONIUM GPAL 238
                2 GPAL 239   GPAL 240   GPAL 241   GPAL 242 '/' SAMPLE')
0018          1013 FCRMAT ('1')
                C
                C
0019          PRINT 1013
0020          PRINT 1000,ILAB(M3)
0021          PRINT 1001
0022          J3=1
                C
                C          URANIUM CALCULATIONS
                C          DO LOOP OVER THE SAMPLES
0023          DO 30 K3=3,8
0024          UR=1.
0025          DO 3 LL3=4,7
0026          3 UR=UR + MEAN3(M3 ,K3,LL3)
0027          DO 5 LL3=4,7
0028          5 ATU(K3,LL3)=(MEAN3(M3 ,K3,LL3)/UR)*100.
0029          ATU(K3,8) =(1./UR)*100.
                C
                C          CALCULATIONS FOR CONTROL
0030          SATU=0.
0031          DO 7 LL3=4,8
0032          7 SATU=SATU+ATU(K3,LL3)
0033          PRINT 1003 ,TYP(K3) ,(ATU(K3,LL3) ,LL3=4,8)
0034          30 CCNTINUE
                C
                C
                C          PLUTONIUM CALCULATIONS
0035          PRINT 1004
0036          DO 60 K3=3,8
0037          PR=1.
0038          DO 13 LL3=8,11
0039          13 PR=PR+MEAN3(M3 ,K3,LL3)
0040          DO 15 LL3=8,11
0041          15 ATP(K3,LL3) = (MEAN3(M3 ,K3,LL3)/PR)*100.
0042          ATP(K3,7) =(1./PR)*100.
                C          IN THIS VECTOR THE ORDER OF ISCTOPES IS : 7-239,8-238,9-240,10-241
                C
0043          SATP=0.
0044          DO 17 LL3=7,11
0045          17 SATP=SATP+ATP(K3,LL3)
0046          PRINT 1005,TYP(K3),ATP(K3,8),ATP(K3,7) ,(ATP(K3,LL3) ,LL3=9,11)
0047          60 CCNTINUE
                C
                C          CONNECTION WITH ALPHA MEASUREMENTS

```

FORTRAN IV G1 RELEASE 2.0

LAB3

DATE = 75015

13/37/58

```

C      POSSIBLE INPUT OF OTHER MEASUREMENTS
0048  CALL ALPHA (M3,WALPHA,NALPHA,MEAN3,ILAB,TYP)
0049  IF (NALPHA.EQ.0) GO TO 62
0050  PRINT 1011
0051  PRINT 1007
0052  DO 61 K3 = 3,8
0053  PRAL = WALPHA (M3,K3)+1.
0054  DO 18 LL3 = 9,11
0055  18 PRAL = PRAL + MEAN3(M3 ,K3,LL3)
C      CALCULATION OF ATOM PERCENTAGES WITH INCLUSION OF ALPHA-MEASUREMENTS
0056  ATAL(K3,7) = (1./PRAL)*100.
0057  ATAL(K3,8) = (WALPHA(M3,K3)/PRAL)*100.
0058  SATPAL = ATAL(K3,7) + ATAL(K3,8)
0059  DO 19 LL3 = 9,11
0060  ATAL(K3,LL3) = (MEAN3(M3 ,K3,LL3)/PRAL)*100.
0061  19 SATPAL = SATPAL + ATAL (K3,LL3)
0062  PRINT 1005, TYP(K3),ATAL(K3,8),ATAL(K3,7),(ATAL(K3,LL3),LL3=9,11)
0063  61 CCNTINUE
0064  GO TO 63
0065  62 PRINT 1009,ILAB(M3)
C
C
C      NEXT PART : WEIGHT PERCENTAGES
C      AGAIN URANIUM CALCULATIONS
0066  63 PRINT 1006
0067  DO 70 K3=3,8
0068  GATU(K3,4)=ATU(K3,4) * 233.0395
0069  GATU(K3,5)=ATU(K3,5) * 234.0409
0070  GATU(K3,6)=ATU(K3,6) * 235.0439
0071  GATU(K3,7)=ATU(K3,7) * 236.0457
0072  GATU(K3,8)=ATU(K3,8) * 238.0508
C      CALCULATIONS OF WEIGHT PERCENTAGES
C      AND CONTROL VALUE
0073  GUS=0.
0074  DO 21 LL3=4,8
0075  21 GUS=GUS+GATU(K3,LL3)
0076  SGU=0.
0077  DO 23 LL3=4,8
0078  GU(K3,LL3)=GATU(K3,LL3)/GUS *100.
0079  23 SGU=SGU+GU(K3,LL3)
0080  PRINT 1002,TYP(K3),(GU(K3,LL3),LL3=4,8)
0081  70 CONTINUE
C
C
C      CORRESPONDING PLUTONIUM CALCULATIONS
0082  PRINT 1008
0083  DO 80 K3=3,8
0084  GATP(K3,8)= ATP(K3,8) * 238.0495

```

FORTRAN IV G1 RELEASE 2.0

LAB3

DATE = 75015

13/37/58

```

0085      GATP(K3,7)= ATP(K3,7) * 239.0522
0086      GATP(K3,9)= ATP(K3,9) * 240.0540
0087      GATP(K3,10)=ATP(K3,10)* 241.0567
0088      GATP(K3,11)=ATP(K3,11)* 242.0587
0089      GPS=0.
0090      DO 31 LL3=7,11
0091 31 GPS=GPS+GATP(K3,LL3)
0092      SGP=0.
0093      DO 33 LL3=7,11
0094      GP(K3,LL3)=GATP(K3,LL3)/GPS *100.
0095 33 SGP=SGP+GP(K3,LL3)
0096      PRINT 1002 ,TYP(K3) ,GP(K3,8) ,GP(K3,7) , (GP(K3,LL3) ,LL3=9,11)
0097 80 CONTINUE

C
C
C      PARALLEL CALCULATIONS FOR ALPHA RESULTS
0098      IF(NALPHA.EQ.0) GO TO 91
0099      PRINT 1011
0100      PRINT 1012
0101      DO 90 K3=3,8
0102      GPALT(K3,8)=ATAL(K3,8)*238.0495
0103      GPALT(K3,7)=ATAL(K3,7)*239.0522
0104      GPALT(K3,9)=ATAL(K3,9)*240.0540
0105      GPALT(K3,10)=ATAL(K3,10)*241.0567
0106      GPALT(K3,11)=ATAL(K3,11)*242.0587
0107      GPALS=0.
0108      DO 82 LL3=7,11
0109 82 GPALS = GPALS + GPALT(K3,LL3)
0110      SGPAL = 0.
0111      DO 84 LL3 = 7,11
0112      GPAL(K3,LL3) = (GPALT(K3,LL3)/GPALS)*100.
0113 84 SGPAL = SGPAL + GPAL(K3,LL3)
0114      PRINT 1002, TYP (K3) ,GPAL(K3,8) ,GPAL(K3,7) , (GPAL(K3,LL3) ,LL3=9,11)
0115 90 CONTINUE
0116      GO TO 92
0117 91 PRINT 1009,ILAB(M3)
0118 92 CONTINUE
0119      RETURN
0120      END

```

FORTRAN IV G1 RELEASE 2.0

LAB4

DATE = 75015

13/37/58

```

0001      SUBROUTINE LAB4 (MEAN2,M4,MEANV,NALPHA,WALPHA,ILAB,TYP)
C
C
C
C
C      THIS SUBROUTINE CALCULATES THE U AND PU CONCENTRATIONS AND U/PU RA
C      EACH LAB.
C      THE CALCULATION OF Z IS DESCRIBED IN INTERNAL NOTES NO. 2 AND 3.
C
0002      CCMCN /DRYA/ KO(21,3,5,2),KOAL(21,3,5,2)
0003      REAL*8 MEAN2(21, 8,11),R(5,2)/10*0.0/,MEANV(21,8,3,11),C(3,5,2),
1CGN(5,2), COM(5,2)/10*0.0/, CSIG(5,2)/10*0.0/, CSD(5,2)/10*0.0/
0004      REAL*8 KO      ,KON(5,2),LAURA(5,2),Z(5,2),CRSD(5,2)/10*0.0/,KOM(
15,2)/10*0.0/,WALPHA(21,8),MEA,KOC(4,2,2),CC(4,2,2),CCAL(4,2,2),CKOA
2L(4,2,2)
0005      DIMENSION ILAB(21),TYP(8)
0006      REAL*8 LAUPA(5,2),D2,D3
0007      REAL*8 RAL(5,2)/10*0.0/,CAL(3,5,2),CONAL(5,2),
1      COMAL(5,2)/10*0.0/,CSIGAL(5,2)/10*0.0/,
2      CSDAL(5,2)/10*0.0/,CRSDAL(5,2)/10*0.0/,
3      KOAL      ,KOMAL(5,2)/10*0.0/,KONAL(5,2),
4      LAUPAL(5,2),PURU(21,5),PURUAL(21,5)
C
0008      1000 FORMAT('RESULTS OF PROGRAM LAB4/' LABORATORY CODE NO.:',I4/////
1' URANIUM CONCENTRATIONS'//21X,'U-238 CONCENTRATIONS IN ATOMS/G SO
2LUTION'//21X,'SAMPLE A',4X
3  ,'SAMPLE B',4X,'SAMPLE R',6X,'SAMPLE A-CALIBRATED   SAMPLE B-CAL
4IBRATED')
0009      1001 FORMAT (//21X,'URANIUM CONCENTRATIONS IN MG URANIUM/G SOLUTION'//)
0010      1012 FORMAT (//21X,'PLUTONIUM CONCENTRATIONS IN MICROGRAMS PLUTONIUM/G
1SOLUTION'//)
0011      1013 FORMAT (
1' C :      1ST RUN   ',3D12.5,D19.5,D23.5/' C :      2ND RUN   '
2,3D12.5,D19.5,D23.5/' C :      3RD RUN   ',3D12.5,D19.5,
3D23.5)
0012      1002 FORMAT (' CONCENTRATION 1-3   ',3D12.5,D19.5,D23.5,
1/' SIGMA',14X,3 D 12.5,' STANDARD DEVIATION ' ,3 D 12.5,/
2' RSD OF MEAN   ',3D12.5//)
0013      1003 FORMAT (' NUMBER OF ATOMS 1   ',3D12.5,D19.5,D23.5/' NUMBER OF ATOM
1S 2   ',3D12.5,D19.5,D23.5/' NUMBER OF ATOMS 3   ',3D12.5,D19.5,D23.
25)
0014      1004 FORMAT (' NO. OF ATOMS 1-3   ',3D12.5,D19.5,D23.5/' RSD OF MEAN',8
1X,3D12.5)
0015      1005 FORMAT (////////' PLUTONIUM CONCENTRATIONS'
1///21X,'PU-239 CONCENTRATIONS IN ATOMS/G SOLUTION'//
2      21X,'SAMPLE A',4X,'SAMPLE B',4X,'SAMPLE R',5X,'SAMPLE A-C

```



FORTRAN IV G1 RELEASE 2.0

LAB4

DATE = 75015

13/37/58

```

3ALIBRATED SAMPLE B-CALIBRATED')
0016 1006 FORMAT ('1'// ' PLUTONIUM CONCENTRATIONS WITH INCLUSION OF THE
RESULTS OF ALPHA SPECTROMETRY'///21X,'PU-239 CONCENTRATIONS IN ATO
2MS/G SOLUTION '// 21X,'SAMPLE A',4X,'SAMPLE B',4X,'S
SAMPLE R',5X,'SAMPLE A-CALIBRATED SAMPLE B-CALIBRATED')
0017 1007 FORMAT (' MEAN2(' ,2I3,' 4),MEAN2(' ,2I3,' 4) AND MEA ARE',3F10.6)
0018 1008 FORMAT (' FOR M4,KK4,L4 =' ,3I5,' MEANV AND MEAN2 ARE=' ,2F10.6)
0019 1009 FCRMAT(///// ' PU/U-RATIOS'/21X,' SAMPLE A',4X,' SAMPLE B',4X,' SA
MPLE R'//)
0020 1010 FORMAT (' MS-MEASUREMENTS',5X,3D12.5/' ALPHA-MEASUREMENTS',2X,3D12
1.5)
0021 1011 FORMAT (' NUMBER OF ATOMS 1 ',3D12.5,D19.5,D23.5/' NUMBER OF ATOM
1S 2 ',3D12.5;D19.5,D23.5/' NUMBER OF ATOMS 3 ',3D12.5,D19.5,D23.
25)
0022 1014 FORMAT (//' (IF ANY OF THESE RATIOS COULD NOT BE CALCULATED,IT IS
1 PUT = 0.))'
0023 1028 FORMAT (' FOR LAB. NO.',I3,' AND SAMPLE',A6,' THE RATIO PU238/239
1 WAS NOT MEASURED BY ALPHA-SPECTROMETRY.')
0024 1029 FORMAT (' FOR LAB NO.',I3,' AND SAMPLE ',A6,' THE RATIO PU 238/23
19 WAS NCT MEASURED BY MASS SPECTROMETRY.')
0025 1030 FORMAT ( ' MISSING MEASUREMENT RESULTS OR'
1. ' WRONG DENOMINATOR,M4,K4,L4 ARE ',3I4,' MEANV AND MEAN2 A
2RE ',2F10.6)
0026 1031 FORMAT (' COM= ',D12.5,' KOM= ',D12.5)
0027 1032 FCRMAT (' CSIG= ',D12.5)
C
C K4 INDICATES THE SAMPLE TYPE;ORDER CF SAMPLE TYPES:US,PS,AU,BU,RU,
C THE INDICES OF MEAN2 STAND FOR:
C FIRST(M4): LAB,
C SECOND(K4): SAMPLE TYPE,
C THIRD(LL): ISOTOPIC RATIO
C THE FIRST INDEX OF R STANDS FOR THE SAMPLE: A, B CR R.
C THE SECOND INDEX STANDS FOR U(1) OR PU(2)
C ISOTOPIC RATIOS ARE GIVEN IN THE ORDER 83,92,02,38,48,58,68,89,09
0028 Z(3,1) = 3.1398E-03
0029 Z(4,1) = 2.9598E-03
0030 Z(5,1) = 3.5275E-03
C DO OVER THE SAMPLES
0031 DO 10 K4 = 3,5
0032 R(K4,1) = 233.0395 * MEAN2(M4 ,K4,4)
1 + 234.0409 * MEAN2(M4 ,K4,5)
2 + 235.0439 * MEAN2(M4 ,K4,6)
3 + 236.0457 * MEAN2(M4 ,K4,7)
4 + 238.0508
0033 LAURA(K4,1)=6.022045E+20/R(K4,1)
C DO OVER THE RUNS
0034 COM(K4,1)=0.
0035 KOM(K4,1)=0.

```

```

FORTRAN IV G1  RELEASE 2.0          LAB4          DATE = 75015          13/37/58

0036          CSIG(K4,1)=0.
0037          DO 9 L4 = 1,3
0038          D3=3.
0039          IF(( MEANV(M4,K4+3,L4,4)- MEAN2(M4,3,4)).LE.0.) GO TO 12
0040          C(L4,K4,1)=Z(K4,1)*R(K4,1)*((1-MEANV(M4,K4+3,L4,4)*MEAN2(M4 ,1,1)
1) / (MEANV(M4,K4+3,L4,4)-MEAN2(M4 ,3,4)))
0041          GO TO 13
0042          12 KK4=K4+3
0043          PRINT 1008,M4, KK4,L4,MEANV(M4, KK4,L4,4),MEAN2(M4,3,4)
0044          C( L4,K4,1)=0.
0045          D3=D3-1.
C          THE VALUES KO DEFINE THE CONCENTRATION OF THE REFERENCE ISOTOPE ,E
C          IN "NUMBER OF ATOMS PER GRAM OF SOLUTION"
C          THEIR MEAN VALUE IS KOM.
0046          13 KO(M4,L4,K4,1) = C(L4,K4,1)*LAURA(K4,1)
0047          IF(D3.LE.0.) GO TO 9
0048          CCM(K4,1)=COM(K4,1)+(1./D3)*C(L4,K4,1)
0049          KOM(K4,1)=KOM(K4,1)+(1./D3)*KO(M4,L4, K4,1)
0050          9 CONTINUE
0051          DO 8 L4=1,3
0052          D2=1./6.
0053          IF(( MEANV(M4,K4+3,L4,4)- MEAN2(M4,3,4)).LE.0.) D2=C.
0054          CSIG(K4,1)=CSIG(K4,1)+D2 *(C(L4,K4,1)-CGM(K4,1))*2
0055          8 CONTINUE
0056          CSD(K4,1)=DSQRT(CSIG(K4,1))
0057          IF(COM(K4,1).LE.0.) GO TO 81
0058          CRSD(K4,1)=(CSD(K4,1)/CCM(K4,1))*100.
0059          GO TO 82
0060          81 CRSD(K4,1)=0.
C          THE CONCENTRATIONS C HAVE THREE INDICES:
C          THE FIRST(L4) STANDS FOR THE RUN,
C          THE SECOND (K4) FOR THE SAMPLE,
C          THE THIRD FOR U(1) OR PU(2).
C          IN THE FORMULAE FOR C THE RESULTS MEANV FROM LAB1 AND MAIN ARE USE
C          MEANV HAS 4 INDICES:
C          THE FIRST (M) STANDS FOR THE LAB
C          THE SECOND (K) STANDS FOR THE SAMPLE TYPE
C          THE THIRD (L) STANDS FOR THE RUN
C          THE FOURTH (LL) STANDS FOR THE ISOTOPIC RATIO
0061          82 MEA=MEAN2(M4,K4+3,4)-MEAN2(M4,K4,4)
0062          IF(MEA.LE.0.) GO TO 11
0063          CCN(K4,1) = Z(K4,1)*R(K4,1)*((1-MEAN2(M4 ,K4+3,4)*MEAN2(M4, 1,1)
1) / (MEAN2(M4 ,K4+3,4)-MEAN2(M4 ,K4,4)))
0064          KCN(K4,1) = CON(K4,1)*LAURA(K4,1)
C          IN THE VECTORS C AND CCN THE INDICES K4=3,4,5 STAND FOR THE ORIGIN
C          SAMPLES A,B,R.
0065          GO TO 10
0066          11 KK4=K4+3

```

```

FORTRAN IV G1  RELEASE 2.0                LAB4                DATE = 75015                13/37/58

0067          PRINT 1007,M4, KK4, M4, K4, MEAN2(M4, KK4, 4), MEAN2(M4, K4, 4), MEA
0068          CON (K4,1) =0.
0069          KON(K4,1)=0.
0070          10 CONTINUE
0071          DO 15 II=1,2
0072          DO 14 I=1,3
C             THE INDEX I STANDS FOR THE RUN, THE INDEX II FOR THE SAMPLE.
C             THE THIRD INDEX STANDS FOR U (1) OR PU (2).
0073          KOC(I,II,1)=(KO(M4,I,II+2,1) * 2.786E+18)/KO(M4,I,5,1)
0074          CC(I,II,1)=(KOC(I,II,1) * C(I,II+2,1))/KO(M4,I,II+2,1)
0075          14 CCNTINUE
0076          KOC(4,II,1)=(KON(II+2,1) * 2.786E+18)/KON(5,1)
0077          CC(4,II,1)=(KOC(4,II,1) * CON(II+2,1))/KO(M4,I,II+2,1)
0078          15 CONTINUE
0079          PRINT 1000, ILAB(M4)
0080          PRINT 1003 , ((KO(M4,L4,K4,1),K4=3,5),(KOC(L4,II,1),II=1,2),L4=1,3)
0081          PRINT 1004 , (KON(K4,1),K4=3,5),(KOC(4,II,1),II=1,2),(CRSD(K4,1),K4=
1=3,5)
0082          PRINT 1001
0083          PRINT 1013,((C(L4,K4,1),K4=3,5),(CC(L4,II,1),II=1,2),L4=1,3)
0084          PRINT 1002,(CON(K4,1),K4=3,5),(CC(4,II,1),II=1,2),(CSIG(K4,1),K4=
13,5),(CSD(K4,1),K4=3,5),(CRSD(K4,1),K4=3,5)

C
C
C
C
C          CALCULATION OF PLUTONIUM CONCENTRATIONS
C          PART A : RESULTS OF MASS SPECTROMETRY
C          PART B : RESULTS OF ALPHA SPECTROMETRY
C
C
C
C          PART A :
C          LAB 23 HAS NOT PERFORMED PU MEASUREMENTS
0085          IF(ILAB(M4).EQ.23) GO TO 100
C          LAB 22 HAS NOT MEASURED THE PU SPIKE, THE BCMN VALUE IS TAKEN 9.1
0086          IF(ILAB(M4).EQ.22) MEAN2(M4,2,2)=2.E-04
C          THE MEASUREMENT RESULTS ON THE PURE PU -SPIKE HAVE TO BE REPLACED
C          BY THE BCMN RESULTS FOR THE MIXED SPIKE 19.2.74
0087          MEAN2(M4,2,2) = 1.34E-02
C          ALSO THE Z-VALUES HAVE BEEN CORRECTED 19.2.74
0088          Z(3,2) = 3.6893E-02
0089          Z(4,2) = 3.4778E-02
0090          Z(5,2) = 4.1449E-02
0091          DO 40 K4 = 3,5
0092          KK4=K4+3
0093          IF (MEAN2(M4 ,K4,8).LE.0.000000) GO TO 28
0094          GO TO 29

```

```

FORTRAN IV G1  RELEASE 2.0          LAB4          DATE = 75015          13/31/58

0095      28 PRINT 1029 ,ILAB(M4),TYP(K4)
0096      29 R(K4,2) = 238.0495*MEAN2(M4 ,K4,8)+239.0522
          1+240.0540*MEAN2(M4 ,K4,9)+241.0567*MEAN2(M4 ,K4,10)+242.0587*MEA
          1N2(M4 ,K4,11)
0097      LAUPA(K4,2) = 6.022045E+17/R(K4,2)
0098      COM(K4,2)=0.
0099      KCM(K4,2)=0.
0100      CSIG(K4,2)=0.
0101      D3=3.
0102      DO 30 L4 = 1,3

          C
          C SPECIAL CASES OF MISSING MEASUREMENTS OF RATIO 242/239 AND MISSING
          C THE MISSING RATIO 242/239 FOR SAMPLE RU AS SUCH IS NEGLIGIBLE -
          C THE RATIO IS VERY LOW FOR THIS SAMPLE
          C IF(MEANV(M4, KK4, L4, 11).LE.0.) GO TO 31
0103      IF(MEANV(M4, KK4, 3, 11).LE.0.) D3=2.
0104      IF((MEANV(M4, KK4, L4, 11)-MEAN2(M4, K4, 11)).LE.0.) GO TO 31
0105      C(L4, K4, 2) = Z(K4, 2)*R(K4, 2)*((1.-MEANV(M4, K4+3, L4, 11))*MEAN2(M4 , 2,
          1, 2))/(MEANV(M4, K4+3, L4, 11)- MEAN2(M4 , K4, 11))
          GO TO 32
0106      31 PRINT 1030, M4, K4, L4, MEANV(M4, KK4, L4, 11) , MEAN2(M4, K4, 11)
0107      C(L4, K4, 2)=0.
0108      32 KO(M4, L4, K4, 2) = C(L4, K4, 2)*LAUPA(K4, 2)
0109      CCM (K4, 2) = COM(K4, 2)+(1./D3)*C(L4, K4, 2)
0110      KOM(K4, 2) = KOM(K4, 2)+(1./D3)*KO(M4, L4, K4, 2)
0111      C IF(ILAB(M4).NE.21) GO TO 30
          C PRINT 1031, CCM(K4, 2), KOM(K4, 2)

0112      30 CONTINUE
0113      DO 38 L4=1,3
0114      D2=1./6.

          C SPECIAL CASE OF MISSING RUN 3 FOR R-SAMPLE
0115      IF(MEANV(M4, KK4, 3, 11).LE.0.) D2=0.5
0116      IF(MEANV(M4, KK4, L4, 11).LE.0.) GO TO 38
0117      CSIG(K4, 2) = CSIG(K4, 2)+D2 *(C(L4, K4, 2)-COM(K4, 2))*2
          C IF(ILAB(M4).NE.21) GO TO 38
          C PRINT 1032, CSIG(K4, 2)

0118      38 CONTINUE
0119      CSD(K4, 2) = DSQRT(CSIG(K4, 2))
0120      IF (CCM(K4, 2).LE.0.) GO TO 33
0121      CRSD(K4, 2) = (CSD(K4, 2)/COM(K4, 2))*100.
0122      GO TO 34
0123      33 CRSD(K4, 2)=0.
0124      34 CON(K4, 2) = Z(K4, 2)*R(K4, 2)*((1.-MEAN2(M4 , K4+3, 11))*MEAN2(M4 , 2,
          12))/(MEAN2(M4 , K4+3, 11)-MEAN2(M4 , K4, 11))
          KON(K4, 2) = CON(K4, 2)*LAUPA(K4, 2)

0125      40 CONTINUE
0126      PRINT 1005
0127      DO 45 II=1,2
0128

```

```

FORTRAN IV G1  RELEASE 2.0                LAB4                DATE = 75015                13/37/58

0129          DO 44 I=1,3
              C      SPECIAL CASE OF MISSING RUN 3 FOR R-SAMPLE
0130          IF(KO(M4,I,5,2).LE.0.) GO TO 43
0131          KOC(I,II,2)=(KO(M4,I,II+2,2) * 2.11E+16) /KO(M4,I,5,2)
0132          CC(I,II,2)=(C(I,II+2,2) * KOC(I,II,2))/KO(M4,I,II+2,2)
0133          GO TO 44
0134          43 KOC(I,II,2)=0.
0135          CC(I,II,2)=0.
0136          44 CONTINUE
0137          KOC(4,II,2)=(KON(II+2,2) * 2.11E+16)/KON(5,2)
0138          CC(4,II,2)=(KOC(4,II,2) * CON(II+2,2))/KO(M4,I,II+2,2)
0139          45 CONTINUE
0140          PRINT 1011,((KO(M4,L4,K4,2),K4=3,5),(KOC(L4,II,2),II=1,2),L4=1,3)
0141          PRINT 1004,(KON(K4,2),K4=3,5),(KOC(4,II,2),II=1,2),(CRSD(K4,2),K4
              1=3,5)
0142          PRINT 1012
0143          PRINT 1013,((C(L4,K4,2),K4=3,5),(CC(L4,II,2),II=1,2),L4=1,3)
0144          PRINT 1002,(CON(K4,2),K4=3,5),(CC(4,II,2),II=1,2),(CSIG(K4,2),K4=
              13,5),(CSD(K4,2),K4=3,5),(CRSD(K4,2),K4=3,5)
0145          60 CONTINUE
              C
              C      PART B:
              C
0146          IF ( NALPHA.EQ.0) GO TO 91
              C
              C
0147          DO 70 K4=3,5
0148          RAL(K4,2)=238.0495*WALPHA(M4,K4)
              1      +239.0522
              2      +240.0540*MEAN2(M4 ,K4,9)
              3      +241.0567*MEAN2(M4 ,K4,10)
              4      +242.0587*MEAN2(M4 ,K4,11)
0149          LAUPAL(K4,2)=6.022045E+17/RAL(K4,2)
0150          COMAL(K4,2)=0.
0151          KOMAL(K4,2)=0.
0152          CSIGAL(K4,2)=0.
0153          D3=3.
0154          KK4=K4+3
0155          DO 65 L4=1,3
0156          IF(MEANV(M4,KK4,3,11).LE.0.) D3=2.
0157          IF ((MEANV(M4,KK4,L4,11)- MEAN2(M4,K4,11)).LE.0.) GO TO 61
0158          CAL(L4,K4,2) = Z(K4,2)*RAL(K4,2)*((1.-MEANV(M4,K4+3,L4,11)*MEAN2(M
              14,2,2))/(MEANV(M4,K4+3,L4,11)-MEAN2(M4 ,K4,11)))
              GO TO 62
0159          61 PRINT 1030,M4,K4,L4,MEANV(M4,KK4,L4,11), MEAN2(M4,K4,11)
0160          CAL(L4,K4,2) =0.
0161          62 KOAL(M4,L4,K4,2)=CAL(L4,K4,2)*LAUPAL(K4,2)
0162          COMAL(K4,2)=COMAL(K4,2)+(1./D3)*CAL(L4,K4,2)
0163

```

FORTRAN IV G1 RELEASE 2.0

LAB4

DATE = 75015

13/37/58

```

0164      KCMAL(K4,2)=KOMAL(K4,2)+(1./D3)*KOAL(M4,L4,K4,2)
0165      65 CONTINUE
0166      DO 68 L4=1,3
0167      D2=1./6.
0168      IF(MEANV(M4,KK4,3,11).LE.0.) D2=0.5
0169      CSIGAL(K4,2)=CSIGAL(K4,2)+D2*(CAL(L4,K4,2)-COMAL(K4,2))**2
0170      68 CONTINUE
0171      CSDAL(K4,2)=DSQRT(CSIGAL(K4,2))
0172      IF(CCMAL(K4,2).LE.0.) GO TO 66
0173      CRSDAL(K4,2)=(CSDAL(K4,2)/COMAL(K4,2))*100.
0174      GO TO 67
0175      66 CRSDAL(K4,2)=0.
0176      67 CONAL(K4,2)=Z(K4,2)*RAL(K4,2)*((1.-MEAN2(M4,K4+3,11))*MEAN2(M4,2,2
1)))/(MEAN2(M4,K4+3,11)-MEAN2(M4,K4,11))
      KONAL(K4,2)=CONAL(K4,2)*LAUPAL(K4,2)
0177      70 CONTINUE
0178      GO TO 71
0179      91 PRINT 1028,ILAB(M4),TYP(3)
C
0181      71 IF(NALPHA.EQ.0) GO TO 90
0182      PRINT 1006
0183      DO 75 II=1,2
0184      DO 74 I=1,3
0185      IF(KOAL(M4,I,5,2).LE.0.) GO TO 73
0186      CKOAL(I,II,2)=(KOAL(M4,I,II+2,2) * 2.11E+16)/KOAL(M4,I,5,2)
0187      CCAL(I,II,2)=(CKOAL(I,II,2) * CAL(I,II+2,2))/KOAL(M4,I,II+2,2)
0188      GO TO 74
0189      73 CKOAL(I,II,2)=0.
0190      CCAL(I,II,2)=0.
0191      74 CONTINUE
0192      CKOAL(4,II,2)=(KCNAL(II+2,2) * 2.11E+16)/KONAL(5,2)
0193      CCAL(4,II,2)=(CKOAL(4,II,2) * CONAL(II+2,2))/KOAL(M4,I,II+2,2)
0194      75 CONTINUE
0195      PRINT 1011,((KOAL(M4,L4,K4,2),K4=3,5),(CKOAL(L4,II,2),II=1,2),L4=1
1,3)
0196      PRINT 1004,((KCNAL(K4,2),K4=3,5),(CKOAL(4,II,2),II=1,2),(CRSDAL(K4,
12),K4=3,5)
0197      PRINT 1012
0198      PRINT 1013,((CAL(L4,K4,2),K4=3,5),(CCAL(L4,II,2),II=1,2),L4=1,3)
0199      PRINT 1002,((CONAL(K4,2),K4=3,5),(CCAL(4,II,2),II=1,2),(CSIGAL(K4,
12),K4=3,5),(CSDAL(K4,2),K4=3,5),(CRSDAL(K4,2),K4=3,5)
0200      90 CONTINUE
C
C
C      CALCULATION OF THE PU/U RATIO
0201      92 DO 94 K4=3,5
0202      IF(CON(K4,1).LE.0.) GO TO 95
0203      PURU(M4,K4)=CON(K4,2)/CON(K4,1)

```

FORTRAN IV G1 RELEASE 2.0

LAB4

DATE = 75015

13/37/58

```
0204          GO TO 96
0205          95 PURU(M4,K4) =0.
0206          96 IF(CCN(K4,1).LE.0.) GO TO 93
0207             IF(NALPHA.EQ.0) GO TO 93
0208             PURUAL(M4,K4)=CONAL(K4,2)/CON(K4,1)
0209             GO TO 94
0210          93 PURUAL(M4,K4) =0.
0211          94 CCNTINUE
0212             PRINT 1009
0213             PRINT 1010,(PURU(M4,K4),K4=3,5),(PURUAL(M4,K4),K4=3,5)
0214             PRINT 1014
0215          100 CONTINUE
0216             RETURN
0217             END
```

FORTRAN IV G1 RELEASE 2.0

LAB5

DATE = 75015

13/37/58

0001

SUBROUTINE LAB5 (NLAB,MVALV,IFILIV,MEANV,SUMV,IR,TYP,ILAB)

C  
C  
C  
C  
C

0002

DIMENSION IR(11),TYP(8),ILAB(21)

0003

INTEGER \* 2 MM(8,11)/88\*0/,SS(8,11)/88\*0/,MVALV(21,8,3,11),IFILIV(

121,8,11),NP1(21)/21\*0/

0004

REAL \* 8 D,F,Y2P,Y3P,MEANV (21,8,3,11), MMEAN, E,

IG,H,IH,SUMV(21,8,3,11),YIP(3),EE

0005

REAL\*8 K11,K12,K22,SQA,SQB,SQR,MCA,MCB,MQR,SDE,RSDE,SIGB,SDB,RSDB,  
1 SIGA,SDA,RSCA,R1,R2,R3,R4,R5,R6,R7

C

0006

1000 FORMAT ('RESULTS OF SUBROUTINE LAB5'/' ANALYSIS OF VARIANCES IN  
1 THREE STEPS: INTERLAB,RUN AND SCAN COMPONENTS' / I5,' LABS ARE TAKE  
2 N INTO ACCOUNT AT MAXIMUM '///' (IF THE RUN OR SCAN COMPONENT OF AN  
3 Y VARIANCE IS NEGATIVE AND THEREFORE CONSIDERED AS NOT SIGNIFICANT  
4,'/' SD AND RSD ARE PUT = 0. CORRESPONDINGLY.)')

0007

1001 FORMAT (' SAMP RAT LABS RUNS SCANS MMEAN SIG-SCAN SD-SCAN R  
1 SD-SCAN SIG-RUN SD-RUN RSD-RUN SIG-LABS SD-LABS RSD-LAB  
3S ' /)

0008

1002 FORMAT (2X,A2,3X,I2,1X,I3,I4,I7,1X,1P10D10.3)

0009

1003 FORMAT (' MQR IS NEGATIVE, MQR,SQR,NN AND SS(K,LL) ARE:',2D15.6,  
1 2I5)

0010

1004 FCRMAT (' SIGB IS NEGATIVE, SIGB, MCB, SQR AND K22 ARE:',4D15.6)

0011

1005 FORMAT (' SIGA IS NEGATIVE, SIGA,K11,MQA, K12, K22, MCB, AND MQR  
1 ARE:',7D15.4)

0012

1006 FORMAT (///)

0013

1007 FORMAT (' NP1,Y2P,D,E,G,H,NN ARE:',I4,5D15.4,I5)

0014

1008 FCRMAT (' YIP(' , I2, ')=' ,F10.2, ' NVALVV=N(I,J)\*\*2= ',I3)

0015

1009 FORMAT (' F,Y3P,IH =' ,3D15.4, ' MM(K,LL) AND SS(K,LL) ARE',2I4)

0016

1010 FORMAT (' COMBINATIONS OF MM=' ,I3, ' ,SS=' ,I3, ' AND NN=' ,I3, ' ARE  
1 ZERO')

0017

1011 FORMAT (///' NN=' ,I5)

0018

1012 FORMAT (' IH=' ,C10.4)

0019

1013 FORMAT (' MMEAN=' ,D10.4)

0020

1016 FORMAT (' MEAN2=' ,D10.4)

0021

1019 FCRMAT (' M,K,L,LL AND SUMV(M,K,L,LL) ARE:',4I4,D15.4)

0022

1020 FORMAT (' IN DO 60 M,K,L,LL,NVAL AND MEANV ARE:',5I5,D15.4)

0023

1021 FCRMAT (' K,LL AND SS ARE:',3I5)

0024

1023 FORMAT (' NP1(M=' ,I2, ') IS',I5, ' Y2P IS ',D15.4)

0025

1024 FCRMAT (' FOR LAB',I3, ' SAMPLE ',A3, ' AND RATIO',I3, ' IFILIV= ',I3)

0026

1025 FORMAT (' FOR LAB',I3, ' SAMPLE ',A3, ' RATIO',I3, ' AND RUN',I2, ' THE  
1 NO. OF SCANS IS',I2)C  
C



```

FORTRAN IV G1  RELEASE 2.0          LAB5          DATE = 75015          13/37/58

      C      MM(K,LL) IS THE NUMBER OF LABS, FOR WHICH STANDARD EXPERIMENT RESUL
      C      AVAILABLE.
      C      SS(K,LL) IS THE NUMBER OF RUNS IN ALL LABS, IN WHICH THE ISOTOPIC R
      C      LL OF THE SAMPLE K HAS BEEN MEASURED.
      C      NN IS THE OVERALL NUMBER OF MEASUREMENTS
      C
0027      PRINT 1000,NLAB
0028      PRINT 1006
0029      PRINT 1001
      C      DO OVER THE SAMPLES  US,PS,AU,BU,RU,AS,BS,RS
0030      DO 90 K=1,8
      C      DO OVER THE ISOTOPIC RATIOS
0031      DO 80 LL=1,11
0032      MM(K,LL) =0
0033      SS(K,LL) =0
0034      LLMAX=11
0035      LLMIN=1
      C      THE SEQUENCE OF ISOTOPIC RATIOS IS 83,92,02, 38,48,58,68, 89,09,19,29
      C      FOR THE US SAMPLE ONLY ONE ISOTOPIC RATIO AND FOR THE PS SAMPLE ON
      C      ISOTOPIC RATIOS ARE MEASURED
0036      IF(K.EQ.1) GO TO 10
0037      IF(K.EQ.2) GO TO 11
0038      GO TO 12
0039      10 LLMAX=1
0040      GO TO 15
0041      11 LLMIN=2
0042      LLMAX=3
0043      GO TO 15
0044      12 LLMIN=4
0045      15 CONTINUE
0046      IF (LL.LT.LLMIN) GO TO 80
0047      IF (LL.GT.LLMAX) GO TO 80
0048      D=0.
0049      E=0.
0050      F=0.
0051      G=0.
0052      H=0.
0053      IH=0.
0054      MMEAN =5.5
0055      K11=5.5
0056      K12=5.5
0057      K22=5.5
0058      SQA=5.5
0059      MQA=5.5
0060      MQB=5.5
0061      MQR=-5.5
      C      DO OVER THE LABS

```

FORTRAN IV G1 RELEASE 2.0

LAB5

DATE = 75015

13/37/58

```

0062          NN=0
0063          Y3P=0.
0064          DO 70 M=1,NLAB
C             SOME SPECIAL LABS SHALL BE OMITTED
0065             IF(K.EQ.1) GO TO 50
0066             IF(K.EQ.2) GO TO 51
0067             IF (K.EQ.3) GO TO 52
0068             IF (K.EQ.4) GO TO 53
0069             IF(K.EQ.5) GO TO 54
C070             IF(K.EQ.6) GO TO 55
0071             IF(K.EQ.7) GO TO 56
0072             IF(K.EQ.8) GO TO 57
0073             GO TO 58
C             UNSPIKED SAMPLES
0074             50 IF(LL.NE.1) GO TO 58
0075                 IF(ILAB(M).EQ.5) GO TO 70
0076                 IF(ILAB(M).EQ.8) GO TO 70
0077                 IF(ILAB(M).EQ.20) GO TO 70
0078                 IF(ILAB(M).EQ.21) GO TO 70
C079                 GO TO 58
0080             51 IF(LL.NE.2) GO TO 41
0081                 IF(ILAB(M).EQ.16) GO TO 70
0082                 IF(ILAB(M).EQ.21) GO TO 70
0083                 GO TO 58
C084             41 IF(LL.NE.3) GO TO 58
0085                 IF(ILAB(M).EQ.8) GO TO 70
0086                 IF(ILAB(M).EQ.16) GO TO 70
0087                 IF(ILAB(M).EQ.21) GO TO 70
0088                 GO TO 58
C089             52 IF (LL.NE.7) GO TO 522
0090                 IF(ILAB(M).EQ.4) GO TO 70
0091                 GO TO 58
0092             522 IF(LL.NE.8) GO TO 42
0093                 IF(ILAB(M).EQ.19) GO TO 70
C094                 GO TO 58
0095             42 IF(LL.NE.9) GO TO 32
0096                 IF(ILAB(M).EQ.4) GO TO 70
0097                 IF(ILAB(M).EQ.5) GO TO 70
0098                 GO TO 58
C099             32 IF(LL.NE.10) GO TO 22
0100                 IF(ILAB(M).EQ.2) GO TO 70
0101                 GO TO 58
0102             22 IF(LL.NE.11) GO TO 58
0103                 IF(ILAB(M).EQ.8) GO TO 70
0104                 GO TO 58
0105             53 IF (LL.NE.7) GO TO 433
0106                 IF(ILAB(M).EQ.18) GO TO 70
0107                 GO TO 58

```

FORTRAN IV G1 RELEASE 2.0

LAB5

DATE = 75015

13/37/58

```

0108      433 IF(LL.NE.8) GO TO 43
0109          IF(ILAB(M).EQ.19) GO TO 70
0110          GO TO 58
0111      43 IF(LL.NE.10) GO TO 33
0112          IF(ILAB(M).EQ.2) GO TO 70
0113          GO TO 58
0114      33 IF(LL.NE.11) GO TO 58
C          ONE VERSION WITH AND ONE WITHOUT INCLUSION OF LAB NO. 20
0115          IF(ILAB(M).EQ.20) GO TO 70
0116          GO TO 58
0117      54 IF(LL.NE.5) GO TO 44
0118          IF(ILAB(M).EQ.5) GO TO 70
0119          GO TO 58
0120      44 IF(LL.NE.6) GO TO 34
0121          IF(ILAB(M).EQ.21) GO TO 70
0122          GO TO 58
0123      34 IF(LL.NE.9) GO TO 24
0124          IF(ILAB(M).EQ.4) GO TO 70
0125          IF(ILAB(M).EQ.5) GO TO 70
0126          IF(ILAB(M).EQ.8) GO TO 70
0127          IF(ILAB(M).EQ.21) GO TO 70
0128          GO TO 58
0129      24 IF(LL.NE.10) GO TO 14
0130          IF(ILAB(M).EQ.2) GO TO 70
0131          IF(ILAB(M).EQ.8) GO TO 70
0132          IF(ILAB(M).EQ.21) GO TO 70
0133          GO TO 58
0134      14 IF(LL.NE.11) GO TO 58
0135          IF(ILAB(M).EQ.6) GO TO 70
0136          IF(ILAB(M).EQ.8) GO TO 70
0137          GO TO 58
C
C
C          SPIKED SAMPLES
0138      55 IF(LL.NE.7) GO TO 45
0139          IF(ILAB(M).EQ.20) GO TO 70
0140          GO TO 58
0141      45 IF(LL.NE.8) GO TO 35
0142          IF(ILAB(M).EQ.19) GO TO 70
0143          GO TO 58
0144      35 IF(LL.NE.10) GO TO 25
0145          IF(ILAB(M).EQ.2) GO TO 70
0146          GO TO 58
0147      25 IF(LL.NE.11) GO TO 58
0148          IF(ILAB(M).EQ.4) GO TO 70
0149          IF(ILAB(M).EQ.8) GO TO 70
0150          GO TO 58
C

```

FORTRAN IV G1 RELEASE 2.0

LAB5

DATE = 75015

13/37/58

```

0151          56 IF(LL.NE.7) GO TO 46
0152          IF(ILAB(M).EQ.4) GO TO 70
0153          IF(ILAB(M).EQ.18) GO TO 70
0154          IF(ILAB(M).EQ.20) GO TO 70
0155          GO TO 58
0156          46 IF(LL.NE.8) GO TO 36
0157          IF(ILAB(M).EQ.19) GO TO 70
0158          GO TO 58
0159          36 IF(LL.NE.10) GO TO 26
0160          IF(ILAB(M).EQ.2) GO TO 70
0161          GO TO 58
0162          26 IF(LL.NE.11) GO TO 58
0163          IF(ILAB(M).EQ.4) GO TO 70
0164          IF(ILAB(M).EQ.6) GO TO 70
0165          IF(ILAB(M).EQ.8) GO TO 70
0166          IF(ILAB(M).EQ.21) GO TO 70
0167          GO TO 58
C
0168          57 IF(LL.NE.5) GO TO 47
0169          IF(ILAB(M).EQ.7) GO TO 70
0170          GO TO 58
0171          47 IF(LL.NE.6) GO TO 37
0172          IF(ILAB(M).EQ.21) GO TO 70
0173          GO TO 58
0174          37 IF(LL.NE.9) GO TO 27
0175          IF(ILAB(M).EQ.8) GO TO 70
0176          IF(ILAB(M).EQ.21) GO TO 70
0177          GO TO 58
0178          27 IF(LL.NE.10) GO TO 17
0179          IF(ILAB(M).EQ.2) GO TO 70
0180          IF(ILAB(M).EQ.8) GO TO 70
0181          IF(ILAB(M).EQ.20) GO TO 70
0182          IF(ILAB(M).EQ.21) GO TO 70
0183          GO TO 58
0184          17 IF(LL.NE.11) GO TO 58
0185          IF(ILAB(M).EQ.4) GO TO 70
0186          IF(ILAB(M).EQ.6) GO TO 70
0187          IF(ILAB(M).EQ.8) GO TO 70
0188          IF(ILAB(M).EQ.21) GO TO 70
C
C
C          END OF THE SELECTION FOR OMISSION OF SINGLE LABS
0189          58 EE=0.
0190          NP1(M) =0
0191          IFILV=IFILIV(M,K,LL)
C          PRINT 1024,ILAB(M),TYP(K),IR(LL),IFILV
C
C          CHECK OF THE NUMBER OF RUNS

```

```

FORTRAN IV G1  RELEASE 2.0                LAB5                DATE = 75015                13/37/58

0192                IF(IFILV.LE.0) GO TO 701
                   C   DO OVER THE RUNS
0193                   Y2P=0.
0194                   DO 59 L=1,3
0195                   59 YIP(L)=0.
0196                   DO 60 L=1,IFILV
0197                   NVALV=MVALV(M,K,L,LL)
0198                   NVALVV=NVALV**2
                   C   PRINT 1020,M,K,L,LL,NVALV,MEANV(M,K,L,LL)
                   C
                   C   CHECK OF THE NUMBER OF SCANS
0199                   IF(NVALV.LE.1) GO TO 601
0200                   IF (L.GT.1) GO TO 19
0201                   MM(K,LL)=MM(K,LL) +1
0202                   19 SS(K,LL)= SS(K,LL) +1
                   C   PRINT 1021,K,LL,SS(K,LL)
0203                   YIP(L)=YIP(L) + NVALV*MEANV(M,K,L,LL)
                   C   PRINT 1008 ,L,YIP(L),NVALVV
0204                   NP1(M)=NP1(M) + NVALV(M,K,L,LL)
0205                   Y2P=Y2P + YIP(L)
                   C   PRINT 1023,M,NP1(M),Y2P
0206                   D=D+NVALVV
0207                   EE= EE+ NVALVV
0208                   62 G=G+SUMV(M,K,L,LL)
0209                   H=H +(YIP(L)**2)/NVALV
0210                   GO TO 60
                   C 601 PRINT 1025,ILAB(M),TYP(K),IR(LL),L,NVALV
0211                   601 CONTINUE
0212                   60 CCNTINUE
0213                   NN=NN + NP1(M)
0214                   IF (NP1(M).LE.0) GO TO 69
0215                   E=E + EE/NP1(M)
0216                   F=F+NP1(M) **2
0217                   IH=IH + (Y2P**2)/NP1(M)
0218                   69 Y3P=Y3P + Y2P
                   C   PRINT 1007,NP1(M), Y2P,D,E,G,H ,NN
                   C   PRINT 1009,F,Y3P,IH,MM(K,LL),SS(K,LL)
0219                   GO TO 70
                   C 701 PRINT 1024,ILAB(M),TYP(K),IR(LL),IFILV
0220                   701 CONTINUE
0221                   70 CONTINUE
0222                   SQB=H-IH
0223                   SQR=G-H
0224                   IF(NN.LE.0) GO TO 172
0225                   IF( MM(K,LL).EQ.1) GO TO 172
0226                   IF((SS(K,LL)-MM(K,LL)).EQ.0) GO TO 172
0227                   IF ((NN-SS(K,LL)).EQ.0) GO TO 172
0228                   MMEAN=Y3P/NN

```

```

FORTRAN IV G1  RELEASE 2.0          LAB 5          DATE = 75015          13/37/58

0229          171 K11=(1./(MM(K,LL)-1))*(NN-F/NN)
0230          K12=(1./(MM(K,LL)-1))*(E-D/NN)
0231          K22=(1./(SS(K,LL)-MM(K,LL)))*(NN-E)
0232          SQA=IH-((Y3P**2)/NN)
C             PRINT 1011,NN
C             PRINT 1012,IH
C             PRINT 1013 ,MMEAN
0233          174 MQA=SQA/(MM(K,LL)-1)
0234          MQB=SQB/(SS(K,LL)-MM(K,LL))
0235          MQR=SQR/(NN-SS(K,LL))
0236          GO TO 173
0237          172 PRINT 1010,MM(K,LL),SS(K,LL),NN
0238          173 IF ( MQR.LT.0.) GO TO 71
0239          SDE=DSQRT(MQR)
0240          GO TO 72
0241          71 PRINT 1003,MQR,SQR,NN,SS(K,LL)
0242          SDE=0.
0243          72 RSDE=SDE/MMEAN *100.
0244          SIGB=(MQB-MQR)/K22
0245          IF (SIGB.LT.0.) GO TO 73
0246          SCB=DSQRT(SIGB)
0247          GO TO 74
0248          73 CONTINUE
C             73 PRINT 1004,SIGB,MQB,SQR,K22
0249          SCB=0.
0250          74 RSDB=SCB/MMEAN *100.
0251          SIGA=(1./K11)*(MQA-((K12/K22)*MQB)-((1.-(K12/K22))*MQR))
0252          IF (SIGA.LT.0.) GO TO 75
0253          SCA=DSQRT(SIGA)
0254          GO TO 76
0255          75 CONTINUE
C             75 PRINT 1005,SIGA,K11,MQA,K12,K22,MQB,MQR
0256          SDA=0.
0257          76 RSDA=SDA/MMEAN *100.
0258          PRINT 1002,TYP(K),IR(LL),MM(K,LL),SS(K,LL),NN,MMEAN,MQR,SDE,RSDE,
             1SIGB,SDB,R SDB,SIGA,SDA,R SDA
0259          80 CONTINUE
0260          90 CONTINUE
0261          RETURN
0262          END

```

FORTRAN IV G1 RELEASE 2.0

LAB6

DATE = 75015

13/37/58

0001

SUBROUTINE LAB6 (MEANV, IR,TYP,ILAB, MEAN2,IFILIV,NLAB)

C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C  
C

THIS SUBROUTINE CALCULATES INTERLAB AND RUN COMPONENTS OF THE VARIANCES FOR ALL THE ISOTOPIC RATIOS OF EACH MEASURED SAMPLE.

M - LAB  
K2 - SAMPLE TYPE  
L2 - RUN  
LL2 - ISOTOPIC RATIO

0002

REAL \* 8 SQIV( 8,3,11),MEANV(21,8,3,11),MEANA,MEAN2(21, 8,11),SUMA,SQL,SG2,SDG,RSDG,SQA,SQR,SIGA2,SDA,SIGE2,SDE,RSDE,RSDCON,DI2FF,RSDA,NLABV(8,11),MITTEV(8,11),MITTEL  
DIMENSION ILAB(21),IR(11),TYP( 8)  
INTEGER \* 2 IFILIV(21,8,11),MVAL(8,3,11)  
CCMMCN /HANNA / LINKS,MX,NVAL  
1/ITEMS/SG2V(21,8,11),SDGV(21,8,11),RSDGV(21,8,11),RSDCOV(21,8,11),SIGA2V(21,8,11),SDAV(21,8,11),RSDAV(21,8,11),SIGE2V(21,8,11),SDEV2(21,8,11),RSDEV(21,8,11),DIFV(21,8,11)  
REAL\*8 SG2V,SDGV,RSDGV,RSDCOV,SIGA2V,SDAV,RSDAV,SIGE2V,SDEV,RSDEV,1DIFVC  
C  
C

0007

1000 FORMAT ('RESULTS OF THE SUBROUTINE LAB6 - TWOFOLD ANALYSIS OF VARIANCES'//)

0008

1001 FCRMAT (' K2,LL2 AND NLABB ARE',3I5)

0009

1002 FORMAT (I3,2X,A2,2X,D10.4,2X, D10.4,1X, D10.4,1X,D10.4,2X,1010.4,1X, D10.4,1X,D10.4,2X, D10.4,1X, D10.4,1X,D10.4)

0010

1003 FORMAT (//)

0011

1004 FORMAT (' SQA=',D11.5,' SQR=',D11.5,' NLAB= ',I3)

0012

1005 FCRMAT (' ALARM ALARM ALARM ')

0013

1006 FORMAT (' THE INTERLAB COMPONENT OF THE VARIANCE IS NEGATIVE,IT IS  
1 PUT = -5.')

0014

1007 FORMAT (' THE RUN COMPONENT OF THE VARIANCE IS NEGATIVE,IT IS PUT  
1=-4.')

0015

1008 FCRMAT (' L=',I3)

0016

1009 FCRMAT (' K2=',I3)

0017

1010 FORMAT (' M,K2,L2,LL2, ARE',4I4,' MEANV IS',D14.4)

0018

1011 FORMAT (' SUM= ',D15.8)

0019

1012 FORMAT (' SUMA=SQL FROM DO 7 = ',D15.8)

0020

1013 FCRMAT (' SUMA FROM DO 8 =',D15.8)

```

FORTRAN IV G1  RELEASE 2.0          LAB6          DATE = 75015          13/37/58

0021      1014 FORMAT (' M,SG2V AND IFILIV ARE :',I4,D15.8,I4,' MEANV IS',D15.8)
C
C      THE DO LOOPS ARE ARRANGED IN THE ORDER WANTED FOR THE OUTPUT LIST
C      THE NORMAL ORDER OF ISOTOPIIC RATIOS IS: 83,92,02,38,48,58,68,89,09,19,29
C      THE ORDER WISHED HERE IS:          83,38,48,58,68,92,02,89,09,19,29
C
0022      PRINT 1000
0023      L2=1
C      DO-LOOP OVER THE ISOTOPIIC RATIOS
0024      DO 999 L=1,13
C      PRINT 1008,L
0025      LL2=L
0026      IF(L.EQ.2) GO TO 999
0027      IF(L.EQ.3) GO TO 999
0028      IF(L.EQ.8) LL2=2
0029      IF(L.EQ.9) LL2=3
0030      IF(L.GT.9) LL2=L-2
C      DO-LOOP OVER THE SAMPLE TYPES
C      THE SAMPLE TYPES ARE US,PS,AU,BU,RU,AS,BS,RS
0031      DO 998 K2=1,8
C      PRINT 1009,K2
C      TRICK FOR ORGANISATION OF THE OUTPUT LIST
C      LAB 10 WITH THE INDEX M=8 HAS MEASURED EVERYTHING
0032      M=8
C      PRINT 1010,M,K2,L2,LL2,MEANV(M,K2,L2,LL2)
0033      IF(MEANV(M,K2,L2,LL2).LT.-2.) GO TO 998
0034      NLAB = 19
0035      NLABB=19
0036      SUM=0.
C      CALCULATION OF THE MEAN OVER ALL THE LABS
0037      DO 5 M=1,NLAB
C      TEST,IF A LAB IS MISSING IN THE SENSE THAT ONLY ONE RUN WAS MEASURED
0038      IF (IFILIV(M,K2,LL2).LT.2) GO TO 6
0039      SUM=SUM+MEAN2(M ,K2,LL2)
0040      GO TO 5
0041      6 NLABB=NLABB-1
0042      5 CONTINUE
C      PRINT 1011,SUM
0043      NLABV(K2,LL2) = NLABB
C      PRINT 1001,K2,LL2,NLABB
0044      84 IF(NLABB.EQ.0) GO TO 86
0045      MITTEL=SUM/NLABB
0046      GO TO 87
0047      86 MITTEL=0.
0048      87 CONTINUE
0049      SQL=0.
0050      SG2=0.
0051      SDG=0.

```



```

FORTRAN IV G1  RELEASE 2.0          LAB6          DATE = 75015          13/37/58

0052          RSDG=0.
              C
0053          SQA=0.
0054          SIGA2=0.
0055          SDA=0.
0056          RSDA=0.
              C
0057          SQR = 0.
0058          SIGE2=0.
0059          SDE=0.
0060          RSDE=0.
              C
0061          CALCULATION OF THE VARIANCE
0062          SUMA=0.
0063          DO 7 M=1,NLAB
0064          IFILI=IFILIV(M,K2,LL2)
0065          IF(IFILI.LT.2) GO TO 7
0066          SUMA=SUMA+(MEAN2(M ,K2,LL2)-MITTEL)**2
0067          7 CONTINUE
              C
0068          PRINT 1012,SQL
0069          SUMA=0.
0070          DO 8 M=1,NLAB
0071          IF(MEAN2(M,K2,LL2).LE.0.) GO TO 8
0072          SUMA=SUMA + (MEAN2(M,K2,LL2)-MITTEL)**2
              C
0073          8 CCNTINUE
0074          PRINT 1013,SUMA
0075          IF(NLABB.LE.1) GO TO 89
0076          SG2=SQL/(NLABB*(NLABB-1))
0077          SDG=DSQRT(SG2)
0078          IF(MITTEL.LE.0.) GO TO 89
0079          RSDG=(SDG/MITTEL)*100.
0080          89 CONTINUE
              C
0081          C THE FOLLOWING CALCULATION IS O.K. FOR THE NORMAL CASE WITH 3 RUNS
0082          SQA=3.*SQL
0083          DO 9 M=1,NLAB
0084          PRINT 1014,M,SG2V(M,K2,LL2),IFILIV(M,K2,LL2),MEAN2(M,K2,LL2)
0085          9 SQR=SQR+SG2V(M,K2,LL2) * 6.
0086          PRINT 1004,SQA,SQR,NLABB
0087          IFILI=3
0088          IF(NLABB.LE.1) GO TO 15
0089          SIGA2=(1./IFILI)*((SQA/(NLABB-1))-(SQR/(NLABB*(IFILI-1))))
0090          GO TO 18
              C
0091          15 SIGA2=-5.
0092          18 IF (SIGA2.LT.0.00000000) GO TO 10
0093          SDA=DSQRT(SIGA2)
0094          IF(MITTEL. LT.0.) GO TO 11
0095          RSDA=(SCA/MITTEL)*100.

```

FORTRAN IV G1 RELEASE 2.0

LAB6

DATE = 75015

13/37/58

```
0091          GO TO 12
C 10 PRINT 1006
0092          10 CONTINUE
0093          SDA =-5.
0094          11 RSDA = -5.
0095          12 NUM=NLABB*IFILI
0096          IF((NUM-NLABB).LE.0) GO TO 16
0097          SIGE2=SQR/(NUM-NLABB)
0098          GO TO 19
0099          16 SIGE2=-4.
0100          19 IF(SIGE2.LT.0.00000000) GO TO 20
0101          SDE=DSQRT(SIGE2)
0102          IF(MITTEL. LT.0.) GO TO 13
0103          RSDE=(SDE/MITTEL)*100.
0104          GO TO 14
0105          20 PRINT 1007
0106          SDE=-4.
0107          13 RSDE=-4.
C          PRINTING OF THE RESULTS
0108          14 IF(L.EQ.8) PRINT 1003
C          EXTRA BLOCK FOR PLUTONIUM RESULTS
0109          PRINT 1002,IR(LL2),TYP(K2),MITTEL,SG2,SDG,RSDG,SIGA2,SDA,RSDA,
          1 SIGE2,SDE,RSDE
0110          998 CONTINUE
0111          999 CONTINUE
0112          RETURN
0113          END
```

FORTRAN IV G1 RELEASE 2.0

DRY

DATE = 75015

13/37/58

0001

SUBROUTINE DRY(M,MEAN2,ILAB,DICO)

C  
C  
C  
C  
C

0002

0003

```

COMMON/HANNA/LINKS,MX,NVAL/DRYA/ KO(21,3,5,2),KOAL(21,3,5,2)
REAL*8 SU,SPU,MEAN2(21,8,11),RSU,RPU,RSPU,RPPU,SUM,
1SIGMA(2,2,3),SD(2,2,3),RSD(2,2,3),DX(10),GS(21,2),GP(21,2),
2R(2,2,3)/12*0./,RR(2,2)/4*0./,D,SQI,XR(2,2,3),XRMEAN(2,2)
3,KO,KOAL,SIGLAB(2,2),SDLAB(2,2),RSDLAB(2,2),A,SUMA,SQILAB
4,DY(10),DICO(2),SUMIT(2,2),XMIT(2,2),DUCC(2)

```

C

0004

DIMENSION ILAB(21),ISAMP(21),IRUN(21),IREL(21),IEXP(21)

C

0005

1000 FORMAT (I1,I2,1X,I1,I2,1X,I2,5F10.6)

0006

1001 FORMAT (10X,7F10.6)

0007

1002 FORMAT (/10F10.6,5X,2F10.6,I3)

0008

1003 FORMAT (///'RESULTS OF THE DRY SPIKE EXPERIMENT'///,21X,  
1' SAMPLE A-1',2X,' SAMPLE A-2')

0009

1004 FORMAT (/ FOR URANIUM')

0010

1005 FORMAT(/// RESULTS OF THE STANDARD EXPERIMENT FOR COMPARISON'  
1/21X,' SAMPLE A')

0011

1006 FORMAT (/// RUN AND LAB MEANS,VALUES OF THE DRY SPIKE EXPERIMENT'  
1/' IN THE ORDER: SAMPLE 1-U233/238, SAMPLE 2-U233/238, SAMPLE 1-PU2  
242/239, SAMPLE 2-PU242/239' / ' VARIANCE ETC. OF SINGLE MEASUREMENT  
3S ARE GIVEN'/// RUN NO. RUN MEAN SIGMA SD RSD'  
43(I3,7X,D11.5,3D11.3/))

0012

1007 FORMAT (/ FOR PLUTONIUM')

0013

1008 FORMAT(' URANIUM RESULTS:/' NUMBER OF ATOMS 1 ',2D13.5/  
1' NUMBER OF ATOMS 2 ',2D13.5/' NUMBER OF ATOMS 3 ',2D13.5//  
2' PLUTONIUM RESULTS:/' NUMBER OF ATOMS 1 ',2D13.5/' NUMBER OF AT  
3OMS 2 ',2D13.5/' NUMBER OF ATOMS 3 ',2D13.5)

0014

1009 FORMAT (' MEAN NUMBER OF ATOMS/' FOR URANIUM:',8X,2D12.5/  
1' FOR PLUTONIUM:',6X,2D12.5)

0015

1010 FORMAT (/// CALCULATION ON THE BASIS OF THE THREE DIFFERENT ISOTOP  
1IC RATIOS MEASURED IN THE DRY SPIKE EXPERIMENT')

0016

1011 FORMAT (/ NUMBER OF ATOMS 1 ',D12.5/' NUMBER OF ATOMS 2 ',  
1D12.5/' NUMBER OF ATOMS 3 ',D12.5)

0017

1012 FORMAT ('RESULTS OF THE SUBROUTINE DRY FOR THE LAB WITH CODE NO.:  
1',I3)

0018

1013 FORMAT (' THE DIXON OUTPUT VALUES ARE:',10F10.6)

0019

1014 FORMAT (/// CALCULATION ON THE BASIS OF THREE RESULTING NUMBERS OF  
1 ATOMS CALCULATED FOR THE DRY SPIKE EXPERIMENT')

0020

1015 FORMAT (' MEAN NUMBER OF ATOMS'/18X,'SAMPLE A-1 SD(MEAN) RSD(M  
1EAN) SAMPLE A-2 SD(MEAN) RSD(MEAN)'/ URANIUM',9X,6D12.5/' P  
2LUTONIUM',7X,6D12.5)

```

FORTRAN IV GI  RELEASE 2.0          DRY          DATE = 75015          13/37/58

0021      1016 FORMAT (4I4,7D12.5)
0022      1017 FORMAT (//' *** XR=8.888 ***')
          C
0023      PRINT 1012 ,ILAB(M)
          C
0024      NOUT = 0
0025      NVAL = 8
0026      MX=10
          C
0027      SU=2.0931E+18
0028      SPU=2.4594E+16
0029      RSU=MEAN2(M,1,1)
0030      RPU=MEAN2(M,3,4)
0031      RSPU=MEAN2(M,2,2)
0032      RPPU=MEAN2(M,3,11)
          C      DO OVER THE RATIOS U233/238 AND PU242/239
0033      DO 50 ID=1,2
          C
          C      DO OVER THE SAMPLES 1 AND 2
0034      DO 40 JD=1,2
0035      SUMA=0.
          C
          C      DO OVER THE THREE RUNS
0036      DO 30 KD=1,3
0037      READ (5,1000) IEXP(M),ILAB(M),ISAMP(M),IRUN(M),IREL(M),(DX(I),I=1,
          15)
0038      ISA=ISAMP(M)
0039      READ (5,1001) (DX(I),I=6,10),GS(M,ISA),GP(M,ISA)
0040      PRINT 1002,(DX(I),I=1,10),GS(M,ISA),GP(M,ISA),ISA
          C      MASS DISCRIMINATION CORRECTIONS
0041      IF(ILAB(M).NE.13) GO TO 17
0042      DUCO(1)=0.9871
0043      DUCO(2)=1.00847
0044      DO 16 LC=1,10
0045      16 DX(LC)=DX(LC) * DUCO(ID)
0046      17 IF(ILAB(M).NE.18) GO TO 19
0047      DO 18 LC=1,10
0048      18 DX(LC)=DX(LC)* DICO(ID)
0049      19 CALL DIXON(DX,DY,NOUT)
0050      SUM=0.
0051      DO 20 LD=1,8
0052      20 SUM=SUM+DY(LD)
0053      IF(ILAB(M).NE.18) GO TO 21
0054      PRINT 1013,(DY(I),I=1,8),SUM
0055      21 SQI=0.
0056      R(ID,JC,KD)=SUM/8.
          C      PRINT 1013,(DY(I),I=1,8),SUM,R(ID,JD,KD)
0057      DO 10 LD=1,8

```

```

FORTRAN IV G1  RELEASE 2.0                DRY                DATE = 75015                13/37/58

0058          D=DY(LD)-R(ID,JD,KD)
0059          10 SQI=SQI+D**2
0060          SIGMA(ID,JD,KD)=SQI/7.
0061          A=SIGMA(ID,JD,KD)
0062          SD(ID,JD,KD)=DSQRT(A)
0063          RSD(ID,JD,KD)=(SD(ID,JD,KD)/R(ID,JD,KD))*100.
0064          SUMA=SUMA+R(ID,JD,KD)

C
C          TEST FOR SPECIAL COMBINATION OF NUMBERS
0065          IF((R(1,JD,KD)-RPU).LE.0.) GO TO 11

C
C          NO. OF ATOMS FOR URANIUM
0066          XR(1,JD,KD)=(SU*GS(M,JD)*(1.-R(1,JD,KD)*RSU))/(GP(M,JD)*(R(1,
          1JD,KD)-RPU))
C          PRINT 1016, ID, JD, KD, M, XR(1,JD,KD), SU, GS(M,JD), R(1,JD,KD), RSU, GP
C          1(M,JD), RPU
0067          GO TO 12
0068          11 XR(1,JD,KD)=8.888
0069          PRINT 1017
0070          12 IF(ID.EQ.1) GO TO 30

C
C          NO. OF ATOMS FOR PLUTONIUM
0071          XR(2,JD,KD)=(SPU*GS(M,JD)*(1.-R(2,JD,KD)*RSPU))/
          1(GP(M,JD)*(R(2,JD,KD)-RPPU))
C          PRINT 1016, ID, JD, KD, M, XR(2,JD,KD), SPU, GS(M,JD), R(2,JD,KD), RSPU, GP
C          1(M,JD), RPPU

0072          30 CONTINUE
0073          RR(ID,JD)=SUMA/3.
0074          SQILAB=0.
0075          SUMIT(ID,JD)=0.
0076          DO 29 KD=1,3
0077          SUMIT(ID,JD)=SUMIT(ID,JD) + XR(ID,JD,KD)
0078          D=R(ID,JD,KD) - RR(ID,JD)
0079          29 SQILAB=SQILAB+ D**2
0080          SIGLAB(ID,JD)=SQILAB/6.
0081          SDLAB(ID,JD)=DSQRT(SIGLAB(ID,JD))
0082          RSDLAB(ID,JD)=(SDLAB(ID,JD)/RR(ID,JD))*100.
0083          XMIT(ID,JD)=SUMIT(ID,JD)/3.
0084          XRMEAN(1,JD)=(SU*GS(M,JD)*(1.-RR(1,JD)*RSU))/(GP(M,JD)*RR(1,
          1JD)-RPU)
0085          IF(ID.EQ.1) GO TO 40
0086          XRMEAN(2,JD)=(SPU*GS(M,JD)*(1.-RR(2,JD)*RSPU))/(GP(M,JD)
          1 *RR(2,JD)-RPPU)

0087          40 CONTINUE
0088          50 CONTINUE
0089          PRINT 1006,(((KD,R(ID,JD,KD),SIGMA(ID,JD,KD),SD(ID,JD,KD),RSD
          1(ID,JD,KD),KD=1,3),JD=1,2),ID=1,2)
0090          PRINT 1005

```

FORTRAN IV G1 RELEASE 2.0

DRY

DATE = 75015

13/37/58

```
0091      PRINT 1004
0092      PRINT 1011 , (KO(M,L4,3,1),L4=1,3)
0093      PRINT 1007
0094      PRINT 1011, (KO(M,L4,3,2),L4=1,3)
0095      PRINT 1003
0096      PRINT 1008 , (((XR(ID,JD,KD),JD=1,2),KD=1,3),ID=1,2)
0097      PRINT 1010
0098      PRINT 1009 , ((XRMEAN(ID,JD),JD=1,2),ID=1,2)
0099      PRINT 1014
0100      PRINT 1015, (((XHIT(ID,JC),SDLAB(ID,JD),RSELAB(ID,JD)),JD=1,2),ID=1
1,2)
0101      RETURN
0102      END
```

12. Example of Evaluation-Results for One Laboratory  
and the Results for the Group of Laboratories

12.1 Results of One Laboratory for the Standard Experiment

We are very grateful to the laboratory with code number 15  
for the allowance to publish their data.

## NEW SERIES OF MEASURING RESULTS

15 US1 83  
 THE ORDERED VALUES S(I) ARE: 0.021770 0.021820 0.021910 0.021960 0.021990 0.021990 0.022140 0.022170 0.022250 0.022280  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.021990 0.021820 0.022170 0.021990 0.022280 0.022140 0.021770 0.021910

15 US2 83  
 THE ORDERED VALUES S(I) ARE: 0.021330 0.021640 0.021650 0.021690 0.021730 0.021820 0.021870 0.021890 0.022030 0.022230  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.021330 0.022030 0.022230 0.021690 0.021890 0.021640 0.021870 0.021820

15 US3 83  
 THE ORDERED VALUES S(I) ARE: 0.021710 0.021760 0.021800 0.021810 0.021820 0.021840 0.021910 0.021920 0.021930 0.021940  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.021710 0.021800 0.021810 0.021840 0.021930 0.021940 0.021910 0.021760

15 PS1 92  
 THE ORDERED VALUES S(I) ARE: 0.000197 0.000203 0.000211 0.000220 0.000222 0.000223 0.000224 0.000229 0.000229 0.000255  
 THE MAXIMUM VALUE S(10)= 0.000255 IS CANCELLED  
 THE RESULTING VALUES X(I) ARE 0.000222 0.000223 0.000211 0.000229 0.000203 0.000229 0.000224 0.000220  
 THE NUMBER OF OUTLIERS IS 1

15 PS1 2  
 THE ORDERED VALUES S(I) ARE: 0.000780 0.000790 0.000790 0.000810 0.000810 0.000830 0.000840 0.000850 0.000890 0.000900  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000810 0.000810 0.000850 0.000900 0.000790 0.000890 0.000840 0.000830

15 PS2 92  
 THE ORDERED VALUES S(I) ARE: 0.000190 0.000204 0.000205 0.000211 0.000223 0.000223 0.000233 0.000236 0.000238 0.000246  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000211 0.000190 0.000204 0.000205 0.000223 0.000233 0.000233 0.000246

15 PS2 2  
 THE ORDERED VALUES S(I) ARE: 0.000843 0.000847 0.000849 0.000850 0.000860 0.000862 0.000864 0.000875 0.000880 0.000890  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000890 0.000880 0.000875 0.000843 0.000847 0.000850 0.000864 0.000860

15 PS3 92  
 THE ORDERED VALUES S(I) ARE: 0.000201 0.000224 0.000226 0.000231 0.000231 0.000232 0.000237 0.000243 0.000250 0.000256  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000232 0.000224 0.000237 0.000201 0.000250 0.000231 0.000226 0.000231

15 PS3 2  
 THE ORDERED VALUES S(I) ARE: 0.000831 0.000833 0.000836 0.000840 0.000845 0.000850 0.000852 0.000853 0.000860 0.000879  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000852 0.000831 0.000833 0.000840 0.000845 0.000850 0.000853 0.000879

15 AU1 38  
 A SERIES OF X(I) COULD NOT BE MEASURED

15 AU1 48  
 THE ORDERED VALUES S(I) ARE: 0.000162 0.000162 0.000164 0.000166 0.000168 0.000168 0.000171 0.000173 0.000174 0.000176  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000168 0.000162 0.000176 0.000174 0.000162 0.000173 0.000168 0.000171

15 AU1 58  
 THE ORDERED VALUES S(I) ARE: 0.022300 0.022300 0.022300 0.022400 0.022400 0.022400 0.022600 0.022600 0.022700 0.022700  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.022300 0.022300 0.022600 0.022700 0.022400 0.022700 0.022400 0.022600

15 AU1 68  
 THE ORDERED VALUES S(I) ARE: 0.003900 0.003930 0.003930 0.003950 0.003990 0.004000 0.004000 0.004010 0.004030  
 THE RESULTING VALUES X(I) ARE 0.003930 0.003950 0.004000 0.004000 0.003990 0.003930 0.003900 0.004010



15 AU2 38  
A SERIES OF X(I) COULD NOT BE MEASURED

15 AU2 48  
THE ORDERED VALUES S(I) ARE: 0.000162 0.000164 0.000167 0.000168 0.000169 0.000170 0.000177 0.000179 0.000183 0.000187  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.000179 0.000164 0.000162 0.000170 0.000168 0.000187 0.000169 0.000177

15 AU2 58  
THE ORDERED VALUES S(I) ARE: 0.021800 0.021900 0.022100 0.022200 0.022300 0.022400 0.022600 0.022600 0.022700 0.022700  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.021800 0.022100 0.021900 0.022300 0.022700 0.022400 0.022600 0.022600

15 AU2 68  
THE ORDERED VALUES S(I) ARE: 0.003840 0.003930 0.003940 0.003940 0.003980 0.003980 0.004000 0.004010 0.004070 0.004100  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.003930 0.003840 0.003980 0.003940 0.003980 0.003940 0.004010 0.004070

15 AU3 38  
A SERIES OF X(I) COULD NOT BE MEASURED

15 AU3 48  
THE ORDERED VALUES S(I) ARE: 0.000153 0.000154 0.000159 0.000165 0.000168 0.000170 0.000181 0.000186 0.000196 0.000204  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.000153 0.000154 0.000165 0.000186 0.000159 0.000196 0.000168 0.000170

15 AU3 58  
THE ORDERED VALUES S(I) ARE: 0.022200 0.022200 0.022200 0.022300 0.022400 0.022400 0.022400 0.022400 0.022600 0.022600  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.022400 0.022200 0.022400 0.022200 0.022600 0.022400 0.022600 0.022200

15 AU3 68  
THE ORDERED VALUES S(I) ARE: 0.003850 0.003890 0.003920 0.003920 0.003970 0.004010 0.004010 0.004010 0.004060 0.004150  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.003920 0.004010 0.003920 0.004010 0.003890 0.004150 0.003970 0.003850

15 BU1 38  
A SERIES OF X(I) COULD NOT BE MEASURED

15 BU1 48  
THE ORDERED VALUES S(I) ARE: 0.000159 0.000166 0.000170 0.000176 0.000185 0.000185 0.000189 0.000191 0.000204 0.000213  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.000191 0.000159 0.000204 0.000185 0.000170 0.000176 0.000213 0.000166

15 BU1 58  
THE ORDERED VALUES S(I) ARE: 0.021900 0.022000 0.022100 0.022200 0.022300 0.022300 0.022400 0.022500 0.022500 0.022500  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.022500 0.022000 0.022400 0.022200 0.022300 0.022100 0.021900 0.022500

15 BU1 68  
THE ORDERED VALUES S(I) ARE: 0.003850 0.003910 0.003920 0.003920 0.003940 0.003940 0.003960 0.003960 0.003990 0.004010  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.003920 0.003910 0.003850 0.003990 0.003940 0.003940 0.004010 0.003920

15 BU2 38  
A SERIES OF X(I) COULD NOT BE MEASURED

15 BU2 48  
THE ORDERED VALUES S(I) ARE: 0.000158 0.000159 0.000166 0.000176 0.000181 0.000184 0.000188 0.000192 0.000197 0.000197  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.000188 0.000158 0.000159 0.000176 0.000184 0.000192 0.000166 0.000197

15 BU2 58  
THE ORDERED VALUES S(I) ARE: 0.022100 0.022100 0.022200 0.022300 0.022400 0.022400 0.022500 0.022500 0.022600 0.022600  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.022100 0.022100 0.022500 0.022500 0.022400 0.022300 0.022200 0.022400

15 BU2 68  
 THE ORDERED VALUES S(I) ARE: 0.003930 0.003960 0.003960 0.003980 0.003980 0.004010 0.004010 0.004020 0.004020 0.004020  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.003930 0.003960 0.004020 0.004020 0.003980 0.004010 0.004010 0.003960

15 BU3 38  
 A SERIES OF X(I) COULD NOT BE MEASURED

15 BU3 48  
 THE ORDERED VALUES S(I) ARE: 0.000155 0.000168 0.000171 0.000175 0.000175 0.000177 0.000178 0.000179 0.000179 0.000186  
 THE MINIMUM VALUE S(1)= 0.000155 IS CANCELLED  
 THE RESULTING VALUES X(I) ARE 0.000171 0.000178 0.000186 0.000175 0.000168 0.000179 0.000179 0.000175  
 THE NUMBER OF OUTLIERS IS 1

15 BU3 58  
 THE ORDERED VALUES S(I) ARE: 0.022200 0.022300 0.022300 0.022400 0.022400 0.022400 0.022500 0.022500 0.022500 0.022600  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.022600 0.022300 0.022200 0.022500 0.022400 0.022400 0.022500 0.022400

15 BU3 68  
 THE ORDERED VALUES S(I) ARE: 0.003890 0.003920 0.003920 0.003930 0.003950 0.003960 0.003970 0.003970 0.003980 0.003990  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.003960 0.003980 0.003970 0.003990 0.003970 0.003950 0.003920 0.003930

15 RU1 38  
 A SERIES OF X(I) COULD NOT BE MEASURED

15 RU1 48  
 THE ORDERED VALUES S(I) ARE: 0.000059 0.000062 0.000062 0.000062 0.000063 0.000065 0.000065 0.000065 0.000068 0.000068  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000068 0.000068 0.000065 0.000065 0.000063 0.000062 0.000062 0.000065

15 RU1 58  
 THE ORDERED VALUES S(I) ARE: 0.007280 0.007280 0.007290 0.007330 0.007340 0.007340 0.007370 0.007380 0.007390 0.007420  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.007420 0.007390 0.007280 0.007340 0.007330 0.007380 0.007370 0.007280

15 RU1 68  
 A SERIES OF X(I) COULD NOT BE MEASURED

15 RU2 38  
 A SERIES OF X(I) COULD NOT BE MEASURED

15 RU2 48  
 THE ORDERED VALUES S(I) ARE: 0.000033 0.000042 0.000045 0.000047 0.000050 0.000057 0.000058 0.000059 0.000067 0.000075  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000050 0.000058 0.000042 0.000075 0.000045 0.000057 0.000047 0.000059

15 RU2 58  
 THE ORDERED VALUES S(I) ARE: 0.007190 0.007200 0.007220 0.007260 0.007290 0.007310 0.007370 0.007430 0.007430 0.007460  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.007370 0.007260 0.007430 0.007290 0.007310 0.007460 0.007220 0.007430

15 RU2 68  
 A SERIES OF X(I) COULD NOT BE MEASURED

15 RU3 38  
 A SERIES OF X(I) COULD NOT BE MEASURED

15 RU3 48  
 THE ORDERED VALUES S(I) ARE: 0.000070 0.000080 0.000081 0.000083 0.000084 0.000086 0.000098 0.000104 0.000109 0.000113  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000081 0.000086 0.000084 0.000070 0.000098 0.000080 0.000083 0.000109

15 RU3 58

THE ORDERED VALUES S(I) ARE: 0.007290 0.007320 0.007330 0.007330 0.007390 0.007430 0.007430 0.007470 0.007490 0.007500  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.007430 0.007330 0.007330 0.007290 0.007500 0.007490 0.007390 0.007320

15 RU3 68

A SERIES OF X(I) COULD NOT BE MEASURED

15 AU1 89

A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 AU1 9

THE ORDERED VALUES S(I) ARE: 0.228900 0.228900 0.229100 0.230000 0.230000 0.230500 0.230700 0.231200 0.231400 0.231400  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.228900 0.229100 0.231400 0.230000 0.230500 0.230700 0.231400 0.228900

15 AU1 19

THE ORDERED VALUES S(I) ARE: 0.125900 0.126000 0.126500 0.126600 0.126600 0.127100 0.127300 0.127300 0.128100 0.128300  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.125900 0.126600 0.127300 0.127100 0.126000 0.127300 0.128100 0.128300

15 AU1 29

THE ORDERED VALUES S(I) ARE: 0.022700 0.022800 0.022900 0.022900 0.022900 0.023000 0.023000 0.023000 0.023100 0.023300  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.023000 0.022900 0.022700 0.022800 0.023000 0.023300 0.022900 0.023000

15 AU2 89

A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 AU2 9

THE ORDERED VALUES S(I) ARE: 0.229600 0.229800 0.229800 0.230400 0.230600 0.230800 0.230800 0.231400 0.231600 0.231800  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.231800 0.230800 0.230600 0.231600 0.231400 0.229800 0.230800 0.229800

15 AU2 19

THE ORDERED VALUES S(I) ARE: 0.124600 0.125000 0.125200 0.125500 0.125600 0.126200 0.126200 0.127300 0.127300 0.127500  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.127300 0.125600 0.125500 0.127500 0.127300 0.126200 0.126200 0.125200

15 AU2 29

THE ORDERED VALUES S(I) ARE: 0.021900 0.022600 0.022700 0.022900 0.022900 0.023000 0.023100 0.023100 0.023300 0.023700  
 THE MINIMUM VALUE S(1)= 0.021900 IS CANCELLED  
 THE RESULTING VALUES X(I) ARE 0.023100 0.022700 0.022900 0.023100 0.023000 0.022900 0.022600 0.023700  
 THE NUMBER OF OUTLIERS IS 1

15 AU3 89

A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 AU3 9

THE ORDERED VALUES S(I) ARE: 0.230300 0.230500 0.230800 0.230900 0.231600 0.232000 0.232500 0.232700 0.233400 0.235200  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.230900 0.230500 0.232000 0.235200 0.231600 0.233400 0.230300 0.230800

15 AU3 19

THE ORDERED VALUES S(I) ARE: 0.125400 0.125900 0.126600 0.127000 0.127100 0.127600 0.128500 0.128700 0.128900 0.129400  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.125900 0.125400 0.126600 0.128700 0.127000 0.128900 0.127600 0.128500

19 AU3 29

THE ORDERED VALUES S(I) ARE: 0.022400 0.022900 0.023100 0.023200 0.023200 0.023300 0.023400 0.023500 0.023600 0.023800  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.023100 0.023200 0.022400 0.022900 0.023500 0.023200 0.023600 0.023800

15 BU1 89

A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 BU1 9  
 THE ORDERED VALUES S(I) ARE: 0.227700 0.228300 0.228900 0.230200 0.230600 0.230600 0.231000 0.231300 0.231300 0.232100  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.231300 0.231300 0.228300 0.230600 0.228900 0.230600 0.231000 0.230200

15 BU1 19  
 THE ORDERED VALUES S(I) ARE: 0.125300 0.125800 0.126100 0.126300 0.126700 0.126900 0.127600 0.127800 0.128100 0.128500  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.125300 0.126100 0.126300 0.128100 0.127800 0.126700 0.126900 0.125800

15 BU1 29  
 THE ORDERED VALUES S(I) ARE: 0.022700 0.022700 0.022800 0.022800 0.022900 0.023000 0.023300 0.023300 0.023300 0.023400  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.023300 0.022800 0.022900 0.023300 0.022800 0.023400 0.022700 0.022700

15 BU2 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 BU2 9  
 THE ORDERED VALUES S(I) ARE: 0.229400 0.229700 0.229700 0.230500 0.231300 0.231700 0.231800 0.231800 0.232200 0.232400  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.229700 0.229700 0.230500 0.232400 0.231700 0.232200 0.231800 0.229400

15 BU2 19  
 THE ORDERED VALUES S(I) ARE: 0.125900 0.126100 0.126500 0.126500 0.126900 0.126900 0.127100 0.127400 0.127500 0.128000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.126500 0.125900 0.127400 0.126900 0.127500 0.126500 0.128000 0.127100

15 BU2 29  
 THE ORDERED VALUES S(I) ARE: 0.022700 0.022700 0.022700 0.022800 0.022900 0.023000 0.023000 0.023100 0.023200 0.023400  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.022700 0.022900 0.022700 0.022800 0.023000 0.023200 0.023000 0.023100

15 BU3 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 BU3 9  
 THE ORDERED VALUES S(I) ARE: 0.229300 0.229400 0.229700 0.229700 0.229900 0.230100 0.230300 0.230600 0.230800 0.231500  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.229400 0.229900 0.230100 0.230600 0.230300 0.231500 0.229700 0.230800

15 BU3 19  
 THE ORDERED VALUES S(I) ARE: 0.124900 0.125500 0.125600 0.125800 0.125900 0.125900 0.126200 0.126500 0.127000 0.127000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.124900 0.125600 0.125900 0.125900 0.127000 0.125800 0.127000 0.126500

15 BU3 29  
 THE ORDERED VALUES S(I) ARE: 0.022600 0.022900 0.022900 0.023000 0.023000 0.023000 0.023100 0.023200 0.023200 0.023200  
 THE MINIMUM VALUE S(I)= 0.022600 IS CANCELLED  
 THE RESULTING VALUES X(I) ARE 0.023000 0.023000 0.022900 0.023000 0.023200 0.022900 0.023200 0.023100  
 THE NUMBER OF OUTLIERS IS 1

15 RU1 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 RU1 9  
 THE ORDERED VALUES S(I) ARE: 0.026100 0.026200 0.026300 0.026300 0.026400 0.026600 0.026600 0.026800 0.026800 0.026900  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.026100 0.026800 0.026800 0.026400 0.026300 0.026600 0.026900 0.026200

15 RU1 19  
 THE ORDERED VALUES S(I) ARE: 0.000732 0.000747 0.000748 0.000748 0.000753 0.000760 0.000767 0.000781 0.000783 0.000787  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000783 0.000748 0.000767 0.000760 0.000753 0.000787 0.000748 0.000781

15 RU1 29  
 THE ORDERED VALUES S(I) ARE: 0.000015 0.000017 0.000018 0.000020 0.000021 0.000021 0.000022 0.000024 0.000027 0.000034

THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000018 0.000024 0.000021 0.000020 0.000015 0.000022 0.000027 0.000034

15 RU2 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 RU2 9  
 THE ORDERED VALUES S(I) ARE: 0.026100 0.026100 0.026200 0.026300 0.026400 0.026500 0.026500 0.026600 0.026600 0.026700  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.026600 0.026100 0.026300 0.026500 0.026600 0.026700 0.026400

15 RU2 19  
 THE ORDERED VALUES S(I) ARE: 0.000713 0.000717 0.000719 0.000734 0.000748 0.000754 0.000755 0.000759 0.000764 0.000766  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000755 0.000754 0.000719 0.000713 0.000766 0.000734 0.000759 0.000748

15 RU2 29  
 THE ORDERED VALUES S(I) ARE: 0.000035 0.000040 0.000043 0.000044 0.000046 0.000050 0.000051 0.000053 0.000054 0.000057  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000044 0.000053 0.000054 0.000035 0.000050 0.000046 0.000057 0.000040

15 RU3 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 RU3 9  
 THE ORDERED VALUES S(I) ARE: 0.025700 0.025900 0.026000 0.026000 0.026000 0.026000 0.026100 0.026200 0.026300 0.026500  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.026000 0.026500 0.026100 0.026000 0.026000 0.026000 0.025900 0.026200

15 RU3 19  
 THE ORDERED VALUES S(I) ARE: 0.000742 0.000747 0.000756 0.000763 0.000765 0.000767 0.000768 0.000791 0.000798 0.000814  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000768 0.000756 0.000742 0.000767 0.000763 0.000791 0.000814 0.000798

15 RU3 29  
 THE ORDERED VALUES S(I) ARE: 0.000069 0.000071 0.000073 0.000075 0.000077 0.000077 0.000080 0.000081 0.000085 0.000087  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000071 0.000075 0.000073 0.000080 0.000087 0.000085 0.000081 0.000077

15 AS1 38  
 THE ORDERED VALUES S(I) ARE: 0.646400 0.647200 0.647400 0.647800 0.648100 0.649300 0.649900 0.650000 0.651200 0.651900  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.647800 0.647200 0.646400 0.649900 0.647400 0.649300 0.648100 0.650000

15 AS1 48  
 THE ORDERED VALUES S(I) ARE: 0.000246 0.000261 0.000275 0.000283 0.000283 0.000284 0.000290 0.000317 0.000318 0.000319  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000246 0.000275 0.000283 0.000283 0.000261 0.000317 0.000319 0.000318

15 AS1 58  
 THE ORDERED VALUES S(I) ARE: 0.022400 0.022440 0.022440 0.022460 0.022460 0.022470 0.022500 0.022510 0.022510 0.022570  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.022440 0.022400 0.022510 0.022440 0.022510 0.022460 0.022470 0.022460

15 AS1 68  
 THE ORDERED VALUES S(I) ARE: 0.003900 0.003940 0.003970 0.003980 0.003980 0.003990 0.004000 0.004000 0.004000 0.004030  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.003970 0.003900 0.003940 0.003980 0.003990 0.004030 0.004000 0.003980

15 AS2 38  
 THE ORDERED VALUES S(I) ARE: 0.634000 0.634000 0.638000 0.641000 0.641000 0.642000 0.642000 0.643000 0.644000 0.645000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.644000 0.642000 0.641000 0.641000 0.634000 0.643000 0.645000 0.642000

15 AS2 48  
 THE ORDERED VALUES S(I) ARE: 0.000356 0.000381 0.000405 0.000405 0.000409 0.000432 0.000434 0.000450 0.000454 0.000462

THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.000381 0.000356 0.000405 0.000409 0.000405 0.000432 0.000434 0.000454

15 AS2 58  
THE ORDERED VALUES S(I) ARE: 0.022200 0.022300 0.022300 0.022300 0.022400 0.022400 0.022500 0.022500 0.022700 0.022800  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.022200 0.022700 0.022500 0.022500 0.022300 0.022800 0.022400 0.022300

15 AS2 68  
THE ORDERED VALUES S(I) ARE: 0.003880 0.003890 0.003920 0.003970 0.003980 0.003980 0.004000 0.004030 0.004040 0.004060  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.003920 0.003890 0.004030 0.003880 0.003980 0.003980 0.004040 0.003970

15 AS3 38  
THE ORDERED VALUES S(I) ARE: 0.642000 0.642000 0.645000 0.645000 0.646000 0.646000 0.646000 0.647000 0.648000 0.649000  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.645000 0.647000 0.649000 0.648000 0.646000 0.646000 0.646000 0.645000

15 AS3 48  
THE ORDERED VALUES S(I) ARE: 0.000275 0.000286 0.000289 0.000292 0.000292 0.000295 0.000296 0.000298 0.000308 0.000308  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.000308 0.000308 0.000289 0.000292 0.000296 0.000275 0.000295 0.000292

15 AS3 58  
THE ORDERED VALUES S(I) ARE: 0.022000 0.022000 0.022100 0.022200 0.022300 0.022300 0.022300 0.022400 0.022400 0.022600  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.022000 0.022300 0.022100 0.022600 0.022200 0.022300 0.022400 0.022400

15 AS3 68  
THE ORDERED VALUES S(I) ARE: 0.003890 0.003890 0.003900 0.003910 0.003910 0.003920 0.003930 0.003950 0.003970 0.003970  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.003970 0.003950 0.003910 0.003910 0.003890 0.003930 0.003920 0.003890

15 BS1 38  
THE ORDERED VALUES S(I) ARE: 0.633000 0.634000 0.634000 0.636000 0.637000 0.637000 0.639000 0.639000 0.639000 0.640000  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.637000 0.640000 0.636000 0.634000 0.639000 0.639000 0.634000 0.633000

15 BS1 48  
THE ORDERED VALUES S(I) ARE: 0.000297 0.000298 0.000299 0.000302 0.000308 0.000312 0.000320 0.000332 0.000335 0.000359  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.000359 0.000332 0.000335 0.000299 0.000308 0.000320 0.000312 0.000298

15 BS1 58  
THE ORDERED VALUES S(I) ARE: 0.021900 0.022200 0.022300 0.022300 0.022300 0.022400 0.022400 0.022500 0.022700 0.022800  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.022700 0.022300 0.021900 0.022400 0.022300 0.022400 0.022500 0.022300

15 BS1 68  
THE ORDERED VALUES S(I) ARE: 0.003800 0.003840 0.003850 0.003910 0.003910 0.003970 0.003980 0.003990 0.004000 0.004000  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.003910 0.003840 0.003850 0.003910 0.003980 0.004000 0.003970 0.003800

15 BS2 38  
THE ORDERED VALUES S(I) ARE: 0.625000 0.625000 0.628000 0.629000 0.630000 0.630000 0.631000 0.631000 0.632000 0.634000  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.634000 0.631000 0.632000 0.630000 0.630000 0.631000 0.629000 0.625000

15 BS2 48  
THE ORDERED VALUES S(I) ARE: 0.000274 0.000275 0.000277 0.000278 0.000278 0.000283 0.000284 0.000287 0.000288 0.000298  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.000287 0.000288 0.000283 0.000274 0.000275 0.000278 0.000278 0.000298

15 BS2 58  
THE ORDERED VALUES S(I) ARE: 0.021900 0.022000 0.022100 0.022100 0.022200 0.022300 0.022300 0.022400 0.022400 0.022400  
THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
0.022300 0.022400 0.021900 0.022200 0.022400 0.022100 0.022300 0.022400

15 BS2 68  
 THE ORDERED VALUES S(I) ARE: 0.003810 0.003910 0.003960 0.003980 0.003980 0.003990 0.004010 0.004040 0.004100 0.004160  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.003980 0.003910 0.003810 0.004040 0.004100 0.004160 0.003960 0.003980

15 BS3 38  
 THE ORDERED VALUES S(I) ARE: 0.627000 0.628000 0.631000 0.633000 0.636000 0.637000 0.638000 0.639000 0.640000 0.640000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.638000 0.640000 0.631000 0.628000 0.627000 0.633000 0.639000 0.640000

15 BS3 48  
 THE ORDERED VALUES S(I) ARE: 0.000290 0.000300 0.000302 0.000304 0.000307 0.000311 0.000318 0.000335 0.000338 0.000340  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000304 0.000338 0.000335 0.000318 0.000311 0.000300 0.000290 0.000302

15 BS3 58  
 THE ORDERED VALUES S(I) ARE: 0.022000 0.022100 0.022300 0.022300 0.022300 0.022400 0.022400 0.022400 0.022600 0.022700  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.022400 0.022100 0.022300 0.022300 0.022000 0.022400 0.022600 0.022700

15 BS3 68  
 THE ORDERED VALUES S(I) ARE: 0.003840 0.003910 0.003930 0.003930 0.003940 0.003950 0.004010 0.004050 0.004050 0.004100  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.003930 0.003940 0.003840 0.003910 0.003950 0.003930 0.004010 0.004050

15 RS1 38  
 THE ORDERED VALUES S(I) ARE: 0.755000 0.755000 0.757000 0.758000 0.758000 0.759000 0.760000 0.760000 0.760000 0.762000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.760000 0.759000 0.755000 0.755000 0.758000 0.762000 0.760000 0.758000

15 RS1 48  
 THE ORDERED VALUES S(I) ARE: 0.000278 0.000286 0.000288 0.000291 0.000295 0.000300 0.000306 0.000308 0.000311 0.000314  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000286 0.000308 0.000288 0.000306 0.000295 0.000314 0.000300 0.000278

15 RS1 58  
 THE ORDERED VALUES S(I) ARE: 0.007650 0.007650 0.007670 0.007700 0.007700 0.007710 0.007720 0.007730 0.007750 0.007760  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.007650 0.007760 0.007700 0.007670 0.007710 0.007750 0.007650 0.007700

15 RS1 68  
 THE ORDERED VALUES S(I) ARE: 0.000078 0.000093 0.000096 0.000098 0.000111 0.000113 0.000114 0.000119 0.000119 0.000121  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000114 0.000119 0.000113 0.000121 0.000111 0.000119 0.000098 0.000096

15 RS2 38  
 THE ORDERED VALUES S(I) ARE: 0.760000 0.760000 0.761000 0.761000 0.762000 0.762000 0.763000 0.764000 0.764000 0.766000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.761000 0.760000 0.766000 0.764000 0.762000 0.764000 0.760000 0.761000

15 RS2 48  
 THE ORDERED VALUES S(I) ARE: 0.000221 0.000225 0.000227 0.000228 0.000231 0.000234 0.000236 0.000239 0.000242 0.000252  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000221 0.000236 0.000234 0.000252 0.000231 0.000225 0.000242 0.000239

15 RS2 58  
 THE ORDERED VALUES S(I) ARE: 0.007350 0.007400 0.007400 0.007420 0.007510 0.007550 0.007600 0.007610 0.007620 0.007650  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.007400 0.007420 0.007510 0.007610 0.007650 0.007620 0.007600 0.007550

15 RS2 68  
 THE ORDERED VALUES S(I) ARE: 0.000030 0.000031 0.000032 0.000033 0.000033 0.000035 0.000036 0.000037 0.000038 0.000039  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000032 0.000030 0.000031 0.000035 0.000039 0.000038 0.000036 0.000033

15 RS3 38  
 THE ORDERED VALUES S(I) ARE: 0.758000 0.759000 0.759000 0.761000 0.761000 0.762000 0.762000 0.762000 0.762000 0.763000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.762000 0.763000 0.762000 0.759000 0.759000 0.758000 0.761000 0.762000

15 RS3 48  
 THE ORDERED VALUES S(I) ARE: 0.000177 0.000177 0.000186 0.000187 0.000194 0.000195 0.000196 0.000196 0.000196 0.000211  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.000177 0.000177 0.000187 0.000194 0.000196 0.000195 0.000196 0.000196

15 RS3 58  
 THE ORDERED VALUES S(I) ARE: 0.007550 0.007570 0.007580 0.007590 0.007590 0.007600 0.007610 0.007620 0.007620 0.007630  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.007610 0.007630 0.007620 0.007550 0.007590 0.007570 0.007600 0.007580

15 RS3 68  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 AS1 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 AS1 9  
 THE ORDERED VALUES S(I) ARE: 0.232300 0.232700 0.233000 0.233000 0.233000 0.233300 0.233400 0.234400 0.234500 0.234500  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.234400 0.233000 0.233000 0.234500 0.232300 0.232700 0.233300 0.233400

15 AS1 19  
 THE ORDERED VALUES S(I) ARE: 0.125300 0.125800 0.125900 0.126000 0.126000 0.126000 0.126400 0.126400 0.127600 0.128000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.126400 0.125800 0.128000 0.127600 0.126000 0.126000 0.125900 0.126000

15 AS1 29  
 THE ORDERED VALUES S(I) ARE: 1.286000 1.289000 1.291000 1.292000 1.295000 1.295000 1.299000 1.301000 1.303000 1.305000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.292000 1.305000 1.303000 1.301000 1.295000 1.289000 1.291000 1.299000

15 AS2 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 AS2 9  
 THE ORDERED VALUES S(I) ARE: 0.232700 0.233200 0.233400 0.233500 0.233900 0.233900 0.234000 0.234700 0.235400 0.235700  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.233200 0.233900 0.233900 0.235700 0.233400 0.233500 0.232700 0.234000

15 AS2 19  
 THE ORDERED VALUES S(I) ARE: 0.125500 0.125700 0.125800 0.125900 0.126100 0.126200 0.126700 0.126700 0.127000 0.127400  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.126100 0.125900 0.127000 0.126700 0.125800 0.126200 0.126700 0.127400

15 AS2 29  
 THE ORDERED VALUES S(I) ARE: 1.283000 1.288000 1.293000 1.295000 1.297000 1.297000 1.300000 1.300000 1.301000 1.301000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.288000 1.293000 1.301000 1.301000 1.283000 1.297000 1.300000 1.300000

15 AS3 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 AS3 9  
 THE ORDERED VALUES S(I) ARE: 0.232300 0.232800 0.233000 0.233100 0.233800 0.234100 0.234300 0.235000 0.235600 0.236600  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.234300 0.233100 0.233000 0.235000 0.236600 0.234100 0.233800 0.232800

15 AS3 19  
 THE ORDERED VALUES S(I) ARE: 0.125700 0.125700 0.126000 0.126600 0.126600 0.126600 0.126900 0.127300 0.127300 0.127700  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.127700 0.126600 0.125700 0.125700 0.126000 0.127300 0.126600 0.126600



15 AS3 29  
 THE ORDERED VALUES S(I) ARE: 1.282000 1.285000 1.286000 1.287000 1.288000 1.288000 1.288000 1.289000 1.289000 1.292000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.286000 1.287000 1.288000 1.282000 1.289000 1.288000 1.288000 1.289000

15 BS1 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 BS1 9  
 THE ORDERED VALUES S(I) ARE: 0.226000 0.228000 0.229000 0.230000 0.230000 0.231000 0.231000 0.231000 0.231000 0.232000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.228000 0.231000 0.229000 0.226000 0.230000 0.231000 0.232000 0.231000

15 BS1 19  
 THE ORDERED VALUES S(I) ARE: 0.123500 0.124000 0.124600 0.125200 0.125500 0.125700 0.126200 0.126500 0.126600 0.127100  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.124000 0.125700 0.126200 0.126500 0.124600 0.125500 0.123500 0.127100

15 BS1 29  
 THE ORDERED VALUES S(I) ARE: 1.263000 1.263000 1.264000 1.266000 1.266000 1.267000 1.270000 1.278000 1.279000 1.286000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.270000 1.286000 1.266000 1.279000 1.267000 1.264000 1.263000 1.278000

15 BS2 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 BS2 9  
 THE ORDERED VALUES S(I) ARE: 0.232100 0.232200 0.233800 0.234200 0.234500 0.235400 0.235400 0.235600 0.236000 0.236600  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.235400 0.232100 0.233800 0.234500 0.235600 0.236000 0.236600 0.234200

15 BS2 19  
 THE ORDERED VALUES S(I) ARE: 0.123300 0.124300 0.124800 0.125200 0.125300 0.125300 0.125600 0.126600 0.126800 0.127200  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.125200 0.125300 0.124300 0.124800 0.126800 0.127200 0.126600 0.125600

15 BS2 29  
 THE ORDERED VALUES S(I) ARE: 1.254000 1.265000 1.267000 1.267000 1.267000 1.274000 1.280000 1.283000 1.285000 1.294000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.267000 1.267000 1.265000 1.274000 1.283000 1.294000 1.285000 1.280000

15 BS3 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 BS3 9  
 THE ORDERED VALUES S(I) ARE: 0.230400 0.231000 0.232300 0.232500 0.233100 0.233300 0.233300 0.233800 0.236100 0.236800  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.236800 0.233800 0.232300 0.233100 0.233300 0.233300 0.231000 0.230400

15 BS3 19  
 THE ORDERED VALUES S(I) ARE: 0.125300 0.125500 0.125800 0.126100 0.126200 0.126900 0.127700 0.128000 0.128000 0.128400  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.128400 0.128000 0.127700 0.128000 0.126900 0.126100 0.125800 0.125500

15 BS3 29  
 THE ORDERED VALUES S(I) ARE: 1.271000 1.272000 1.273000 1.277000 1.278000 1.278000 1.281000 1.283000 1.286000 1.286000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.286000 1.281000 1.277000 1.283000 1.278000 1.271000 1.273000 1.272000

15 RS1 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 RS1 9  
 THE ORDERED VALUES S(I) ARE: 0.031900 0.032600 0.032800 0.032900 0.032900 0.032900 0.033100 0.033200 0.033500 0.034500

THE MAXIMUM VALUE S(I)= 0.034500 IS CANCELLED  
 THE MINIMUM VALUE S(I)= 0.031900 IS CANCELLED  
 THE RESULTING VALUES X(I) ARE 0.032900 0.032900 0.032800 0.033100 0.033200 0.032900 0.033500 0.032600  
 THE NUMBER OF OUTLIERS IS 2

15 RS1 19  
 THE ORDERED VALUES S(I) ARE: 0.001660 0.001840 0.001850 0.001950 0.002000 0.002110 0.002130 0.002190 0.002260 0.002380  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.002190 0.002130 0.002260 0.002110 0.002380 0.002000 0.001950 0.001850

15 RS1 29  
 THE ORDERED VALUES S(I) ARE: 1.145000 1.147000 1.147000 1.147000 1.154000 1.165000 1.166000 1.171000 1.179000 1.181000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.145000 1.154000 1.166000 1.179000 1.165000 1.147000 1.147000 1.147000

15 RS2 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 RS2 9  
 THE ORDERED VALUES S(I) ARE: 0.032300 0.032400 0.032500 0.032500 0.032700 0.032700 0.032800 0.032900 0.032900 0.033000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.032500 0.032700 0.032700 0.032900 0.033000 0.032300 0.032500 0.032400

15 RS2 19  
 THE ORDERED VALUES S(I) ARE: 0.001800 0.001800 0.001830 0.001830 0.001870 0.001910 0.001920 0.001920 0.001960 0.001970  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.001970 0.001870 0.001830 0.001920 0.001920 0.001910 0.001800 0.001800

15 RS2 29  
 THE ORDERED VALUES S(I) ARE: 1.163000 1.165000 1.166000 1.168000 1.168000 1.168000 1.169000 1.169000 1.171000 1.172000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.168000 1.165000 1.166000 1.169000 1.168000 1.172000 1.169000 1.168000

15 RS3 89  
 A SERIES OF X(I) WAS NOT MEASURED BY PURPOSE

15 RS3 9  
 THE ORDERED VALUES S(I) ARE: 0.032800 0.032800 0.032900 0.033000 0.033000 0.033000 0.033100 0.033100 0.033200 0.033200  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.033000 0.032900 0.032800 0.033000 0.033100 0.033100 0.033200 0.033200

15 RS3 19  
 THE ORDERED VALUES S(I) ARE: 0.001980 0.001990 0.002000 0.002020 0.002040 0.002050 0.002050 0.002060 0.002080 0.002100  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 0.002040 0.002050 0.001990 0.002100 0.002020 0.001980 0.002000 0.002080

15 RS3 29  
 THE ORDERED VALUES S(I) ARE: 1.144000 1.149000 1.150000 1.151000 1.153000 1.154000 1.154000 1.157000 1.161000 1.164000  
 THE REMAINING X(I) ARE TAKING THE FIRST 8 (NO OUTLIERS):  
 1.164000 1.157000 1.154000 1.149000 1.154000 1.161000 1.144000 1.151000

STANDARD EXPERIMENT - RESULTS OF PROGRAM LAB1  
 MEAN VALUE, VARIANCE, SD AND RSD FOR THE ISOTOPIC RATIOS OF EACH RUN  
 (8 SCANS/RUN WERE CONSIDERED)

LABORATORY CODE NO: 15 TOTAL NUMBER OF OUTLIERS FOR THIS LAB= 6  
 THE PU-241 VALUES OF THE UNSPIKED SAMPLES HAVE BEEN CORRECTED FOR AM-241 DECAY

ISOT.RATIO	SAMPLE	RUN	RUN-MEAN	SCAN-VARIANCE	SCAN-SD	SCAN-RSD(%)	NO.OF MEASUREMENTS	NO.OF OUTLIERS
83	US	1	0.22008750D-01	0.31412500D-07	0.17723572D-03	0.80529661D+00	10	0
83	US	2	0.21812500D-01	0.72650000D-07	0.26953664D-03	0.12356981D+01	10	0
83	US	3	0.21837500D-01	0.69642857D-08	0.83452296D-04	0.38215133D+00	10	0
92	PS	1	0.22012500D-03	0.80125000D-10	0.89512569D-05	0.40664427D+01	10	1
92	PS	2	0.21687500D-03	0.32098214D-09	0.17915975D-04	0.82609681D+01	10	0
92	PS	3	0.22900000D-03	0.19142857D-09	0.13835771D-04	0.60418216D+01	10	0
2	PS	1	0.84000000D-03	0.15142857D-08	0.38913824D-04	0.46325981D+01	10	0
2	PS	2	0.86362500D-03	0.28483929D-09	0.16877182D-04	0.19542258D+01	10	0
2	PS	3	0.84787500D-03	0.22755357D-09	0.15084879D-04	0.17791395D+01	10	0
38	AU	1	0.0	0.0	0.0	0.0	0	0
38	AU	2	0.0	0.0	0.0	0.0	0	0
38	AU	3	0.0	0.0	0.0	0.0	0	0
38	BU	1	0.0	0.0	0.0	0.0	0	0
38	BU	2	0.0	0.0	0.0	0.0	0	0
38	BU	3	0.0	0.0	0.0	0.0	0	0
38	RU	1	0.0	0.0	0.0	0.0	0	0
38	RU	2	0.0	0.0	0.0	0.0	0	0
38	RU	3	0.0	0.0	0.0	0.0	0	0
38	AS	1	0.64826250D+00	0.17655357D-05	0.13287346D-02	0.20496861D+00	10	0
38	AS	2	0.64150000D+00	0.11142857D-04	0.33380918D-02	0.52035726D+00	10	0
38	AS	3	0.64650000D+00	0.20000000D-05	0.14142136D-02	0.21874920D+00	10	0
38	BS	1	0.63650000D+00	0.71428571D-05	0.26726124D-02	0.41989197D+00	10	0
38	BS	2	0.63025000D+00	0.67857143D-05	0.26049404D-02	0.41331858D+00	10	0
38	BS	3	0.63450000D+00	0.29428571D-04	0.54248107D-02	0.85497411D+00	10	0
38	RS	1	0.75837500D+00	0.59821429D-05	0.24458420D-02	0.32251089D+00	10	0
38	RS	2	0.76225000D+00	0.47857143D-05	0.21876275D-02	0.28699607D+00	10	0
38	RS	3	0.76075000D+00	0.33571429D-05	0.18322508D-02	0.24084795D+00	10	0
48	AU	1	0.16925000D-03	0.27642857D-10	0.52576475D-05	0.31064387D+01	10	0
48	AU	2	0.17200000D-03	0.70285714D-10	0.83836576D-05	0.48742195D+01	10	0
48	AU	3	0.16887500D-03	0.23098214D-09	0.15198097D-04	0.89996131D+01	10	0
48	BU	1	0.18300000D-03	0.35600000D-09	0.18867962D-04	0.10310362D+02	10	0
48	BU	2	0.17750000D-03	0.22857143D-09	0.15118579D-04	0.85175093D+01	10	0
48	BU	3	0.17637500D-03	0.30267857D-10	0.55016231D-05	0.31192761D+01	10	1
48	RU	1	0.64750000D-04	0.56428571D-11	0.23756499D-05	0.36686793D+01	10	0
48	RU	2	0.54125000D-04	0.11155357D-09	0.10561892D-04	0.19513889D+02	10	0
48	RU	3	0.86375000D-04	0.14312500D-09	0.11963486D-04	0.13850635D+02	10	0
48	AS	1	0.28775000D-03	0.77335714D-09	0.27809300D-04	0.96643960D+01	10	0
48	AS	2	0.40950000D-03	0.97171429D-09	0.31172332D-04	0.76122911D+01	10	0
48	AS	3	0.29437500D-03	0.11283929D-09	0.10622584D-04	0.36085210D+01	10	0
48	BS	1	0.32037500D-03	0.43169643D-09	0.20777306D-04	0.64853080D+01	10	0
48	BS	2	0.28262500D-03	0.65696429D-10	0.81053333D-05	0.28678756D+01	10	0
48	BS	3	0.31225000D-03	0.29050000D-09	0.17044061D-04	0.54584661D+01	10	0
48	RS	1	0.29687500D-03	0.15241071D-09	0.12345473D-04	0.41584753D+01	10	0
48	RS	2	0.23500000D-03	0.95428571D-10	0.97687549D-05	0.41569170D+01	10	0
48	RS	3	0.18975000D-03	0.70785714D-10	0.84134246D-05	0.44339524D+01	10	0
58	AU	1	0.22500000D-01	0.28571429D-07	0.16903085D-03	0.75124823D+00	10	0
58	AU	2	0.22300000D-01	0.11428571D-06	0.33806170D-03	0.15159718D+01	10	0
58	AU	3	0.22375000D-01	0.27857143D-07	0.16690459D-03	0.74594231D+00	10	0
58	BU	1	0.22237500D-01	0.51250000D-07	0.22638463D-03	0.10180309D+01	10	0
58	BU	2	0.22312500D-01	0.26964286D-07	0.16420806D-03	0.73594647D+00	10	0
58	BU	3	0.22412500D-01	0.15535714D-07	0.12464235D-03	0.55612870D+00	10	0
58	RU	1	0.73487500D-02	0.25839286D-08	0.50832358D-04	0.69171434D+00	10	0
58	RU	2	0.73462500D-02	0.79125000D-08	0.88952234D-04	0.12108523D+01	10	0
58	RU	3	0.73850000D-02	0.65142857D-08	0.80711125D-04	0.10929062D+01	10	0

58	AS	1	0.224612500-01	0.135535710-08	0.368151750-04	0.163905280+00	10	0
58	AS	2	0.224625000-01	0.426785710-07	0.206587930-03	0.919701400+00	10	0
58	AS	3	0.222875000-01	0.355357140-07	0.188509190-03	0.845806790+00	10	0
58	BS	1	0.223500000-01	0.514285710-07	0.226778680-03	0.101466970+01	10	0
58	BS	2	0.222500000-01	0.314285710-07	0.177281050-03	0.796766770+00	10	0
58	BS	3	0.223500000-01	0.542857140-07	0.232992950-03	0.104247400+01	10	0
58	RS	1	0.769875000-02	0.172678570-08	0.415546110-04	0.539757900+00	10	0
58	RS	2	0.754500000-02	0.882857140-08	0.939604780-04	0.124533440+01	10	0
58	RS	3	0.759375000-02	0.712500000-09	0.266926960-04	0.351508750+00	10	0
68	AU	1	0.396375000-02	0.171250000-08	0.413823630-04	0.104402050+01	9	0
68	AU	2	0.396125000-02	0.449821430-08	0.670687280-04	0.169312030+01	10	0
68	AU	3	0.396500000-02	0.874285710-08	0.935032470-04	0.235821560+01	10	0
68	BU	1	0.393500000-02	0.242857140-08	0.492805380-04	0.125236440+01	10	0
68	BU	2	0.398625000-02	0.114107140-08	0.337797490-04	0.847406680+00	10	0
68	BU	3	0.395875000-02	0.583928570-09	0.241646140-04	0.610410210+00	10	0
68	RU	1	0.0	0.0	0.0	0.0	0	0
68	RU	2	0.0	0.0	0.0	0.0	0	0
68	RU	3	0.0	0.0	0.0	0.0	0	0
68	AS	1	0.397375000-02	0.154107140-08	0.392564830-04	0.987895130+00	10	0
68	AS	2	0.396125000-02	0.358392860-08	0.598659220-04	0.151128860+01	10	0
68	AS	3	0.392125000-02	0.783928570-09	0.279987240-04	0.714025490+00	10	0
68	BS	1	0.390750000-02	0.530714290-08	0.728501400-04	0.186436700+01	10	0
68	BS	2	0.399250000-02	0.119071430-07	0.109119860-03	0.273312100+01	10	0
68	BS	3	0.394500000-02	0.400000000-08	0.632455530-04	0.160318260+01	10	0
68	RS	1	0.111375000-03	0.905535710-10	0.951596400-05	0.854407540+01	10	0
68	RS	2	0.341500000-04	0.106257140-10	0.325971080-05	0.954527310+01	10	0
68	RS	3	0.0	0.0	0.0	0.0	0	0
89	AU	1	0.0	0.0	0.0	0.0	0	0
89	AU	2	0.0	0.0	0.0	0.0	0	0
89	AU	3	0.0	0.0	0.0	0.0	0	0
89	BU	1	0.0	0.0	0.0	0.0	0	0
89	BU	2	0.0	0.0	0.0	0.0	0	0
89	BU	3	0.0	0.0	0.0	0.0	0	0
89	RU	1	0.0	0.0	0.0	0.0	0	0
89	RU	2	0.0	0.0	0.0	0.0	0	0
89	RU	3	0.0	0.0	0.0	0.0	0	0
89	AS	1	0.0	0.0	0.0	0.0	0	0
89	AS	2	0.0	0.0	0.0	0.0	0	0
89	AS	3	0.0	0.0	0.0	0.0	0	0
89	BS	1	0.0	0.0	0.0	0.0	0	0
89	BS	2	0.0	0.0	0.0	0.0	0	0
89	BS	3	0.0	0.0	0.0	0.0	0	0
89	RS	1	0.0	0.0	0.0	0.0	0	0
89	RS	2	0.0	0.0	0.0	0.0	0	0
89	RS	3	0.0	0.0	0.0	0.0	0	0
9	AU	1	0.230112500+00	0.111267860-05	0.105483580-02	0.458400040+00	10	0
9	AU	2	0.230825000+00	0.576428570-06	0.759228930-03	0.328919720+00	10	0
9	AU	3	0.231837500+00	0.284839290-05	0.168771820-02	0.727974650+00	10	0
9	BU	1	0.230275000+00	0.123357140-05	0.111066260-02	0.482320100+00	10	0
9	BU	2	0.230925000+00	0.152500000-05	0.123490890-02	0.534766220+00	10	0
9	BU	3	0.230287500+00	0.449821430-06	0.670687280-03	0.291239120+00	10	0
9	RU	1	0.265125000-01	0.926785710-07	0.304431550-03	0.114825670+01	10	0
9	RU	2	0.264125000-01	0.526785710-07	0.229518130-03	0.868975400+00	10	0
9	RU	3	0.260875000-01	0.355357140-07	0.188509190-03	0.722603500+00	10	0
9	AS	1	0.233325000+00	0.599285710-06	0.774135460-03	0.331784190+00	10	0
9	AS	2	0.233787500+00	0.784107140-06	0.885498250-03	0.378762010+00	10	0
9	AS	3	0.234087500+00	0.158410710-05	0.125861320-02	0.537667830+00	10	0
9	BS	1	0.229750000+00	0.392857140-05	0.198206240-02	0.862703990+00	10	0
9	BS	2	0.234775000+00	0.205928570-05	0.143502120-02	0.611232520+00	10	0
9	BS	3	0.233000000+00	0.378857140-05	0.194642530-02	0.835375660+00	10	0
9	RS	1	0.329875000-01	0.755357140-07	0.274837610-03	0.833156850+00	10	2
9	RS	2	0.326250000-01	0.592857140-07	0.243486580-03	0.746319020+00	10	0
9	RS	3	0.330375000-01	0.198214290-07	0.140788600-03	0.426147850+00	10	0
19	AU	1	0.127731790+00	0.781588550-06	0.884074970-03	0.692133850+00	10	0
19	AU	2	0.127003040+00	0.834271810-06	0.913384810-03	0.719183400+00	10	0
19	AU	3	0.127983080+00	0.176019800-05	0.132672460-02	0.103664060+01	10	0

19	BU	1	0.12727947D+00	0.93169974D-06	0.96524595D-03	0.75836738D+00	10	0
19	BU	2	0.12763127D+00	0.45249863D-06	0.67268019D-03	0.52704965D+00	10	0
19	BU	3	0.12672662D+00	0.52466748D-06	0.72433934D-03	0.57157630D+00	10	0
19	RU	1	0.78488539D-03	0.27139836D-09	0.16474173D-04	0.20989271D+01	10	0
19	RU	2	0.76195500D-03	0.39459757D-09	0.19864480D-04	0.26070411D+01	10	0
19	RU	3	0.79911696D-03	0.60969830D-09	0.24692070D-04	0.30899194D+01	10	0
19	AS	1	0.12724406D+00	0.73163868D-06	0.85535880D-03	0.67221907D+00	10	0
19	AS	2	0.12725663D+00	0.32324437D-06	0.56854584D-03	0.44677108D+00	10	0
19	AS	3	0.12730694D+00	0.52572406D-06	0.72506831D-03	0.56954342D+00	10	0
19	BS	1	0.12622588D+00	0.15995707D-05	0.12647414D-02	0.10019668D+01	10	0
19	BS	2	0.12672488D+00	0.10820070D-05	0.10401957D-02	0.82082988D+00	10	0
19	BS	3	0.12789950D+00	0.12942787D-05	0.11376637D-02	0.88949819D+00	10	0
19	RS	1	0.21239177D-02	0.30069007D-07	0.17340417D-03	0.81643545D+01	10	0
19	RS	2	0.18910044D-02	0.39925352D-08	0.63186511D-04	0.33414259D+01	10	0
19	RS	3	0.20471193D-02	0.18767089D-08	0.43320999D-04	0.21161932D+01	10	0
29	AU	1	0.22950000D-01	0.31428571D-07	0.17728105D-03	0.77246646D+00	10	0
29	AU	2	0.23000000D-01	0.11142857D-06	0.33380918D-03	0.14513443D+01	10	1
29	AU	3	0.23212500D-01	0.19267857D-06	0.43895167D-03	0.18910142D+01	10	0
29	BU	1	0.22987500D-01	0.86964286D-07	0.29489708D-03	0.12828584D+01	10	0
29	BU	2	0.22925000D-01	0.33571429D-07	0.18322508D-03	0.79923697D+00	10	0
29	BU	3	0.23037500D-01	0.14107143D-07	0.11877349D-03	0.51556590D+00	10	1
29	RU	1	0.22625000D-04	0.34267857D-10	0.58538754D-05	0.25873482D+02	10	0
29	RU	2	0.47375000D-04	0.56553571D-10	0.75202109D-05	0.15873796D+02	10	0
29	RU	3	0.78625000D-04	0.31982143D-10	0.56552757D-05	0.71927194D+01	10	0
29	AS	1	0.12968750D+01	0.35553571D-04	0.59626816D-02	0.45977304D+00	10	0
29	AS	2	0.12953750D+01	0.45982143D-04	0.67810134D-02	0.52347879D+00	10	0
29	AS	3	0.12871250D+01	0.52678571D-05	0.22951813D-02	0.17831845D+00	10	0
29	BS	1	0.12716250D+01	0.69982143D-04	0.83655330D-02	0.65786164D+00	10	0
29	BS	2	0.12768750D+01	0.10726786D-03	0.10357020D-01	0.81112244D+00	10	0
29	BS	3	0.12776250D+01	0.29696429D-04	0.54494430D-02	0.42652914D+00	10	0
29	RS	1	0.11562500D+01	0.15392857D-03	0.12406795D-01	0.10730201D+01	10	0
29	RS	2	0.11681250D+01	0.44107143D-05	0.21001701D-02	0.17978984D+00	10	0
29	RS	3	0.11542500D+01	0.41642857D-04	0.64531277D-02	0.55907539D+00	10	0

## RESULTS OF PROGRAM LAB2 FOR THE STANDARD EXPERIMENT

FOR EACH LAB AND FOR EACH SAMPLE AND ISOTOPIIC RATIO THE FOLLOWING VALUES ARE GIVEN:  
 LAB MEAN(MEAN) WITH VARIANCE(SG2),STANDARD DEVIATION(SDG) AND RELATIVE STANDARD DEVIATION(RSDG).  
 IN ADDITION VARIANCE COMPONENTS FOR RUN AND SCAN ARE GIVEN SEPARATELY .

(IF THE RUN OR SCAN COMPONENT OF ANY VARIANCE IS NEGATIVE AND THEREFORE HAS TO BE CONSIDERED AS NEGLIGIBLE,  
 IT IS PUT = -5.0 (RUN) OR = -4.0 (SCAN),RESPECTIVELY.)

LABORATORY CODE NO.: 15

RAT	SAMP	LAB	MEAN	VARIANCE	STAND. DEV.	RSD(%)	VARIANCE	SD	RSD(%)	VARIANCE	SD	RSD(%)
			LAB MEAN	LAB MEAN	LAB MEAN	LAB MEAN	RUN	RUN	RUN	SCAN	SCAN	SCAN
83	US	0.	2189D-01	0.3804D-08	0.617D-04	0.282D+00	0.6785D-0E	0.224D-04	0.376D+00	0.3701D-07	0.192D-03	0.879D+00
38	AU	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
38	RU	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
38	RS	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
38	AS	0.	6454D+00	0.4102D-05	0.203D-02	0.314D+00	0.1169D-04	0.342D-02	0.530D+00	0.4969D-05	0.223D-02	0.345D+00
38	BS	0.	6337D+00	0.3396D-05	0.184D-02	0.291D+00	0.8381D-05	0.289D-02	0.457D+00	0.1445D-04	0.380D-02	0.600D+00
38	RS	0.	7605D+00	0.1273D-05	0.113D-02	0.148D+00	0.3229D-05	0.180D-02	0.236D+00	0.4708D-05	0.217D-02	0.285D+00
48	AU	0.	1700D-03	0.9705D-12	0.985D-06	0.579D+00	-0.1079D-10	-0.500D+01	-0.500D+01	0.1096D-09	0.105D-04	0.616D+01
48	BU	0.	1790D-03	0.4189D-11	0.205D-05	0.114D+01	-0.1305D-10	-0.500D+01	-0.500D+01	0.2049D-09	0.143D-04	0.800D+01
48	RU	0.	6842D-04	0.9003D-10	0.949D-05	0.139D+02	0.2593D-09	0.161D-04	0.235D+02	0.8677D-10	0.932D-05	0.136D+02
48	AS	0.	3305D-03	0.1562D-08	0.395D-04	0.120D+02	0.4609D-08	0.679D-04	0.205D+02	0.6193D-09	0.249D-04	0.753D+01
48	RS	0.	3051D-03	0.1316D-09	0.115D-04	0.376D+01	0.3620D-09	0.190D-04	0.624D+01	0.2626D-09	0.162D-04	0.531D+01
48	BS	0.	2405D-03	0.9640D-09	0.310D-04	0.129D+02	0.2879D-08	0.537D-04	0.223D+02	0.1062D-09	0.103D-04	0.428D+01
58	AU	0.	2230D-01	0.3403D-08	0.583D-04	0.261D+00	0.3095D-08	0.556D-04	0.248D+00	0.5690D-07	0.239D-03	0.107D+01
58	BU	0.	2232D-01	0.2560D-08	0.507D-04	0.227D+00	0.3802D-08	0.617D-04	0.276D+00	0.3125D-07	0.177D-03	0.792D+00
58	RU	0.	7360D-02	0.1568D-09	0.125D-04	0.170D+00	-0.2385D-09	-0.500D+01	-0.500D+01	0.5670D-08	0.793D-04	0.102D+01
58	AS	0.	2240D-01	0.3379D-08	0.581D-04	0.259D+00	0.6821D-08	0.826D-04	0.369D+00	0.2652D-07	0.163D-03	0.727D+00
58	BS	0.	2232D-01	0.1111D-08	0.333D-04	0.149D+00	-0.2381D-08	-0.500D+01	-0.500D+01	0.4571D-07	0.214D-03	0.958D+00
58	RS	0.	7612D-02	0.2058D-08	0.454D-04	0.596D+00	0.5704D-08	0.755D-04	0.992D+00	0.3756D-08	0.613D-04	0.805D+00
68	AU	0.	3963D-02	0.1215D-11	0.110D-05	0.278D-01	-0.6194D-09	-0.500D+01	-0.500D+01	0.4985D-08	0.706D-04	0.178D+01
68	BU	0.	3960D-02	0.2193D-09	0.148D-04	0.374D+00	0.4847D-09	0.220D-04	0.556D+00	0.1385D-08	0.372D-04	0.940D+00
68	RU	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
68	AS	0.	3552D-02	0.2567D-09	0.158D-04	0.401D+00	0.5056D-05	0.225D-04	0.569D+00	0.1970D-08	0.444D-04	0.112D+01
68	BS	0.	3548D-02	0.6049D-09	0.246D-04	0.623D+00	0.9307D-09	0.305D-04	0.773D+00	0.7071D-08	0.841D-04	0.213D+01
68	RS	0.	7276D-04	0.1491D-08	0.386D-04	0.531D+02	0.2931D-08	0.541D-04	0.744D+02	0.4068D-09	0.202D-04	0.277D+02
92	PS	0.	2220D-03	0.1313D-10	0.362D-05	0.163D+01	0.1470D-10	0.383D-05	0.173D+01	0.1975D-09	0.141D-04	0.633D+01
2	PS	0.	8505D-03	0.4823D-10	0.695D-05	0.817D+00	0.6026D-10	0.776D-05	0.913D+00	0.6756D-09	0.260D-04	0.306D+01
89	AU	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
89	BU	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
89	RU	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
89	AS	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
89	BS	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
89	RS	0.	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
9	AU	0.	2309D+00	0.2505D-06	0.500D-03	0.217D+00	0.5623D-06	0.750D-03	0.325D+00	0.1512D-05	0.123D-02	0.533D+00
9	BU	0.	2305D+00	0.4606D-07	0.215D-03	0.931D-01	0.4494D-08	0.670D-04	0.291D-01	0.1069D-05	0.103D-02	0.449D+00
9	RU	0.	2634D-01	0.1646D-07	0.128D-03	0.487D+00	0.4184D-07	0.205D-03	0.777D+00	0.6030D-07	0.246D-03	0.932D+00
9	AS	0.	2337D+00	0.4918D-07	0.222D-03	0.949D-01	0.2391D-07	0.155D-03	0.662D-01	0.9892D-06	0.995D-03	0.426D+00
9	BS	0.	2225D+00	0.2165D-05	0.147D-02	0.633D+00	0.6087D-05	0.247D-02	0.106D+01	0.3259D-05	0.181D-02	0.776D+00
9	RS	0.	3288D-01	0.1689D-07	0.130D-03	0.395D+00	0.4423D-07	0.210D-03	0.640D+00	0.5155D-07	0.227D-03	0.690D+00
19	AU	0.	1276D+00	0.8637D-07	0.294D-03	0.230D+00	0.1184D-06	0.344D-03	0.270D+00	0.1125D-05	0.106D-02	0.832D+00
19	BU	0.	1272D+00	0.6932D-07	0.263D-03	0.207D+00	0.1284D-06	0.358D-03	0.282D+00	0.6363D-06	0.798D-03	0.627D+00
19	RU	0.	7820D-03	0.1172D-09	0.108D-04	0.138D+01	0.2984D-09	0.173D-04	0.221D+01	0.4252D-09	0.206D-04	0.264D+01
19	AS	0.	1273D+00	0.3691D-09	0.192D-04	0.151D-01	-0.6475D-07	-0.500D+01	-0.500D+01	0.5269D-06	0.726D-03	0.570D+00
19	BS	0.	1270D+00	0.2461D-06	0.496D-03	0.391D+00	0.5726D-06	0.757D-03	0.596D+00	0.1325D-05	0.115D-02	0.907D+00
19	RS	0.	2021D-02	0.4695D-08	0.685D-04	0.339D+01	0.1259D-07	0.112D-03	0.555D+01	0.1198D-07	0.109D-03	0.542D+01
29	AU	0.	2305D-01	0.6476D-08	0.805D-04	0.349D+00	0.5446D-08	0.738D-04	0.320D+00	0.1118D-06	0.334D-03	0.145D+01
29	BU	0.	2298D-01	0.1059D-08	0.325D-04	0.142D+00	-0.2433D-08	-0.500D+01	-0.500D+01	0.4488D-07	0.212D-03	0.922D+00
29	RU	0.	4954D-04	0.2625D-09	0.162D-04	0.327D+02	0.7824D-09	0.280D-04	0.565D+02	0.4093D-10	0.640D-05	0.129D+02
29	AS	0.	1293D+01	0.9187D-05	0.303D-02	0.234D+00	0.2395D-04	0.489D-02	0.378D+00	0.2893D-04	0.538D-02	0.416D+00
29	BS	0.	1275D+01	0.3562D-05	0.189D-02	0.148D+00	0.2065D-05	0.144D-02	0.113D+00	0.6898D-04	0.831D-02	0.651D+00
29	RS	0.	1160D+01	0.1875D-04	0.433D-02	0.373D+00	0.4792D-04	0.692D-02	0.597D+00	0.6666D-04	0.816D-02	0.704D+00

## RESULTS OF PROGRAM LAB3 FOR THE STANDARD EXPERIMENT

LABORATORY CODE NO.: 15

## ISOTOPIC COMPOSITION IN ATOM PERCENTAGES

URANIUM SAMPLE	ATU233	ATU234	ATU235	ATU236	ATU238
AU	0.0	0.16565D-01	0.21813D+01	0.38609D+00	0.97416D+02
BU	0.0	0.17435D-01	0.21745D+01	0.38579D+00	0.97422D+02
RU	0.0	0.67912D-02	0.73057D+00	0.0	0.99263D+02
AS	0.38599D+02	0.19768D-01	0.13399D+01	0.23635D+00	0.59805D+02
BS	0.38170D+02	0.18375D-01	0.13441D+01	0.23781D+00	0.60229D+02
RS	0.43003D+02	0.13602D-01	0.43048D+00	0.41146D-02	0.56549D+02

PLUTONIUM SAMPLE	ATP238	ATP239	ATP240	ATP241	ATP242
AU	0.0	0.72382D+02	0.16715D+02	0.92340D+01	0.16687D+01
BU	0.0	0.72427D+02	0.16694D+02	0.92137D+01	0.16646D+01
RU	0.0	0.97355D+02	0.25641D+01	0.76130D-01	0.48231D-02
AS	0.0	0.37677D+02	0.88064D+01	0.47951D+01	0.48721D+02
BS	0.0	0.37953D+02	0.88244D+01	0.48181D+01	0.48404D+02
RS	0.0	0.45570D+02	0.14985D+01	0.92082D-01	0.52840D+02

## RESULTS OF SUBROUTINE ALPHA

LABORATORY CODE NO.: 15

SAMPLE	ALPHA-ACTIVITY MEAN VALUE	VARIANCE OF SINGLE MEASUREMENT	SD	RSD(%)	RSD(%) OF MEAN	RATIO 89
AU	0.22032D+01	0.14303D-04	0.3782D-02	0.1717D+00	0.9911D-01	0.1470D-01
BU	0.22178D+01	0.25056D-02	0.5006D-01	0.2257D+01	0.1303D+01	0.1478D-01
RU	0.10140D-01	0.12000D-06	0.3464D-03	0.3416D+01	0.1972D+01	0.4001D-04
AS	0.21636D+01	0.41429D-02	0.6437D-01	0.2975D+01	0.1718D+01	0.1452D-01
BS	0.21719D+01	0.57173D-02	0.7561D-01	0.3481D+01	0.2010D+01	0.1454D-01
FOR SAMPLE RS	NO ALPHA MEASUREMENTS; GIVEN VALUES ARE DUMMIES					
RS	0.99000D+01	0.99000D+01	0.9900D+01	0.9900D+01	0.9900D+01	0.3992D-01

## ISOTOPIC COMPOSITION IN ATOM PERCENTAGES, CALCULATED IN PROGRAM LAB3 INCLUDING ALPHA RESULTS

PLUTONIUM SAMPLE	ATAL38	ATAL39	ATAL40	ATAL41	ATAL42
AU	0.10528D+01	0.71620D+02	0.16539D+02	0.91368D+01	0.16511D+01
BU	0.10595D+01	0.71660D+02	0.16517D+02	0.91161D+01	0.16470D+01
RU	0.38946D-02	0.97351D+02	0.25640D+01	0.76127D-01	0.48229D-02
AS	0.54398D+00	0.37472D+02	0.87585D+01	0.47691D+01	0.48456D+02
BS	0.54869D+00	0.37745D+02	0.87760D+01	0.47917D+01	0.48139D+02
RS	0.17867D+01	0.44755D+02	0.14717D+01	0.90436D-01	0.51896D+02

## ISOTOPIC COMPOSITION IN WEIGHT PERCENTAGES

URANIUM SAMPLE	GU233	GU234	GU235	GU236	GU238
AU	0.0	0.16291D-01	0.21544D+01	0.38296D+00	0.97446D+02
BU	0.0	0.17146D-01	0.21477D+01	0.38266D+00	0.97452D+02
RU	0.0	0.66774D-02	0.72141D+00	0.0	0.99272D+02
AS	0.38104D+02	0.19598D-01	0.13340D+01	0.23633D+00	0.60306D+02
BS	0.37677D+02	0.18215D-01	0.13382D+01	0.23776D+00	0.60729D+02
RS	0.42485D+02	0.13496D-01	0.42895D+00	0.41175D-02	0.57069D+02

PLUTONIUM SAMPLE	GP238	GP239	GP240	GP241	GP242
AU	0.0	0.72261D+02	0.16757D+02	0.92958D+01	0.16869D+01
BU	0.0	0.72306D+02	0.16736D+02	0.92753D+01	0.16827D+01
RU	0.0	0.97344D+02	0.25745D+01	0.76760D-01	0.48832D-02
AS	0.0	0.37419D+02	0.87827D+01	0.48022D+01	0.48996D+02
BS	0.0	0.37694D+02	0.88010D+01	0.48254D+01	0.48679D+02
RS	0.0	0.45266D+02	0.14947D+01	0.92234D-01	0.53147D+02

ISOTOPIC COMPOSITION IN WEIGHT PERCENTAGES, CALCULATED IN PROGRAM LAB3  
INCLUDING ALPHA RESULTS

PLUTONIUM SAMPLE	GPAL 238	GPAL 239	GPAL 240	GPAL 241	GPAL 242
AU	0.10467D+01	0.71504D+02	0.16581D+02	0.91985D+01	0.16692D+01
BU	0.10533D+01	0.71544D+02	0.16560D+02	0.91776D+01	0.16650D+01
RU	0.38778D-02	0.97340D+02	0.25744D+01	0.76757D-01	0.48830D-02
AS	0.53802D+00	0.37218D+02	0.87355D+01	0.47764D+01	0.48732D+02
BS	0.54269D+00	0.37490D+02	0.87532D+01	0.47992D+01	0.48415D+02
RS	0.17677D+01	0.44465D+02	0.14683D+01	0.90604D-01	0.52208D+02



RESULTS OF PROGRAM LAB4  
LABORATORY CODE NO.: 15

## URANIUM CONCENTRATIONS

## U-238 CONCENTRATIONS IN ATOMS/G SOLUTION

	SAMPLE A	SAMPLE B	SAMPLE R	SAMPLE A-CALIBRATED	SAMPLE B-CALIBRATED
NUMBER OF ATOMS 1	0.28753D+19	0.27613D+19	0.27546D+19	0.29081D+19	0.27928D+19
NUMBER OF ATOMS 2	0.29061D+19	0.27891D+19	0.27404D+19	0.29545D+19	0.28355D+19
NUMBER OF ATOMS 3	0.28833D+19	0.27701D+19	0.27459D+19	0.29254D+19	0.28106D+19
NO. OF ATOMS 1-3	0.28882D+19	0.27735D+19	0.27469D+19	0.29293D+19	0.28129D+19
RSD OF MEAN	0.31903D+00	0.29536D+00	0.15094D+00		

## URANIUM CONCENTRATIONS IN MG URANIUM/G SOLUTION

C : 1ST RUN	0.11664D+01	0.11201D+01	0.10969D+01	0.11797D+01	0.11328D+01
C : 2ND RUN	0.11789D+01	0.11313D+01	0.10912D+01	0.11985D+01	0.11502D+01
C : 3RD RUN	0.11696D+01	0.11237D+01	0.10934D+01	0.11867D+01	0.11401D+01
CONCENTRATION 1-3	0.11716D+01	0.11250D+01	0.10938D+01	0.11903D+01	0.11424D+01
SIGMA	0.13972D-04	0.11042D-04	0.27258D-05		
STANDARD DEVIATION	0.37379D-02	0.33229D-02	0.16510D-02		
RSD OF MEAN	0.31903D+00	0.29536D+00	0.15094D+00		

FOR LAB NO. 15 AND SAMPLE AU THE RATIO PU 238/239 WAS NOT MEASURED BY MASS SPECTROMETRY.  
FOR LAB NO. 15 AND SAMPLE BU THE RATIO PU 238/239 WAS NOT MEASURED BY MASS SPECTROMETRY.  
FOR LAB NO. 15 AND SAMPLE RU THE RATIO PU 238/239 WAS NOT MEASURED BY MASS SPECTROMETRY.

## PLUTONIUM CONCENTRATIONS

## PU-239 CONCENTRATIONS IN ATOMS/G SOLUTION

	SAMPLE A	SAMPLE B	SAMPLE R	SAMPLE A-CALIBRATED	SAMPLE B-CALIBRATED
NUMBER OF ATOMS 1	0.17138D+17	0.16487D+17	0.21254D+17	0.17014D+17	0.16368D+17
NUMBER OF ATOMS 2	0.17159D+17	0.16417D+17	0.21035D+17	0.17212D+17	0.16468D+17
NUMBER OF ATOMS 3	0.17273D+17	0.16407D+17	0.21292D+17	0.17117D+17	0.16259D+17
NO. OF ATOMS 1-3	0.17190D+17	0.16437D+17	0.21193D+17	0.17114D+17	0.16365D+17
RSD OF MEAN	0.24332D+00	0.15350D+00	0.37806D+00		

## PLUTONIUM CONCENTRATIONS IN MICROGRAMS PLUTONIUM/G SOLUTION

C : 1ST RUN	0.94148D+01	0.90515D+01	0.86673D+01	0.93466D+01	0.89859D+01
C : 2ND RUN	0.94261D+01	0.90130D+01	0.85778D+01	0.94554D+01	0.90410D+01
C : 3RD RUN	0.94887D+01	0.90075D+01	0.86826D+01	0.94034D+01	0.89265D+01
CONCENTRATION 1-3	0.94431D+01	0.90240D+01	0.86423D+01	0.93566D+01	0.90008D+01
SIGMA	0.52794D-03	0.19187D-03	0.10676D-02		
STANDARD DEVIATION	0.22977D-01	0.13852D-01	0.32674D-01		
RSD OF MEAN	0.24332D+00	0.15350D+00	0.37806D+00		

## PLUTONIUM CONCENTRATIONS WITH INCLUSION OF THE RESULTS OF ALPHA SPECTROMETRY

## PU-239 CONCENTRATIONS IN ATOMS/G SOLUTION

	SAMPLE A	SAMPLE B	SAMPLE R	SAMPLE A-CALIBRATED	SAMPLE B-CALIBRATED
NUMBER OF ATOMS 1	0.17138D+17	0.16487D+17	0.21254D+17	0.17014D+17	0.16368D+17
NUMBER OF ATOMS 2	0.17159D+17	0.16417D+17	0.21035D+17	0.17212D+17	0.16468D+17
NUMBER OF ATOMS 3	0.17273D+17	0.16407D+17	0.21292D+17	0.17117D+17	0.16259D+17
NO. OF ATOMS 1-3	0.17190D+17	0.16437D+17	0.21193D+17	0.17114D+17	0.16365D+17
RSD OF MEAN	0.24332D+00	0.15350D+00	0.37806D+00		

## PLUTONIUM CONCENTRATIONS IN MICROGRAMS PLUTONIUM/G SOLUTION

C :	1ST RUN	0.95144D+01	0.91479D+01	0.86676D+01	0.94454D+01	0.90815D+01
C :	2ND RUN	0.95258D+01	0.91089D+01	0.85781D+01	0.95554D+01	0.91372D+01
C :	3RD RUN	0.95891D+01	0.91034D+01	0.86829D+01	0.95028D+01	0.90215D+01
	CONCENTRATION 1-3	0.95430D+01	0.91200D+01	0.86426D+01	0.94555D+01	0.90967D+01
	SIGMA	0.53917D-03	0.19598D-03	0.10677D-02		
	STANDARD DEVIATION	0.23220D-01	0.13999D-01	0.32676D-01		
	RSD OF MEAN	0.24332D+00	0.15350D+00	0.37806D+00		

## PU/U-RATIOS

	SAMPLE A	SAMPLE B	SAMPLE R
MS-MEASUREMENTS	0.80599D+01	0.80212D+01	0.79010D+01
ALPHA-MEASUREMENTS	0.81452D+01	0.81066D+01	0.79013D+01

(IF ANY OF THESE RATIOS COULD NOT BE CALCULATED, IT IS PUT = 0.)

12.2 Results of One Laboratory for the Dry Spike  
and Aluminium Capsule Experiments

## RESULTS OF THE SUBROUTINE DRY FOR THE LAB WITH CODE NO.: 15

0.592900	0.593800	0.595300	0.594900	0.592800	0.594200	0.593700	0.591600	0.592100	0.594200	1.169430	1.401480	1
0.595200	0.595800	0.594500	0.593300	0.594700	0.593400	0.593200	0.594400	0.594600	0.594400	1.169430	1.401480	1
0.589300	0.590900	0.592600	0.594900	0.596000	0.593200	0.592900	0.598100	0.599700	0.599800	1.169430	1.401480	1
0.592000	0.592700	0.592000	0.592300	0.591200	0.591500	0.590400	0.589200	0.588100	0.587400	1.160090	1.397150	2
0.593900	0.592800	0.593300	0.590500	0.591800	0.597100	0.594100	0.594300	0.593900	0.594200	1.160090	1.397150	2
0.588600	0.587200	0.589300	0.589800	0.588900	0.590200	0.591200	0.588400	0.592900	0.588100	1.160090	1.397150	2
1.200000	1.201000	1.208000	1.197000	1.190000	1.199000	1.202000	1.197000	1.196000	1.200000	1.169430	1.401480	2
1.195000	1.198000	1.194000	1.198000	1.201000	1.199000	1.199000	1.201000	1.201000	1.198000	1.169430	1.401480	1
1.197000	1.193000	1.189000	1.194000	1.194000	1.196000	1.196000	1.200000	1.196000	1.191000	1.169430	1.401480	1
1.190000	1.184000	1.180000	1.188000	1.190000	1.192000	1.193000	1.196000	1.194000	1.192000	1.160090	1.397150	2
1.190000	1.190000	1.189000	1.191000	1.193000	1.195000	1.193000	1.192000	1.191000	1.190000	1.160090	1.397150	2
1.193000	1.194000	1.201000	1.196000	1.197000	1.194000	1.194000	1.200000	1.193000	1.190000	1.160090	1.397150	2

RUN AND LAB MEANS, VALUES OF THE DRY SPIKE EXPERIMENT  
 IN THE ORDER: SAMPLE 1-U233/238, SAMPLE 2-U233/238, SAMPLE 1-PU242/239, SAMPLE 2-PU242/239  
 VARIANCE ETC. OF SINGLE MEASUREMENTS ARE GIVEN

RUN NO.	RUN MEAN	SIGMA	SD	RSD
1	0.59365D+00	0.144D-05	0.120D-02	0.202D+00
2	0.59431D+00	0.898D-06	0.948D-03	0.159D+00
3	0.59349D+00	0.786D-05	0.280D-02	0.472D+00
1	0.59141D+00	0.130D-05	0.114D-02	0.193D+00
2	0.59307D+00	0.176D-05	0.133D-02	0.223D+00
3	0.58920D+00	0.149D-05	0.122D-02	0.207D+00
1	0.11990D+01	0.457D-05	0.214D-02	0.178D+00
2	0.11981D+01	0.641D-05	0.253D-02	0.211D+00
3	0.11949D+01	0.104D-04	0.323D-02	0.270D+00
1	0.11891D+01	0.261D-04	0.511D-02	0.430D+00
2	0.11916D+01	0.398D-05	0.200D-02	0.167D+00
3	0.11961D+01	0.898D-05	0.300D-02	0.251D+00

RESULTS OF THE STANDARD EXPERIMENT FOR COMPARISON  
SAMPLE A

## FOR URANIUM

NUMBER OF ATOMS 1	0.28753D+19
NUMBER OF ATOMS 2	0.29061D+19
NUMBER OF ATOMS 3	0.28833D+19

## FOR PLUTONIUM

NUMBER OF ATOMS 1	0.17138D+17
NUMBER OF ATOMS 2	0.17159D+17
NUMBER OF ATOMS 3	0.17273D+17

## RESULTS OF THE DRY SPIKE EXPERIMENT

	SAMPLE A-1	SAMPLE A-2
URANIUM RESULTS:		
NUMBER OF ATOMS 1	0.29038D+19	0.29006D+19
NUMBER OF ATOMS 2	0.29005D+19	0.28924D+19
NUMBER OF ATOMS 3	0.29046D+19	0.29116D+19

PLUTONIUM RESULTS:		
NUMBER OF ATOMS 1	0.17171D+17	0.17234D+17
NUMBER OF ATOMS 2	0.17184D+17	0.17196D+17
NUMBER OF ATOMS 3	0.17232D+17	0.17129D+17

CALCULATION ON THE BASIS OF THE THREE DIFFERENT ISOTOPIIC RATIOS MEASURED IN THE DRY SPIKE EXPERIMENT  
MEAN NUMBER OF ATOMS

FOR URANIUM:	0.29030D+19	0.29015D+19
FOR PLUTONIUM:	0.17100D+17	0.17090D+17

CALCULATION ON THE BASIS OF THREE RESULTING NUMBERS OF ATOMS CALCULATED FOR THE DRY SPIKE EXPERIMENT  
MEAN NUMBER OF ATOMS

	SAMPLE A-1	SD(MEAN)	RSD(MEAN)	SAMPLE A-2	SD(MEAN)	RSD(MEAN)
URANIUM	0.29030D+19	0.25232D-03	0.42490D-01	0.29015D+19	0.11224D-02	0.18984D+00
PLUTONIUM	0.17196D+17	0.12549D-02	0.10480D+00	0.17186D+17	0.20480D-02	0.17177D+00

## RESULTS OF THE SUBROUTINE DRY FOR THE LAB WITH CODE NO.: 15

## ALUMINIUM CAPSULE EXPERIMENT

0.696400	0.694600	0.695500	0.697600	0.697600	0.695800	0.694100	0.694300	0.697000	0.697800	1.128250	1.147090	1
0.698700	0.698200	0.699000	0.698300	0.701100	0.701200	0.700500	0.702300	0.695700	0.698000	1.128250	1.147090	1
0.696200	0.694300	0.694500	0.695700	0.695400	0.697800	0.695800	0.698700	0.699200	0.701100	1.128250	1.147090	1
0.687900	0.684900	0.684900	0.682600	0.684700	0.685700	0.684500	0.682000	0.682900	0.680600	1.131580	1.172000	2
0.684100	0.679600	0.683000	0.678200	0.679800	0.682000	0.684700	0.683800	0.684200	0.683200	1.131580	1.172000	2
0.682600	0.688700	0.693600	0.687100	0.687000	0.687600	0.687900	0.684400	0.685900	0.690400	1.131580	1.172000	2
1.411100	1.409200	1.408100	1.408100	1.402200	1.409000	1.408500	1.408100	1.403400	1.403900	1.128250	1.147090	1
1.387700	1.385900	1.399900	1.407000	1.401300	1.393400	1.402100	1.398100	1.395400	1.407300	1.128250	1.147090	1
1.402600	1.402400	1.403300	1.401300	1.406800	1.406600	1.408900	1.406300	1.401400	1.398200	1.128250	1.147090	1
1.376900	1.375100	1.372900	1.385100	1.382500	1.389800	1.385700	1.386400	1.390500	1.388000	1.131580	1.172000	2
1.389300	1.395100	1.382800	1.383400	1.389400	1.387300	1.386500	1.387200	1.393800	1.389600	1.131580	1.172000	2
1.393100	1.389000	1.384700	1.373500	1.380000	1.384300	1.370800	1.381400	1.385300	1.383600	1.131580	1.172000	2

RUN AND LAB MEANS, VALUES OF THE DRY SPIKE EXPERIMENT  
IN THE ORDER: SAMPLE 1-U233/238, SAMPLE 2-U233/238, SAMPLE 1-PU242/239, SAMPLE 2-PU242/239  
VARIANCE ETC. OF SINGLE MEASUREMENTS ARE GIVEN

RUN NO.	RUN MEAN	SIGMA	SD	RSD
1	0.69574D+00	0.193D-05	0.139D-02	0.199D+00
2	0.69991D+00	0.242D-05	0.156D-02	0.222D+00
3	0.69605D+00	0.231D-05	0.152D-02	0.218D+00
1	0.68465D+00	0.329D-05	0.181D-02	0.265D+00
2	0.68190D+00	0.584D-05	0.242D-02	0.354D+00
3	0.68736D+00	0.104D-04	0.322D-02	0.469D+00
1	0.14080D+01	0.657D-05	0.256D-02	0.182D+00
2	0.13969D+01	0.538D-04	0.734D-02	0.525D+00
3	0.14048D+01	0.734D-05	0.271D-02	0.193D+00
1	0.13818D+01	0.371D-04	0.609D-02	0.441D+00
2	0.13876D+01	0.149D-04	0.386D-02	0.278D+00
3	0.13821D+01	0.553D-04	0.743D-02	0.538D+00

RESULTS OF THE STANDARD EXPERIMENT FOR COMPARISON  
SAMPLE A

## FOR URANIUM

NUMBER OF ATOMS 1	0.28753D+19
NUMBER OF ATOMS 2	0.29061D+19
NUMBER OF ATOMS 3	0.28833D+19

## FOR PLUTONIUM

NUMBER OF ATOMS 1	0.17138D+17
NUMBER OF ATOMS 2	0.17159D+17
NUMBER OF ATOMS 3	0.17273D+17

## RESULTS OF THE DRY SPIKE EXPERIMENT

	SAMPLE A-1	SAMPLE A-2
URANIUM RESULTS:		
NUMBER OF ATOMS 1	0.29140D+19	0.29075D+19
NUMBER OF ATOMS 2	0.28963D+19	0.29194D+19
NUMBER OF ATOMS 3	0.29127D+19	0.28959D+19

PLUTONIUM RESULTS:		
NUMBER OF ATOMS 1	0.17136D+17	0.17153D+17
NUMBER OF ATOMS 2	0.17278D+17	0.17078D+17
NUMBER OF ATOMS 3	0.17178D+17	0.17149D+17

CALCULATION ON THE BASIS OF THE THREE DIFFERENT ISOTOPIC RATIOS MEASURED IN THE DRY SPIKE EXPERIMENT  
 MEAN NUMBER OF ATOMS  
 FOR URANIUM: 0.29076D+19 0.29076D+19  
 FOR PLUTONIUM: 0.17160D+17 0.17084D+17

CALCULATION ON THE BASIS OF THREE RESULTING NUMBERS OF ATOMS CALCULATED FOR THE DRY SPIKE EXPERIMENT  
 MEAN NUMBER OF ATOMS

	SAMPLE A-1	SD(MEAN)	RSD(MEAN)	SAMPLE A-2	SD(MEAN)	RSD(MEAN)
URANIUM	0.29077D+19	0.13426D-02	0.19256D+00	0.29076D+19	0.15769D-02	0.23033D+00
PLUTONIUM	0.17197D+17	0.32978D-02	0.23501D+00	0.17127D+17	0.18936D-02	0.13684D+00

**12.3 Results of the Group of Laboratories**



## RESULTS OF LAB2 FOR ALL THE LABS

LAB	MEAN	VARIANCE LAB MEAN	STAND.DEV. LAB MEAN	RSD(%) LAB MEAN	VARIANCE RUN	SD RUN	RSD(%) RUN	VARIANCE SCAN	SD SCAN	RSD(%) SCAN
RATIO 83 SAMPLE US										
LAB 2	0.22410-01	0.15090-07	0.1230-03	0.5480+00	0.42990-07	0.2070-03	0.9250+00	0.18230-07	0.1350-03	0.6020+00
LAB 3	0.22510-01	0.27100-09	0.1650-04	0.7310-01	- .11510-08	- .5000+01	- .5000+01	0.15710-07	0.1250-03	0.5570+00
LAB 4	0.21940-01	0.22570-09	0.1500-04	0.6850-01	- .12720-08	- .5000+01	- .5000+01	0.15600-07	0.1250-03	0.5690+00
LAB 5	0.25720-01	0.10810-06	0.3290-03	0.1280+01	0.29610-06	0.5440-03	0.2120+01	0.22560-06	0.4750-03	0.1850+01
LAB 6	0.22120-01	0.11510-08	0.3390-04	0.1530+00	0.18280-08	0.4280-04	0.1930+00	0.13010-07	0.1140-03	0.5160+00
LAB 7	0.22140-01	0.17360-10	0.4170-05	0.1880-01	- .60040-08	- .5000+01	- .5000+01	0.48450-07	0.2200-03	0.9940+00
LAB 8	0.23610-01	0.13550-06	0.3680-03	0.1560+01	0.32390-06	0.5690-03	0.2410+01	0.66150-06	0.8130-03	0.3450+01
LAB 10	0.21660-01	0.12730-07	0.1130-03	0.5210+00	0.35980-07	0.1900-03	0.8760+00	0.17570-07	0.1330-03	0.6120+00
LAB 12	0.22330-01	0.12970-08	0.3600-04	0.1610+00	- .31360-08	- .5000+01	- .5000+01	0.56220-07	0.2370-03	0.1060+01
LAB 13	0.21990-01	0.12600-08	0.3550-04	0.1610+00	- .95870-08	- .5000+01	- .5000+01	0.10690-06	0.3270-03	0.1490+01
LAB 14	0.22170-01	0.52080-10	0.7220-05	0.3250-01	- .45390-09	- .5000+01	- .5000+01	0.48810-08	0.6990-04	0.3150+00
LAB 15	0.21890-01	0.38040-08	0.6170-04	0.2820+00	0.67850-08	0.8240-04	0.3760+00	0.37010-07	0.1920-03	0.8790+00
LAB 16	0.22120-01	0.25520-10	0.5050-05	0.2280-01	- .18910-09	- .5000+01	- .5000+01	0.21250-08	0.4610-04	0.2080+00
LAB 17	0.22280-01	0.15540-08	0.3940-04	0.1770+00	0.78720-09	0.2810-04	0.1260+00	0.30990-07	0.1760-03	0.7900+00
LAB 18	0.21990-01	0.54060-08	0.7350-04	0.3340+00	0.13870-07	0.1180-03	0.5350+00	0.18810-07	0.1370-03	0.6240+00
LAB 19	0.22300-01	0.30280-09	0.1740-04	0.7800+01	0.14260-09	0.1190-04	0.5360-01	0.61260-08	0.7830-04	0.3510+00
LAB 20	0.24860-01	0.0	0.0	0.0	- .50000+01	- .5000+01	- .5000+01	0.29950-06	0.5470-03	0.2200+01
LAB 21	0.23080-01	0.86290-09	0.2940-04	0.1270+00	0.13740-08	0.3710-04	0.1610+00	0.28130-08	0.5300-04	0.2300+00
LAB 23	0.22010-01	0.21890-07	0.1480-03	0.6720+00	0.60160-07	0.2450-03	0.1110+01	0.44170-07	0.2100-03	0.9550+00
RATIO 38 SAMPLE AU										
LAB 2	0.82210-06	0.26400-13	0.1620-06	0.1980+02	0.63340-13	0.2520-06	0.3060+02	0.12680-12	0.3560-06	0.4330+02
LAB 3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 8	0.67500-04	0.13960-09	0.1180-04	0.1750+02	0.39050-09	0.1980-04	0.2930+02	0.22620-09	0.1500-04	0.2230+02
LAB 10	0.22080-05	0.17360-12	0.4170-06	0.1890+02	0.27310-12	0.5230-06	0.2370+02	0.19820-11	0.1410-05	0.6380+02
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 15	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 38 SAMPLE BU										
LAB 2	0.12010-05	0.21940-12	0.4680-06	0.3900+02	0.63320-12	0.7960-06	0.6630+02	0.19890-12	0.4460-06	0.3710+02
LAB 3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 8	0.74580-04	0.19600-09	0.1400-04	0.1880+02	0.46070-09	0.2150-04	0.2880+02	0.10180-08	0.3190-04	0.4280+02
LAB 10	0.22500-05	0.88020-12	0.9380-06	0.4170+02	0.25320-11	0.1590-05	0.7070+02	0.86900-12	0.9320-06	0.4140+02
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 15	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 38 SAMPLE RU										
LAB 2	0.55190-06	0.14100-15	0.1190-07	0.2150+01	- .27880-12	- .5000+01	- .5000+01	0.22320-11	0.1490-05	0.2710+03
LAB 3	0.37710-04	0.63870-11	0.2530-05	0.6700+01	0.16620-10	0.4080-05	0.1080+02	0.20300-10	0.4510-05	0.1190+02

LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 8	0.9542D-04	0.7413D-10	0.861D-05	0.902D+01	0.2138D-09	0.146D-04	0.153D+02	0.6845D-10	0.827D-05	0.867D+01
LAB 10	0.2333D-05	0.1372D-12	0.370D-06	0.159D+02	0.2403D-12	0.490D-06	0.210D+02	0.1369D-11	0.117D-05	0.501D+02
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 15	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 38 SAMPLE AS										
LAB 2	0.6368D+00	0.1303D-05	0.114D-02	0.179D+00	0.3525D-05	0.188D-02	0.295D+00	0.3064D-05	0.175D-02	0.275D+00
LAB 3	0.6322D+00	0.6518D-06	0.807D-03	0.128D+00	0.1866D-05	0.137D-02	0.216D+00	0.7127D-06	0.844D-03	0.134D+00
LAB 4	0.6416D+00	0.3836D-04	0.619D-02	0.965D+00	0.1014D-03	0.101D-01	0.157D+01	0.1094D-03	0.105D-01	0.163D+01
LAB 5	0.6476D+00	0.4837D-09	0.220D-04	0.340D-02	-4.855D-06	-5.00D+01	-5.00D+01	0.3896D-05	0.197D-02	0.305D+00
LAB 6	0.6395D+00	0.3063D-05	0.175D-02	0.274D+00	0.7861D-05	0.280D-02	0.438D+00	0.1062D-04	0.326D-02	0.510D+00
LAB 7	0.6379D+00	0.6730D-06	0.820D-03	0.129D+00	-1.273D-04	-5.00D+01	-5.00D+01	0.1180D-03	0.109D-01	0.170D+01
LAB 8	0.6392D+00	0.1357D-04	0.368D-02	0.576D+00	0.3331D-04	0.577D-02	0.903D+00	0.5935D-04	0.770D-02	0.121D+01
LAB 10	0.6368D+00	0.4060D-05	0.202D-02	0.316D+00	0.1204D-04	0.347D-02	0.545D+00	0.1113D-05	0.105D-02	0.166D+00
LAB 12	0.6376D+00	0.6480D-06	0.805D-03	0.126D+00	0.1595D-05	0.126D-02	0.198D+00	0.2789D-05	0.167D-02	0.262D+00
LAB 13	0.6446D+00	0.3458D-07	0.186D-03	0.288D-01	-1.260D-06	-5.00D+01	-5.00D+01	0.1838D-05	0.136D-02	0.210D+00
LAB 14	0.6414D+00	0.1608D-05	0.127D-02	0.198D+00	0.3737D-05	0.193D-02	0.301D+00	0.8710D-05	0.295D-02	0.460D+00
LAB 15	0.6454D+00	0.4102D-05	0.203D-02	0.314D+00	0.1169D-04	0.342D-02	0.530D+00	0.4969D-05	0.223D-02	0.345D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.6339D+00	0.5243D-05	0.229D-02	0.361D+00	0.1361D-04	0.369D-02	0.582D+00	0.1695D-04	0.412D-02	0.650D+00
LAB 18	0.6420D+00	0.4141D-05	0.203D-02	0.317D+00	0.1018D-04	0.319D-02	0.497D+00	0.1794D-04	0.424D-02	0.660D+00
LAB 19	0.6416D+00	0.2556D-06	0.506D-03	0.788D-01	0.4892D-06	0.699D-03	0.109D+00	0.2222D-05	0.149D-02	0.232D+00
LAB 20	0.6373D+00	0.2280D-06	0.477D-03	0.749D-01	-9.661D-05	-5.00D+01	-5.00D+01	0.8276D-04	0.910D-02	0.143D+01
LAB 21	0.6300D+00	0.6391D-07	0.253D-03	0.401D-01	-3.877D-06	-5.00D+01	-5.00D+01	0.4636D-05	0.215D-02	0.342D+00
LAB 23	0.6519D+00	0.4376D-06	0.661D-03	0.101D+00	0.5215D-06	0.722D-03	0.111D+00	0.6329D-05	0.252D-02	0.386D+00
RATIO 38 SAMPLE BS										
LAB 2	0.6309D+00	0.5116D-05	0.226D-02	0.359D+00	0.1479D-04	0.385D-02	0.610D+00	0.4427D-05	0.210D-02	0.334D+00
LAB 3	0.6235D+00	0.1126D-06	0.336D-03	0.538D-01	-1.552D-05	-5.00D+01	-5.00D+01	0.1512D-04	0.389D-02	0.624D+00
LAB 4	0.6326D+00	0.3139D-05	0.177D-02	0.280D+00	0.4815D-05	0.219D-02	0.347D+00	0.3683D-04	0.607D-02	0.959D+00
LAB 5	0.6408D+00	0.8678D-07	0.295D-03	0.460D-01	-1.203D-06	-5.00D+01	-5.00D+01	0.3045D-05	0.175D-02	0.272D+00
LAB 6	0.6331D+00	0.1511D-05	0.123D-02	0.194D+00	-9.953D-06	-5.00D+01	-5.00D+01	0.4375D-04	0.661D-02	0.104D+01
LAB 7	0.6248D+00	0.1960D-05	0.140D-02	0.224D+00	-3.335D-04	-5.00D+01	-5.00D+01	0.3139D-03	0.177D-01	0.284D+01
LAB 8	0.6340D+00	0.2357D-04	0.486D-02	0.766D+00	0.5108D-04	0.715D-02	0.113D+01	0.1571D-03	0.125D-01	0.198D+01
LAB 10	0.6322D+00	0.1182D-05	0.109D-02	0.172D+00	0.3388D-05	0.184D-02	0.291D+00	0.1265D-05	0.112D-02	0.178D+00
LAB 12	0.6335D+00	0.5008D-06	0.708D-03	0.112D+00	0.4450D-06	0.667D-03	0.105D+00	0.8460D-05	0.291D-02	0.459D+00
LAB 13	0.6352D+00	0.4574D-07	0.214D-03	0.337D-01	-1.407D-06	-5.00D+01	-5.00D+01	0.2224D-05	0.149D-02	0.235D+00
LAB 14	0.6299D+00	0.1783D-05	0.134D-02	0.212D+00	0.4559D-05	0.214D-02	0.339D+00	0.6334D-05	0.252D-02	0.400D+00
LAB 15	0.6337D+00	0.3396D-05	0.184D-02	0.291D+00	0.8381D-05	0.289D-02	0.457D+00	0.1445D-04	0.380D-02	0.600D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.6315D+00	0.4961D-06	0.704D-03	0.112D+00	0.1206D-05	0.110D-02	0.174D+00	0.2259D-05	0.150D-02	0.238D+00
LAB 18	0.6306D+00	0.1074D-05	0.104D-02	0.164D+00	0.1936D-05	0.139D-02	0.221D+00	0.1029D-04	0.321D-02	0.509D+00
LAB 19	0.6348D+00	0.6834D-06	0.827D-03	0.130D+00	0.1467D-05	0.121D-02	0.191D+00	0.4663D-05	0.216D-02	0.340D+00
LAB 20	0.6327D+00	0.1782D-05	0.134D-02	0.211D+00	-5.735D-05	-5.00D+01	-5.00D+01	0.8866D-04	0.942D-02	0.149D+01
LAB 21	0.6261D+00	0.3821D-06	0.618D-03	0.987D-01	0.9631D-06	0.981D-03	0.157D+00	0.1466D-05	0.121D-02	0.193D+00
LAB 23	0.6472D+00	0.1591D-05	0.126D-02	0.195D+00	0.3902D-05	0.198D-02	0.305D+00	0.6976D-05	0.264D-02	0.408D+00
RATIO 38 SAMPLE RS										
LAB 2	0.7540D+00	0.1898D-07	0.138D-03	0.183D-01	-2.641D-06	-5.00D+01	-5.00D+01	0.2569D-05	0.160D-02	0.213D+00
LAB 3	0.7432D+00	0.3309D-07	0.182D-03	0.245D-01	-4.315D-06	-5.00D+01	-5.00D+01	0.4246D-05	0.206D-02	0.277D+00
LAB 4	0.7542D+00	0.4634D-05	0.215D-02	0.285D+00	0.7196D-06	0.848D-03	0.112D+00	0.1055D-03	0.103D-01	0.136D+01
LAB 5	0.7624D+00	0.4260D-06	0.653D-03	0.856D-01	0.7260D-06	0.852D-03	0.112D+00	0.4417D-05	0.210D-02	0.276D+00
LAB 6	0.7499D+00	0.1450D-05	0.120D-02	0.161D+00	0.6827D-06	0.826D-03	0.110D+00	0.2933D-04	0.542D-02	0.722D+00
LAB 7	0.7541D+00	0.2382D-05	0.154D-02	0.205D+00	0.1509D-05	0.123D-02	0.163D+00	0.4509D-04	0.671D-02	0.890D+00
LAB 8	0.7595D+00	0.1161D-05	0.108D-02	0.142D+00	-9.267D-05	-5.00D+01	-5.00D+01	0.1020D-03	0.101D-01	0.133D+01
LAB 10	0.7519D+00	0.7474D-05	0.273D-02	0.364D+00	0.2155D-04	0.464D-02	0.617D+00	0.6949D-05	0.264D-02	0.351D+00

LAB 12	0.7511D+00	0.1197D-05	0.109D-02	0.146D+00	0.2647D-05	0.163D-02	0.217D+00	0.7555D-05	0.275D-02	0.366D+00
LAB 13	0.7550D+00	0.1074D-06	0.328D-03	0.434D-01	-2069D-06	-500D+01	-500D+01	0.423D-05	0.206D-02	0.273D+00
LAB 14	0.7500D+00	0.6859D-07	0.262D-03	0.349D-01	-5359D-06	-500D+01	-500D+01	0.5934D-05	0.244D-02	0.325D+00
LAB 15	0.7605D+00	0.1273D-05	0.113D-02	0.148D+00	0.3229D-05	0.180D-02	0.236D+00	0.4708D-05	0.217D-02	0.285D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.7544D+00	0.6034D-06	0.777D-03	0.103D+00	0.1181D-05	0.109D-02	0.144D+00	0.5032D-05	0.224D-02	0.297D+00
LAB 18	0.7581D+00	0.2565D-05	0.160D-02	0.211D+00	0.6681D-05	0.258D-02	0.341D+00	0.8116D-05	0.285D-02	0.376D+00
LAB 19	0.7540D+00	0.5554D-06	0.745D-03	0.988D-01	0.8187D-06	0.905D-03	0.120D+00	0.6779D-05	0.260D-02	0.345D+00
LAB 20	0.7536D+00	0.1268D-05	0.113D-02	0.149D+00	-1029D-04	-500D+01	-500D+01	0.1127D-03	0.106D-01	0.141D+01
LAB 21	0.7399D+00	0.3710D-07	0.193D-03	0.260D-01	-6572D-06	-500D+01	-500D+01	0.6148D-05	0.248D-02	0.335D+00
LAB 23	0.7558D+00	0.2816D-05	0.168D-02	0.222D+00	0.7136D-05	0.267D-02	0.353D+00	0.1050D-04	0.324D-02	0.429D+00
RATIO 48 SAMPLE AU										
LAB 2	0.1659D-03	0.3872D-12	0.622D-06	0.375D+00	-1005D-10	-500D+01	-500D+01	0.8968D-10	0.947D-05	0.571D+01
LAB 3	0.1780D-03	0.2146D-11	0.146D-05	0.823D+00	0.2027D-11	0.142D-05	0.800D+00	0.3529D-10	0.594D-05	0.334D+01
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.2537D-03	0.8473D-09	0.291D-04	0.115D+02	0.2431D-08	0.493D-04	0.194D+02	0.8893D-09	0.298D-04	0.118D+02
LAB 6	0.1710D-03	0.1901D-10	0.436D-05	0.255D+01	0.3919D-10	0.626D-05	0.366D+01	0.1426D-09	0.119D-04	0.699D+01
LAB 7	0.1387D-03	0.4115D-10	0.641D-05	0.462D+01	0.8601D-10	0.927D-05	0.668D+01	0.2994D-09	0.173D-04	0.125D+02
LAB 8	0.1933D-03	0.1892D-10	0.435D-05	0.225D+01	-4338D-10	-500D+01	-500D+01	0.8012D-09	0.283D-04	0.146D+02
LAB 10	0.1658D-03	0.1835D-11	0.135D-05	0.817D+00	0.1168D-12	0.342D-06	0.206D+00	0.4311D-10	0.657D-05	0.396D+01
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.1844D-03	0.7525D-10	0.867D-05	0.471D+01	0.2072D-09	0.144D-04	0.781D+01	0.1482D-09	0.122D-04	0.660D+01
LAB 14	0.2000D-03	0.0	0.0	0.0	-4100D-40	-500D+01	-500D+01	0.3280D-39	0.181D-19	0.906D-14
LAB 15	0.1700D-03	0.9705D-12	0.985D-06	0.579D+00	-1079D-10	-500D+01	-500D+01	0.1096D-09	0.105D-04	0.616D+01
LAB 16	0.1829D-03	0.9028D-11	0.300D-05	0.164D+01	0.2076D-10	0.456D-05	0.249D+01	0.5060D-10	0.711D-05	0.389D+01
LAB 17	0.1579D-03	0.2569D-10	0.507D-05	0.321D+01	0.7374D-10	0.859D-05	0.544D+01	0.2679D-10	0.518D-05	0.328D+01
LAB 18	0.1936D-03	0.1055D-10	0.325D-05	0.168D+01	0.8723D-11	0.295D-05	0.153D+01	0.1833D-09	0.135D-04	0.700D+01
LAB 19	0.1754D-03	0.1336D-10	0.366D-05	0.208D+01	0.3937D-10	0.627D-05	0.358D+01	0.5738D-11	0.240D-05	0.137D+01
LAB 20	0.1517D-03	0.5897D-10	0.768D-05	0.506D+01	0.1747D-09	0.132D-04	0.872D+01	0.1758D-10	0.419D-05	0.276D+01
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 48 SAMPLE BU										
LAB 2	0.1696D-03	0.3536D-11	0.188D-05	0.111D+01	0.2884D-11	0.170D-05	0.100D+01	0.6180D-10	0.786D-05	0.463D+01
LAB 3	0.1740D-03	0.3667D-10	0.606D-05	0.348D+01	0.1097D-09	0.105D-04	0.602D+01	0.241D-11	0.155D-05	0.893D+00
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.1996D-03	0.0	0.0	0.0	-5904D-40	-500D+01	-500D+01	0.4723D-39	0.217D-19	0.109D-13
LAB 6	0.1640D-03	0.1332D-10	0.365D-05	0.223D+01	0.1598D-10	0.400D-05	0.244D+01	0.1919D-09	0.139D-04	0.845D+01
LAB 7	0.1533D-03	0.2726D-10	0.522D-05	0.340D+01	0.5186D-10	0.720D-05	0.470D+01	0.2393D-09	0.155D-04	0.101D+02
LAB 8	0.1850D-03	0.1563D-11	0.125D-05	0.676D+00	-5677D-10	-500D+01	-500D+01	0.4917D-09	0.222D-04	0.120D+02
LAB 10	0.1657D-03	0.5220D-11	0.228D-05	0.138D+01	0.8217D-11	0.287D-05	0.173D+01	0.5956D-10	0.772D-05	0.466D+01
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.1989D-03	0.5106D-10	0.715D-05	0.359D+01	0.1266D-09	0.113D-04	0.566D+01	0.2127D-09	0.146D-04	0.733D+01
LAB 14	0.1633D-03	0.1215D-11	0.110D-05	0.675D+00	0.8185D-12	0.905D-06	0.554D+00	0.2262D-10	0.478D-05	0.291D+01
LAB 15	0.1790D-03	0.4189D-11	0.205D-05	0.114D+01	-1305D-10	-500D+01	-500D+01	0.2049D-09	0.143D-04	0.800D+01
LAB 16	0.1658D-03	0.2257D-11	0.150D-05	0.906D+00	-5580D-11	-500D+01	-500D+01	0.9881D-10	0.994D-05	0.599D+01
LAB 17	0.1621D-03	0.5382D-11	0.232D-05	0.143D+01	0.1176D-10	0.343D-05	0.212D+01	0.3512D-10	0.593D-05	0.366D+01
LAB 18	0.1794D-03	0.1588D-09	0.126D-04	0.702D+01	0.4668D-09	0.216D-04	0.120D+02	0.7563D-10	0.870D-05	0.485D+01
LAB 19	0.1710D-03	0.5641D-11	0.238D-05	0.139D+01	0.1447D-10	0.380D-05	0.222D+01	0.1958D-10	0.443D-05	0.259D+01
LAB 20	0.1380D-03	0.8211D-10	0.906D-05	0.657D+01	0.2448D-09	0.156D-04	0.113D+02	0.1227D-10	0.350D-05	0.254D+01
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 48 SAMPLE RU										
LAB 2	0.5932D-04	0.8368D-12	0.915D-06	0.154D+01	-5177D-12	-500D+01	-500D+01	0.2422D-10	0.492D-05	0.830D+01
LAB 3	0.6683D-04	0.1604D-10	0.401D-05	0.599D+01	0.4544D-10	0.674D-05	0.101D+02	0.2149D-10	0.464D-05	0.694D+01
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.9980D-04	0.9184D-40	0.958D-20	0.960D-14	0.1952D-39	0.140D-19	0.140D-13	0.6428D-39	0.254D-19	0.254D-13
LAB 6	0.6670D-04	0.2127D-11	0.146D-05	0.219D+01	-2122D-11	-500D+01	-500D+01	0.5803D-10	0.825D-05	0.124D+02
LAB 7	0.3500D-04	0.9896D-11	0.315D-05	0.899D+01	0.2537D-10	0.504D-05	0.144D+02	0.3452D-10	0.588D-05	0.168D+02
LAB 8	0.5625D-04	0.3646D-11	0.191D-05	0.339D+01	-9226D-11	-500D+01	-500D+01	0.1613D-09	0.127D-04	0.226D+02
LAB 10	0.5267D-04	0.6476D-12	0.805D-06	0.153D+01	0.1394D-11	0.118D-05	0.224D+01	0.4393D-11	0.210D-05	0.398D+01
LAB 12	0.5567D-04	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 14	0.5083D-04	0.6944D-12	0.833D-06	0.164D+01	0.1190D-11	0.109D-05	0.215D+01	0.7143D-11	0.267D-05	0.526D+01
LAB 15	0.6842D-04	0.9003D-10	0.949D-05	0.139D+02	0.2593D-09	0.161D-04	0.235D+02	0.8677D-10	0.932D-05	0.136D+02
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.5162D-04	0.2172D-11	0.147D-05	0.285D+01	0.4031D-11	0.201D-05	0.389D+01	0.1987D-10	0.446D-05	0.864D+01

LAB 18	0.3920D-04	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 19	0.5936D-04	0.5146D-13	0.227D-06	0.382D+00	-.1025D-11	-.500D+01	-.500D+01	0.9433D-11	0.307D-05	0.517D+01	
LAB 20	0.4871D-04	0.8939D-11	0.299D-05	0.614D+01	0.265D-10	0.515D-05	0.106D+02	0.2565D-11	0.160D-05	0.329D+01	
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
RATIO 48 SAMPLE AS											
LAB 2	0.2894D-03	0.5389D-12	0.734D-06	0.254D+00	-.3059D-12	-.500D+01	-.500D+01	0.1538D-10	0.392D-05	0.136D+01	
LAB 3	0.3119D-03	0.2031D-09	0.143D-04	0.457D+01	0.5957D-09	0.244D-04	0.782D+01	0.1091D-09	0.104D-04	0.335D+01	
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 5	0.3160D-03	0.2767D-09	0.166D-04	0.526D+01	0.7115D-09	0.267D-04	0.844D+01	0.9486D-09	0.308D-04	0.975D+01	
LAB 6	0.2790D-03	0.2307D-09	0.152D-04	0.544D+01	0.6389D-09	0.253D-04	0.906D+01	0.4257D-09	0.206D-04	0.740D+01	
LAB 7	0.1971D-03	0.2929D-09	0.171D-04	0.868D+01	0.7058D-09	0.266D-04	0.135D+02	0.1383D-08	0.372D-04	0.189D+02	
LAB 8	0.2708D-03	0.2778D-09	0.167D-04	0.615D+01	0.6176D-09	0.249D-04	0.918D+01	0.1726D-08	0.415D-04	0.153D+02	
LAB 10	0.2934D-03	0.4676D-10	0.684D-05	0.233D+01	0.1261D-09	0.112D-04	0.383D+01	0.1132D-09	0.106D-04	0.363D+01	
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 13	0.2923D-03	0.8642D-10	0.930D-05	0.318D+01	0.2010D-09	0.142D-04	0.485D+01	0.4657D-09	0.216D-04	0.738D+01	
LAB 14	0.2879D-03	0.8507D-11	0.292D-05	0.101D+01	0.2321D-10	0.482D-05	0.167D+01	0.1845D-10	0.430D-05	0.149D+01	
LAB 15	0.3305D-03	0.1562D-08	0.395D-04	0.120D+02	0.4609D-08	0.679D-04	0.205D+02	0.6193D-09	0.249D-04	0.753D+01	
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 17	0.2825D-03	0.2099D-09	0.145D-04	0.513D+01	0.5679D-09	0.238D-04	0.844D+01	0.4940D-09	0.222D-04	0.787D+01	
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 19	0.3540D-03	0.7830D-09	0.280D-04	0.790D+01	0.2342D-08	0.484D-04	0.137D+02	0.5758D-10	0.759D-05	0.214D+01	
LAB 20	0.2550D-03	0.6123D-08	0.783D-04	0.307D+02	0.1834D-07	0.135D-03	0.531D+02	0.2798D-09	0.167D-04	0.656D+01	
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
RATIO 48 SAMPLE BS											
LAB 2	0.2997D-03	0.6978D-10	0.835D-05	0.279D+01	0.1257D-09	0.114D-04	0.380D+01	0.6371D-09	0.252D-04	0.842D+01	
LAB 3	0.3318D-03	0.1482D-08	0.385D-04	0.116D+02	0.4427D-08	0.665D-04	0.201D+02	0.1610D-09	0.127D-04	0.382D+01	
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 5	0.3327D-03	0.1107D-08	0.333D-04	0.100D+02	0.3320D-08	0.576D-04	0.173D+02	0.0	0.0	0.0	
LAB 6	0.2761D-03	0.7971D-12	0.893D-06	0.323D+00	-.7766D-10	-.500D+01	-.500D+01	0.6404D-09	0.253D-04	0.917D+01	
LAB 7	0.2246D-03	0.1231D-09	0.111D-04	0.494D+01	0.2616D-09	0.162D-04	0.720D+01	0.8613D-09	0.293D-04	0.131D+02	
LAB 8	0.3000D-03	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 10	0.2992D-03	0.1792D-10	0.423D-05	0.141D+01	0.2132D-10	0.462D-05	0.154D+01	0.2595D-09	0.161D-04	0.538D+01	
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 13	0.2904D-03	0.1336D-09	0.116D-04	0.398D+01	0.3797D-09	0.195D-04	0.671D+01	0.1682D-09	0.130D-04	0.447D+01	
LAB 14	0.2854D-03	0.6944D-12	0.833D-06	0.292D+00	-.1265D-11	-.500D+01	-.500D+01	0.2679D-10	0.518D-05	0.181D+01	
LAB 15	0.3051D-03	0.1316D-09	0.115D-04	0.376D+01	0.3620D-09	0.190D-04	0.624D+01	0.2626D-09	0.162D-04	0.531D+01	
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 17	0.2900D-03	0.9896D-11	0.315D-05	0.108D+01	0.1109D-10	0.333D-05	0.115D+01	0.1488D-09	0.122D-04	0.421D+01	
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 19	0.3823D-03	0.7592D-09	0.276D-04	0.721D+01	0.2273D-08	0.477D-04	0.125D+02	0.3743D-10	0.612D-05	0.160D+01	
LAB 20	0.3944D-03	0.5347D-08	0.731D-04	0.185D+02	-.1086D-05	-.500D+01	-.500D+01	0.8774D-05	0.296D-02	0.751D+03	
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
RATIO 48 SAMPLE RS											
LAB 2	0.1999D-03	0.4715D-13	0.217D-06	0.109D+00	-.2783D-12	-.500D+01	-.500D+01	0.3358D-11	0.183D-05	0.916D+00	
LAB 3	0.1891D-03	0.2934D-10	0.542D-05	0.286D+01	0.8238D-10	0.908D-05	0.480D+01	0.4504D-10	0.671D-05	0.355D+01	
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 5	0.2287D-03	0.2248D-09	0.150D-04	0.656D+01	0.4447D-09	0.211D-04	0.922D+01	0.1838D-08	0.429D-04	0.187D+02	
LAB 6	0.2072D-03	0.4372D-10	0.661D-05	0.319D+01	0.5343D-10	0.731D-05	0.353D+01	0.6218D-09	0.249D-04	0.120D+02	
LAB 7	0.1012D-03	0.2453D-09	0.157D-04	0.155D+02	0.5914D-09	0.243D-04	0.240D+02	0.1157D-08	0.340D-04	0.336D+02	
LAB 8	0.2083D-03	0.6944D-10	0.833D-05	0.400D+01	0.1190D-09	0.109D-04	0.524D+01	0.7143D-09	0.267D-04	0.128D+02	
LAB 10	0.2047D-03	0.1023D-11	0.101D-05	0.494D+00	-.6836D-11	-.500D+01	-.500D+01	0.7923D-10	0.890D-05	0.435D+01	
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 13	0.2020D-03	0.5252D-10	0.725D-05	0.359D+01	0.1354D-09	0.116D-04	0.576D+01	0.1769D-09	0.133D-04	0.659D+01	
LAB 14	0.2037D-03	0.2353D-39	0.153D-19	0.753D-14	-.3348D-11	-.500D+01	-.500D+01	0.2679D-10	0.518D-05	0.254D+01	
LAB 15	0.2405D-03	0.9640D-09	0.310D-04	0.129D+02	0.2879D-08	0.537D-04	0.223D+02	0.1062D-09	0.103D-04	0.428D+01	
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 17	0.1871D-03	0.5382D-11	0.232D-05	0.124D+01	0.6399D-11	0.253D-05	0.135D+01	0.7798D-10	0.883D-05	0.472D+01	
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 20	0.2205D-03	0.5803D-08	0.762D-04	0.346D+02	0.1739D-07	0.132D-03	0.598D+02	0.1220D-09	0.110D-04	0.501D+01	
LAB 21	0.2082D-03	0.2438D-10	0.494D-05	0.237D+01	0.6995D-10	0.836D-05	0.402D+01	0.2542D-10	0.504D-05	0.242D+01	
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
RATIO 58 SAMPLE AU											

LAB 2	0.2218D-01	0.5658D-09	0.238D-04	0.107D+00	-.2413D-08	-.500D+01	-.500D+01	0.3288D-07	0.181D-03	0.817D+00
LAB 3	0.2235D-01	0.2486D-09	0.158D-04	0.705D-01	-.5083D-09	-.500D+01	-.500D+01	0.1003D-07	0.100D-03	0.448D+00
LAB 4	0.2237D-01	0.1595D-07	0.126D-03	0.565D+00	0.3746D-07	0.194D-03	0.865D+00	0.8321D-07	0.288D-03	0.129D+01
LAB 5	0.2253D-01	0.3138D-07	0.177D-03	0.786D+00	0.9105D-07	0.302D-03	0.134D+01	0.2481D-07	0.158D-03	0.699D+00
LAB 6	0.2237D-01	0.2756D-09	0.166D-04	0.742D-01	-.8089D-09	-.500D+01	-.500D+01	0.1309D-07	0.114D-03	0.511D+00
LAB 7	0.2201D-01	0.1580D-08	0.397D-04	0.181D+00	-.5275D-08	-.500D+01	-.500D+01	0.8012D-07	0.283D-03	0.129D+01
LAB 8	0.2263D-01	0.1304D-07	0.114D-03	0.504D+00	0.2038D-07	0.143D-03	0.631D+00	0.1499D-06	0.387D-03	0.171D+01
LAB 10	0.2211D-01	0.2685D-08	0.518D-04	0.234D+00	0.5317D-08	0.729D-04	0.330D+00	0.2190D-07	0.148D-03	0.669D+00
LAB 12	0.2226D-01	0.7043D-09	0.265D-04	0.119D+00	-.1955D-08	-.500D+01	-.500D+01	0.3255D-07	0.180D-03	0.810D+00
LAB 13	0.2222D-01	0.6106D-09	0.247D-04	0.111D+00	-.3539D-09	-.500D+01	-.500D+01	0.1748D-07	0.132D-03	0.595D+00
LAB 14	0.2213D-01	0.1319D-08	0.363D-04	0.164D+00	0.3088D-08	0.556D-04	0.251D+00	0.6964D-08	0.835D-04	0.377D+00
LAB 15	0.2239D-01	0.3403D-08	0.583D-04	0.261D+00	0.3095D-08	0.556D-04	0.248D+00	0.569D-07	0.239D-03	0.107D+01
LAB 16	0.2231D-01	0.2708D-08	0.520D-04	0.233D+00	0.6004D-08	0.775D-04	0.347D+00	0.1696D-07	0.130D-03	0.584D+00
LAB 17	0.2208D-01	0.5514D-08	0.743D-04	0.336D+00	0.1570D-07	0.125D-03	0.567D+00	0.6758D-08	0.822D-04	0.372D+00
LAB 18	0.2230D-01	0.8303D-08	0.911D-04	0.409D+00	0.2067D-07	0.144D-03	0.645D+00	0.3393D-07	0.184D-03	0.826D+00
LAB 19	0.2204D-01	0.1873D-08	0.433D-04	0.196D+00	0.5099D-08	0.714D-04	0.324D+00	0.4165D-08	0.645D-04	0.293D+00
LAB 20	0.2196D-01	0.2787D-08	0.528D-04	0.240D+00	-.1260D-07	-.500D+01	-.500D+01	0.1677D-06	0.410D-03	0.186D+01
LAB 21	0.2222D-01	0.7898D-09	0.281D-04	0.127D+00	0.2019D-08	0.449D-04	0.202D+00	0.2804D-08	0.530D-04	0.238D+00
LAB 23	0.2262D-01	0.1580D-08	0.397D-04	0.176D+00	0.2336D-08	0.483D-04	0.214D+00	0.1923D-07	0.139D-03	0.613D+00
RATIO 58 SAMPLE BU										
LAB 2	0.2212D-01	0.9674D-09	0.311D-04	0.141D+00	0.2051D-08	0.453D-04	0.205D+00	0.6813D-08	0.825D-04	0.373D+00
LAB 3	0.2195D-01	0.5048D-08	0.710D-04	0.324D+00	0.1463D-07	0.121D-03	0.551D+00	0.4137D-08	0.643D-04	0.293D+00
LAB 4	0.2235D-01	0.7075D-07	0.266D-03	0.119D+01	0.2079D-06	0.456D-03	0.204D+01	0.3446D-07	0.186D-03	0.831D+00
LAB 5	0.2231D-01	0.1518D-07	0.123D-03	0.552D+00	0.4347D-07	0.209D-03	0.934D+00	0.1656D-07	0.129D-03	0.577D+00
LAB 6	0.2238D-01	0.4877D-09	0.221D-04	0.987D-01	-.6136D-08	-.500D+01	-.500D+01	0.6079D-07	0.247D-03	0.110D+01
LAB 7	0.2230D-01	0.2778D-09	0.167D-04	0.755D-01	-.9643D-08	-.500D+01	-.500D+01	0.8381D-07	0.289D-03	0.131D+01
LAB 8	0.2239D-01	0.1755D-07	0.132D-03	0.592D+00	-.3199D-08	-.500D+01	-.500D+01	0.4468D-06	0.668D-03	0.299D+01
LAB 10	0.2204D-01	0.4936D-08	0.703D-04	0.319D+00	0.1358D-07	0.117D-03	0.529D+00	0.9849D-08	0.992D-04	0.450D+00
LAB 12	0.2228D-01	0.5858D-08	0.765D-04	0.344D+00	0.1427D-07	0.119D-03	0.536D+00	0.2645D-07	0.163D-03	0.730D+00
LAB 13	0.2220D-01	0.3003D-08	0.548D-04	0.247D+00	0.7591D-08	0.871D-04	0.392D+00	0.1135D-07	0.107D-03	0.480D+00
LAB 14	0.2216D-01	0.2104D-08	0.459D-04	0.207D+00	0.4025D-08	0.634D-04	0.286D+00	0.1829D-07	0.135D-03	0.610D+00
LAB 15	0.2232D-01	0.2569D-08	0.507D-04	0.227D+00	0.3802D-08	0.617D-04	0.276D+00	0.3125D-07	0.177D-03	0.792D+00
LAB 16	0.2242D-01	0.3974D-07	0.199D-03	0.889D+00	0.1170D-06	0.342D-03	0.153D+01	0.1798D-07	0.134D-03	0.598D+00
LAB 17	0.2214D-01	0.1036D-09	0.102D-04	0.460D-01	-.8630D-09	-.500D+01	-.500D+01	0.9392D-08	0.969D-04	0.438D+00
LAB 18	0.2244D-01	0.1482D-08	0.385D-04	0.172D+00	0.3300D-08	0.574D-04	0.256D+00	0.9162D-08	0.957D-04	0.427D+00
LAB 19	0.2202D-01	0.4075D-09	0.202D-04	0.917D-01	0.8323D-09	0.288D-04	0.131D+00	0.3121D-08	0.559D-04	0.254D+00
LAB 20	0.2190D-01	0.2399D-07	0.155D-03	0.707D+00	0.6958D-07	0.264D-03	0.120D+01	0.1908D-07	0.138D-03	0.631D+00
LAB 21	0.2227D-01	0.2251D-08	0.474D-04	0.213D+00	0.5883D-08	0.767D-04	0.344D+00	0.6968D-08	0.835D-04	0.375D+00
LAB 23	0.2252D-01	0.6771D-09	0.260D-04	0.116D+00	-.1853D-08	-.500D+01	-.500D+01	0.3107D-07	0.176D-03	0.783D+00
RATIO 58 SAMPLE RU										
LAB 2	0.7267D-02	0.5081D-10	0.713D-05	0.981D-01	-.2677D-09	-.500D+01	-.500D+01	0.3361D-08	0.580D-04	0.798D+00
LAB 3	0.7257D-02	0.8073D-10	0.899D-05	0.124D+00	0.1451D-09	0.120D-04	0.166D+00	0.7771D-09	0.279D-04	0.384D+00
LAB 4	0.7259D-02	0.9497D-10	0.975D-05	0.134D+00	0.2395D-09	0.155D-04	0.213D+00	0.3631D-09	0.191D-04	0.262D+00
LAB 5	0.7381D-02	0.6405D-09	0.253D-04	0.343D+00	0.1640D-08	0.405D-04	0.549D+00	0.2255D-08	0.475D-04	0.643D+00
LAB 6	0.7328D-02	0.1726D-10	0.415D-05	0.567D-01	-.1031D-08	-.500D+01	-.500D+01	0.8664D-08	0.931D-04	0.127D+01
LAB 7	0.7150D-02	0.2083D-09	0.144D-04	0.202D+00	0.2083D-09	0.144D-04	0.202D+00	0.3333D-08	0.577D-04	0.807D+00
LAB 8	0.7317D-02	0.1396D-09	0.118D-04	0.161D+00	-.2831D-08	-.500D+01	-.500D+01	0.2600D-07	0.161D-03	0.220D+01
LAB 10	0.7217D-02	0.5178D-09	0.228D-04	0.315D+00	0.1253D-08	0.354D-04	0.490D+00	0.2401D-08	0.490D-04	0.679D+00
LAB 12	0.7232D-02	0.5382D-11	0.232D-05	0.321D-01	-.3533D-09	-.500D+01	-.500D+01	0.2956D-08	0.544D-04	0.752D+00
LAB 13	0.7263D-02	0.6383D-09	0.253D-04	0.348D+00	0.9728D-09	0.312D-04	0.429D+00	0.7537D-08	0.868D-04	0.120D+01
LAB 14	0.7267D-02	0.5052D-10	0.711D-05	0.978D-01	0.7165D-10	0.846D-05	0.116D+00	0.6393D-09	0.253D-04	0.348D+00
LAB 15	0.7360D-02	0.1568D-09	0.125D-04	0.170D+00	-.2385D-09	-.500D+01	-.500D+01	0.5670D-08	0.753D-04	0.102D+01
LAB 16	0.7260D-02	0.5057D-10	0.711D-05	0.979D-01	0.4317D-10	0.657D-05	0.905D-01	0.8683D-09	0.295D-04	0.406D+00
LAB 17	0.7236D-02	0.3141D-09	0.177D-04	0.245D+00	0.6786D-09	0.260D-04	0.360D+00	0.2109D-08	0.459D-04	0.635D+00
LAB 18	0.7320D-02	0.5942D-09	0.244D-04	0.333D+00	0.1488D-08	0.386D-04	0.527D+00	0.2360D-08	0.486D-04	0.664D+00
LAB 19	0.7187D-02	0.3627D-11	0.190D-05	0.265D-01	-.3173D-09	-.500D+01	-.500D+01	0.2625D-08	0.512D-04	0.713D+00
LAB 20	0.7137D-02	0.9977D-08	0.999D-04	0.140D+01	0.2925D-07	0.171D-03	0.240D+01	0.5490D-08	0.741D-04	0.104D+01
LAB 21	0.2912D-01	0.0	0.0	0.0	-.5000D+01	-.500D+01	-.500D+01	0.1247D-06	0.353D-03	0.121D+01
LAB 23	0.7195D-02	0.2006D-08	0.448D-04	0.623D+00	0.5557D-08	0.745D-04	0.104D+01	0.3699D-08	0.608D-04	0.845D+00
RATIO 58 SAMPLE AS										
LAB 2	0.2208D-01	0.1176D-08	0.343D-04	0.155D+00	-.1937D-08	-.500D+01	-.500D+01	0.4371D-07	0.209D-03	0.947D+00
LAB 3	0.2215D-01	0.4608D-09	0.215D-04	0.969D-01	0.7093D-09	0.266D-04	0.120D+00	0.5385D-08	0.734D-04	0.331D+00
LAB 4	0.2261D-01	0.8247D-08	0.908D-04	0.402D+00	-.1944D-07	-.500D+01	-.500D+01	0.3535D-06	0.595D-03	0.263D+01
LAB 5	0.2231D-01	0.3029D-08	0.550D-04	0.247D+00	0.6039D-08	0.777D-04	0.348D+00	0.2439D-07	0.156D-03	0.700D+00
LAB 6	0.2222D-01	0.1694D-08	0.412D-04	0.185D+00	-.1414D-09	-.500D+01	-.500D+01	0.4179D-07	0.204D-03	0.920D+00
LAB 7	0.2204D-01	0.1974D-07	0.140D-03	0.638D+00	0.4274D-07	0.207D-03	0.938D+00	0.1318D-06	0.363D-03	0.165D+01

LAB 8	0.22170-01	0.34080-07	0.185D-03	0.8330+00	0.84300-07	0.2900-03	0.131D+01	0.1435D-06	0.379D-03	0.171D+01
LAB 10	0.22050-01	0.86270-07	0.294D-04	0.133D+00	0.2126D-08	0.461D-04	0.209D+00	0.3699D-08	0.608D-04	0.276D+00
LAB 12	0.2212D-01	0.1167D-07	0.108D-03	0.488D+00	0.3026D-07	0.174D-03	0.786D+00	0.3797D-07	0.195D-03	0.881D+00
LAB 13	0.2228D-01	0.3751D-10	0.612D-05	0.275D-01	-1.996D-08	-5.00D+01	-5.00D+01	0.1686D-07	0.130D-03	0.583D+00
LAB 14	0.2228D-01	0.1436D-08	0.379D-04	0.170D+00	0.3535D-08	0.595D-04	0.267D+00	0.6173D-08	0.786D-04	0.353D+00
LAB 15	0.2240D-01	0.3379D-08	0.581D-04	0.259D+00	0.6821D-08	0.826D-04	0.369D+00	0.2652D-07	0.163D-03	0.727D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.2223D-01	0.9719D-08	0.986D-04	0.444D+00	0.1819D-07	0.135D-03	0.607D+00	0.8771D-07	0.296D-03	0.133D+01
LAB 18	0.2251D-01	0.1005D-07	0.100D-03	0.445D+00	0.2307D-07	0.152D-03	0.675D+00	0.5658D-07	0.238D-03	0.106D+01
LAB 19	0.2237D-01	0.7569D-10	0.870D-05	0.389D-01	-1.019D-08	-5.00D+01	-5.00D+01	0.9965D-08	0.998D-04	0.446D+00
LAB 20	0.2163D-01	0.1805D-07	0.134D-03	0.621D+00	0.4602D-07	0.215D-03	0.992D+00	0.6512D-07	0.255D-03	0.118D+01
LAB 21	0.2211D-01	0.1225D-07	0.111D-03	0.501D+00	0.3630D-07	0.191D-03	0.862D+00	0.3721D-08	0.610D-04	0.276D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 58 SAMPLE BS										
LAB 2	0.2211D-01	0.3670D-08	0.606D-04	0.274D+00	0.9476D-08	0.973D-04	0.440D+00	0.1228D-07	0.111D-03	0.501D+00
LAB 3	0.2202D-01	0.2015D-08	0.449D-04	0.204D+00	0.5392D-08	0.734D-04	0.334D+00	0.5218D-08	0.722D-04	0.328D+00
LAB 4	0.2209D-01	0.5453D-07	0.234D-03	0.106D+01	0.1424D-06	0.377D-03	0.171D+01	0.1699D-06	0.412D-03	0.187D+01
LAB 5	0.2243D-01	0.1449D-07	0.120D-03	0.537D+00	0.3754D-07	0.194D-03	0.864D+00	0.4742D-07	0.218D-03	0.971D+00
LAB 6	0.2233D-01	0.3153D-07	0.178D-03	0.795D+00	0.8918D-07	0.299D-03	0.134D+01	0.4327D-07	0.208D-03	0.932D+00
LAB 7	0.2207D-01	0.8125D-08	0.901D-04	0.408D+00	-4.524D-08	-5.00D+01	-5.00D+01	0.2312D-06	0.481D-03	0.218D+01
LAB 8	0.2243D-01	0.1887D-07	0.137D-03	0.612D+00	0.3457D-07	0.186D-03	0.829D+00	0.1764D-06	0.420D-03	0.187D+01
LAB 10	0.2203D-01	0.5828D-09	0.241D-04	0.110D+00	-5.863D-10	-5.00D+01	-5.00D+01	0.1446D-07	0.120D-03	0.546D+00
LAB 12	0.2221D-01	0.2203D-08	0.469D-04	0.211D+00	0.1514D-08	0.389D-04	0.175D+00	0.4076D-07	0.202D-03	0.909D+00
LAB 13	0.2229D-01	0.2771D-09	0.166D-04	0.747D-01	0.1282D-09	0.113D-04	0.508D-01	0.5625D-08	0.750D-04	0.336D+00
LAB 14	0.2210D-01	0.9960D-09	0.316D-04	0.143D+00	0.2166D-08	0.465D-04	0.211D+00	0.6576D-08	0.811D-04	0.367D+00
LAB 15	0.2232D-01	0.1111D-08	0.333D-04	0.149D+00	-2.381D-08	-5.00D+01	-5.00D+01	0.4571D-07	0.214D-03	0.958D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.2207D-01	0.2040D-08	0.452D-04	0.205D+00	0.3821D-08	0.618D-04	0.280D+00	0.1838D-07	0.136D-03	0.614D+00
LAB 18	0.2260D-01	0.5952D-07	0.244D-03	0.108D+01	0.1634D-06	0.404D-03	0.179D+01	0.1212D-06	0.348D-03	0.154D+01
LAB 19	0.2237D-01	0.3065D-08	0.554D-04	0.247D+00	0.8306D-08	0.911D-04	0.407D+00	0.7108D-08	0.843D-04	0.377D+00
LAB 20	0.2186D-01	0.1964D-07	0.140D-03	0.641D+00	0.3534D-07	0.188D-03	0.860D+00	0.1885D-06	0.434D-03	0.199D+01
LAB 21	0.2219D-01	0.3027D-08	0.550D-04	0.248D+00	0.7532D-08	0.868D-04	0.391D+00	0.1240D-07	0.111D-03	0.502D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 58 SAMPLE RS										
LAB 2	0.7515D-02	0.3956D-08	0.629D-04	0.837D+00	0.1153D-07	0.107D-03	0.143D+01	0.2737D-08	0.523D-04	0.696D+00
LAB 3	0.7453D-02	0.6615D-10	0.813D-05	0.109D+00	-8.960D-09	-5.00D+01	-5.00D+01	0.8755D-08	0.936D-04	0.126D+01
LAB 4	0.7462D-02	0.1130D-07	0.106D-03	0.142D+01	0.2894D-07	0.170D-03	0.228D+01	0.3970D-07	0.199D-03	0.267D+01
LAB 5	0.7518D-02	0.1212D-09	0.110D-04	0.146D+00	-8.457D-09	-5.00D+01	-5.00D+01	0.9674D-08	0.984D-04	0.131D+01
LAB 6	0.7592D-02	0.1307D-07	0.114D-03	0.151D+01	0.3852D-07	0.196D-03	0.259D+01	0.5485D-08	0.741D-04	0.975D+00
LAB 7	0.7550D-02	0.9896D-09	0.315D-04	0.417D+00	0.1109D-08	0.333D-04	0.441D+00	0.1488D-07	0.122D-03	0.162D+01
LAB 8	0.7692D-02	0.6424D-09	0.253D-04	0.330D+00	-3.594D-08	-5.00D+01	-5.00D+01	0.4417D-07	0.210D-03	0.273D+01
LAB 10	0.7420D-02	0.4156D-09	0.204D-04	0.275D+00	0.7771D-09	0.279D-04	0.376D+00	0.3757D-08	0.613D-04	0.826D+00
LAB 12	0.7500D-02	0.1229D-08	0.351D-04	0.468D+00	-1.458D-09	-5.00D+01	-5.00D+01	0.3067D-07	0.175D-03	0.234D+01
LAB 13	0.7392D-02	0.6524D-09	0.255D-04	0.346D+00	0.1759D-08	0.419D-04	0.567D+00	0.1589D-08	0.399D-04	0.539D+00
LAB 14	0.7462D-02	0.2552D-10	0.505D-05	0.677D-01	-1.168D-10	-5.00D+01	-5.00D+01	0.7360D-09	0.266D-04	0.356D+00
LAB 15	0.7612D-02	0.2058D-08	0.454D-04	0.596D+00	0.5704D-08	0.755D-04	0.992D+00	0.3756D-08	0.613D-04	0.805D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.7390D-02	0.2146D-09	0.146D-04	0.198D+00	0.5062D-09	0.225D-04	0.304D+00	0.1100D-08	0.332D-04	0.449D+00
LAB 18	0.7544D-02	0.4148D-08	0.644D-04	0.854D+00	0.1138D-07	0.107D-03	0.141D+01	0.8480D-08	0.921D-04	0.122D+01
LAB 19	0.7661D-02	0.9334D-09	0.306D-04	0.399D+00	0.2355D-08	0.485D-04	0.633D+00	0.3565D-08	0.597D-04	0.779D+00
LAB 20	0.7346D-02	0.1875D-08	0.433D-04	0.589D+00	0.4487D-08	0.670D-04	0.912D+00	0.9103D-08	0.954D-04	0.130D+01
LAB 21	0.8439D-01	0.4610D-05	0.215D-02	0.254D+01	0.1382D-04	0.372D-02	0.440D+01	0.1160D-06	0.341D-03	0.404D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 68 SAMPLE AU										
LAB 2	0.3919D-02	0.1046D-09	0.102D-04	0.261D+00	0.8489D-10	0.921D-05	0.235D+00	0.1832D-08	0.428D-04	0.109D+01
LAB 3	0.4023D-02	0.3038D-08	0.551D-04	0.137D+01	0.8935D-08	0.945D-04	0.235D+01	0.1433D-08	0.379D-04	0.941D+00
LAB 4	0.5267D-02	0.4498D-06	0.671D-03	0.127D+02	0.1333D-05	0.115D-02	0.219D+02	0.1354D-06	0.368D-03	0.699D+01
LAB 5	0.4017D-02	0.1369D-08	0.370D-04	0.921D+00	0.3431D-08	0.586D-04	0.146D+01	0.5406D-08	0.735D-04	0.183D+01
LAB 6	0.3951D-02	0.1130D-09	0.106D-04	0.269D+00	0.5292D-10	0.727D-05	0.184D+00	0.2288D-08	0.478D-04	0.121D+01
LAB 7	0.3896D-02	0.4861D-09	0.220D-04	0.566D+00	0.1064D-08	0.326D-04	0.837D+00	0.3155D-08	0.562D-04	0.144D+01
LAB 8	0.3867D-02	0.4288D-08	0.655D-04	0.169D+01	0.1044D-07	0.102D-03	0.264D+01	0.1940D-07	0.139D-03	0.360D+01
LAB 10	0.3929D-02	0.1435D-09	0.120D-04	0.305D+00	0.2398D-09	0.155D-04	0.394D+00	0.1526D-08	0.391D-04	0.994D+00
LAB 12	0.3967D-02	0.2603D-08	0.510D-04	0.129D+01	0.6504D-08	0.806D-04	0.203D+01	0.1344D-07	0.102D-03	0.258D+01
LAB 13	0.3905D-02	0.1775D-09	0.133D-04	0.341D+00	0.2607D-09	0.161D-04	0.414D+00	0.2175D-08	0.466D-04	0.119D+01
LAB 14	0.3933D-02	0.4861D-09	0.220D-04	0.561D+00	0.1280D-08	0.358D-04	0.910D+00	0.1429D-08	0.378D-04	0.961D+00
LAB 15	0.3963D-02	0.1215D-11	0.110D-05	0.278D-01	-6.194D-09	-5.00D+01	-5.00D+01	0.4985D-08	0.706D-04	0.178D+01

LAB 16	0.3918D-02	0.2257D-11	0.150D-05	0.383D-01	-.1247D-09	-.500D+01	-.500D+01	0.1052D-08	0.324D-04	0.828D+00
LAB 17	0.3907D-02	0.1130D-09	0.106D-04	0.272D+00	0.1806D-09	0.134D-04	0.344D+00	0.1268D-08	0.356D-04	0.911D+00
LAB 18	0.3844D-02	0.4344D-09	0.208D-04	0.542D+00	0.8670D-09	0.294D-04	0.766D+00	0.3491D-08	0.591D-04	0.154D+01
LAB 19	0.3920D-02	0.5980D-10	0.773D-05	0.197D+00	0.1343D-09	0.116D-04	0.296D+00	0.3613D-09	0.190D-04	0.485D+00
LAB 20	0.3868D-02	0.6318D-09	0.251D-04	0.650D+00	0.1626D-08	0.403D-04	0.104D+01	0.2160D-08	0.465D-04	0.120D+01
LAB 21	0.3963D-02	0.2694D-09	0.164D-04	0.414D+00	0.5883D-09	0.243D-04	0.612D+00	0.1760D-08	0.420D-04	0.106D+01
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 68 SAMPLE BU										
LAB 2	0.3921D-02	0.3425D-10	0.585D-05	0.149D+00	-.5639D-10	-.500D+01	-.500D+01	0.1273D-08	0.357D-04	0.910D+00
LAB 3	0.3955D-02	0.2456D-09	0.157D-04	0.396D+00	0.7015D-09	0.265D-04	0.670D+00	0.2822D-09	0.168D-04	0.425D+00
LAB 4	0.4117D-02	0.7569D-08	0.870D-04	0.211D+01	0.2253D-07	0.150D-03	0.365D+01	0.1429D-08	0.378D-04	0.918D+00
LAB 5	0.3917D-02	0.2772D-09	0.167D-04	0.425D+00	0.2005D-09	0.142D-04	0.361D+00	0.5050D-08	0.711D-04	0.181D+01
LAB 6	0.3952D-02	0.1355D-10	0.368D-05	0.931D-01	-.4351D-09	-.500D+01	-.500D+01	0.3806D-08	0.617D-04	0.156D+01
LAB 7	0.3900D-02	0.1562D-09	0.125D-04	0.321D+00	-.5580D-09	-.500D+01	-.500D+01	0.8214D-08	0.906D-04	0.232D+01
LAB 8	0.3829D-02	0.3403D-08	0.583D-04	0.152D+01	0.9338D-08	0.966D-04	0.252D+01	0.6964D-08	0.835D-04	0.218D+01
LAB 10	0.3895D-02	0.4943D-09	0.222D-04	0.571D+00	0.1289D-08	0.359D-04	0.922D+00	0.1550D-08	0.394D-04	0.101D+01
LAB 12	0.3960D-02	0.6564D-09	0.256D-04	0.647D+00	0.8277D-09	0.288D-04	0.727D+00	0.9133D-08	0.956D-04	0.241D+01
LAB 13	0.3897D-02	0.5728D-09	0.239D-04	0.614D+00	0.1456D-08	0.382D-04	0.979D+00	0.2104D-08	0.459D-04	0.118D+01
LAB 14	0.3962D-02	0.5208D-10	0.722D-05	0.182D+00	-.1637D-09	-.500D+01	-.500D+01	0.2560D-08	0.506D-04	0.128D+01
LAB 15	0.3960D-02	0.2193D-09	0.148D-04	0.374D+00	0.4847D-09	0.220D-04	0.556D+00	0.1385D-08	0.372D-04	0.940D+00
LAB 16	0.3975D-02	0.9236D-10	0.961D-05	0.242D+00	0.1059D-09	0.103D-04	0.259D+00	0.1370D-08	0.370D-04	0.931D+00
LAB 17	0.3931D-02	0.1229D-08	0.351D-04	0.892D+00	0.3550D-08	0.596D-04	0.152D+01	0.1090D-08	0.330D-04	0.840D+00
LAB 18	0.3904D-02	0.4170D-07	0.204D-03	0.523D+01	0.1248D-06	0.353D-03	0.905D+01	0.2414D-08	0.491D-04	0.126D+01
LAB 19	0.3906D-02	0.1629D-09	0.128D-04	0.327D+00	0.4344D-09	0.208D-04	0.534D+00	0.4339D-09	0.208D-04	0.533D+00
LAB 20	0.3807D-02	0.8441D-09	0.291D-04	0.763D+00	0.2415D-08	0.491D-04	0.129D+01	0.9368D-09	0.306D-04	0.804D+00
LAB 21	0.3942D-02	0.4691D-09	0.217D-04	0.549D+00	0.1283D-08	0.358D-04	0.909D+00	0.9943D-09	0.315D-04	0.800D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 68 SAMPLE RU										
LAB 2	0.7579D-05	0.8172D-12	0.904D-06	0.119D+02	-.5899D-12	-.500D+01	-.500D+01	0.2433D-10	0.493D-05	0.651D+02
LAB 3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 10	0.1292D-05	0.5330D-12	0.730D-06	0.565D+02	0.1555D-11	0.125D-05	0.965D+02	0.3512D-12	0.593D-06	0.459D+02
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 15	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 21	0.1965D-02	0.0	0.0	0.0	-.5000D+01	-.500D+01	-.500D+01	0.4457D-08	0.668D-04	0.340D+01
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 68 SAMPLE AS										
LAB 2	0.3873D-02	0.1351D-09	0.116D-04	0.300D+00	0.3462D-09	0.186D-04	0.480D+00	0.4724D-09	0.217D-04	0.561D+00
LAB 3	0.3835D-02	0.2340D-08	0.484D-04	0.126D+01	0.6985D-08	0.836D-04	0.218D+01	0.2871D-09	0.169D-04	0.442D+00
LAB 4	0.3842D-02	0.1867D-06	0.432D-03	0.112D+02	0.5535D-06	0.744D-03	0.194D+02	0.5226D-07	0.229D-03	0.595D+01
LAB 5	0.3909D-02	0.3639D-09	0.191D-04	0.488D+00	0.6832D-09	0.261D-04	0.669D+00	0.3267D-08	0.572D-04	0.146D+01
LAB 6	0.3912D-02	0.7191D-09	0.268D-04	0.686D+00	0.1500D-08	0.387D-04	0.990D+00	0.5255D-08	0.725D-04	0.185D+01
LAB 7	0.3858D-02	0.3038D-08	0.551D-04	0.143D+01	0.8564D-08	0.925D-04	0.240D+01	0.4405D-08	0.664D-04	0.172D+01
LAB 8	0.3837D-02	0.2083D-09	0.144D-04	0.376D+00	-.1912D-08	-.500D+01	-.500D+01	0.2030D-07	0.142D-03	0.371D+01
LAB 10	0.3855D-02	0.6148D-10	0.784D-05	0.203D+00	-.2611D-10	-.500D+01	-.500D+01	0.1684D-08	0.410D-04	0.106D+01
LAB 12	0.3885D-02	0.2621D-08	0.512D-04	0.132D+01	0.6376D-08	0.799D-04	0.206D+01	0.1189D-07	0.109D-03	0.281D+01
LAB 13	0.3859D-02	0.1478D-08	0.384D-04	0.996D+00	0.4301D-08	0.656D-04	0.170D+01	0.1355D-08	0.325D-04	0.842D+00
LAB 14	0.3900D-02	0.1927D-10	0.439D-05	0.113D+00	0.4985D-11	0.223D-05	0.572D-01	0.4226D-09	0.206D-04	0.527D+00
LAB 15	0.3952D-02	0.2507D-09	0.158D-04	0.401D+00	0.5059D-09	0.225D-04	0.569D+00	0.1970D-08	0.444D-04	0.112D+01
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.3952D-02	0.1523D-08	0.390D-04	0.988D+00	0.4248D-08	0.652D-04	0.165D+01	0.2577D-08	0.508D-04	0.128D+01
LAB 18	0.3871D-02	0.2402D-07	0.155D-03	0.400D+01	0.7161D-07	0.268D-03	0.691D+01	0.3620D-08	0.602D-04	0.155D+01
LAB 19	0.3899D-02	0.1551D-09	0.125D-04	0.319D+00	0.3366D-09	0.183D-04	0.471D+00	0.1030D-08	0.321D-04	0.823D+00
LAB 20	0.3569D-02	0.4390D-08	0.663D-04	0.186D+01	0.1301D-07	0.114D-03	0.320D+01	0.1274D-08	0.357D-04	0.100D+01
LAB 21	0.3866D-02	0.4502D-10	0.671D-05	0.174D+00	0.9140D-10	0.956D-05	0.247D+00	0.3493D-09	0.187D-04	0.483D+00

LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 68 SAMPLE BS										
LAB 2	0.3873D-02	0.4455D-10	0.667D-05	0.172D+00	0.5880D-10	0.767D-05	0.198D+00	0.5987D-09	0.245D-04	0.632D+00
LAB 3	0.3787D-02	0.2248D-09	0.150D-04	0.396D+00	0.6392D-09	0.253D-04	0.668D+00	0.2830D-09	0.168D-04	0.444D+00
LAB 4	0.3654D-02	0.1862D-06	0.431D-03	0.118D+02	0.5542D-06	0.744D-03	0.204D+02	0.3446D-07	0.186D-03	0.508D+01
LAB 5	0.3946D-02	0.4678D-09	0.216D-04	0.548D+00	0.7055D-09	0.266D-04	0.673D+00	0.5584D-08	0.747D-04	0.189D+01
LAB 6	0.3932D-02	0.1063D-09	0.103D-04	0.262D+00	-1.523D-09	-5.00D+01	-5.00D+01	0.3768D-08	0.614D-04	0.156D+01
LAB 7	0.3812D-02	0.6771D-09	0.260D-04	0.683D+00	0.1116D-08	0.334D-04	0.876D+00	0.7321D-08	0.856D-04	0.224D+01
LAB 8	0.3929D-02	0.1944D-08	0.441D-04	0.112D+01	0.4903D-08	0.700D-04	0.178D+01	0.7440D-08	0.863D-04	0.220D+01
LAB 10	0.3860D-02	0.4379D-09	0.209D-04	0.542D+00	0.9274D-09	0.305D-04	0.789D+00	0.3091D-08	0.556D-04	0.144D+01
LAB 12	0.3864D-02	0.1981D-09	0.141D-04	0.364D+00	-5.050D-09	-5.00D+01	-5.00D+01	0.8794D-08	0.938D-04	0.243D+01
LAB 13	0.3917D-02	0.8045D-09	0.284D-04	0.724D+00	0.2326D-08	0.482D-04	0.123D+01	0.7001D-09	0.265D-04	0.675D+00
LAB 14	0.3872D-02	0.2054D-09	0.143D-04	0.370D+00	0.5764D-09	0.240D-04	0.620D+00	0.3179D-09	0.178D-04	0.460D+00
LAB 15	0.3948D-02	0.6049D-09	0.246D-04	0.623D+00	0.9307D-09	0.305D-04	0.773D+00	0.7071D-08	0.841D-04	0.213D+01
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.3863D-02	0.2658D-09	0.163D-04	0.422D+00	0.6702D-09	0.259D-04	0.670D+00	0.1018D-08	0.319D-04	0.826D+00
LAB 18	0.3532D-02	0.2819D-07	0.168D-03	0.475D+01	0.8241D-07	0.287D-03	0.813D+01	0.1731D-07	0.132D-03	0.372D+01
LAB 19	0.3875D-02	0.1663D-08	0.408D-04	0.105D+01	0.4856D-08	0.697D-04	0.180D+01	0.1059D-08	0.325D-04	0.840D+00
LAB 20	0.3547D-02	0.1029D-08	0.321D-04	0.905D+00	0.2844D-08	0.533D-04	0.150D+01	0.1949D-08	0.441D-04	0.124D+01
LAB 21	0.3866D-02	0.3366D-11	0.183D-05	0.475D-01	-1.893D-09	-5.00D+01	-5.00D+01	0.1595D-08	0.399D-04	0.103D+01
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 68 SAMPLE RS										
LAB 2	0.1803D-04	0.7241D-13	0.269D-06	0.149D+01	0.1651D-12	0.406D-06	0.225D+01	0.4170D-12	0.646D-06	0.358D+01
LAB 3	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 6	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 10	0.1787D-04	0.4818D-11	0.219D-05	0.123D+02	0.1296D-10	0.360D-05	0.201D+02	0.1197D-10	0.346D-05	0.194D+02
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 14	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 15	0.7276D-04	0.1491D-08	0.386D-04	0.531D+02	0.2931D-08	0.541D-04	0.744D+02	0.4068D-09	0.202D-04	0.277D+02
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 19	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 20	0.7112D-04	0.5381D-09	0.232D-04	0.326D+02	0.1610D-08	0.401D-04	0.564D+02	0.3630D-10	0.603D-05	0.847D+01
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0

RATIO 92 SAMPLE PS										
LAB 2	0.1340D-01	0.1406D-10	0.375D-05	0.180D+01	-2.455D-10	-5.00D+01	-5.00D+01	0.4214D-09	0.205D-04	0.983D+01
LAB 3	0.1340D-01	0.3703D-10	0.409D-05	0.297D+01	0.8857D-10	0.941D-05	0.460D+01	0.1801D-09	0.134D-04	0.656D+01
LAB 4	0.1340D-01	0.2533D-11	0.159D-05	0.823D+00	0.7240D-11	0.269D-05	0.139D+01	0.2875D-11	0.170D-05	0.876D+00
LAB 5	0.1340D-01	0.3299D-09	0.182D-04	0.739D+01	0.7292D-09	0.270D-04	0.110D+02	0.2083D-08	0.456D-04	0.186D+02
LAB 6	0.1340D-01	0.4328D-10	0.658D-05	0.223D+01	0.8209D-10	0.906D-05	0.308D+01	0.3821D-09	0.195D-04	0.664D+01
LAB 7	0.1340D-01	0.4861D-11	0.220D-05	0.132D+01	0.1064D-10	0.326D-05	0.195D+01	0.3155D-10	0.562D-05	0.336D+01
LAB 8	0.1340D-01	0.0	0.0	0.0	-5.000D+01	-5.00D+01	-5.00D+01	0.3679D-07	0.192D-03	0.548D+02
LAB 10	0.1340D-01	0.2965D-11	0.172D-05	0.866D+00	-6.257D-11	-5.00D+01	-5.00D+01	0.1212D-09	0.110D-04	0.554D+01
LAB 12	0.1340D-01	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.1340D-01	0.1642D-10	0.405D-05	0.197D+01	0.3701D-10	0.608D-05	0.296D+01	0.9797D-10	0.990D-05	0.482D+01
LAB 14	0.1340D-01	0.1562D-11	0.125D-05	0.621D+00	0.2381D-11	0.154D-05	0.767D+00	0.1845D-10	0.430D-05	0.213D+01
LAB 15	0.1340D-01	0.1313D-10	0.362D-05	0.163D+01	0.1470D-10	0.383D-05	0.173D+01	0.1975D-09	0.141D-04	0.633D+01
LAB 16	0.2745D-02	0.7413D-10	0.861D-05	0.314D+00	0.1710D-09	0.131D-04	0.476D+00	0.4113D-09	0.203D-04	0.739D+00
LAB 17	0.1340D-01	0.1736D-12	0.417D-06	0.211D+00	-1.786D-11	-5.00D+01	-5.00D+01	0.1845D-10	0.430D-05	0.217D+01
LAB 18	0.1340D-01	0.2299D-10	0.479D-05	0.168D+01	0.6557D-10	0.810D-05	0.283D+01	0.2714D-10	0.521D-05	0.182D+01
LAB 19	0.1340D-01	0.4631D-10	0.681D-05	0.317D+01	0.1312D-09	0.115D-04	0.533D+01	0.6163D-10	0.785D-05	0.365D+01
LAB 20	0.1340D-01	0.1548D-10	0.393D-05	0.215D+01	0.2334D-10	0.483D-05	0.264D+01	0.1849D-09	0.136D-04	0.743D+01
LAB 21	0.1340D-01	0.2000D-07	0.141D-03	0.268D+01	0.3983D-07	0.200D-03	0.378D+01	0.1449D-08	0.381D-04	0.722D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 2 SAMPLE PS										







LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.2187D-03	0.2726D-08	0.522D-04	0.239D+02	0.8104D-08	0.900D-04	0.412D+02	0.5827D-09	0.241D-04	0.110D+02
LAB 18	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 19	0.6949D-01	0.2262D-02	0.476D-01	0.684D+02	0.6757D-02	0.822D-01	0.118D+03	0.2325D-03	0.152D-01	0.219D+02
LAB 20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 9	SAMPLE AU									
LAB 2	0.2306D+00	0.1207D-06	0.347D-03	0.151D+00	0.2967D-06	0.545D-03	0.236D+00	0.5223D-06	0.723D-03	0.313D+00
LAB 3	0.2318D+00	0.6558D-07	0.256D-03	0.110D+00	0.1468D-06	0.383D-03	0.165D+00	0.3994D-06	0.632D-03	0.273D+00
LAB 4	0.2305D+00	0.2454D-05	0.157D-02	0.680D+00	0.7065D-05	0.266D-02	0.115D+01	0.2363D-05	0.154D-02	0.667D+00
LAB 5	0.2305D+00	0.4323D-05	0.208D-02	0.902D+00	0.1108D-04	0.333D-02	0.144D+01	0.1510D-04	0.389D-02	0.169D+01
LAB 6	0.2310D+00	0.1257D-07	0.112D-03	0.485D-01	-0.9013D-07	-0.500D+01	-0.500D+01	0.1023D-05	0.101D-02	0.438D+00
LAB 7	0.2307D+00	0.3803D-06	0.617D-03	0.267D+00	0.9188D-06	0.959D-03	0.415D+00	0.1776D-05	0.133D-02	0.578D+00
LAB 8	0.2297D+00	0.2726D-06	0.522D-03	0.227D+00	-0.1765D-05	-0.500D+01	-0.500D+01	0.2066D-04	0.455D-02	0.190D+01
LAB 10	0.2298D+00	0.4779D-07	0.219D-03	0.951D-01	0.1183D-06	0.344D-03	0.150D+00	0.2006D-06	0.448D-03	0.195D+00
LAB 12	0.2306D+00	0.1590D-07	0.126D-03	0.547D-01	-0.3190D-07	-0.500D+01	-0.500D+01	0.6369D-06	0.798D-03	0.346D+00
LAB 13	0.2305D+00	0.9281D-07	0.305D-03	0.132D+00	0.2594D-06	0.509D-03	0.221D+00	0.1523D-06	0.390D-03	0.169D+00
LAB 14	0.2301D+00	0.2478D-06	0.498D-03	0.216D+00	0.6618D-06	0.814D-03	0.354D+00	0.6532D-06	0.808D-03	0.351D+00
LAB 15	0.2309D+00	0.2505D-06	0.500D-03	0.217D+00	0.5623D-06	0.750D-03	0.325D+00	0.1512D-05	0.123D-02	0.533D+00
LAB 16	0.2321D+00	0.1063D-06	0.326D-03	0.141D+00	0.2950D-06	0.543D-03	0.234D+00	0.1919D-06	0.438D-03	0.189D+00
LAB 17	0.2308D+00	0.1539D-06	0.392D-03	0.170D+00	0.3480D-06	0.590D-03	0.256D+00	0.9092D-06	0.954D-03	0.413D+00
LAB 18	0.2305D+00	0.1513D-07	0.123D-03	0.534D-01	0.9711D-08	0.985D-04	0.428D-01	0.2854D-06	0.534D-03	0.232D+00
LAB 19	0.2302D+00	0.1349D-07	0.116D-03	0.505D-01	0.9769D-08	0.988D-04	0.429D-01	0.2456D-06	0.496D-03	0.215D+00
LAB 20	0.2317D+00	0.7669D-06	0.876D-03	0.378D+00	0.1425D-05	0.119D-02	0.515D+00	0.7006D-05	0.265D-02	0.114D+01
LAB 21	0.2308D+00	0.1587D-06	0.398D-03	0.173D+00	0.3435D-06	0.586D-03	0.254D+00	0.1360D-05	0.103D-02	0.446D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 9	SAMPLE BU									
LAB 2	0.2306D+00	0.2189D-07	0.148D-03	0.642D-01	0.3602D-07	0.190D-03	0.823D-01	0.2373D-06	0.487D-03	0.211D+00
LAB 3	0.2306D+00	0.1925D-07	0.139D-03	0.602D-01	0.4199D-07	0.205D-03	0.889D-01	0.1261D-06	0.355D-03	0.154D+00
LAB 4	0.2293D+00	0.4870D-06	0.698D-03	0.304D+00	0.1048D-05	0.102D-02	0.446D+00	0.3303D-05	0.182D-02	0.793D+00
LAB 5	0.2308D+00	0.5090D-07	0.226D-03	0.978D-01	-0.2691D-07	-0.500D+01	-0.500D+01	0.1437D-05	0.120D-02	0.519D+00
LAB 6	0.2310D+00	0.6892D-08	0.830D-04	0.359D-01	-0.1523D-06	-0.500D+01	-0.500D+01	0.1384D-05	0.118D-02	0.509D+00
LAB 7	0.2299D+00	0.2799D-07	0.167D-03	0.728D-01	-0.2417D-06	-0.500D+01	-0.500D+01	0.2605D-05	0.161D-02	0.702D+00
LAB 8	0.2275D+00	0.4288D-06	0.655D-03	0.288D+00	-0.1257D-05	-0.500D+01	-0.500D+01	0.2035D-04	0.451D-02	0.198D+01
LAB 10	0.2293D+00	0.6895D-08	0.830D-04	0.362D-01	-0.2091D-07	-0.500D+01	-0.500D+01	0.3328D-06	0.577D-03	0.252D+00
LAB 12	0.2306D+00	0.6911D-07	0.263D-03	0.114D+00	-0.1637D-08	-0.500D+01	-0.500D+01	0.1672D-05	0.129D-02	0.561D+00
LAB 13	0.2300D+00	0.6294D-08	0.793D-04	0.345D-01	-0.3939D-07	-0.500D+01	-0.500D+01	0.4662D-06	0.683D-03	0.297D+00
LAB 14	0.2309D+00	0.1004D-06	0.317D-03	0.137D+00	0.2151D-06	0.464D-03	0.201D+00	0.6884D-06	0.830D-03	0.359D+00
LAB 15	0.2305D+00	0.4606D-07	0.215D-03	0.931D-01	0.4494D-08	0.670D-04	0.291D-01	0.1069D-05	0.103D-02	0.449D+00
LAB 16	0.2316D+00	0.3663D-08	0.605D-04	0.261D-01	-0.1153D-07	-0.500D+01	-0.500D+01	0.1802D-06	0.424D-03	0.183D+00
LAB 17	0.2307D+00	0.7180D-06	0.847D-03	0.367D+00	0.1843D-05	0.136D-02	0.589D+00	0.2489D-05	0.158D-02	0.684D+00
LAB 18	0.2303D+00	0.7652D-08	0.875D-04	0.380D-01	0.8232D-08	0.907D-04	0.394D-01	0.1178D-06	0.343D-03	0.149D+00
LAB 19	0.2299D+00	0.4424D-07	0.210D-03	0.915D-01	0.1160D-06	0.341D-03	0.148D+00	0.1340D-06	0.366D-03	0.159D+00
LAB 20	0.2295D+00	0.9637D-07	0.310D-03	0.135D+00	-0.3718D-06	-0.500D+01	-0.500D+01	0.5287D-05	0.230D-02	0.100D+01
LAB 21	0.2306D+00	0.1525D-06	0.391D-03	0.169D+00	0.3403D-06	0.583D-03	0.253D+00	0.9374D-06	0.968D-03	0.420D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 9	SAMPLE RU									
LAB 2	0.2639D-01	0.3177D-10	0.564D-05	0.214D-01	-0.5282D-08	-0.500D+01	-0.500D+01	0.4302D-07	0.207D-03	0.786D+00
LAB 3	0.2655D-01	0.8807D-09	0.297D-04	0.112D+00	-0.1214D-08	-0.500D+01	-0.500D+01	0.3085D-07	0.176D-03	0.662D+00
LAB 4	0.2342D-01	0.3658D-07	0.191D-03	0.817D+00	0.1083D-06	0.329D-03	0.141D+01	0.1161D-07	0.108D-03	0.460D+00
LAB 5	0.2769D-01	0.5404D-06	0.735D-03	0.265D+01	0.1599D-05	0.126D-02	0.457D+01	0.1751D-06	0.418D-03	0.151D+01
LAB 6	0.2652D-01	0.4918D-09	0.222D-04	0.836D-01	-0.4994D-08	-0.500D+01	-0.500D+01	0.5176D-07	0.228D-03	0.858D+00
LAB 7	0.2590D-01	0.7465D-09	0.273D-04	0.106D+00	-0.1890D-08	-0.500D+01	-0.500D+01	0.3304D-07	0.182D-03	0.702D+00
LAB 8	0.2734D-01	0.5172D-07	0.227D-03	0.832D+00	0.1353D-07	0.116D-03	0.425D+00	0.1133D-05	0.106D-02	0.389D+01
LAB 10	0.2648D-01	0.2636D-08	0.513D-04	0.194D+00	0.4665D-08	0.683D-04	0.258D+00	0.2594D-07	0.161D-03	0.608D+00
LAB 12	0.2634D-01	0.7741D-09	0.278D-04	0.106D+00	0.3809D-09	0.195D-04	0.741D-01	0.1553D-07	0.125D-03	0.473D+00
LAB 13	0.2616D-01	0.4906D-09	0.222D-04	0.847D-01	-0.8715D-09	-0.500D+01	-0.500D+01	0.1875D-07	0.137D-03	0.523D+00
LAB 14	0.2650D-01	0.2707D-07	0.165D-03	0.621D+00	0.8056D-07	0.284D-03	0.107D+01	0.5068D-08	0.712D-04	0.269D+00
LAB 15	0.2634D-01	0.1646D-07	0.128D-03	0.487D+00	0.4184D-07	0.205D-03	0.777D+00	0.6030D-07	0.246D-03	0.932D+00
LAB 16	0.2619D-01	0.8127D-09	0.285D-04	0.109D+00	0.2045D-08	0.452D-04	0.173D+00	0.3142D-08	0.561D-04	0.214D+00
LAB 17	0.2625D-01	0.6041D-08	0.777D-04	0.296D+00	0.1106D-07	0.105D-03	0.401D+00	0.5647D-07	0.238D-03	0.905D+00
LAB 18	0.2623D-01	0.3694D-09	0.192D-04	0.733D-01	0.9737D-09	0.312D-04	0.119D+00	0.1075D-08	0.328D-04	0.125D+00
LAB 19	0.2631D-01	0.1844D-08	0.429D-04	0.163D+00	0.3541D-08	0.595D-04	0.226D+00	0.1592D-07	0.126D-03	0.480D+00
LAB 20	0.2609D-01	0.3753D-07	0.194D-03	0.742D+00	0.7484D-07	0.274D-03	0.105D+01	0.3019D-06	0.549D-03	0.211D+01
LAB 21	0.2797D-01	0.5845D-09	0.242D-04	0.864D-01	-0.1852D-09	-0.500D+01	-0.500D+01	0.1551D-07	0.125D-03	0.445D+00

LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 9 SAMPLE AS											
LAB 2	0.23510+00	0.95900-07	0.3100-03	0.1320+00	0.17700-06	0.4210-03	0.1790+00	0.88550-06	0.9410-03	0.4000+00	
LAB 3	0.23550+00	0.20280-06	0.4500-03	0.1910+00	0.56790-06	0.7540-03	0.3200+00	0.32460-06	0.5700-03	0.2420+00	
LAB 4	0.22860+00	0.13440-05	0.1160-02	0.5070+00	0.30390-05	0.1740-02	0.7630+00	0.79450-05	0.2820-02	0.1230+01	
LAB 5	0.23400+00	0.22250-06	0.4720-03	0.2020+00	0.50980-06	0.7140-03	0.3050+00	0.12620-05	0.1120-02	0.4800+00	
LAB 6	0.23530+00	0.59430-07	0.2440-03	0.1040+00	- .14230-06	- .5000+01	- .5000+01	0.25640-05	0.1600-02	0.6810+00	
LAB 7	0.23510+00	0.14430-06	0.3800-03	0.1620+00	0.18200-06	0.4270-03	0.1810+00	0.20070-05	0.1420-02	0.6020+00	
LAB 8	0.23280+00	0.57310-05	0.2390-02	0.1030+01	0.89290-05	0.2990-02	0.1280+01	0.66110-04	0.8130-02	0.3490+01	
LAB 10	0.23490+00	0.11290-06	0.3360-03	0.1430+00	0.27490-06	0.5240-03	0.2230+00	0.51060-06	0.7150-03	0.3040+00	
LAB 12	0.23450+00	0.26780-06	0.5170-03	0.2210+00	0.43640-06	0.6610-03	0.2820+00	0.29350-05	0.1710-02	0.7300+00	
LAB 13	0.23370+00	0.17940-07	0.1340-03	0.5730-01	- .15690-07	- .5000+01	- .5000+01	0.55610-06	0.7460-03	0.3190+00	
LAB 14	0.23410+00	0.94110-07	0.3070-03	0.1310+00	0.23070-06	0.4800-03	0.2050+00	0.41280-06	0.6420-03	0.2740+00	
LAB 15	0.23370+00	0.49180-07	0.2220-03	0.9490-01	0.23910-07	0.1550-03	0.6620-01	0.98920-06	0.9950-03	0.4260+00	
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 17	0.23480+00	0.10690-05	0.1030-02	0.4400+00	0.30120-05	0.1740-02	0.7390+00	0.15600-05	0.1250-02	0.5320+00	
LAB 18	0.23410+00	0.11440-06	0.3380-03	0.1450+00	0.32330-06	0.5690-03	0.2430+00	0.15920-06	0.3990-03	0.1700+00	
LAB 19	0.23510+00	0.13350-07	0.1160-03	0.4920-01	- .93750-07	- .5000+01	- .5000+01	0.10700-05	0.1030-02	0.4400+00	
LAB 20	0.23360+00	0.33680-06	0.5800-03	0.2480+00	0.82240-06	0.9070-03	0.3880+00	0.15040-05	0.1230-02	0.5250+00	
LAB 21	0.23440+00	0.13000-06	0.3610-03	0.1540+00	0.20890-06	0.4570-03	0.1950+00	0.14490-05	0.1200-02	0.5140+00	
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
RATIO 9 SAMPLE BS											
LAB 2	0.23500+00	0.15990-07	0.1260-03	0.5380-01	- .16340-06	- .5000+01	- .5000+01	0.16910-05	0.1300-02	0.5530+00	
LAB 3	0.23640+00	0.21740-06	0.4660-03	0.1970+00	0.61830-06	0.7860-03	0.3330+00	0.27090-06	0.5200-03	0.2200+00	
LAB 4	0.23480+00	0.62690-06	0.7920-03	0.3370+00	0.10040-05	0.1000-02	0.4270+00	0.70140-05	0.2650-02	0.1130+01	
LAB 5	0.23360+00	0.58690-06	0.7660-03	0.3280+00	0.15250-05	0.1230-02	0.5290+00	0.18900-05	0.1370-02	0.5890+00	
LAB 6	0.23330+00	0.27180-05	0.1650-02	0.7070+00	0.76690-05	0.2770-02	0.1190+01	0.38770-05	0.1970-02	0.8440+00	
LAB 7	0.23270+00	0.32190-06	0.5670-03	0.2440+00	0.66460-06	0.8150-03	0.3500+00	0.24090-05	0.1550-02	0.6670+00	
LAB 8	0.23750+00	0.20850-05	0.1440-02	0.6080+00	- .34350-05	- .5000+01	- .5000+01	0.77520-04	0.8800-02	0.3710+01	
LAB 10	0.23400+00	0.39170-07	0.1980-03	0.8460-01	0.91840-07	0.3030-03	0.1300+00	0.20530-06	0.4530-03	0.1940+00	
LAB 12	0.23420+00	0.10640-07	0.1030-03	0.4400-01	- .10840-06	- .5000+01	- .5000+01	0.11230-05	0.1060-02	0.4520+00	
LAB 13	0.23410+00	0.42540-08	0.6520-04	0.2790-01	- .24630-07	- .5000+01	- .5000+01	0.29910-06	0.5470-03	0.2340+00	
LAB 14	0.23390+00	0.15350-06	0.3920-03	0.1670+00	0.42940-06	0.6550-03	0.2800+00	0.24800-06	0.4980-03	0.2130+00	
LAB 15	0.23250+00	0.21650-05	0.1470-02	0.6330+00	0.60870-05	0.2470-02	0.1060+01	0.32590-05	0.1810-02	0.7760+00	
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 17	0.23450+00	0.50020-07	0.2240-03	0.9540-01	0.99170-07	0.3150-03	0.1340+00	0.40710-06	0.6380-03	0.2720+00	
LAB 18	0.23350+00	0.11280-06	0.3360-03	0.1440+00	0.29790-06	0.5460-03	0.2340+00	0.32270-06	0.5680-03	0.2430+00	
LAB 19	0.23510+00	0.28960-07	0.1700-03	0.7240-01	0.16550-07	0.1290-03	0.5470-01	0.56260-06	0.7500-03	0.3190+00	
LAB 20	0.23500+00	0.13040-04	0.3610-02	0.1530+01	0.36930-04	0.6080-02	0.2580+01	0.17610-04	0.4200-02	0.1780+01	
LAB 21	0.23420+00	0.79650-07	0.2820-03	0.1210+00	0.19330-06	0.4400-03	0.1880+00	0.36550-06	0.6050-03	0.2580+00	
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
RATIO 9 SAMPLE RS											
LAB 2	0.33240-01	0.27650-08	0.5260-04	0.1580+00	- .10550-07	- .5000+01	- .5000+01	0.15070-06	0.3880-03	0.1170+01	
LAB 3	0.33240-01	0.27870-08	0.5280-04	0.1590+00	0.56090-08	0.7490-04	0.2250+00	0.22030-07	0.1480-03	0.4460+00	
LAB 4	0.32920-01	0.15360-06	0.3920-03	0.1190+01	0.45120-06	0.6720-03	0.2040+01	0.77440-07	0.2780-03	0.8450+00	
LAB 5	0.33380-01	0.62990-07	0.2510-03	0.7520+00	0.16580-06	0.4070-03	0.1220+01	0.18550-06	0.4310-03	0.1290+01	
LAB 6	0.33500-01	0.14720-08	0.3840-04	0.1150+00	- .17130-07	- .5000+01	- .5000+01	0.17240-06	0.4150-03	0.1240+01	
LAB 7	0.32750-01	0.88770-07	0.2980-03	0.9100+00	0.24480-06	0.4950-03	0.1510+01	0.17230-06	0.4150-03	0.1270+01	
LAB 8	0.35870-01	0.24530-05	0.1570-02	0.4370+01	0.68070-05	0.2610-02	0.7270+01	0.44230-05	0.2100-02	0.5860+01	
LAB 10	0.32890-01	0.95950-09	0.3100-04	0.9420-01	- .89360-09	- .5000+01	- .5000+01	0.30180-07	0.1740-03	0.5280+00	
LAB 12	0.32900-01	0.37320-08	0.6110-04	0.1860+00	0.56250-08	0.7500-04	0.2280+00	0.44580-07	0.2110-03	0.6420+00	
LAB 13	0.32770-01	0.11810-07	0.1090-03	0.3320+00	0.30780-07	0.1750-03	0.5350+00	0.37220-07	0.1930-03	0.5890+00	
LAB 14	0.33120-01	0.56420-08	0.7510-04	0.2270+00	0.16160-07	0.1270-03	0.3840+00	0.61310-08	0.7830-04	0.2360+00	
LAB 15	0.32880-01	0.16890-07	0.1300-03	0.3950+00	0.44230-07	0.2100-03	0.6400+00	0.51550-07	0.2270-03	0.6900+00	
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
LAB 17	0.32940-01	0.73380-08	0.8570-04	0.2600+00	0.21220-07	0.1460-03	0.4420+00	0.63440-08	0.7960-04	0.2420+00	
LAB 18	0.32910-01	0.61410-09	0.2480-04	0.7530-01	0.15060-08	0.3880-04	0.1180+00	0.26930-08	0.5190-04	0.1580+00	
LAB 19	0.33030-01	0.17690-08	0.4210-04	0.1270+00	0.44450-08	0.6670-04	0.2020+00	0.69030-08	0.8310-04	0.2520+00	
LAB 20	0.32270-01	0.17860-07	0.1340-03	0.4140+00	0.27480-07	0.1660-03	0.5140+00	0.20890-06	0.4570-03	0.1420+01	
LAB 21	0.66620-01	0.10400-06	0.3220-03	0.4840+00	- .23160-06	- .5000+01	- .5000+01	0.35170-05	0.1880-02	0.2820+01	
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	
RATIO 19 SAMPLE AU											
LAB 2	0.13280+00	0.23850-07	0.1540-03	0.1160+00	- .16900-06	- .5000+01	- .5000+01	0.19250-05	0.1390-02	0.1040+01	
LAB 3	0.12920+00	0.35020-06	0.5920-03	0.4580+00	0.10390-05	0.1020-02	0.7890+00	0.93000-07	0.3050-03	0.2360+00	
LAB 4	0.12840+00	0.45730-06	0.6760-03	0.5260+00	0.10760-05	0.1040-02	0.8080+00	0.23670-05	0.1540-02	0.1200+01	
LAB 5	0.12510+00	0.14590-05	0.1210-02	0.9650+00	0.36180-05	0.1900-02	0.1520+01	0.60740-05	0.2460-02	0.1970+01	

LAB 6	0.12770+00	0.9342D-08	0.967D-04	0.757D-01	-.3327D-07	-.500D+01	-.500D+01	0.4904D-06	0.700D-03	0.548D+00
LAB 7	0.1271D+00	0.6425D-06	0.802D-03	0.630D+00	0.1829D-05	0.135D-02	0.106D+01	0.7899D-06	0.889D-03	0.699D+00
LAB 8	0.1292D+00	0.5340D-06	0.731D-03	0.566D+00	0.6387D-07	0.253D-03	0.196D+00	0.1230D-04	0.351D-02	0.272D+01
LAB 10	0.1272D+00	0.3542D-07	0.188D-03	0.148D+00	0.9513D-07	0.308D-03	0.242D+03	0.889D-07	0.298D-03	0.234D+00
LAB 12	0.1277D+00	0.4772D-07	0.218D-03	0.171D+00	0.7745D-08	0.880D-04	0.689D-01	0.1083D-05	0.104D-02	0.815D+00
LAB 13	0.1268D+00	0.8683D-08	0.932D-04	0.735D-01	-.5728D-08	-.500D+01	-.500D+01	0.2542D-06	0.504D-03	0.398D+00
LAB 14	0.1262D+00	0.2075D-07	0.144D-03	0.114D+00	0.6954D-08	0.998D-04	0.790D-01	0.4184D-06	0.647D-03	0.512D+00
LAB 15	0.1276D+00	0.8637D-07	0.294D-03	0.230D+00	0.1184D-06	0.344D-03	0.270D+00	0.1125D-05	0.106D-02	0.832D+00
LAB 16	0.1282D+00	0.7094D-07	0.266D-03	0.208D+00	0.1965D-06	0.443D-03	0.346D+00	0.1308D-06	0.362D-03	0.282D+00
LAB 17	0.1272D+00	0.2534D-07	0.159D-03	0.125D+00	0.3544D-07	0.188D-03	0.148D+00	0.3245D-06	0.570D-03	0.448D+00
LAB 18	0.1276D+00	0.1795D-06	0.424D-03	0.332D+00	0.5258D-06	0.725D-03	0.568D+00	0.1012D-06	0.318D-03	0.249D+00
LAB 19	0.1263D+00	0.8453D-07	0.291D-03	0.230D+00	0.2346D-06	0.484D-03	0.383D+00	0.1520D-06	0.390D-03	0.309D+00
LAB 20	0.1273D+00	0.2257D-06	0.475D-03	0.373D+00	0.2546D-06	0.505D-03	0.396D+00	0.3379D-05	0.184D-02	0.144D+01
LAB 21	0.1261D+00	0.3461D-06	0.588D-03	0.466D+00	0.8685D-06	0.932D-03	0.739D+00	0.1358D-05	0.117D-02	0.924D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 19 SAMPLE BU										
LAB 2	0.1307D+00	0.1037D-05	0.102D-02	0.779D+00	0.2965D-05	0.172D-02	0.132D+01	0.1165D-05	0.108D-02	0.826D+00
LAB 3	0.1277D+00	0.8324D-08	0.912D-04	0.714D-01	0.1645D-07	0.128D-03	0.100D+00	0.6819D-07	0.261D-03	0.204D+00
LAB 4	0.1272D+00	0.6437D-08	0.802D-04	0.631D-01	-.9169D-07	-.500D+01	-.500D+01	0.8880D-06	0.942D-03	0.741D+00
LAB 5	0.1266D+00	0.8152D-07	0.286D-03	0.226D+00	0.1722D-06	0.415D-03	0.328D+00	0.5788D-06	0.761D-03	0.601D+00
LAB 6	0.1276D+00	0.2034D-07	0.143D-03	0.112D+00	-.8533D-07	-.500D+01	-.500D+01	0.1171D-05	0.108D-02	0.848D+00
LAB 7	0.1273D+00	0.1491D-06	0.386D-03	0.303D+00	0.1818D-06	0.426D-03	0.335D+00	0.2124D-05	0.146D-02	0.115D+01
LAB 8	0.1258D+00	0.4962D-07	0.223D-03	0.177D+00	-.1053D-05	0.358D+01	-.500D+01	0.9616D-05	0.310D-02	0.246D+01
LAB 10	0.1268D+00	0.2709D-08	0.520D-04	0.411D-01	-.2021D-07	-.500D+01	-.500D+01	0.2267D-06	0.476D-03	0.376D+00
LAB 12	0.1276D+00	0.3115D-07	0.176D-03	0.138D+00	0.3771D-07	0.194D-03	0.152D+00	0.4460D-06	0.668D-03	0.524D+00
LAB 13	0.1269D+00	0.3196D-07	0.179D-03	0.141D+00	0.6559D-07	0.256D-03	0.202D+00	0.2422D-06	0.492D-03	0.388D+00
LAB 14	0.1274D+00	0.1177D-06	0.343D-03	0.269D+00	0.3283D-06	0.573D-03	0.450D+00	0.1983D-06	0.445D-03	0.350D+00
LAB 15	0.1272D+00	0.6932D-07	0.263D-03	0.207D+00	0.1284D-06	0.358D-03	0.282D+00	0.6363D-06	0.798D-03	0.627D+00
LAB 16	0.1273D+00	0.5539D-08	0.744D-04	0.585D-01	-.2788D-08	-.500D+01	-.500D+01	0.1552D-06	0.394D-03	0.310D+00
LAB 17	0.1276D+00	0.1555D-07	0.125D-03	0.977D-01	-.5343D-07	-.500D+01	-.500D+01	0.8005D-06	0.895D-03	0.701D+00
LAB 18	0.1270D+00	0.8934D-09	0.299D-04	0.235D-01	-.8974D-08	-.500D+01	-.500D+01	0.9323D-07	0.305D-03	0.241D+00
LAB 19	0.1264D+00	0.3300D-07	0.182D-03	0.144D+00	0.6361D-07	0.252D-03	0.200D+00	0.2830D-06	0.532D-03	0.412D+00
LAB 20	0.1262D+00	0.4096D-06	0.640D-03	0.507D+00	0.6313D-06	0.795D-03	0.630D+00	0.4781D-05	0.219D-02	0.173D+01
LAB 21	0.1262D+00	0.1101D-06	0.332D-03	0.263D+00	0.2361D-06	0.486D-03	0.385D+00	0.7533D-06	0.868D-03	0.687D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 19 SAMPLE RU										
LAB 2	0.8246D-03	0.9155D-09	0.303D-04	0.367D+01	0.2317D-08	0.481D-04	0.584D+01	0.3439D-08	0.586D-04	0.711D+01
LAB 3	0.7314D-03	0.8067D-11	0.284D-05	0.388D+00	-.9618D-11	-.500D+01	-.500D+01	0.2706D-09	0.164D-04	0.225D+01
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.8439D-03	0.7972D-09	0.282D-04	0.335D+01	0.1899D-08	0.436D-04	0.516D+01	0.3941D-08	0.628D-04	0.744D+01
LAB 6	0.7909D-03	0.1710D-09	0.131D-04	0.165D+01	0.3219D-09	0.179D-04	0.227D+01	0.1528D-08	0.391D-04	0.494D+01
LAB 7	0.7984D-03	0.1669D-09	0.129D-04	0.162D+01	0.2543D-09	0.159D-04	0.200D+01	0.1971D-08	0.444D-04	0.556D+01
LAB 8	0.9920D-03	0.1383D-08	0.372D-04	0.375D+01	-.1106D-09	-.500D+01	-.500D+01	0.3408D-07	0.185D-03	0.186D+02
LAB 10	0.7441D-03	0.2276D-10	0.477D-05	0.641D+00	0.3949D-10	0.628D-05	0.845D+00	0.2304D-09	0.152D-04	0.204D+01
LAB 12	0.7439D-03	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.7138D-03	0.3296D-10	0.574D-05	0.804D+00	-.9635D-11	-.500D+01	-.500D+01	0.8682D-09	0.295D-04	0.413D+01
LAB 14	0.7593D-03	0.1939D-09	0.139D-04	0.183D+01	0.5780D-09	0.240D-04	0.317D+01	0.2921D-10	0.540D-05	0.712D+00
LAB 15	0.7820D-03	0.1172D-09	0.108D-04	0.138D+01	0.2984D-09	0.173D-04	0.221D+01	0.4252D-09	0.206D-04	0.264D+01
LAB 16	0.7375D-03	0.7006D-10	0.837D-05	0.113D+01	0.2018D-09	0.142D-04	0.193D+01	0.6665D-10	0.816D-05	0.111D+01
LAB 17	0.7245D-03	0.3191D-10	0.565D-05	0.780D+00	0.8715D-10	0.934D-05	0.129D+01	0.6870D-10	0.829D-05	0.114D+01
LAB 18	0.7446D-03	0.3461D-11	0.186D-05	0.250D+00	0.5021D-11	0.224D-05	0.301D+00	0.4290D-10	0.655D-05	0.880D+00
LAB 19	0.7476D-03	0.2822D-10	0.531D-05	0.711D+00	0.7703D-10	0.878D-05	0.117D+01	0.6102D-10	0.781D-05	0.104D+01
LAB 20	0.6795D-03	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 21	0.1333D-02	0.3417D-09	0.185D-04	0.139D+01	0.8982D-09	0.300D-04	0.225D+01	0.1016D-08	0.319D-04	0.239D+01
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 19 SAMPLE AS										
LAB 2	0.1299D+00	0.1446D-05	0.120D-02	0.925D+00	0.4117D-05	0.203D-02	0.156D+01	0.1765D-05	0.133D-02	0.102D+01
LAB 3	0.1263D+00	0.1920D-06	0.438D-03	0.347D+00	0.5718D-06	0.756D-03	0.599D+00	0.3305D-07	0.182D-03	0.144D+00
LAB 4	0.1230D+00	0.1419D-04	0.377D-02	0.306D+01	0.4218D-04	0.649D-02	0.528D+01	0.3203D-05	0.179D-02	0.146D+01
LAB 5	0.1257D+00	0.3858D-06	0.621D-03	0.494D+00	0.9948D-06	0.997D-03	0.793D+03	0.1301D-05	0.114D-02	0.907D+00
LAB 6	0.1270D+00	0.4761D-07	0.218D-03	0.172D+00	-.4716D-06	-.500D+01	-.500D+01	0.4915D-05	0.222D-02	0.175D+01
LAB 7	0.1274D+00	0.8955D-07	0.299D-03	0.235D+00	0.1408D-06	0.375D-03	0.295D+00	0.1023D-05	0.101D-02	0.794D+00
LAB 8	0.1289D+00	0.2859D-05	0.169D-02	0.131D+01	0.2949D-05	0.172D-02	0.133D+01	0.4503D-04	0.671D-02	0.521D+01
LAB 10	0.1281D+00	0.4069D-06	0.638D-03	0.498D+00	0.1197D-05	0.109D-02	0.854D+00	0.1924D-06	0.439D-03	0.342D+00
LAB 12	0.1268D+00	0.2538D-07	0.159D-03	0.126D+00	-.7488D-07	-.500D+01	-.500D+01	0.1208D-05	0.110D-02	0.867D+00
LAB 13	0.1261D+00	0.7087D-07	0.266D-03	0.211D+00	0.1938D-06	0.440D-03	0.349D+00	0.1502D-06	0.388D-03	0.307D+00

LAB 14	0.1279D+00	0.4670D-07	0.216D-03	0.169D+00	0.1058D-06	0.325D-03	0.254D+00	0.2740D-06	0.523D-03	0.409D+00
LAB 15	0.1273D+00	0.3691D-09	0.192D-04	0.151D-01	-.6475D-07	-.500D+01	-.500D+01	0.5269D-06	0.726D-03	0.570D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.1263D+00	0.2544D-06	0.504D-03	0.399D+00	0.7074D-06	0.841D-03	0.666D+00	0.4460D-06	0.668D-03	0.529D+00
LAB 18	0.1267D+00	0.5429D-07	0.233D-03	0.184D+00	0.1531D-06	0.391D-03	0.309D+00	0.7825D-07	0.280D-03	0.221D+00
LAB 19	0.1268D+00	0.5917D-07	0.243D-03	0.192D+00	0.1126D-06	0.336D-03	0.265D+00	0.5194D-06	0.721D-03	0.569D+00
LAB 20	0.1251D+00	0.4483D-06	0.670D-03	0.535D+00	0.1072D-05	0.104D-02	0.827D+00	0.2182D-05	0.148D-02	0.118D+01
LAB 21	0.1255D+00	0.5313D-07	0.230D-03	0.184D+00	0.7942D-07	0.282D-03	0.225D+00	0.6397D-06	0.800D-03	0.637D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 19 SAMPLE BS										
LAB 2	0.1344D+00	0.3277D-06	0.572D-03	0.426D+00	-.4153D-06	-.500D+01	-.500D+01	0.1119D-04	0.334D-02	0.249D+01
LAB 3	0.1269D+00	0.2864D-07	0.169D-03	0.133D+00	0.7554D-07	0.275D-03	0.217D+00	0.8299D-07	0.288D-03	0.227D+00
LAB 4	0.1264D+00	0.1501D-06	0.387D-03	0.307D+00	0.2496D-06	0.500D-03	0.395D+00	0.1606D-05	0.127D-02	0.100D+01
LAB 5	0.1251D+00	0.3566D-06	0.597D-03	0.477D+00	0.9216D-06	0.960D-03	0.767D+00	0.1185D-05	0.109D-02	0.870D+00
LAB 6	0.1262D+00	0.1512D-05	0.123D-02	0.274D+00	0.4010D-05	0.200D-02	0.159D+01	0.4203D-05	0.205D-02	0.162D+01
LAB 7	0.1256D+00	0.2781D-06	0.527D-03	0.420D+00	0.6811D-06	0.825D-03	0.657D+00	0.1227D-05	0.111D-02	0.882D+00
LAB 8	0.1276D+00	0.3029D-05	0.174D-02	0.136D+01	0.5831D-05	0.241D-02	0.189D+01	0.2606D-04	0.510D-02	0.400D+01
LAB 10	0.1270D+00	0.8822D-07	0.297D-03	0.234D+00	0.2587D-06	0.509D-03	0.401D+00	0.4735D-07	0.218D-03	0.171D+00
LAB 12	0.1268D+00	0.1314D-07	0.115D-03	0.904D-01	-.2110D-07	-.500D+01	-.500D+01	0.4840D-06	0.696D-03	0.549D+00
LAB 13	0.1259D+00	0.1081D-06	0.329D-03	0.261D+00	0.2993D-06	0.547D-03	0.434D+00	0.1994D-06	0.447D-03	0.350D+00
LAB 14	0.1275D+00	0.6534D-07	0.256D-03	0.201D+00	0.1822D-06	0.427D-03	0.335D+00	0.1108D-06	0.333D-03	0.261D+00
LAB 15	0.1270D+00	0.2461D-06	0.496D-03	0.391D+00	0.5726D-06	0.757D-03	0.596D+00	0.1325D-05	0.115D-02	0.907D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.1267D+00	0.3710D-07	0.193D-03	0.152D+00	0.7905D-07	0.281D-03	0.222D+00	0.2581D-06	0.508D-03	0.401D+00
LAB 18	0.1262D+00	0.2358D-07	0.154D-03	0.122D+00	0.4815D-07	0.219D-03	0.174D+00	0.1808D-06	0.425D-03	0.337D+00
LAB 19	0.1267D+00	0.4035D-07	0.201D-03	0.159D+00	0.6614D-07	0.250D-03	0.203D+00	0.4393D-06	0.663D-03	0.523D+00
LAB 20	0.1262D+00	0.8482D-05	0.291D-02	0.231D+01	0.2494D-04	0.499D-02	0.396D+01	0.4088D-05	0.202D-02	0.160D+01
LAB 21	0.1247D+00	0.9483D-07	0.308D-03	0.247D+00	0.2467D-06	0.497D-03	0.398D+00	0.3024D-06	0.550D-03	0.441D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 19 SAMPLE RS										
LAB 2	0.2463D-02	0.1442D-07	0.120D-03	0.488D+01	0.3594D-07	0.190D-03	0.770D+01	0.5863D-07	0.242D-03	0.983D+01
LAB 3	0.1968D-02	0.8750D-08	0.935D-04	0.475D+01	0.2575D-07	0.160D-03	0.815D+01	0.4008D-08	0.633D-04	0.322D+01
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.2065D-02	0.1092D-08	0.330D-04	0.160D+01	0.2647D-08	0.514D-04	0.249D+01	0.5033D-08	0.709D-04	0.344D+01
LAB 6	0.2010D-02	0.3767D-09	0.194D-04	0.966D+00	0.6219D-09	0.249D-04	0.124D+01	0.4066D-08	0.638D-04	0.317D+01
LAB 7	0.2037D-02	0.7680D-09	0.277D-04	0.136D+01	0.1309D-08	0.362D-04	0.178D+01	0.7961D-08	0.892D-04	0.438D+01
LAB 8	0.2348D-02	0.7055D-08	0.840D-04	0.358D+01	0.1356D-07	0.116D-03	0.496D+01	0.5087D-07	0.247D-03	0.105D+02
LAB 10	0.2026D-02	0.5048D-10	0.710D-05	0.351D+00	-.4039D-10	-.500D+01	-.500D+01	0.1535D-08	0.392D-04	0.193D+01
LAB 12	0.2157D-02	0.4069D-09	0.202D-04	0.935D+00	0.8722D-10	0.934D-05	0.433D+00	0.9067D-08	0.952D-04	0.441D+01
LAB 13	0.1927D-02	0.2300D-08	0.480D-04	0.249D+01	0.6197D-08	0.787D-04	0.409D+01	0.5634D-08	0.751D-04	0.390D+01
LAB 14	0.2044D-02	0.6594D-11	0.257D-05	0.126D+00	-.9891D-10	-.500D+01	-.500D+01	0.9495D-09	0.308D-04	0.151D+01
LAB 15	0.2021D-02	0.4695D-08	0.685D-04	0.339D+01	0.1259D-07	0.112D-03	0.555D+01	0.1198D-07	0.109D-03	0.542D+01
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.2010D-02	0.8635D-10	0.929D-05	0.462D+00	0.1970D-09	0.140D-04	0.698D+00	0.4967D-09	0.223D-04	0.111D+01
LAB 18	0.2011D-02	0.1782D-10	0.422D-05	0.210D+00	0.4254D-10	0.652D-05	0.324D+00	0.8749D-10	0.935D-05	0.465D+00
LAB 19	0.2040D-02	0.1040D-10	0.323D-05	0.158D+00	-.1089D-10	-.500D+01	-.500D+01	0.3368D-09	0.184D-04	0.900D+00
LAB 20	0.1737D-02	0.2161D-08	0.465D-04	0.268D+01	0.6015D-08	0.776D-04	0.447D+01	0.3756D-08	0.613D-04	0.353D+01
LAB 21	0.9146D-02	0.1657D-09	0.129D-04	0.141D+00	-.4282D-06	-.500D+01	-.500D+01	0.3428D-05	0.185D-02	0.202D+02
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 29 SAMPLE AU										
LAB 2	0.2345D-01	0.1058D-07	0.103D-03	0.439D+00	0.1681D-07	0.130D-03	0.553D+00	0.1195D-06	0.346D-03	0.147D+01
LAB 3	0.2314D-01	0.1580D-08	0.397D-04	0.172D+00	0.3142D-08	0.561D-04	0.242D+00	0.1278D-07	0.113D-03	0.489D+00
LAB 4	0.2271D-01	0.1915D-06	0.438D-03	0.193D+01	0.5646D-06	0.751D-03	0.331D+01	0.7792D-07	0.279D-03	0.123D+01
LAB 5	0.2266D-01	0.8929D-07	0.299D-03	0.132D+01	0.1727D-06	0.416D-03	0.183D+01	0.7615D-06	0.873D-03	0.385D+01
LAB 6	0.2308D-01	0.1773D-09	0.133D-04	0.577D-01	-.9080D-08	-.500D+01	-.500D+01	0.7690D-07	0.277D-03	0.120D+01
LAB 7	0.2305D-01	0.8507D-09	0.292D-04	0.127D+00	-.5060D-08	-.500D+01	-.500D+01	0.6089D-07	0.247D-03	0.107D+01
LAB 8	0.2783D-01	0.2230D-05	0.148D-02	0.533D+01	0.6503D-05	0.255D-02	0.916D+01	0.8535D-06	0.924D-03	0.332D+01
LAB 10	0.2287D-01	0.1090D-08	0.330D-04	0.144D+00	0.1334D-08	0.365D-04	0.160D+00	0.1548D-07	0.124D-03	0.544D+00
LAB 12	0.2289D-01	0.5613D-08	0.749D-04	0.327D+00	0.9538D-08	0.977D-04	0.427D+00	0.5841D-07	0.242D-03	0.106D+01
LAB 13	0.2261D-01	0.9065D-08	0.952D-04	0.421D+00	0.2634D-07	0.162D-03	0.718D+00	0.6857D-08	0.828D-04	0.366D+00
LAB 14	0.2280D-01	0.2168D-07	0.147D-03	0.646D+00	0.6381D-07	0.253D-03	0.111D+01	0.9940D-07	0.997D-04	0.437D+00
LAB 15	0.2305D-01	0.6476D-08	0.805D-04	0.349D+00	0.5446D-08	0.738D-04	0.320D+00	0.1118D-06	0.334D-03	0.145D+01
LAB 16	0.2304D-01	0.1215D-11	0.110D-05	0.479D-02	-.4180D-09	-.500D+01	-.500D+01	0.3373D-08	0.581D-04	0.252D+00
LAB 17	0.2284D-01	0.1580D-08	0.397D-04	0.174D+00	0.4055D-08	0.637D-04	0.279D+00	0.5474D-08	0.740D-04	0.324D+00
LAB 18	0.2297D-01	0.1603D-08	0.400D-04	0.174D+00	0.3171D-08	0.563D-04	0.245D+00	0.1309D-07	0.114D-03	0.498D+00
LAB 19	0.2255D-01	0.5866D-09	0.242D-04	0.107D+00	-.7164D-09	-.500D+01	-.500D+01	0.1981D-07	0.141D-03	0.624D+00

LAB 20	0.2269D-01	0.7911D-07	0.281D-03	0.124D+01	0.1988D-06	0.446D-03	0.197D+01	0.3080D-06	0.555D-03	0.245D+01
LAB 21	0.2416D-01	0.8561D-07	0.293D-03	0.121D+01	0.2248D-06	0.474D-03	0.196D+01	0.2560D-06	0.506D-03	0.209D+01
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 29 SAMPLE BU										
LAB 2	0.2331D-01	0.6059D-08	0.778D-04	0.334D+00	0.7299D-08	0.854D-04	0.367D+00	0.8702D-07	0.295D-03	0.127D+01
LAB 3	0.2330D-01	0.3012D-08	0.549D-04	0.236D+00	0.6175D-08	0.786D-04	0.337D+00	0.2289D-07	0.151D-03	0.649D+00
LAB 4	0.2247D-01	0.8130D-07	0.285D-03	0.127D+01	0.2390D-06	0.489D-03	0.218D+01	0.3917D-07	0.198D-03	0.881D+00
LAB 5	0.2262D-01	0.1580D-08	0.397D-04	0.176D+00	-0.2421D-07	-0.500D+01	-0.500D+01	0.2316D-06	0.481D-03	0.213D+01
LAB 6	0.2305D-01	0.4319D-08	0.657D-04	0.285D+00	0.5108D-08	0.715D-04	0.310D+00	0.6279D-07	0.251D-03	0.109D+01
LAB 7	0.2288D-01	0.1783D-07	0.134D-03	0.584D+00	0.4513D-07	0.212D-03	0.929D+00	0.6685D-07	0.259D-03	0.113D+01
LAB 8	0.2373D-01	0.3991D-07	0.200D-03	0.842D+00	0.6404D-07	0.253D-03	0.107D+01	0.4456D-06	0.668D-03	0.281D+01
LAB 10	0.2287D-01	0.4882D-09	0.221D-04	0.966D-01	-0.7091D-10	-0.500D+01	-0.500D+01	0.1228D-07	0.111D-03	0.485D+00
LAB 12	0.2303D-01	0.1776D-08	0.421D-04	0.183D+00	-0.9090D-08	-0.500D+01	-0.500D+01	0.1153D-06	0.340D-03	0.147D+01
LAB 13	0.2278D-01	0.4437D-08	0.666D-04	0.292D+00	0.8681D-08	0.932D-04	0.409D+00	0.3704D-07	0.192D-03	0.845D+00
LAB 14	0.2301D-01	0.5440D-08	0.738D-04	0.321D+00	0.1419D-07	0.119D-03	0.518D+00	0.1703D-07	0.131D-03	0.567D+00
LAB 15	0.2298D-01	0.1059D-08	0.325D-04	0.142D+00	-0.2433D-08	-0.500D+01	-0.500D+01	0.4488D-07	0.212D-03	0.922D+00
LAB 16	0.2286D-01	0.1813D-08	0.426D-04	0.186D+00	0.4532D-08	0.673D-04	0.294D+00	0.7254D-08	0.852D-04	0.373D+00
LAB 17	0.2290D-01	0.1684D-08	0.410D-04	0.179D+00	0.1806D-08	0.425D-04	0.186D+00	0.2596D-07	0.161D-03	0.704D+00
LAB 18	0.2298D-01	0.2677D-08	0.517D-04	0.225D+00	0.7075D-08	0.841D-04	0.366D+00	0.7654D-08	0.875D-04	0.381D+00
LAB 19	0.2255D-01	0.1627D-09	0.128D-04	0.566D-01	-0.1587D-08	-0.500D+01	-0.500D+01	0.1660D-07	0.129D-03	0.571D+00
LAB 20	0.2131D-01	0.4429D-06	0.666D-03	0.312D+01	0.1178D-05	0.109D-02	0.509D+01	0.1205D-05	0.110D-02	0.515D+01
LAB 21	0.2343D-01	0.3051D-08	0.552D-04	0.236D+00	-0.2703D-08	-0.500D+01	-0.500D+01	0.9485D-07	0.308D-03	0.131D+01
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 29 SAMPLE RU										
LAB 2	0.9509D-04	0.1122D-09	0.106D-04	0.192D+02	0.3303D-09	0.182D-04	0.330D+02	0.4987D-10	0.706D-05	0.128D+02
LAB 3	0.3117D-04	0.7575D-11	0.275D-05	0.883D+01	0.2027D-10	0.450D-05	0.144D+02	0.1961D-10	0.443D-05	0.142D+02
LAB 4	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 6	0.9997D-04	0.7958D-10	0.892D-05	0.892D+01	0.2055D-09	0.143D-04	0.143D+02	0.2562D-09	0.163D-04	0.163D+02
LAB 7	0.2417D-04	0.1163D-10	0.341D-05	0.141D+02	0.2046D-10	0.452D-05	0.187D+02	0.1155D-09	0.107D-04	0.445D+02
LAB 8	0.3000D-03	0.3333D-08	0.577D-04	0.192D+02	0.1000D-07	0.100D-03	0.333D+02	0.1093D-39	0.105D-19	0.349D-14
LAB 10	0.3229D-04	0.1580D-12	0.397D-06	0.123D+01	-0.1403D-11	-0.500D+01	-0.500D+01	0.1502D-10	0.388D-05	0.120D+02
LAB 12	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 13	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 14	0.2917D-04	0.6944D-12	0.833D-06	0.286D+01	0.1190D-11	0.109D-05	0.374D+01	0.7143D-11	0.267D-05	0.916D+01
LAB 15	0.4954D-04	0.2625D-09	0.162D-04	0.327D+02	0.7824D-09	0.280D-04	0.565D+02	0.4093D-10	0.640D-05	0.129D+02
LAB 16	0.3450D-04	0.2505D-11	0.158D-05	0.459D+01	0.7041D-11	0.265D-05	0.769D+01	0.3798D-11	0.195D-05	0.565D+01
LAB 17	0.3037D-04	0.1005D-11	0.100D-05	0.330D+01	0.2246D-11	0.150D-05	0.493D+01	0.6161D-11	0.248D-05	0.817D+01
LAB 18	0.4611D-04	0.3444D-10	0.587D-05	0.127D+02	0.1024D-09	0.101D-04	0.219D+02	0.7229D-11	0.269D-05	0.583D+01
LAB 19	0.5076D-04	0.4810D-12	0.694D-06	0.137D+01	0.1289D-11	0.114D-05	0.224D+01	0.1232D-11	0.111D-05	0.219D+01
LAB 20	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 21	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 29 SAMPLE AS										
LAB 2	0.1302D+01	0.8778D-05	0.296D-02	0.228D+00	0.2371D-04	0.487D-02	0.374D+00	0.2102D-04	0.458D-02	0.352D+00
LAB 3	0.1307D+01	0.7415D-06	0.861D-03	0.659D-01	0.1562D-05	0.125D-02	0.956D-01	0.5302D-05	0.230D-02	0.176D+00
LAB 4	0.8175D+00	0.5200D-02	0.721D-01	0.882D+01	0.1560D-01	0.125D+00	0.153D+02	0.1848D-04	0.430D-02	0.526D+00
LAB 5	0.1298D+01	0.1068D-03	0.103D-01	0.796D+00	0.3148D-03	0.177D-01	0.137D+01	0.4322D-04	0.657D-02	0.506D+00
LAB 6	0.1301D+01	0.1713D-04	0.414D-02	0.318D+00	-0.6156D-04	-0.500D+01	-0.500D+01	0.9035D-03	0.301D-01	0.231D+01
LAB 7	0.1304D+01	0.7511D-05	0.274D-02	0.210D+00	0.1366D-04	0.370D-02	0.283D+00	0.7095D-04	0.842D-02	0.646D+00
LAB 8	0.9347D+00	0.1073D-02	0.328D-01	0.350D+01	0.3038D-02	0.551D-01	0.590D+01	0.1455D-02	0.381D-01	0.408D+01
LAB 10	0.1321D+01	0.3191D-04	0.565D-02	0.427D+00	0.9397D-04	0.969D-02	0.734D+00	0.1413D-04	0.376D-02	0.284D+00
LAB 12	0.1304D+01	0.5648D-05	0.238D-02	0.182D+00	0.1079D-04	0.328D-02	0.252D+00	0.4925D-04	0.702D-02	0.538D+00
LAB 13	0.1302D+01	0.2119D-07	0.146D-03	0.112D-01	-0.1214D-05	-0.500D+01	-0.500D+01	0.1022D-04	0.320D-02	0.245D+00
LAB 14	0.1309D+01	0.8988D-06	0.948D-03	0.724D-01	0.3075D-06	0.555D-03	0.424D-01	0.1911D-04	0.437D-02	0.334D+00
LAB 15	0.1293D+01	0.9187D-05	0.303D-02	0.234D+00	0.2395D-04	0.489D-02	0.378D+00	0.2893D-04	0.538D-02	0.416D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.1297D+01	0.3007D-05	0.173D-02	0.134D+00	0.2917D-05	0.171D-02	0.132D+00	0.4883D-04	0.699D-02	0.539D+00
LAB 18	0.1304D+01	0.2212D-06	0.470D-03	0.361D-01	0.4559D-06	0.675D-03	0.518D-01	0.1663D-05	0.129D-02	0.989D-01
LAB 19	0.1306D+01	0.1657D-05	0.129D-02	0.986D-01	0.3894D-05	0.197D-02	0.151D+00	0.8608D-05	0.293D-02	0.225D+00
LAB 20	0.1310D+01	0.5330D-06	0.730D-03	0.557D-01	-0.1915D-04	-0.500D+01	-0.500D+01	0.1660D-03	0.129D-01	0.984D+00
LAB 21	0.1277D+01	0.7731D-05	0.278D-02	0.218D+00	0.2097D-04	0.458D-02	0.359D+00	0.1775D-04	0.421D-02	0.330D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 29 SAMPLE BS										
LAB 2	0.1302D+01	0.1807D-04	0.425D-02	0.327D+00	0.5206D-04	0.722D-02	0.554D+00	0.1733D-04	0.416D-02	0.320D+00
LAB 3	0.1295D+01	0.4784D-06	0.692D-03	0.534D-01	0.1274D-05	0.113D-02	0.872D-01	0.1290D-05	0.114D-02	0.877D-01

LAB 4	0.8977D+00	0.6698D-02	0.818D-01	0.912D+01	0.2009D-01	0.142D+00	0.158D+02	0.5336D-04	0.730D-02	0.814D+00
LAB 5	0.1281D+01	0.4871D-04	0.698D-02	0.545D+00	0.1236D-03	0.111D-01	0.868D+00	0.1807D-03	0.134D-01	0.105D+01
LAB 6	0.1166D+01	0.1530D-04	0.391D-02	0.335D+00	-0.6790D-04	-0.500D+01	-0.500D+01	0.9104D-03	0.302D-01	0.259D+01
LAB 7	0.1286D+01	0.728D-05	0.270D-02	0.210D+00	-0.2286D-05	-0.500D+01	-0.500D+01	0.1931D-03	0.139D-01	0.108D+01
LAB 8	0.1122D+01	0.1368D-02	0.370D-01	0.330D+01	0.3954D-02	0.629D-01	0.561D+01	0.1198D-02	0.346D-01	0.309D+01
LAB 10	0.1300D+01	0.7713D-05	0.278D-02	0.214D+00	0.2246D-04	0.474D-02	0.365D+00	0.5435D-05	0.233D-02	0.179D+00
LAB 12	0.1288D+01	0.3783D-05	0.194D-02	0.151D+00	0.7431D-05	0.273D-02	0.212D+00	0.3135D-04	0.560D-02	0.435D+00
LAB 13	0.1291D+01	0.1289D-06	0.359D-03	0.278D-01	-0.3208D-06	-0.500D+01	-0.500D+01	0.5660D-05	0.238D-02	0.184D+00
LAB 14	0.1291D+01	0.4188D-05	0.205D-02	0.158D+00	0.1126D-04	0.336D-02	0.260D+00	0.1043D-04	0.323D-02	0.250D+00
LAB 15	0.1275D+01	0.3562D-05	0.189D-02	0.148D+00	0.2065D-05	0.144D-02	0.113D+00	0.6898D-04	0.831D-02	0.651D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.1296D+01	0.1016D-04	0.319D-02	0.246D+00	0.2833D-04	0.532D-02	0.411D+00	0.1726D-04	0.415D-02	0.321D+00
LAB 18	0.1287D+01	0.6826D-06	0.826D-03	0.642D-01	-0.7447D-06	-0.500D+01	-0.500D+01	0.2234D-04	0.473D-02	0.367D+00
LAB 19	0.1292D+01	0.8880D-06	0.942D-03	0.729D-01	-0.7239D-06	-0.500D+01	-0.500D+01	0.2710D-04	0.521D-02	0.403D+00
LAB 20	0.1298D+01	0.1903D-03	0.138D-01	0.106D+01	0.5499D-03	0.234D-01	0.181D+01	0.1689D-03	0.130D-01	0.100D+01
LAB 21	0.1206D+01	0.1489D-04	0.386D-02	0.320D+00	0.4331D-04	0.658D-02	0.546D+00	0.1078D-04	0.328D-02	0.272D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
RATIO 29 SAMPLE RS										
LAB 2	0.1181D+01	0.2116D-04	0.460D-02	0.390D+00	0.5916D-04	0.769D-02	0.651D+00	0.3466D-04	0.589D-02	0.499D+00
LAB 3	0.1173D+01	0.6391D-06	0.799D-03	0.681D-01	0.1740D-05	0.132D-02	0.112D+00	0.1421D-05	0.119D-02	0.102D+00
LAB 4	0.1115D+01	0.9621D-04	0.981D-02	0.880D+00	0.2863D-03	0.169D-01	0.152D+01	0.1897D-04	0.436D-02	0.391D+00
LAB 5	0.1154D+01	0.5989D-05	0.245D-02	0.212D+00	0.1412D-04	0.376D-02	0.326D+00	0.3076D-04	0.555D-02	0.480D+00
LAB 6	0.1033D+01	0.1353D-03	0.116D-01	0.113D+01	0.2623D-03	0.162D-01	0.157D+01	0.1148D-02	0.339D-01	0.328D+01
LAB 7	0.1166D+01	0.3622D-05	0.190D-02	0.163D+00	-0.1445D-04	-0.500D+01	-0.500D+01	0.2025D-03	0.142D-01	0.122D+01
LAB 8	0.1067D+01	0.1310D-02	0.362D-01	0.339D+01	0.3839D-02	0.620D-01	0.581D+01	0.7274D-03	0.270D-01	0.253D+01
LAB 10	0.1174D+01	0.1090D-04	0.330D-02	0.281D+00	0.3194D-04	0.565D-02	0.481D+00	0.6219D-05	0.249D-02	0.212D+00
LAB 12	0.1168D+01	0.1616D-05	0.127D-02	0.109D+00	0.7693D-06	0.877D-03	0.751D-01	0.3264D-04	0.571D-02	0.489D+00
LAB 13	0.1168D+01	0.2348D-06	0.485D-03	0.415D-01	-0.1233D-06	-0.500D+01	-0.500D+01	0.6623D-05	0.257D-02	0.220D+00
LAB 14	0.1171D+01	0.3960D-05	0.199D-02	0.170D+00	0.1058D-04	0.325D-02	0.278D+00	0.1041D-04	0.323D-02	0.276D+00
LAB 15	0.1160D+01	0.1875D-04	0.433D-02	0.373D+00	0.4792D-04	0.692D-02	0.597D+00	0.6666D-04	0.816D-02	0.704D+00
LAB 16	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
LAB 17	0.1172D+01	0.4022D-06	0.634D-03	0.541D-01	0.1485D-06	0.385D-03	0.329D-01	0.8466D-05	0.291D-02	0.248D+00
LAB 18	0.1164D+01	0.5336D-07	0.231D-03	0.198D-01	-0.1074D-06	-0.500D+01	-0.500D+01	0.2140D-05	0.146D-02	0.126D+00
LAB 19	0.1163D+01	0.1174D-05	0.108D-02	0.932D-01	0.2638D-05	0.162D-02	0.140D+00	0.7062D-05	0.266D-02	0.229D+00
LAB 20	0.1160D+01	0.1340D-04	0.366D-02	0.315D+00	-0.3235D-05	-0.500D+01	-0.500D+01	0.3474D-03	0.186D-01	0.161D+01
LAB 21	0.7005D+00	0.4727D-06	0.687D-03	0.981D-01	-0.2998D-06	-0.500D+01	-0.500D+01	0.9961D-05	0.316D-02	0.451D+00
LAB 23	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0

0



RESULTS OF SUBROUTINE LAB5  
 ANALYSIS OF VARIANCES IN THREE STEPS: INTERLAB, RUN AND SCAN COMPONENTS  
 19 LABS ARE TAKEN INTO ACCOUNT AT MAXIMUM

(IF THE RUN OR SCAN COMPONENT OF ANY VARIANCE IS NEGATIVE AND THEREFORE CONSIDERED AS NOT SIGNIFICANT,  
 SD AND RSD ARE PUT = 0. CORRESPONDINGLY.)

SAMP	RAT	LABS	RUNS	SCANS	MMEAN	SIG-SCAN	SD-SCAN	RSD-SCAN	SIG-RUN	SD-RUN	RSD-RUN	SIG-LABS	SD-LABS	RSD-LABS
US	83	15	45	360	2.2120-02	2.9050-08	1.7050-04	7.7040-01	9.3830-09	9.6860-05	4.3780-01	4.4670-08	2.1140-04	9.5530-01
PS	92	15	42	336	2.1930-04	3.1800-10	1.7830-05	8.1320+00	8.0530-11	8.9740-06	4.3920+00	1.7160-09	4.1420-05	1.8890+01
PS	2	14	41	328	8.3500-04	6.6210-10	2.5730-05	3.0820+00	4.1360-11	6.4310-06	7.7030-01	6.3510-10	2.5200-05	3.0180+00
AU	38	3	9	72	2.3510-05	7.6100-11	8.7240-06	3.7110+01	1.3030-10	1.1410-05	4.8550+01	1.4050-09	3.7490-05	1.5940+02
AU	48	15	45	359	1.7890-04	1.9010-10	1.3790-05	7.7090+00	2.0100-10	1.4180-05	7.9270+00	6.2640-10	2.5030-05	1.3990+01
AU	58	19	57	456	2.2270-02	4.1130-08	2.0280-04	9.1070-01	9.9100-09	9.9550-05	4.4700-01	3.2510-08	1.8030-04	8.0960-01
AU	68	17	51	407	3.9290-03	3.7850-09	6.1520-05	1.5660+00	2.0550-09	4.5330-05	1.1540+00	1.5410-09	3.9250-05	9.9900-01
AU	89	6	18	144	1.5050-02	5.4300-08	2.3300-04	1.5490+00	8.5290-09	9.2350-05	6.1380-01	3.9450-07	6.2810-04	4.1740+00
AU	9	16	48	384	2.3070-01	2.3270-06	1.5260-03	6.6110-01	2.1930-07	4.6830-04	2.0290-01	2.7470-07	5.2410-04	2.2720-01
AU	19	17	51	408	1.2740-01	1.7960-06	1.3400-03	1.0520+00	5.8430-07	7.6440-04	6.0020-01	8.6550-07	9.3030-04	7.3050-01
AU	29	17	51	408	2.2980-02	1.1280-07	3.3590-04	1.4620+00	7.5260-08	2.7430-04	1.1940+00	1.1530-07	3.3950-04	1.4780+00
BU	38	3	9	72	2.6010-05	3.3980-10	1.8430-05	7.0870+01	1.5460-10	1.2430-05	4.7810+01	1.7040-09	4.1280-05	1.5870+02
BU	48	15	45	356	1.7170-04	1.1670-10	1.0800-05	6.2900+00	6.0860-11	7.8010-06	4.5430+00	2.1140-10	1.4540-05	8.4660+00
BU	58	19	57	456	2.2230-02	4.4600-08	2.1120-04	9.5020-01	2.5590-08	1.6000-04	7.1980-01	2.0900-08	1.4460-04	6.5040-01
BU	68	17	51	404	3.9330-03	2.8900-09	5.3760-05	1.3670+00	2.5340-09	5.0340-05	1.2800+00	3.2020-09	5.6590-05	1.4390+00
BU	89	6	18	144	1.4820-02	4.5150-08	2.1250-04	1.4340+00	6.6520-09	8.1560-05	5.5030-01	1.5960-08	1.2630-04	8.5230-01
BU	9	18	54	432	2.3020-01	2.3790-06	1.5420-03	6.7000-01	8.4970-08	2.9150-04	1.2660-01	7.1340-07	8.4460-04	3.6690-01
BU	19	17	51	408	1.2700-01	1.3570-06	1.1650-03	9.1720-01	3.2120-08	1.7920-04	1.4110-01	2.6350-07	5.1330-04	4.0430-01
BU	29	17	51	408	2.2990-02	7.8520-08	2.8020-04	1.2190+00	2.1350-08	1.4610-04	6.3570-01	9.2230-08	3.0370-04	1.3210+00
RU	38	4	11	88	3.7040-05	2.4590-11	4.9580-06	1.3390+01	6.6060-11	8.1280-06	2.1940+01	1.9910-09	4.4630-05	1.2050+02
RU	48	11	33	264	5.5970-05	3.9980-11	6.3230-06	1.1300+01	3.1840-11	5.6430-06	1.0080+01	8.3670-11	9.1470-05	1.6340+01
RU	58	18	54	432	7.2570-03	4.5060-09	6.7130-05	9.2490-01	2.0280-09	4.5030-05	6.2050-01	3.6170-09	6.0140-05	8.2860-01
RU	68	3	7	56	2.8450-04	2.2280-10	1.4930-05	5.2470+00	2.5830-11	0.0	0.0	7.6880-07	8.7680-04	3.0820+02
RU	89	5	15	120	7.5810-04	1.2470-08	1.1170-04	1.4730+01	6.3770-08	2.5250-04	3.3310+01	1.9740-06	1.4050-03	1.8530+02
RU	9	14	42	336	2.6300-02	4.7340-08	2.1760-04	8.2720-01	1.4690-08	1.2120-04	4.6080-01	2.6830-08	1.6380-04	6.2280-01
RU	19	12	36	288	7.5980-04	7.9180-10	2.8140-05	3.7030+00	3.1190-10	1.7660-05	2.3240+00	1.2550-09	3.5430-05	4.6630+00
RU	29	10	30	240	3.8320-05	2.6650-11	5.1620-06	1.3470+01	1.2660-10	1.1250-05	2.9370+01	7.5780-11	8.7050-06	2.2720+01
AS	38	18	54	432	6.3990-01	2.5290-05	5.0290-03	7.8600-01	9.9130-06	3.1480-03	4.9210-01	2.4800-05	4.9800-03	7.7830-01
AS	48	13	39	312	2.8920-04	5.1200-10	2.2630-05	7.8230+00	2.2670-09	4.7620-05	1.6460+01	6.6590-10	2.5810-05	8.9220+00
AS	58	17	51	408	2.2210-02	6.2260-08	2.4950-04	1.1240+00	1.6210-08	1.2730-04	5.7330-01	4.0240-08	2.0060-04	9.0320-01
AS	68	16	48	384	3.8820-03	6.9280-09	8.3240-05	2.1440+00	4.1070-08	2.0270-04	5.2210+00	1.2640-08	0.0	0.0
AS	89	5	15	120	1.4820-02	1.7550-07	4.1900-04	2.8270+00	7.5490-08	2.7470-04	1.8540+00	1.4510-07	3.8090-04	2.5690+00
AS	9	17	51	408	2.3410-01	5.4260-06	2.3290-03	9.9510-01	1.0870-06	1.0430-03	4.4550-01	1.9110-06	1.3820-03	5.9050-01
AS	19	16	48	384	1.2660-01	3.8580-06	1.9640-03	1.5520+00	3.1150-06	1.7650-03	1.3950+00	6.7430-07	8.2110-04	6.4880-01
AS	29	15	45	360	1.3020+00	9.3900-05	9.6900-03	7.4400-01	2.8610-05	5.3490-03	4.1060-01	7.7860-05	8.8240-03	6.7750-01
BS	38	18	54	432	6.3260-01	4.0070-05	6.3300-03	1.0010+00	3.0610-06	1.7500-03	2.7660-01	2.6520-05	5.1490-03	8.1400-01
BS	48	13	37	296	3.0200-04	2.9790-10	1.7260-05	5.7160+00	9.2160-10	3.0360-05	1.0050+01	1.0410-09	3.2270-05	1.0680+01
BS	58	17	51	408	2.2210-02	6.7430-08	2.5970-04	1.1690+00	3.1400-08	1.7720-04	7.9790-01	2.2140-08	1.4880-04	6.7000-01
BS	68	14	42	336	3.8820-03	3.4750-09	5.8950-05	1.5180+00	1.2050-09	3.4710-05	8.9410-01	1.7680-09	4.2050-05	1.0830+00
BS	89	5	15	120	1.5450-02	1.7560-07	4.1900-04	2.7110+00	3.9570-07	6.2910-04	4.0700+00	1.6220-06	1.2740-03	8.2410+00
BS	9	17	51	408	2.3440-01	7.0050-06	2.6470-03	1.1290+00	3.0530-06	1.7470-03	7.4540-01	2.5550-07	5.0550-04	2.1570-01
BS	19	16	48	384	1.2640-01	2.6120-06	1.6160-03	1.2790+00	2.4020-06	1.5500-03	1.2260+00	2.8760-07	0.0	0.0
BS	29	13	39	312	1.2910+00	5.7680-05	7.5950-03	5.8830-01	6.1100-05	7.8160-03	6.0550-01	3.3990-05	5.8300-03	4.5170-01
RS	38	18	54	432	7.5340-01	2.6210-05	5.1200-03	6.7950-01	1.4020-06	1.1840-03	1.5720-01	2.8960-05	5.3820-03	7.1430-01
RS	48	12	36	288	2.0830-04	3.1970-10	1.7880-05	8.5830+00	1.7640-09	4.2000-05	2.0160+01	3.6790-10	0.0	0.0
RS	58	16	48	384	7.5070-03	1.1760-08	1.0840-04	1.4440+00	6.3490-09	7.9680-05	1.0610+00	7.1770-09	8.4720-05	1.1290+00
RS	68	4	11	88	4.2420-05	2.2480-11	4.7410-06	1.1180+01	8.8860-10	2.9810-05	7.0270+01	6.3930-10	2.5280-05	5.9600+01
RS	89	5	15	120	1.4280-02	4.6550-05	6.8220-03	4.7780+01	1.3520-03	3.6770-02	2.5750+02	5.0050-04	2.2370-02	1.5670+02
RS	9	15	45	360	3.2980-02	7.8330-08	2.7990-04	8.4850-01	6.6010-08	2.5690-04	7.7900-01	6.1720-08	2.4840-04	7.5320-01
RS	19	12	36	288	2.0260-03	4.2630-09	6.5290-05	3.2220+00	4.1080-09	6.4090-05	3.1630+00	1.4820-09	3.8500-05	1.9000+00
RS	29	13	39	312	1.1670+00	5.8230-05	7.6310-03	6.5370-01	1.1620-05	3.4090-03	2.9210-01	4.5460-05	6.7420-03	5.7760-01

## RESULTS OF THE SUBROUTINE LAB6 - TWO-FOLD ANALYSIS OF VARIANCES

83	US	0.22460-01	0.48090-07	0.21930-03	0.97630+00	0.84830-06	0.92110-03	0.41010+01	0.51600-07	0.22710-03	0.10110+01
38	AU	0.23510-04	0.48390-09	0.22000-04	0.93570+02	0.14050-08	0.37490-04	0.15940+03	0.13980-09	0.11820-04	0.50290+02
38	BU	0.26010-04	0.58990-09	0.24290-04	0.93370+02	0.17040-08	0.41280-04	0.15870+03	0.19710-09	0.14040-04	0.53970+02
38	RU	0.34000-04	0.49230-09	0.22190-04	0.65250+02	0.19490-08	0.44150-04	0.12980+03	0.60490-10	0.77780-05	0.22870+02
38	AS	0.63990+00	0.16200-05	0.12730-02	0.19890+00	0.24800-04	0.49800-02	0.77830+00	0.13070-04	0.36160-02	0.56510+00
38	BS	0.63260+00	0.16230-05	0.12740-02	0.20140+00	0.26520-04	0.51490-02	0.81400+00	0.80690-05	0.28410-02	0.44900+00
38	RS	0.75340+00	0.16960-05	0.13020-02	0.17280+00	0.28960-04	0.53820-02	0.71430+00	0.46780-05	0.21630-02	0.28710+00
48	AU	0.17880-03	0.46630-10	0.68290-05	0.38190+01	0.62450-09	0.24990-04	0.13970+02	0.22490-09	0.15000-04	0.83870+01
48	BU	0.17120-03	0.17040-10	0.41280-05	0.24110+01	0.22910-09	0.15140-04	0.88390+01	0.79640-10	0.89240-05	0.52110+01
48	RU	0.59320-04	0.17530-10	0.41870-05	0.70580+01	0.21750-09	0.14750-04	0.24860+02	0.31170-10	0.55830-05	0.94120+01
48	AS	0.28920-03	0.11100-09	0.10540-04	0.36430+01	0.66550-09	0.25810-04	0.89220+01	0.23310-08	0.48280-04	0.16690+02
48	BS	0.30860-03	0.15060-09	0.12270-04	0.39770+01	0.12520-08	0.35380-04	0.11470+02	0.21190-08	0.46030-04	0.14920+02
48	RS	0.20010-03	0.84310-10	0.91820-05	0.45890+01	0.52200-09	0.22850-04	0.11420+02	0.17220-08	0.41500-04	0.20740+02
58	AU	0.22270-01	0.19750-08	0.44440-04	0.19960+00	0.32510-07	0.18030-03	0.80960+00	0.15050-07	0.12270-03	0.55090+00
58	BU	0.22230-01	0.16470-08	0.40580-04	0.18260+00	0.20900-07	0.14460-03	0.65040+00	0.31170-07	0.17650-03	0.79430+00
58	RU	0.22570-02	0.24890-09	0.15780-04	0.21740+00	0.36170-08	0.60140-04	0.82860+00	0.25910-08	0.50900-04	0.70140+00
58	AS	0.22210-01	0.28370-08	0.53270-04	0.23990+00	0.40240-07	0.20060-03	0.90320+00	0.23990-07	0.15490-03	0.69750+00
58	BS	0.22210-01	0.20830-08	0.45640-04	0.20550+00	0.22140-07	0.14880-03	0.67000+00	0.39830-07	0.19960-03	0.89870+00
58	RS	0.12030-01	0.20450-04	0.45230-02	0.37600+02	0.34740-03	0.18640-01	0.15500+03	0.82090-06	0.90600-03	0.75320+01
68	AU	0.40030-02	0.56490-08	0.75160-04	0.18770+01	0.75890-07	0.27550-03	0.68820+01	0.77350-07	0.27810-03	0.69480+01
68	BU	0.39300-02	0.23150-09	0.15210-04	0.38720+00	0.93400-09	0.30560-04	0.77770+00	0.96980-08	0.98480-04	0.25060+01
68	RU	0.44350-05	0.98830-11	0.31440-05	0.70880+02	0.19090-10	0.43690-05	0.98510+02	0.20250-11	0.14230-05	0.32090+02
68	AS	0.38630-02	0.41270-09	0.20320-04	0.52590+00	-0.63990-08	-0.50000+01	-0.50000+01	0.40240-07	0.20060-03	0.51930+01
68	BS	0.38280-02	0.98480-09	0.31380-04	0.81970+00	0.36220-08	0.60180-04	0.15720+01	0.39360-07	0.19840-03	0.51820+01
68	RS	0.44950-04	0.24300-09	0.15590-04	0.34680+02	0.46360-09	0.21530-04	0.47900+02	0.15250-08	0.39060-04	0.86890+02
92	PS	0.12770-01	0.39290-06	0.62680-03	0.49070+01	0.66770-05	0.25840-02	0.20230+02	0.36400-08	0.60340-04	0.47240+00
2	PS	0.90220-03	0.21850-08	0.46740-04	0.51810+01	0.37080-07	0.19260-03	0.21350+02	0.18650-09	0.13650-04	0.15140+01
89	AU	0.17930-01	0.83360-05	0.28870-02	0.16110+02	0.35990-04	0.59990-02	0.33470+02	0.67070-04	0.81900-02	0.45690+02
89	BU	0.16600-01	0.31690-05	0.17800-02	0.10720+02	0.21930-04	0.46820-02	0.28200+02	0.77750-06	0.88170-03	0.53110+01
89	RU	0.75810-03	0.39920-06	0.63180-03	0.83340+02	0.19740-05	0.14050-02	0.18530+03	0.65330-07	0.25560-03	0.33720+02
89	AS	0.21940-01	0.50710-04	0.71210-02	0.32450+02	0.30140-03	0.17360-01	0.79120+02	0.86160-05	0.29350-02	0.13380+02
89	BS	0.29630-01	0.20120-03	0.14190-01	0.47870+02	0.67040-03	0.25890-01	0.87380+02	0.16110-02	0.40140-01	0.13540+03
89	RS	0.14280-01	0.19060-03	0.13810-01	0.96690+02	0.50050-03	0.22370-01	0.15670+03	0.13580-02	0.36840-01	0.25810+03
9	AU	0.23070+00	0.22110-07	0.14870-03	0.64450-01	-0.12970-06	-0.50000+01	-0.50000+01	0.15830-05	0.12580-02	0.54530+00
9	BU	0.23020+00	0.46710-07	0.21610-03	0.93890-01	0.71340-06	0.84460-03	0.36690+00	0.38230-06	0.61830-03	0.26860+00
9	RU	0.26370-01	0.47200-07	0.21730-03	0.82390+00	0.80940-06	0.89960-03	0.34120+01	0.12090-06	0.34770-03	0.13190+01
9	AS	0.23410+00	0.14700-06	0.38340-03	0.16380+00	0.19110-05	0.13820-02	0.59050+00	0.17660-05	0.13290-02	0.56770+00
9	BS	0.23440+00	0.92050-07	0.30340-03	0.12940+00	0.25550-06	0.50550-03	0.21570+00	0.39280-05	0.19820-02	0.84560+00
9	RS	0.35130-01	0.39050-05	0.19760-02	0.56250+01	0.66220-04	0.81380-02	0.23160+02	0.51810-06	0.71980-03	0.20490+01
19	AU	0.12770+00	0.15070-06	0.38820-03	0.30410+00	0.24570-05	0.15680-02	0.12280+01	0.76790-06	0.87630-03	0.68640+00
19	BU	0.12720+00	0.60960-07	0.24690-03	0.19410+00	0.97620-06	0.98800-03	0.77680+00	0.36330-06	0.60280-03	0.47390+00
19	RU	0.81320-03	0.14870-08	0.38560-04	0.47410+01	0.23520-07	0.15340-03	0.18860+02	0.80320-09	0.28340-04	0.34850+01
19	AS	0.12680+00	0.14250-06	0.37750-03	0.29780+00	0.12090-05	0.11000-02	0.86760+00	0.36410-05	0.19080-02	0.12050+01
19	BS	0.12690+00	0.25520-06	0.50520-03	0.39820+00	0.34630-05	0.18610-02	0.14670+01	0.26260-05	0.16210-02	0.12770+01
19	RS	0.25010-02	0.19790-06	0.44490-03	0.17790+02	0.31640-05	0.17790-02	0.71130+02	0.79440-08	0.89130-04	0.35640+01
29	AU	0.23250-01	0.80410-07	0.28360-03	0.12200+01	0.12970-05	0.11390-02	0.48990+01	0.45160-06	0.67200-03	0.28910+01
29	BU	0.22890-01	0.14050-07	0.11850-03	0.51770+00	0.21840-06	0.46740-03	0.20420+01	0.10330-06	0.32130-03	0.14040+01
29	RU	0.65260-04	0.48970-09	0.22130-04	0.33910+02	0.55560-08	0.74540-04	0.11420+03	0.96150-09	0.31010-04	0.47510+02
29	AS	0.12520+01	0.12090-02	0.34780-01	0.27770+01	0.20180-01	0.14200+00	0.11340+02	0.11430-02	0.33800-01	0.26990+01
29	BS	0.12450+01	0.63090-03	0.25120-01	0.20170+01	0.10230-01	0.10120+00	0.81220+01	0.14810-02	0.38480-01	0.30900+01
29	RS	0.11230+01	0.79380-03	0.28180-01	0.25090+01	0.13400-01	0.11580+00	0.10310+02	0.28650-03	0.16930-01	0.15070+01

13. IDA 72-Measurements in One Additional Laboratory

by

E. Drosselmeyer

Institute for Applied Systems Analysis and  
Reactor Physics, Gesellschaft für Kernforschung,  
Karlsruhe, Germany

Due to laboratory difficulties the IDA 72-measurements could not be executed in the foreseen time interval in the laboratory with code no. 22. So the data have been communicated to the evaluation group at a time when the results of the other participants had already been published.

Nevertheless this laboratory has participated fully in the standard experiment. Therefore, in order to allow a comparison of the results of this laboratory with those of the majority, they have been included in the figures and tables, as they are given in the general report on the experiment. In this extra note, short comments are added.

The paper should be considered as an appendix to the general report. Comments are given only for those figures and tables in which the results of laboratory 22 show special features. In all figures the mean value has been calculated without the value from laboratory 22.

In Fig. 13-1 the values for the ratio  $U-235/U-238$  of the unspiked A and B samples lie about 8 % below the mean value. Also, the values for the ratio  $U-236/U-238$  in Fig. 13-2 lie below the mean value, this time about 10 %. The same is true for the results of the R sample where again the ratio  $U-235/U-238$  /Fig. 13-3/ lies about 10 % below the line of the mean value. All these effects make a contamination by depleted uranium during the chemical preparation rather probable. Another possibility is a contamination of the filaments by natural uranium. It was stated, however, at the meeting of IDA participants that this is rather unlikely /Vol. I, Chapter 8/.

It is strange in this context that the concentration value for  $U-238$  /Fig. 13-4/ lies somewhat high for the A and B samples in agreement with the arguments given before - but the value for the R sample lies too low. This may be due to the fact that for the spiked R samples the measured ratios of e.g.  $U-233/U-238$  were measured higher than in the group of the other laboratories /Fig. 13-5/. The effect could not be explained.

In the case of plutonium the results on the ratio Pu-240/Pu-239 of the unspiked A and B samples /Fig. 13-6/ are outliers with values lying more than 2 % below the mean values. For the ratios Pu-241/Pu-239 of the same samples /Fig. 13-7/ the  $1\sigma$ -ranges of the measurements were abnormally large. The same is true for the measurements of the ratio Pu-242/Pu-239 /Fig. 13-8/. The values for the unspiked R sample given in Fig. 13-9 agree well with the group of the other laboratories.

The laboratory 22 has decided to measure the ratio Pu-238/Pu-239 by  $\alpha$ -spectrometry. Only for the unspiked R sample the result is in agreement with the mean value /Fig. 13-10/, the values for the unspiked and spiked A and B samples and for the spiked R sample deviate by about - 9 %, - 30 % and + 70 % /Fig. 13-11, 13-12, 13-13/. Specially for the spiked R sample also other laboratories found highly deviating results - the differences found are as large as 333 %!

The resulting figures on the Pu-239-concentration calculated for laboratory 22 lie well within the band of values around the mean values calculated for the group, see Fig. 13-14.

Comparison of the results of the analysis of variances as they are calculated for 19 laboratories (normal group) or 20 laboratories (with inclusion of the results of laboratory 22) show no striking features.

In most cases the mean values and the estimates of the relative standard deviations remain practically unchanged. In a few cases where addition of this one laboratory changes the mean value considerably ( $\sim 1\%$ ) the estimate for the interlaboratory component of the relative standard deviation becomes larger as one would also expect for these cases. In a few cases the estimates for the scan component of the relative standard deviation becomes a bit smaller which is also understandable since the number of scans is enlarged by 24 when one laboratory is added.

How the numerical results of laboratory 22 for isotopic ratios compare with those of the group can be seen in Tables 13-1, 13-2 and 13-3 for the uranium isotopes of the unspiked A, B and R samples and in Tables 13-4, 13-5 and 13-6 for the plutonium isotopes of the unspiked A, B and R samples.

The results of  $\alpha$ -spectrometry are combined in Tables 13-7, 13-8 and 13-9.

Table 13-10 gives a survey over the concentration of the main isotopes U-238 and Pu-239 as well as the Pu/U ratios for all the laboratories.

As can be seen in the corresponding figures /Fig. 13-4 and 13-5 for uranium, Fig. 13-16 and 13-17 for plutonium/ the results of laboratory 22 lie well inside the band of values obtained by the group.

Table 13-1: IDA-72: Calculated Laboratory Mean Values of the  
Isotopic Composition of Sample A(unspiked)

Lab. Code	Relative Isotopic Abundances [ atom % ]				
	U-233	U-234	U-235	U-236	U-238
2	0.00008	0.0162	2.162	0.3818	97.440
3	-	0.0173	2.178	0.3919	97.413
4	-	-	(2.177)	(0.5125)	(97.311)
5	-	0.0247	2.195	0.3912	97.390
6	-	0.0167	2.179	0.3849	97.419
7	-	0.0135	2.145	0.3797	97.462
8	0.00657	0.0188	2.204	0.3766	97.394
10	0.00022	0.0162	2.155	0.3829	97.446
12	-	-	2.169	0.3866	97.444
13	-	0.0180	2.165	0.3805	97.436
14	-	0.0195	2.156	0.3833	97.441
15	-	0.0166	2.181	0.3861	97.416
16	-	0.0178	2.174	0.3817	97.427
17	-	0.0154	2.152	0.3808	97.452
18	-	0.0189	2.173	0.3746	97.433
19	-	0.0171	2.148	0.3820	97.453
20	-	0.0148	2.141	0.3770	97.468
21	-	-	2.165	0.3862	97.449
22	-	(0.0149)	(2.007)	(0.3479)	(97.631)
23	-	-	(2.212)	-	(97.788)
Mean of means:		0.0174	2.167	0.3828	97.434
SD		0.0026	0.017	0.0047	0.022
RSD [%]		14.7	0.79	1.23	0.02

Values in brackets were not used for the calculation of the mean values.

**Table 13-2: IDA-72: Calculated Laboratory Mean Values of the Isotopic Composition of Sample B(unspiked)**

Relative Isopic Abundances [atom % ]

Lab. Code	U-233	U-234	U-235	U-236	U-238
2	0.00012	0.0165	2.155	0.3821	97.446
3	-	0.0170	2.140	0.3854	97.458
4	-	-	2.177	0.4011	97.422
5	-	0.0194	2.174	0.3816	97.425
6	-	0.0160	2.180	0.3850	97.419
7	-	0.0149	2.152	0.3801	97.453
8	0.00727	0.0180	2.181	0.3730	97.421
10	0.00022	0.0161	2.148	0.3796	97.456
12	-	-	2.171	0.3858	97.443
13	-	0.0194	2.163	0.3798	97.438
14	-	0.0159	2.160	0.3861	97.439
15	-	0.0174	2.175	0.3858	97.422
16	-	0.0162	2.185	0.3873	97.412
17	-	0.0158	2.157	0.3831	97.444
18	-	0.0175	2.186	0.3803	97.416
19	-	0.0167	2.146	0.3807	97.457
20	-	0.0135	2.135	0.3712	97.481
21	-	-	2.170	0.3841	97.446
22	-	(0.0147)	(1.988)	(0.3376)	(97.660)
23	-	-	(2.203)	-	(97.797)
Mean of means:		0.0167	2.164	0.3829	97.439
SD		0.0015	0.016	0.0063	0.019
RSD [%]		9.26	0.73	1.64	0.02

Values in brackets were not used for the calculation of the mean values.



**Table 13-3:** IDA-72: Calculated Laboratory Mean Values of the Isotopic Composition of Sample R(unspiked)

Lab. Code	Relative Isotopic Abundances [atom %]				
	U-233	U-234	U-235	U-236	U-238
2	0.00005	0.0059	0.7214	0.00075	99.272
3	0.00374	0.0066	0.7204	-	99.269
4	-	-	0.7207	-	99.279
5	-	0.0099	0.7326	-	99.257
6	-	0.0066	0.7274	-	99.266
7	-	0.0035	0.7099	-	99.287
8	0.00947	0.0056	0.7263	-	99.259
10	0.00023	0.0052	0.7165	0.00013	99.278
12	-	0.0055	0.7179	-	99.277
13	-	-	0.7211	-	99.279
14	-	0.0050	0.7215	-	99.273
15	-	0.0068	0.7306	-	99.263
16	-	-	0.7208	-	99.279
17	-	0.0051	0.7184	-	99.276
18	-	0.0039	0.7267	-	99.269
19	-	0.0059	0.7135	-	99.281
20	-	0.0048	0.7086	-	99.287
21	-	-	(2.824 )	(0.1906)	(96.985)
22	-	(0.0049)	(0.6506)	-	(99.344)
23	-	-	0.7144	-	99.286
Mean of means :		0.0057	0.7205	-	99.274
SD		0.0015	0.0066		0.009
RSD [%]		26.7	0.91	-	0.01
CBNM		0.0055	0.7203		99.2742

Values in brackets were not used for the calculation of the mean values.

**Table 13-4:** IDA-72 : Calculated Laboratory Mean Values of the Isotopic Composition of Sample A (unspiked)

Relative Isotopic Abundances [atom %]

Lab. Code	Method used for Pu-238 determination	Pu-238	Pu-239	Pu-240	Pu-241	Pu-242	Reason for rejection
2	none	-	(72.106)	(16.628)	(9.576)	(1.691)	no Pu-238 measurement
3	$\alpha$	1.048	71.491	16.571	9.235	1.655	
4	$\alpha$	1.028	71.634	16.510	9.200	1.627	
5	$\alpha$	1.045	71.795	16.551	8.983	1.627	
6	$\alpha$	1.057	71.605	16.538	9.147	1.653	
7	$\alpha$	1.042	71.661	16.535	9.110	1.652	
8	MS	(1.085)	(71.330)	(16.385)	(9.215)	(1.985)	Pu-242/Pu-239 outlier
10	MS	1.046	71.709	16.479	9.125	1.640	
12	none	-	(72.402)	(16.697)	(9.244)	(1.657)	no Pu-238 measurement
13	$\alpha$	1.054	71.704	16.526	9.095	1.621	
14	MS	1.074	71.728	16.507	9.055	1.636	
15	$\alpha$	1.053	71.620	16.539	9.137	1.651	
16	MS	1.159	71.452	16.582	9.162	1.646	
17	$\alpha$	1.053	71.661	16.536	9.113	1.637	
18	$\alpha$	1.048	71.648	16.514	9.144	1.646	
19	$\alpha$	1.058	71.748	16.515	9.062	1.618	
20	none	-	(72.373)	(16.769)	(9.216)	(1.642)	no Pu-238 measurement
21	$\alpha$	(0.655)	(71.930)	(16.605)	(9.072)	(1.738)	Pu-238/Pu-239 outlier
22	$\alpha$	(0.947)	(71.947)	(16.232)	(9.113)	(1.760)	
Mean of means		1.059	71.650	16.531	9.121	1.639	
SD		0.032	0.096	0.027	0.065	0.013	
RSD [%]		3.01	0.13	0.17	0.71	0.77	

Values in brackets were not used for the calculation of mean value, SD and RSD

**Table 13-5:** IDA-72 : Calculated Laboratory Mean Values of the Isotopic Composition of Sample B(unspiked)

Lab. Code	Method used for Pu-238 determination	Relative Isotopic Abundances [atom %]					Reason for rejection
		Pu-238	Pu-239	Pu-240	Pu-241	Pu-242	
2	none	-	(72.219)	(16.655)	(9.443)	(1.683)	no Pu-238 measurement
3	$\alpha$	1.047	71.621	16.514	9.148	1.669	
4	$\alpha$	1.035	71.763	16.457	9.131	1.613	
5	$\alpha$	(0.862)	(71.840)	(16.579)	(9.094)	(1.625)	Pu-238/Pu-239 outlier
6	$\alpha$	1.047	71.619	16.544	9.140	1.651	
7	$\alpha$	1.043	71.704	16.487	9.126	1.641	
8	MS	1.073	71.842	16.341	9.039	1.705	
10	MS	1.047	71.762	16.452	9.098	1.641	
12	none	-	(72.400)	(16.696)	(9.236)	(1.667)	no Pu-238 measurement
13	$\alpha$	1.045	71.724	16.496	9.101	1.634	
14	MS	1.070	71.620	16.540	9.122	1.648	
15	$\alpha$	1.060	71.660	16.517	9.116	1.647	
16	MS	1.054	71.611	16.584	9.115	1.637	
17	$\alpha$	1.049	71.641	16.525	9.145	1.640	
18	$\alpha$	1.021	71.714	16.513	9.104	1.648	
19	$\alpha$	1.055	71.762	16.495	9.070	1.618	
20	none	-	(72.623)	(16.666)	(9.163)	(1.548)	no Pu-238 measurement
21	$\alpha$	(0.716)	(71.929)	(16.589)	(9.081)	(1.685)	Pu-238/Pu-239 outlier
22	$\alpha$	(0.975)	(72.166)	(16.218)	(9.022)	(1.619)	
Mean of means		1.050	71.696	16.497	9.112	1.646	
SD		0.014	0.073	0.059	0.031	0.023	
RSD [%]		1.30	0.10	0.36	0.34	1.38	

Values in brackets were not used for the calculation of mean value, SD and RSD

**Table 13-6:** IDA-72 : Calculated Laboratory Mean Values of the Isotopic Composition of Sample R(unspiked)

Relative Isotopic Abundances [atom %]

Lab. Code	Method used for Pu-238 determination	Pu-238	Pu-239	Pu-240	Pu-241	Pu-242	Reason for rejection
2	none	-	97.345	2.569	0.0803	0.00536	
3	$\alpha$	0.00387	97.338	2.584	0.0712	0.00303	
4	$\alpha$	(0.00383)	(97.708)	(2.288)	-	-	no Pu-241 measurement, Pu-240/Pu-239 outlier
5	$\alpha$	0.00358	97.222	2.692	0.0820	-	
6	$\alpha$	0.00663	97.326	2.581	0.0770	0.00973	
7	$\alpha$	0.00371	97.394	2.522	0.0778	0.00235	
8	MS	(0.02916)	(97.188)	(2.657)	(0.0964)	(0.02916)	Pu-241/Pu-239 outlier
10	MS	0.00697	97.340	2.577	0.0724	0.00314	
12	none	-	97.363	2.565	0.0724	-	
13	$\alpha$	0.00755	97.376	2.547	0.0695	-	
14	MS	0.00933	97.334	2.580	0.0739	0.00284	
15	$\alpha$	0.00389	97.351	2.564	0.0761	0.00482	
16	none	-	97.375	2.550	0.0718	0.00336	
17	$\alpha$	0.00378	97.367	2.556	0.0705	0.00296	
18	$\alpha$	0.00362	97.366	2.554	0.0725	0.00449	
19	none	-	97.361	2.561	0.0728	0.00494	
20	none	-	97.392	2.541	0.0662	-	
21	$\alpha$	(0.02597)	(97.128)	(2.717)	(0.1295)	-	Pu-238/Pu-239 and Pu-241/Pu-239 outlier
22	$\alpha$	(0.0042)	(97.347)	(2.570)	(0.0737)	(0.00592)	
Mean of means		0.00529	97.350	2.570	0.0738	0.00427	
SD		0.00212	0.041	0.038	0.0042	0.00207	
RSD		40.1	0.04	1.47	5.68	48.4	
CBNM		0.0039	97.355	2.565	0.0724	0.0032	

Values in brackets were not used for the calculation of mean value, SD and RSD

**Table 13-7: IDA-72 :  $\alpha$  - Spectrometric Determination of Pu-238 on Unspiked Samples A and B**

Laboratory Code	Sample A, unspiked			Sample B, unspiked		
	Lab. mean of $\alpha$ -activity ratio $\frac{\text{Pu-238}}{(\text{Pu-239}+\text{Pu-240})}$	RSD of single determination [ % ]	Calculated isotopic ratio $\frac{\text{Pu-238}}{\text{Pu-239}}$	Lab. mean of $\alpha$ -activity ratio $\frac{\text{Pu-238}}{(\text{Pu-239}+\text{Pu-240})}$	RSD of single determination [ % ]	Calculated isotopic ratio $\frac{\text{Pu-238}}{\text{Pu-239}}$
3	2.194	1.01	0.01466	2.193	1.39	0.01462
4	2.153	2.18	0.01435	2.169	1.31	0.01443
5	2.182	0.74	0.01455	(1.798)	0.54	(0.01199)
6	2.213	1.55	0.01477	2.192	1.07	0.01462
7	2.180	1.47	0.01454	2.184	0.05	0.01454
13	2.206	0.73	0.01470	2.188	0.43	0.01457
15	2.203	0.17	0.01470	2.218	2.26	0.01478
17	2.204	0.52	0.01470	2.196	0.23	0.01465
18	2.193	0.70	0.01462	2.137	0.71	0.01424
19	2.213	0.29	0.01474	2.208	0.39	0.01470
21	(1.366)	-	(0.00911)	(1.494)	-	(0.00996)
22	(1.995)	2.74	(0.01317)	(2.051)	2.51	(0.01352)
Mean of lab. means	2.194	-	0.01463	2.187	-	0.01457

Values in brackets are considered as outliers and not used for calculating the mean of lab. means.

**Table 13-8: IDA-72 :  $\alpha$  -Spectrometric Determination of Pu-238 on Spiked Samples A and B**

Laboratory Code	Sample A, spiked			Sample B, spiked		
	Lab. mean of $\alpha$ -acti- vity ratio Pu-238	RSD of single determina- tion [ % ]	Calculated isotopic ratio Pu-238	Lab. mean of $\alpha$ -acti- vity ratio Pu-238	RSD of single determina- tion [ % ]	Calculated isotopic ratio
	(Pu-239+Pu-240)		Pu-239	(Pu-239+Pu-240)		$\frac{\text{Pu-238}}{\text{Pu-239}}$
3	2.134	1.46	0.01436	2.146	1.00	0.01448
4	2.113	1.56	0.01404	2.085	1.96	0.01402
5	2.059	0.43	0.01382	2.134	1.06	0.01431
6	2.162	0.55	0.01455	2.156	0.50	0.01446
7	2.162	0.35	0.01455	2.158	0.03	0.01445
13	2.163	1.08	0.01451	2.126	0.17	0.01427
15	2.164	2.98	0.01452	2.172	3.48	0.01454
17	2.175	0.23	0.01463	2.177	0.12	0.01463
18	2.153	1.42	0.01446	2.140	0.47	0.01435
19	2.167	0.61	0.01458	2.160	0.13	0.01453
21	(1.045)	-	(0.00702)	(0.884)	-	(0.00594)
22	(1.670)	1.52	(0.01114)	(1.496)	9.85	(0.00998)
Mean of lab. means	2.145	-	0.01440	2.145	-	0.01440

Values in brackets are considered as outliers and not used for calculating the mean of lab. means.

**Table 13-9:** IDA-72 :  $\alpha$  -Spectrometric Determination of Pu-238 on R-Samples

Laboratory Code	Sample R, unspiked			Sample R, spiked		
	Lab. mean of $\alpha$ -activity ratio Pu-238 (Pu-239+Pu-240)	RSD of single determination [ % ]	Calculated isotopic ratio Pu-238 Pu-239	Lab. mean of $\alpha$ -activity ratio Pu-238 (Pu-239+Pu-240)	RSD of single determination [ % ]	Calculated isotopic ratio Pu-238 Pu-239
3	0.0101	4.89	0.00004	0.0235	2.80	0.00010
4	0.0100	6.26	0.00004	0.0197	4.33	0.00008
5	0.0093	4.69	0.00004	(0.0717)	0.58	(0.00029)
6	0.0173 <sup>1)</sup>	6.97	0.00007	0.0218	7.99	0.00009
7	0.0097	1.58	0.00004	0.0229	6.62	0.00009
13	0.0197	1.40	0.00008	(0.0461)	1.95	(0.00019)
15	0.0101	3.42	0.00004	-	-	-
17	0.0098	0.59	0.00004	0.0294	1.10	0.00012
18	0.0094	3.35	0.00004	0.0286	2.62	0.00012
19	-	-	-	-	-	-
21	(0.0674)	-	(0.00027)	(0.1053)	-	(0.00047)
22	0.0109	6.99	0.00004	(0.0415)	7.8	(0.00017)
Mean of lab. means	0.0117	-	0.00005	0.0243	-	0.00010

Values in brackets are considered as outliers and not used for calculating the mean of lab. means.


<sup>1)</sup> Mean value of 2 single determinations only. The third determination was marked as outlier by the laboratory.

**Table 13-10: IDA 72: U-238 and Pu-239 Concentrations and Resulting Pu-239/U-238 Concentration Ratios Calculated for Samples A,B and R**

Lab. Code	Sample A			Sample B			Sample R		
	Concentration		Atomic ratio $\frac{\text{Pu-239}}{\text{U-238}}$ [x 10 <sup>-2</sup> ]	Concentration		Atomic ratio $\frac{\text{Pu-239}}{\text{U-238}}$ [x 10 <sup>-2</sup> ]	Concentration		Atomic ratio $\frac{\text{Pu-239}}{\text{U-238}}$ [x 10 <sup>-2</sup> ]
	U-238 [x10 <sup>18</sup> atoms g sol.]	Pu-239 [x10 <sup>16</sup> atoms g sol.]		U-238 [x10 <sup>18</sup> atoms g sol.]	Pu-239 [x10 <sup>16</sup> atoms g sol.]		U-238 [x10 <sup>18</sup> atoms g sol.]	Pu-239 [x10 <sup>16</sup> atoms g sol.]	
2	2.944	1.732	0.5883	2.802	1.632	0.5824	2.770	2.081	0.7513
3	2.922	1.713	0.5862	2.794	1.631	0.5838	2.811	2.094	0.7449
4	2.922	(2.644)	-	2.795	(2.261)	-	2.770	(2.206)	-
5	2.928	1.696	0.5792	2.790	1.622	0.5814	2.732	2.129	0.7793
6	2.915	(1.513)	-	2.776	(1.597)	-	2.786	(2.382)	-
7	2.939	1.706	0.5805	2.830	1.632	0.5767	2.770	2.107	0.7606
8	2.955	(2.212)	-	2.809	(1.718)	-	2.747	(2.307)	-
10	2.935	1.695	0.5775	2.787	1.626	0.5834	2.779	2.092	0.7528
12	2.929	1.709	0.5835	2.779	1.632	0.5873	2.781	2.104	0.7566
13	2.911	1.711	0.5878	2.786	1.628	0.5844	2.767	2.103	0.7600
14	2.907	1.706	0.5869	2.791	1.631	0.5844	2.785	2.099	0.7537
15	2.929	1.711	0.5842	2.813	1.637	0.5819	2.747	2.119	0.7714
17	2.959	1.725	0.5830	2.800	1.628	0.5814	2.769	2.096	0.7570
18	2.936	1.704	0.5804	2.818	1.627	0.5774	2.755	2.111	0.7662
19	2.922	1.698	0.5811	2.784	1.619	0.5815	2.770	2.113	0.7628
20	2.941	1.690	0.5746	2.793	1.606	0.5750	2.766	2.118	0.7657
21	2.920	(1.042)	-	2.770	(1.042)	-	2.822	(3.530)	-
22	(2.979)	(1.728)	(0.5801)	(2.839)	(1.617)	(0.5696)	(2.747)	2.113	(0.7692)
23	2.882	-	-	2.737	-	-	2.764	-	-
Mean of means	2.928	1.707	0.5826	2.792	1.627	0.5816	2.772	2.105	0.7602
SD	0.019	0.012	0.0042	0.020	0.008	0.0034	0.021	0.013	0.0091
RSD [%]	0.64	0.69	0.71	0.73	0.48	0.59	0.76	0.62	1.20
						CBNM	2.785	2.109	0.7573

Concentration values given for samples A and B are those obtained after calibration with sample R  
Values in brackets were not used for the calculation of mean value, SD and RSD.





( Mean values per laboratory; error bars indicate  $\pm 1\sigma$  - range of these means )

Fig.13-1 IDA-72: MS-Determinations of Isotopic Ratio U235/U238 of Unspiked Samples A and B




Fig.13-2 IDA-72: MS-Determinations of Isotopic Ratio U236/U238 of Unspiked Samples A and B




Fig 13-3 IDA-72 : MS- Determinations of Isotopic Ratio U235 / U238 of Unspiked Sample R




Fig.13-4 IDA-72 : U238 Concentration of Samples A,B and R




Fig.13-5 IDA-72 : MS- Determinations of Isotopic Ratio U233 / U238 of Spiked Samples A,B and R




Fig.13-6IDA-72: MS-Determinations of Isotopic Ratio Pu240/Pu239 of Unspiked Samples A and B




Fig.13-7 IDA-72: MS - Determinations of Isotopic Ratio Pu 241 / Pu 239  
 of Unspiked Samples A and B






Fig.13 - 8IDA-72: MS - Determinations of Isotopic Ratio Pu 242 / Pu 239 of Unspiked Samples A and B





( Mean values per laboratory; error bars indicate  $\pm 1\sigma$  - range of these means )

Fig.13-9 IDA-72 : MS - Determinations of Isotopic Ratio Pu 240 / Pu 239 of Unspiked Sample R



( Mean values per laboratory; error bars indicate  $\pm 1\sigma$ -range of these means )

Fig.13-10 IDA-72:  $\alpha$  - Spectrometric Determinations of the Activity Ratio Pu 238 / ( Pu 239 + Pu 240 ) on the Unspiked Sample R






Fig.13-11 IDA-72:  $\alpha$  - Spectrometric Determinations of the Activity Ratio  
 Pu 238 / ( Pu 239 + Pu 240 ) on the Unspiked Samples A and B



( Mean values per laboratory; error bars indicate  $\pm 1\sigma$ -range of these means )

Fig.13-12IDA-72:  $\alpha$ - Spectrometric Determinations of the Activity Ratio  
Pu 238 / (Pu 239 + Pu 240) on the Spiked Samples A and B



( Mean values per laboratory; error bars indicate  $\pm 1\sigma$ -range of these means )

Fig.13-13 IDA-72:  $\alpha$ -Spectrometric Determinations of the Activity Ratio Pu 238 / ( Pu 239 + Pu 240 ) on the Spiked Sample R




Fig.13-14 IDA-72: Pu 239 Concentration of Samples A, B and R






Fig.13-15 IDA-72 : U238 Concentration of Samples A and B  
 after "Calibration" with Sample R



( Mean values per laboratory ; error bars indicate  $\pm 1\sigma$ -range of these means )

Fig.13-16 IDA-72 : MS - Determinations of Isotopic Ratio Pu 242 / Pu 239  
of Spiked Samples A,B and R





( Mean values per laboratory; error bars indicating  $\pm 1\sigma$ -range taken as in Fig. 13-14 )

Fig.13-17IDA-72: Pu 239 Concentration of Sample A and B after "Calibration" with Sample R

## References

## Volume II

- /1/ Jones, J.  
Selected Measurement Methods for Plutonium and Uranium  
in the Nuclear Fuel Cycle, USAEC, pp 8, 1963
- /2/ CODATA Bulletin No. 11, International Council of Scientific  
Unions, CODATA Central Office, 6 Frankfurt/Main,  
West Germany, 1973
- /3/ BNL Note 50233 (T-570), 1970
- /4/ Lederer, M. and Hollander, G.M.  
Table of Isotopes, 1967
- /5/ Cabell, M.J.  
Chemical Nuclear Data, Measurements and Applications,  
Proceedings of the International Conference, September 20-22,  
Canterbury, Kent, 1971
- /6/ Rad. Chim. Acta 12, 160, 1969
- /7/ v. Baeckmann, A.  
Determination of Actinide Elements by X-Ray Analysis,  
IAEA : STI/PUB 337, 33, 1972
- /8/ Natrella, Mary Gibbons  
Experimental Statistics  
National Bureau of Standards Handbook 91
- /9/ Dixon, W.J.  
Processing Data for Outliers  
Biometrics 9, 1953
- /10/ Dixon, W.J.  
Ratios Involving Extreme Values  
Anuals of Math. Stat., 22, pp 68-78, 1951

- /11/ Pawwat, L.G.  
Probability and Experimental Errors in Science,  
John Wiley and Sons Inc., 1961
- /12/ Worthing, Archie G., Geffner, Joseph  
Treatment of Experimental Data,  
John Wiley and Sons Inc., New-York-London
- /13/ Pugh, E.M., Wunslow, G.H.  
The Analysis of Physical Measurements,  
Addison - Wesley, 1966
- /14/ Barford, N.C.  
Kleine Einführung in die statistische Analyse von Meß-  
ergebnissen,  
Akademische Verlagsanstalt, Frankfurt/Main, 1970
- /15/ Lyon, A.J.  
Dealing with Data,  
Pergamon Press, 1970
- /16/ Bartsch, H.J.  
Mathematische Formeln,  
VEB-Fachbuchverlag, Leipzig, 1971
- /17/ Jahnke, Emde, Lösch  
Tafeln höherer Funktionen,  
Teubner, Stuttgart, 1966
- /18/ Joint Integral Safeguards Experiment (JEX 70) at the  
EUROCHEMIC Reprocessing Plant, Mol, Belgium,  
Edited by R. Kraemer and W. Beyrich  
EUR 4576e - KFK 1100, pp 7-134, 1971
- /19/ Joint Integral Safeguards Experiment (JEX 70) at the  
EUROCHEMIC Reprocessing Plant, Mol, Belgium,  
Edited by R. Kraemer and W. Beyrich  
EUR 4576e-KFK 1100, pp 7-134, 1971